

Data Path : Z:\HPCHEM1\BNA M\DATA\BM063016\
 Data File : BM006152.D
 Acq On : 30 Jun 2016 08:15
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02054

Quant Time: Jun 30 14:47:41 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM062816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 29 00:57:09 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	146	0.00
2 S	1,4-Dioxane-d8	0.431	0.322	25.3#	137	0.00
3 S	Phenol-d5	1.634	1.331	18.5	144	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	0.963	0.813	15.6	144	0.00
5 S	2-Chlorophenol-d4	1.254	1.006	19.8	144	0.00
6 S	4-Methylphenol-d8	1.315	1.061	19.3	141	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	142	0.00
8 S	Nitrobenzene-d5	0.135	0.111	17.8	146	0.00
9 S	2-Nitrophenol-d4	0.154	0.121	21.4#	138	0.00
10 S	2,4-Dichlorophenol-d3	0.281	0.227	19.2	142	0.00
11	Naphthalene	0.876	0.717	18.2	144	0.00
12 S	4-Chloroaniline-d4	0.222	0.176	20.7#	123	0.00
13	2-Methylnaphthalene	0.665	0.538	19.1	141	0.00
14	1-Methylnaphthalene	0.632	0.508	19.6	142	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	135	0.00
16 S	Dimethylphthalate-d6	1.434	1.120	21.9#	131	0.00
17 S	Acenaphthylene-d8	1.740	1.445	17.0	138	0.00
18	Acenaphthylene	1.822	1.509	17.2	135	0.00
19 C	Acenaphthene	1.162	0.957	17.6	135	0.00
20 S	4-Nitrophenol-d4	0.272	0.216	20.6#	125	0.00
21 S	Fluorene-d10	1.223	0.983	19.6	134	0.00
22	Fluorene	1.323	1.080	18.4	133	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.071	0.076	-7.0	211#	0.00
25	Phenanthrene	0.946	0.772	18.4	131	0.00
26 S	Anthracene-d10	0.802	0.656	18.2	130	0.00
27	Anthracene	0.973	0.805	17.3	131	0.00
28 C	Fluoranthene	1.159	0.915	21.1	123	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
30 S	Pyrene-d10	0.805	0.768	4.6	122	0.00
31	Pyrene	1.055	1.005	4.7	122	0.00
32	Benzo(a)anthracene	1.040	0.848	18.5	102	0.00
33	Chrysene	0.957	0.778	18.7	100	0.00
34 I	Perylene-d12	1.000	1.000	0.0	75	0.00
35	Benzo(b)fluoranthene	1.038	0.858	17.3	79	0.00
36	Benzo(k)fluoranthene	0.974	0.835	14.3	79	0.00
37 S	Benzo(a)pyrene-d12	0.796	0.641	19.5	75	0.00
38 C	Benzo(a)pyrene	0.981	0.799	18.6	75	0.00
39	Indeno(1,2,3-cd)pyrene	1.019	0.896	12.1	77	-0.01
40	Dibenzo(a,h)anthracene	0.843	0.753	10.7	78	0.00
41	Benzo(g,h,i)perylene	0.853	0.778	8.8	81	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM063016\
Data File : BM006152.D
Acq On : 30 Jun 2016 08:15
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02054

Quant Time: Jun 30 14:47:41 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM062816.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 29 00:57:09 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0