

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM063021\
 Data File : BM030803.D
 Acq On : 03 Jul 2021 09:12
 Operator : CG/JU
 Sample : M2869-16
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGDY6

Manual Integrations
 APPROVED

mohammad
 7/6/2021 9:25:54 AM

Quant Time: Jul 05 01:02:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM062221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 05 01:01:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	91302	20.00	ng/ul	0.00
20) Naphthalene-d8	10.551	136	363549	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.398	164	216156	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.133	188	412678	20.00	ng/ul	0.00
79) Chrysene-d12	21.303	240	417638	20.00	ng/ul	0.00
88) Perylene-d12	23.550	264	470531	20.00	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.263	96	4413m	1.96	ng/uL	0.00
4) Pyridine-d5	3.746	84	6873m	1.14	ng/ul	0.07
7) Phenol-d5	6.934	99	72430	9.20	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.104	67	48050	9.71	ng/ul	0.00
11) 2-Chlorophenol-d4	7.298	132	62083	10.21	ng/ul	0.00
15) 4-Methylphenol-d8	8.469	113	57967	9.46	ng/ul	0.00
21) Nitrobenzene-d5	8.922	128	27772	10.24	ng/ul	0.00
24) 2-Nitrophenol-d4	9.639	143	29414	9.76	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.181	165	53242	9.22	ng/ul	0.00
31) 4-Chloroaniline-d4	10.698	131	64106	7.96	ng/ul	0.00
46) Dimethylphthalate-d6	13.810	166	162057	10.86	ng/ul	0.00
49) Acenaphthylene-d8	14.086	160	201029	10.38	ng/ul	0.00
54) 4-Nitrophenol-d4	14.621	143	13162m	4.65	ng/ul	0.02
60) Fluorene-d10	15.392	176	145002	10.95	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.521	200	9403	3.48	ng/ul	0.00
73) Anthracene-d10	17.233	188	221617	11.51	ng/ul	0.00
81) Pyrene-d10	19.521	212	231418	10.57	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.409	264	245471	10.03	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM063021\
Data File : BM030803.D
Acq On : 03 Jul 2021 09:12
Operator : CG/JU
Sample : M2869-16
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGDY6

Manual Integrations
APPROVED

mohammad
7/6/2021 9:25:54 AM

Quant Time: Jul 05 01:02:20 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM062221.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 05 01:01:28 2021
Response via : Initial Calibration

