

Data Path : Z:\HPCHEM1\BNA_M\Data\BM070115\
 Data File : BM001814.D
 Acq On : 01 Jul 2015 19:02
 Operator : TP/UM
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02017

Quant Time: Jul 02 07:54:17 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061515.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 02 07:09:03 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.479	0.443	7.5	88	0.00
3 S	1,4-Dioxane-d8	0.419	0.434	-3.6	96	0.00
4	Benzaldehyde	1.005	1.038	-3.3	93	0.00
5 S	Phenol-d5	1.536	1.538	-0.1	97	0.00
6	Phenol	1.585	1.589	-0.3	96	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.860	0.865	-0.6	94	0.00
8	Bis(2-Chloroethyl)ether	1.191	1.188	0.3	94	0.00
9 S	2-Chlorophenol-d4	1.215	1.250	-2.9	100	0.00
10	2-Chlorophenol	1.258	1.284	-2.1	97	0.00
11	2-Methylphenol	1.174	1.161	1.1	96	0.00
12	2,2'-oxybis(1-Chloropropane	1.458	1.369	6.1	87	0.00
13 S	4-Methylphenol-d8	1.233	1.247	-1.1	98	0.00
14	Acetophenone	1.957	1.988	-1.6	97	0.00
15 P	N-Nitroso-di-n-propylamine	0.956	0.961	-0.5	95	0.00
16	4-Methylphenol	1.275	1.268	0.5	96	0.00
17	Hexachloroethane	0.505	0.509	-0.8	95	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
19 S	Nitrobenzene-d5	0.138	0.148	-7.2	99	0.00
20	Nitrobenzene	0.375	0.386	-2.9	97	0.00
21	Isophorone	0.626	0.648	-3.5	96	0.00
22 S	2-Nitrophenol-d4	0.151	0.164	-8.6	101	0.00
23 C	2-Nitrophenol	0.166	0.176	-6.0	99	0.00
24	2,4-Dimethylphenol	0.383	0.385	-0.5	93	0.00
25	Bis(2-Chloroethoxy)methane	0.391	0.392	-0.3	95	0.00
26 S	2,4-Dichlorophenol-d3	0.324	0.345	-6.5	101	0.00
27 C	2,4-Dichlorophenol	0.314	0.331	-5.4	99	0.00
28	Naphthalene	1.010	1.017	-0.7	95	0.00
29 S	4-Chloroaniline-d4	0.330	0.405	-22.7#	114	0.00
30	4-Chloroaniline	0.335	0.411	-22.7#	113	0.00
31 C	Hexachlorobutadiene	0.241	0.256	-6.2	101	0.00
32	Caprolactam	0.094	0.090	4.3	95	0.00
33 C	4-Chloro-3-methylphenol	0.335	0.342	-2.1	95	0.00
34	2-Methylnaphthalene	0.730	0.734	-0.5	94	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
36	1,2,4,5-Tetrachlorobenzene	0.747	0.777	-4.0	97	0.00
37	Hexachlorocyclopentadiene	0.452	0.421	6.9	90	0.00
38 C	2,4,6-Trichlorophenol	0.435	0.459	-5.5	99	0.00
39	2,4,5-Trichlorophenol	0.476	0.505	-6.1	101	0.00
40	1,1'-Biphenyl	1.632	1.674	-2.6	97	0.00
41	2-Chloronaphthalene	1.280	1.305	-2.0	96	0.00
42	2-Nitroaniline	0.339	0.354	-4.4	96	0.00
43 S	Dimethylphthalate-d6	1.483	1.484	-0.1	93	0.00
44	Dimethylphthalate	1.621	1.627	-0.4	95	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM070115\
 Data File : BM001814.D
 Acq On : 01 Jul 2015 19:02
 Operator : TP/UM
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02017

Quant Time: Jul 02 07:54:17 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061515.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 02 07:09:03 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.305	0.328	-7.5	99	0.00
46 S	Acenaphthylene-d8	1.944	2.023	-4.1	97	0.00
47	Acenaphthylene	1.982	2.030	-2.4	97	0.00
48	3-Nitroaniline	0.285	0.319	-11.9	107	0.00
49 C	Acenaphthene	1.325	1.350	-1.9	95	0.00
50	2,4-Dinitrophenol	0.189	0.166	12.2	94	0.00
51 S	4-Nitrophenol-d4	0.274	0.251	8.4	89	0.00
52	4-Nitrophenol	0.283	0.267	5.7	92	0.00
53	Dibenzofuran	1.897	1.944	-2.5	97	0.00
54	2,4-Dinitrotoluene	0.449	0.468	-4.2	95	0.00
55	2,3,4,6-Tetrachlorophenol	0.411	0.433	-5.4	98	0.00
56	Diethylphthalate	1.541	1.567	-1.7	94	0.00
57 S	Fluorene-d10	1.376	1.410	-2.5	97	0.00
58	Fluorene	1.579	1.624	-2.8	97	0.00
59	4-Chlorophenyl-phenylether	0.851	0.886	-4.1	98	0.00
60	4-Nitroaniline	0.323	0.331	-2.5	97	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.122	0.116	4.9	96	0.00
63	4,6-Dinitro-2-methylphenol	0.130	0.121	6.9	93	0.00
64	N-Nitrosodiphenylamine	0.584	0.597	-2.2	96	0.00
65	4-Bromophenyl-phenylether	0.223	0.232	-4.0	99	0.00
66	Hexachlorobenzene	0.251	0.259	-3.2	98	0.00
67	Atrazine	0.235	0.237	-0.9	96	0.00
68 C	Pentachlorophenol	0.154	0.147	4.5	97	0.00
69	Phenanthrene	1.084	1.098	-1.3	96	0.00
70 S	Anthracene-d10	0.932	0.958	-2.8	96	0.00
71	Anthracene	1.113	1.122	-0.8	94	0.00
72	1,2,3,4-Tetrachlorobenzene	0.314	0.330	-5.1	100	0.00
73	Pentachlorobenzene	0.296	0.309	-4.4	98	0.00
74	Carbazole	0.977	0.972	0.5	94	0.00
75	Di-n-butylphthalate	1.044	1.073	-2.8	96	0.00
76 C	Fluoranthene	1.297	1.290	0.5	95	0.00
77 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
78 S	Pyrene-d10	0.847	0.901	-6.4	94	0.00
79	Pyrene	1.110	1.182	-6.5	94	0.00
80	Butylbenzylphthalate	0.362	0.392	-8.3	96	0.00
81	3,3'-Dichlorobenzidine	0.355	0.386	-8.7	97	0.00
82	Benzo(a)anthracene	1.169	1.215	-3.9	91	0.00
83	Bis(2-ethylhexyl)phthalate	0.544	0.573	-5.3	93	0.00
84	Chrysene	1.108	1.119	-1.0	89	0.00
85 I	Perylene-d12	1.000	1.000	0.0	79	0.00
86	Di-n-octyl phthalate	0.983	0.989	-0.6	90	0.00
87	Benzo(b)fluoranthene	1.164	1.226	-5.3	86	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM070115\
Data File : BM001814.D
Acq On : 01 Jul 2015 19:02
Operator : TP/UM
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02017

Quant Time: Jul 02 07:54:17 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061515.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jul 02 07:09:03 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.125	1.152	-2.4	87	0.00
89 S	Benzo(a)pyrene-d12	0.882	0.909	-3.1	85	0.00
90 C	Benzo(a)pyrene	1.133	1.169	-3.2	85	0.00
91	Indeno(1,2,3-cd)pyrene	1.343	1.378	-2.6	84	0.00
92	Dibenzo(a,h)anthracene	1.133	1.174	-3.6	85	0.00
93	Benzo(g,h,i)perylene	1.129	1.163	-3.0	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0