

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070125\
 Data File : BM050335.D
 Acq On : 01 Jul 2025 20:18
 Operator : RC/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC020EC

Quant Time: Jul 02 10:48:16 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM070125.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 01 15:14:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.503	0.545	-8.3	100	0.00
3 S	1,4-Dioxane-d8	0.485	0.503	-3.7	94	0.00
4 S	Pyridine-d5	1.312	1.407	-7.2	97	0.00
5	Pyridine	1.369	1.495	-9.2	97	0.00
6	Benzaldehyde	0.800	0.967	-20.9	98	0.00
7 S	Phenol-d5	1.450	1.522	-5.0	96	0.00
8	Phenol	1.519	1.630	-7.3	97	0.00
9 S	Bis-(2-Chloroethyl)ether-d8	0.919	1.027	-11.8	100	0.00
10	Bis(2-Chloroethyl)ether	1.193	1.299	-8.9	98	0.00
11 S	2-Chlorophenol-d4	1.181	1.235	-4.6	96	0.00
12	2-Chlorophenol	1.239	1.297	-4.7	96	0.00
13	2-Methylphenol	1.096	1.147	-4.7	96	0.00
14	2,2'-oxybis(1-Chloropropane	1.864	2.095	-12.4	100	0.00
15 S	4-Methylphenol-d8	1.105	1.165	-5.4	96	0.00
16	Acetophenone	1.771	1.976	-11.6	100	0.00
17 P	N-Nitroso-di-n-propylamine	0.919	0.973	-5.9	98	0.00
18	4-Methylphenol	1.191	1.255	-5.4	96	0.00
19	Hexachloroethane	0.511	0.544	-6.5	99	0.00
20 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
21 S	Nitrobenzene-d5	0.138	0.136	1.4	94	0.00
22	Nitrobenzene	0.357	0.374	-4.8	99	0.00
23	Isophorone	0.653	0.667	-2.1	98	0.00
24 S	2-Nitrophenol-d4	0.129	0.123	4.7	93	0.00
25 C	2-Nitrophenol	0.153	0.149	2.6	94	0.00
26	2,4-Dimethylphenol	0.342	0.352	-2.9	98	0.00
27	Bis(2-Chloroethoxy)methane	0.421	0.431	-2.4	97	0.00
28 S	2,4-Dichlorophenol-d3	0.313	0.307	1.9	94	0.00
29 C	2,4-Dichlorophenol	0.312	0.310	0.6	94	0.00
30	Naphthalene	1.029	1.041	-1.2	97	0.00
31 S	4-Chloroaniline-d4	0.416	0.439	-5.5	98	0.00
32	4-Chloroaniline	0.421	0.439	-4.3	97	0.00
33 C	Hexachlorobutadiene	0.219	0.213	2.7	94	0.00
34	Caprolactam	0.091	0.087	4.4	96	0.00
35 C	4-Chloro-3-methylphenol	0.285	0.294	-3.2	98	0.00
36	2-Methylnaphthalene	0.715	0.712	0.4	96	0.00
37	1-Methylnaphthalene	0.708	0.707	0.1	96	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
39	1,2,4,5-Tetrachlorobenzene	0.651	0.621	4.6	93	0.00
40	Hexachlorocyclopentadiene	0.339	0.308	9.1	92	0.00
41 C	2,4,6-Trichlorophenol	0.373	0.369	1.1	95	0.00
42	2,4,5-Trichlorophenol	0.415	0.402	3.1	93	0.00
43	1,1'-Biphenyl	1.473	1.477	-0.3	96	0.00
44	2-Chloronaphthalene	1.180	1.191	-0.9	97	0.00
45	2-Nitroaniline	0.272	0.278	-2.2	97	0.00
46 S	Dimethylphthalate-d6	1.362	1.362	0.0	96	0.00
47	Dimethylphthalate	1.355	1.379	-1.8	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070125\
 Data File : BM050335.D
 Acq On : 01 Jul 2025 20:18
 Operator : RC/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC020EC

Quant Time: Jul 02 10:48:16 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM070125.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 01 15:14:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2,6-Dinitrotoluene	0.221	0.217	1.8	94	0.00
49 S	Acenaphthylene-d8	1.668	1.681	-0.8	97	0.00
50	Acenaphthylene	1.858	1.891	-1.8	97	0.00
51	3-Nitroaniline	0.240	0.245	-2.1	96	0.00
52 C	Acenaphthene	1.202	1.223	-1.7	98	0.00
53	2,4-Dinitrophenol	0.099	0.074	25.3#	95	0.00
54 S	4-Nitrophenol-d4	0.245	0.242	1.2	98	0.00
55	4-Nitrophenol	0.192	0.203	-5.7	98	0.00
56	Dibenzofuran	1.731	1.754	-1.3	97	0.00
57	2,4-Dinitrotoluene	0.330	0.337	-2.1	96	0.00
58	2,3,4,6-Tetrachlorophenol	0.320	0.314	1.9	94	0.00
59	Diethylphthalate	1.274	1.317	-3.4	98	0.00
60 S	Fluorene-d10	1.237	1.253	-1.3	97	0.00
61	Fluorene	1.438	1.460	-1.5	97	0.00
62	4-Chlorophenyl-phenylether	0.735	0.739	-0.5	98	0.00
63	4-Nitroaniline	0.237	0.264	-11.4	98	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
65 S	4,6-Dinitro-2-methylphenol-	0.086	0.068	20.9	92	0.00
66	4,6-Dinitro-2-methylphenol	0.097	0.081	16.5	92	0.00
67	N-Nitrosodiphenylamine	0.615	0.617	-0.3	97	0.00
68	4-Bromophenyl-phenylether	0.218	0.207	5.0	93	0.00
69	Hexachlorobenzene	0.262	0.252	3.8	93	0.00
70	Atrazine	0.210	0.205	2.4	94	0.00
71 C	Pentachlorophenol	0.149	0.137	8.1	94	0.00
72	Phenanthrene	1.138	1.147	-0.8	96	0.00
73 S	Anthracene-d10	0.992	1.007	-1.5	96	0.00
74	Anthracene	1.152	1.167	-1.3	96	0.00
75	1,2,3,4-Tetrachlorobenzene	0.339	0.319	5.9	93	0.00
76	Pentachlorobenzene	0.319	0.310	2.8	95	0.00
77	Carbazole	1.008	1.039	-3.1	96	0.00
78	Di-n-butylphthalate	1.018	1.055	-3.6	95	0.00
79 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
80	Fluoranthene	1.229	1.190	3.2	94	0.00
81 S	Pyrene-d10	1.068	1.051	1.6	94	0.00
82	Pyrene	1.333	1.327	0.5	95	0.00
83	Butylbenzylphthalate	0.376	0.370	1.6	90	0.00
84	3,3'-Dichlorobenzidine	0.363	0.351	3.3	88	0.00
85	Benzo(a)anthracene	1.305	1.310	-0.4	94	0.00
86	Bis(2-ethylhexyl)phthalate	0.588	0.603	-2.6	93	0.00
87	Chrysene	1.234	1.234	0.0	94	0.00
88 I	Perylene-d12	1.000	1.000	0.0	93	0.00
89	Di-n-octyl phthalate	1.001	0.879	12.2	91	0.00
90	Benzo(b)fluoranthene	1.206	1.186	1.7	92	0.00
91	Benzo(k)fluoranthene	1.229	1.242	-1.1	95	0.00
92 S	Benzo(a)pyrene-d12	1.014	1.002	1.2	93	0.00
93 C	Benzo(a)pyrene	1.155	1.172	-1.5	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070125\
Data File : BM050335.D
Acq On : 01 Jul 2025 20:18
Operator : RC/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC020EC

Quant Time: Jul 02 10:48:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM070125.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 01 15:14:38 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
94	Indeno(1,2,3-cd)pyrene	1.428	1.421	0.5	94	0.00
95	Dibenzo(a,h)anthracene	1.186	1.191	-0.4	95	0.00
96	Benzo(g,h,i)perylene	1.148	1.162	-1.2	96	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0