

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM070215\
 Data File : BM001823.D
 Acq On : 02 Jul 2015 17:32
 Operator : TP/UM
 Sample : G2868-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C01M5

Quant Time: Jul 03 02:04:31 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061515.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 03 02:02:31 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	61236	20.00	ng/ul	0.00
18) Naphthalene-d8	10.45	136	251846	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	155316	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.06	188	366941	20.00	ng/ul	0.00
77) Chrysene-d12	21.26	240	428798	20.00	ng/ul	0.00
85) Perylene-d12	23.47	264	433145	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.14	96	1495	1.16	ng/uL	0.00
5) Phenol-d5	6.83	99	25267	5.37	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.00	67	69122	26.24	ng/ul	0.00
9) 2-Chlorophenol-d4	7.19	132	79537	21.38	ng/ul	0.00
13) 4-Methylphenol-d8	8.36	113	48744	12.91	ng/ul	0.00
19) Nitrobenzene-d5	8.82	128	46917	26.91	ng/ul	0.00
22) 2-Nitrophenol-d4	9.54	143	52760	27.68	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.07	165	98197	24.06	ng/ul	0.00
29) 4-Chloroaniline-d4	10.59	131	5629	1.36	ng/ul	0.00
43) Dimethylphthalate-d6	13.73	166	302920	26.30	ng/ul	0.00
46) Acenaphthylene-d8	14.01	160	402792	26.68	ng/ul	0.00
51) 4-Nitrophenol-d4	14.52	143	8203	3.86	ng/ul	0.00
57) Fluorene-d10	15.31	176	294538	27.56	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.43	200	50143	22.37	ng/ul	0.00
70) Anthracene-d10	17.16	188	458080	26.80	ng/ul	0.00
78) Pyrene-d10	19.46	212	518396	28.54	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.33	264	491087	25.72	ng/ul	0.00


Target Compounds					Qvalue	
2) 1,4-Dioxane	2.88	88	4127	2.81	ng/uL#	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM070215\
Data File : BM001823.D
Acq On : 02 Jul 2015 17:32
Operator : TP/UM
Sample : G2868-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C01M5

Quant Time: Jul 03 02:04:31 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061515.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 03 02:02:31 2015
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 2.81 ng/uL
 RT: 2.88 min Scan# 33
 Delta R.T. -0.29 min
 Lab File: BM001823.D
 Acq: 02 Jul 2015 17:32

Instrument :
 BNA_M
 ClientSampleId :
 C01M5

Tgt Ion: 88 Resp: 4127
 Ion Ratio Lower Upper
 88 100
 43 2372.3 24.2 36.4#
 58 29.9 56.2 84.2#

