

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM070216\
 Data File : BM006219.D
 Acq On : 03 Jul 2016 03:08
 Operator : UM/SJ
 Sample : H3819-05
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BDHW0

Quant Time: Jul 03 05:50:07 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM070216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 03 00:59:09 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	326943	20.00	ng/ul	0.00
18) Naphthalene-d8	10.43	136	1562963	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.30	164	976880	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.05	188	2180997	20.00	ng/ul	0.00
75) Chrysene-d12	21.24	240	2012808	20.00	ng/ul	0.00
83) Perylene-d12	23.46	264	1729104	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.16	96	4713	0.62	ng/uL	0.00
5) Phenol-d5	6.82	99	207543	6.78	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.99	67	139421	7.72	ng/ul	0.00
9) 2-Chlorophenol-d4	7.18	132	160198	6.84	ng/ul	0.00
13) 4-Methylphenol-d8	8.35	113	170403	6.88	ng/ul	-0.01
19) Nitrobenzene-d5	8.80	128	83176	6.68	ng/ul	0.00
22) 2-Nitrophenol-d4	9.52	143	90225	6.39	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.05	165	158654	6.20	ng/ul	0.00
29) 4-Chloroaniline-d4	10.57	131	239918	9.03	ng/ul	0.00
43) Dimethylphthalate-d6	13.71	166	578158	7.09	ng/ul	0.00
46) Acenaphthylene-d8	13.99	160	695145	6.89	ng/ul	0.00
51) 4-Nitrophenol-d4	14.51	143	77476	5.01	ng/ul	-0.01
57) Fluorene-d10	15.29	176	498821	7.10	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.43	200	51418	4.02	ng/ul	0.00
70) Anthracene-d10	17.15	188	759642	7.30	ng/ul	0.00
76) Pyrene-d10	19.44	212	806655	8.40	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.32	264	590091	7.35	ng/ul	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM070216\
Data File : BM006219.D
Acq On : 03 Jul 2016 03:08
Operator : UM/SJ
Sample : H3819-05
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BDHW0

Quant Time: Jul 03 05:50:07 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM070216.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Jul 03 00:59:09 2016
Response via : Initial Calibration

