

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM070220\
 Data File : BM026648.D
 Acq On : 02 Jul 2020 11:19
 Operator : JU/CG
 Sample : PB129849BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK49

Quant Time: Jul 02 12:07:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM063020MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 02 11:21:37 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.35	152	85189	20.00	ng/ul	0.00
18) Naphthalene-d8	10.11	136	338912	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.01	164	200792	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.77	188	429423	20.00	ng/ul	0.00
78) Chrysene-d12	20.99	240	377172	20.00	ng/ul	0.00
86) Perylene-d12	23.06	264	391776	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	2.96	96	19913	7.02	ng/uL	0.02
5) Phenol-d5	6.56	99	306566	39.29	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.71	67	202047	38.29	ng/ul	0.00
9) 2-Chlorophenol-d4	6.90	132	238639	39.81	ng/ul	0.00
13) 4-Methylphenol-d8	8.07	113	234914	37.65	ng/ul	0.00
19) Nitrobenzene-d5	8.50	128	102595	38.28	ng/ul	0.00
22) 2-Nitrophenol-d4	9.22	143	116780	40.86	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.75	165	210481	37.61	ng/ul	0.00
29) 4-Chloroaniline-d4	10.26	131	280498	47.60	ng/ul	0.00
44) Dimethylphthalate-d6	13.44	166	623910	38.95	ng/ul	0.00
47) Acenaphthylene-d8	13.70	160	723225	37.37	ng/ul	0.00
52) 4-Nitrophenol-d4	14.26	143	63595	27.51	ng/ul	0.00
58) Fluorene-d10	15.02	176	526127	37.51	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.17	200	25157	9.38	ng/ul	0.00
71) Anthracene-d10	16.87	188	779469	36.86	ng/ul	0.00
79) Pyrene-d10	19.17	212	817980	41.73	ng/ul	0.00
90) Benzo(a)pyrene-d12	22.93	264	806770	38.16	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM070220\
Data File : BM026648.D
Acq On : 02 Jul 2020 11:19
Operator : JU/CG
Sample : PB129849BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK49

Quant Time: Jul 02 12:07:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM063020MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jul 02 11:21:37 2020
Response via : Initial Calibration

