

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM070316\
 Data File : BM006227.D
 Acq On : 03 Jul 2016 10:10
 Operator : UM/SJ
 Sample : PB91683BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK83

Quant Time: Jul 05 04:38:06 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM070216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 04 01:54:46 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	237335	20.00	ng/ul	0.00
18) Naphthalene-d8	10.43	136	1139540	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.30	164	669859	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.04	188	1501957	20.00	ng/ul	0.00
75) Chrysene-d12	21.24	240	1420688	20.00	ng/ul	0.00
83) Perylene-d12	23.46	264	1154589	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.16	96	22185	4.00	ng/uL	0.00
5) Phenol-d5	6.82	99	919702	41.39	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.99	67	602130	45.92	ng/ul	0.00
9) 2-Chlorophenol-d4	7.18	132	699035	41.09	ng/ul	0.00
13) 4-Methylphenol-d8	8.36	113	746921	41.53	ng/ul	0.00
19) Nitrobenzene-d5	8.80	128	376907	41.51	ng/ul	0.00
22) 2-Nitrophenol-d4	9.52	143	424641	41.26	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.05	165	700052	37.53	ng/ul	0.00
29) 4-Chloroaniline-d4	10.57	131	1058365	54.61	ng/ul	0.00
43) Dimethylphthalate-d6	13.71	166	2299863	41.14	ng/ul	0.00
46) Acenaphthylene-d8	13.99	160	2822040	40.77	ng/ul	0.00
51) 4-Nitrophenol-d4	14.52	143	428599	40.45	ng/ul	0.00
57) Fluorene-d10	15.30	176	1888797	39.19	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.43	200	335652	38.14	ng/ul	0.00
70) Anthracene-d10	17.14	188	2864954	39.98	ng/ul	0.00
76) Pyrene-d10	19.45	212	3008625	44.41	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.32	264	2283620	42.57	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM070316\
Data File : BM006227.D
Acq On : 03 Jul 2016 10:10
Operator : UM/SJ
Sample : PB91683BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK83

Quant Time: Jul 05 04:38:06 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM070216.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 04 01:54:46 2016
Response via : Initial Calibration

