



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070325\
Data File : BM050345.D
Acq On : 03 Jul 2025 08:29
Operator : RC/JU
Sample : DFTPP091
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP091

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM070325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jul 03 12:56:29 2025



AutoFind: Scans 2496, 2497, 2498; Background Corrected with Scan 2489

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.0	279692	PASS
68	69	0.00	2	1.5	4008	PASS
69	198	0.00	100	36.1	265301	PASS
70	69	0.00	2	0.6	1495	PASS
127	198	10	80	42.0	309355	PASS
197	198	0.00	2	0.3	2329	PASS
198	198	100	100	100.0	735851	PASS
199	198	5	9	6.9	50635	PASS
275	198	10	60	25.1	184629	PASS
365	198	1	100	3.2	23285	PASS
441	198	0.01	100	10.3	75656	PASS
442	198	50	100	67.0	493056	PASS
443	442	15	24	19.6	96744	PASS