


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM070517\
 Data File : BM010782.D
 Acq On : 05 Jul 2017 10:02
 Operator : SJ/JU
 Sample : DFTPP22
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP22

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM042517.M
 Title : SVOA CALIBRATION
 Last Update : Sat May 27 06:06:59 2017



AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2450

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.2	39954	PASS
68	69	0.00	2	1.7	923	PASS
69	198	0.00	100	37.5	55085	PASS
70	69	0.00	2	0.8	428	PASS
127	198	10	80	43.6	64138	PASS
197	198	0.00	2	0.7	1086	PASS
198	198	100	100	100.0	147058	PASS
199	198	5	9	6.5	9567	PASS
275	198	10	60	27.3	40170	PASS
365	198	1	100	4.5	6556	PASS
441	443	0.01	100	83.9	16258	PASS
442	198	50	100	73.4	107944	PASS
443	442	15	24	17.9	19373	PASS

m/z	Abundance
36.90	362.0
38.00	315.0
39.10	3183.0
49.30	691.0
50.10	12919.0
51.00	45712.0
52.00	2437.0
55.10	555.0
56.00	1205.0
57.00	3588.0
61.00	516.0
62.00	670.0
63.00	2530.0
65.10	940.0
67.10	352.0
68.10	926.0
69.00	61560.0
70.00	559.0
73.10	1312.0
74.10	6988.0
75.00	11098.0
76.00	3512.0
77.10	69400.0
78.00	4754.0
79.00	6632.0
80.00	3829.0
81.00	6213.0
82.00	1677.0
82.90	993.0
83.90	324.0
85.00	1017.0
86.00	1694.0
87.00	532.0
91.00	1644.0
91.90	1575.0
93.00	10943.0
93.90	580.0
95.00	418.0
95.90	366.0
96.90	311.0
98.00	8086.0
99.00	7127.0
100.00	558.0
100.90	3290.0
103.00	1293.0
103.90	2366.0
105.00	1671.0
106.10	439.0
107.00	26256.0
108.00	3756.0
109.00	1067.0
110.00	35680.0
111.00	6285.0
112.00	1279.0
114.80	321.0
116.00	1642.0
117.00	25296.0
117.90	2483.0
119.10	389.0
119.90	391.0
122.00	1411.0
123.00	2553.0
123.90	1415.0
125.00	1185.0
125.80	423.0
126.20	466.0
127.00	66528.0
128.00	4707.0
129.00	29960.0
130.00	3235.0
130.90	658.0
133.00	317.0
133.90	1505.0
134.90	2841.0
136.00	1826.0
137.00	1456.0
138.00	598.0
139.10	399.0
139.90	416.0
140.90	4446.0
141.90	1615.0
143.00	947.0
145.20	427.0
146.00	1024.0
147.00	2745.0
147.90	5906.0
149.00	1312.0
149.90	327.0
150.90	753.0
151.90	600.0
152.90	1442.0

Instrument :
BNA_M
ClientSampleId :
DFTPP22

153.90	1006.0
155.00	2554.0
156.00	3357.0
157.00	603.0
158.00	949.0
158.80	628.0
160.00	1611.0
161.00	1645.0
162.20	692.0
164.90	1572.0
165.90	1171.0
167.00	8961.0
167.90	4805.0
169.10	678.0
169.90	327.0
170.90	492.0
171.80	563.0
172.90	1371.0
173.90	2107.0
175.00	3078.0
176.00	647.0
176.90	1651.0
177.90	647.0
178.90	7146.0
180.00	5194.0
181.00	2018.0
182.00	432.0
183.10	308.0
184.10	772.0
185.00	3280.0
186.00	22680.0
187.00	8209.0
188.00	1034.0
189.10	2420.0
189.90	381.0
191.00	910.0
192.00	2700.0
193.10	2713.0
193.90	902.0
195.10	365.0
195.90	5578.0
197.00	1015.0
197.90	144768.0
198.90	8932.0
200.00	753.0
201.10	577.0
203.00	1438.0
203.90	7500.0
205.00	11425.0
206.00	42744.0
207.10	6237.0
208.00	2451.0
208.90	830.0
210.10	963.0
211.00	2238.0
215.10	737.0
216.00	1563.0
216.90	11142.0
218.00	2114.0
218.80	452.0
220.00	309.0
221.00	4095.0
221.80	1451.0
223.00	3158.0
224.00	25120.0
225.00	6224.0
226.00	632.0
227.00	12104.0
228.00	1723.0
228.90	2196.0
229.90	325.0
231.00	991.0
232.70	311.0
233.90	1001.0
235.10	704.0
235.90	709.0
237.10	929.0
238.80	768.0
240.90	607.0
242.00	1310.0
243.10	1375.0
244.10	16456.0
245.10	2056.0
246.00	4770.0
246.80	808.0
249.00	1054.0
254.10	1064.0
255.00	95544.0
256.00	14109.0
257.00	1158.0
258.00	5554.0
259.10	1374.0
264.10	304.0
264.90	3752.0
266.00	522.0

Instrument :
BNA_M
ClientSampleId :
DFTPP22

273.10	2541.0
274.00	6792.0
275.00	35968.0
276.00	5818.0
276.90	4558.0
278.00	986.0
281.00	355.0
285.00	1120.0
293.00	847.0
293.90	349.0
296.00	14945.0
297.10	1838.0
298.20	371.0
301.80	336.0
302.90	700.0
303.20	745.0
304.00	433.0
314.10	739.0
315.00	1356.0
316.10	552.0
321.10	351.0
321.80	342.0
323.00	3114.0
323.90	861.0
326.90	973.0
334.20	2051.0
335.00	395.0
341.00	533.0
346.00	1224.0
352.00	914.0
352.90	924.0
354.00	561.0
355.00	422.0
365.00	4351.0
365.90	1004.0
372.00	1720.0
372.90	338.0
373.20	320.0
383.10	432.0
401.90	567.0
403.00	746.0
403.90	572.0
421.00	898.0
422.10	647.0
423.10	3610.0
424.00	1155.0
441.00	11616.0
442.10	78216.0
443.10	13583.0
444.10	1677.0

Instrument :
BNA_M
ClientSampleId :
DFTPP22

m/z	Abundance
37.20	378.0
38.20	741.0
39.10	3095.0
48.90	397.0
50.10	15133.0
51.00	47152.0
52.10	2689.0
54.90	433.0
55.90	1163.0
57.00	3357.0
60.90	557.0
62.00	930.0
63.10	3096.0
65.00	1145.0
68.00	1011.0
69.00	65672.0
70.10	300.0
73.10	1883.0
74.10	7799.0
75.10	11856.0
76.10	3621.0
77.10	73568.0
78.10	5234.0
79.00	5724.0
80.00	4820.0
81.00	6088.0
82.00	1661.0
83.00	1787.0
85.00	1154.0
86.00	2011.0
87.00	571.0
88.90	426.0
90.90	1496.0
92.00	1367.0
93.00	10355.0
94.10	891.0
96.10	309.0
96.90	374.0
98.00	9406.0
99.00	6673.0
100.00	695.0
101.00	3385.0
102.90	1419.0
104.00	2021.0
105.00	2260.0
105.90	434.0
107.00	28208.0
108.00	3373.0
109.10	726.0
110.00	38848.0
111.00	7232.0
112.00	868.0
112.90	583.0
115.90	1724.0
117.00	30456.0
118.00	1896.0
120.00	387.0
121.00	305.0
122.00	2812.0
123.00	2662.0
123.90	937.0
124.90	899.0
126.20	541.0
127.00	77736.0
128.00	5101.0
129.00	34728.0
130.00	2540.0
130.90	1246.0
133.10	329.0
134.00	1249.0
135.00	3108.0
135.80	940.0
137.00	1346.0
137.90	390.0
139.00	542.0
139.90	896.0
141.00	5556.0
142.00	1769.0
142.80	1076.0
144.00	437.0
145.10	304.0
145.90	1403.0
147.00	3133.0
148.00	7075.0
148.80	1492.0
150.90	824.0
151.90	447.0
153.00	1664.0
153.90	1386.0
155.10	2857.0
156.00	5027.0

Instrument :
BNA_M
ClientSampleId :
DFTPP22

156.90	776.0
157.90	969.0
159.20	724.0
160.00	1653.0
160.90	1810.0
162.00	825.0
163.80	338.0
165.00	2124.0
166.00	2659.0
167.00	11970.0
168.00	6460.0
169.00	536.0
169.90	316.0
170.90	707.0
171.90	910.0
173.00	1005.0
174.10	2186.0
175.00	4803.0
176.10	1040.0
176.90	1779.0
178.10	846.0
179.00	8631.0
180.00	5110.0
180.90	2696.0
182.00	418.0
183.00	401.0
184.00	815.0
184.90	4780.0
186.00	27264.0
187.00	8339.0
188.10	808.0
189.00	2023.0
190.80	779.0
192.00	2628.0
193.00	2845.0
194.00	674.0
196.00	3608.0
197.00	1606.0
197.90	176064.0
198.90	11562.0
199.90	1146.0
201.30	546.0
201.90	517.0
203.00	1716.0
204.00	8261.0
205.00	13855.0
206.00	53264.0
207.10	8047.0
208.00	1978.0
208.90	1212.0
210.30	918.0
211.00	2577.0
215.00	945.0
216.00	1083.0
216.90	17256.0
217.90	2091.0
219.00	414.0
221.00	5834.0
222.00	2552.0
222.90	3838.0
224.00	29952.0
225.00	7458.0
225.90	863.0
227.00	16504.0
228.00	2442.0
229.00	2806.0
230.10	337.0
230.90	1160.0
234.00	1320.0
234.90	1145.0
235.90	927.0
236.90	1776.0
237.90	348.0
238.90	474.0
240.00	607.0
240.80	702.0
241.00	706.0
242.10	1348.0
243.00	1613.0
244.00	22688.0
245.00	3183.0
246.00	5744.0
247.00	828.0
248.00	353.0
248.90	616.0
253.00	658.0
254.10	1391.0
255.00	125528.0
256.00	19888.0
257.00	1133.0
258.00	8845.0
259.00	1583.0
261.10	439.0
265.00	3501.0
266.00	697.0

Instrument :
BNA_M
ClientSampleId :
DFTPP22

267.00	328.0
271.20	302.0
272.00	830.0
273.00	3739.0
274.00	9798.0
275.00	48008.0
276.00	6103.0
277.00	5248.0
277.90	659.0
282.90	604.0
284.00	472.0
285.00	829.0
289.00	375.0
293.00	1455.0
294.00	321.0
295.10	326.0
296.00	20056.0
297.00	3294.0
302.00	616.0
303.00	2574.0
303.90	1028.0
314.20	620.0
315.00	2658.0
316.00	865.0
321.00	649.0
321.90	420.0
323.00	5155.0
323.90	641.0
327.00	801.0
328.00	472.0
331.90	459.0
332.90	415.0
334.00	3625.0
335.20	578.0
342.10	329.0
345.90	1020.0
347.00	303.0
352.00	1208.0
353.00	1353.0
354.00	1336.0
365.00	7826.0
365.90	1170.0
370.70	355.0
372.10	2455.0
373.00	367.0
383.00	673.0
402.00	846.0
403.10	835.0
404.10	362.0
421.00	1189.0
422.00	831.0
423.10	6474.0
424.10	1549.0
425.00	333.0
440.10	418.0
441.10	19144.0
442.10	128280.0
443.10	24824.0
444.00	1927.0

Instrument :
BNA_M
ClientSampleId :
DFTPP22

m/z	Abundance
38.00	454.0
39.10	1803.0
40.00	436.0
49.00	486.0
50.00	8735.0
51.10	27000.0
52.00	1677.0
56.00	1108.0
57.00	2102.0
61.00	799.0
62.00	526.0
63.10	1637.0
65.10	820.0
68.00	833.0
69.00	38024.0
70.00	425.0
73.00	1519.0
74.10	4443.0
75.00	7108.0
76.10	2062.0
77.10	44944.0
78.00	2871.0
79.00	3164.0
80.00	2316.0
80.90	3128.0
81.90	829.0
83.10	793.0
85.00	735.0
86.10	1376.0
87.10	406.0
90.90	1089.0
92.00	833.0
93.00	6902.0
93.90	353.0
97.20	409.0
98.00	5977.0
99.00	4295.0
100.00	349.0
101.10	1577.0
103.00	785.0
104.00	1343.0
104.90	1419.0
106.00	356.0
107.00	18056.0
108.00	2296.0
109.10	643.0
110.00	23168.0
111.00	3728.0
111.90	577.0
116.10	963.0
117.00	19672.0
118.00	1159.0
119.00	316.0
122.00	1087.0
122.90	2095.0
123.80	1015.0
125.10	900.0
125.90	490.0
127.00	48152.0
128.00	3224.0
129.00	22048.0
129.90	1789.0
130.90	321.0
133.90	752.0
135.10	1985.0
136.00	887.0
137.00	920.0
140.90	4018.0
141.90	895.0
143.00	783.0
146.10	818.0
146.90	2068.0
148.00	4293.0
148.90	770.0
151.00	414.0
151.90	470.0
152.90	1333.0
154.10	812.0
155.00	1802.0
156.00	3140.0
157.00	754.0
157.80	353.0
158.90	566.0
160.10	1541.0
160.90	2381.0
161.90	389.0
164.80	1502.0
166.00	962.0
167.00	7118.0
168.00	3879.0
169.00	566.0

Instrument :
BNA_M
ClientSampleId :
DFTPP22

170.00	493.0
172.00	733.0
173.10	1157.0
173.90	1075.0
175.00	3322.0
175.80	716.0
176.20	463.0
176.90	1141.0
177.90	960.0
179.00	4726.0
179.90	2673.0
181.00	1319.0
184.00	625.0
185.00	3356.0
186.00	20400.0
187.10	5151.0
188.00	485.0
189.00	1309.0
189.90	401.0
191.00	804.0
192.00	2216.0
193.00	2100.0
193.80	747.0
196.10	3133.0
196.70	621.0
197.10	639.0
197.90	120344.0
198.90	8207.0
199.90	665.0
201.20	528.0
201.70	559.0
203.00	2163.0
204.00	6830.0
205.00	8976.0
206.00	36944.0
207.00	5134.0
207.90	1338.0
208.80	329.0
209.10	301.0
209.90	1361.0
211.10	1613.0
216.10	1228.0
217.00	11300.0
217.90	1495.0
219.00	387.0
220.20	414.0
221.00	4155.0
221.90	1187.0
222.90	1896.0
224.00	21056.0
225.00	5604.0
225.90	749.0
227.00	11060.0
228.00	1444.0
228.90	1860.0
229.90	330.0
230.90	886.0
234.00	596.0
235.00	866.0
236.10	304.0
237.10	697.0
239.00	619.0
241.10	921.0
241.90	901.0
243.10	1285.0
244.00	15880.0
245.10	1585.0
245.90	3997.0
246.90	923.0
249.00	784.0
253.20	516.0
254.10	843.0
255.00	87024.0
256.10	14638.0
256.90	827.0
258.00	6469.0
259.00	1004.0
265.10	2311.0
266.10	405.0
273.00	2693.0
274.00	6645.0
275.00	36536.0
276.00	4705.0
277.00	3779.0
277.90	750.0
281.10	458.0
283.00	361.0
285.00	668.0
292.90	1307.0
295.00	454.0
296.00	13501.0
297.00	2574.0
303.10	1314.0
304.00	386.0
307.90	540.0

Instrument :

BNA_M

ClientSampleId :

DFTPP22

314.00	691.0
314.90	1882.0
316.10	781.0
320.90	340.0
323.00	4296.0
324.20	370.0
327.00	530.0
328.00	670.0
332.00	414.0
332.90	500.0
334.00	3256.0
335.00	703.0
341.10	666.0
345.80	904.0
351.90	1376.0
353.00	686.0
354.00	1139.0
355.10	429.0
365.00	7492.0
366.00	1447.0
370.90	322.0
372.10	1974.0
373.00	459.0
383.00	408.0
390.10	433.0
402.00	927.0
403.10	976.0
404.00	724.0
420.90	657.0
422.10	832.0
423.10	6215.0
424.00	974.0
441.10	18016.0
442.10	117336.0
443.10	19712.0
444.10	2223.0

Instrument :

BNA_M

ClientSampleId :

DFTPP22

Scan 2450 (17.298 min): BM010782.D
DFTPP22

m/z	Abundance
48.90	366.0
73.10	895.0
147.00	482.0
193.00	486.0
194.00	438.0
206.90	1141.0
209.00	518.0
221.10	341.0
267.00	302.0
281.00	426.0

Instrument :
BNA_M
ClientSampleId :
DFTPP22