

Data Path : Z:\HPCHEM1\BNA M\DATA\BM070615\
 Data File : BM001875.D
 Acq On : 07 Jul 2015 22:11
 Operator : TP/UM
 Sample : G2861-10RE
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Jul 08 05:35:36 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM061515.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 08 05:33:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	67339	20.00	ng/ul	0.00
18) Naphthalene-d8	10.46	136	236985	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.33	164	135174	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.07	188	299759	20.00	ng/ul	0.00
77) Chrysene-d12	21.27	240	368627	20.00	ng/ul	0.00
85) Perylene-d12	23.50	264	383846	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.15	96	3879	2.75	ng/uL	0.00
5) Phenol-d5	6.83	99	75813	14.66	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.00	67	51651	17.83	ng/ul	0.00
9) 2-Chlorophenol-d4	7.20	132	66660	16.30	ng/ul	0.00
13) 4-Methylphenol-d8	8.37	113	30017	7.23	ng/ul	0.00
19) Nitrobenzene-d5	8.83	128	34284	20.90	ng/ul	0.00
22) 2-Nitrophenol-d4	9.55	143	36151	20.15	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.08	165	62601	16.30	ng/ul	0.00
29) 4-Chloroaniline-d4	10.57	131	165673	42.41	ng/ul	-0.03
43) Dimethylphthalate-d6	13.74	166	184851	18.44	ng/ul	0.00
46) Acenaphthylene-d8	14.02	160	244340	18.59	ng/ul	0.00
51) 4-Nitrophenol-d4	14.53	143	32635	17.65	ng/ul	0.00
57) Fluorene-d10	15.32	176	178309	19.17	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.42	200	2061	1.13	ng/ul	-0.04
70) Anthracene-d10	17.17	188	262781	18.82	ng/ul	0.00
78) Pyrene-d10	19.47	212	307020	19.66	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.36	264	316061	18.68	ng/ul	0.00

Target Compounds

						Ovalue
4) Benzaldehyde	6.80	77	3476	1.03	ng/ul#	67
14) Acetophenone	8.49	105	10762	1.63	ng/ul#	49
15) N-Nitroso-di-n-propylamine	8.43	70	5212	1.62	ng/ul#	77
17) Hexachloroethane	8.77	117	32745	19.26	ng/ul#	1
20) Nitrobenzene	8.85	77	8558	1.93	ng/ul#	36
21) Isophorone	9.37	82	31714	4.27	ng/ul#	1
23) 2-Nitrophenol	9.58	139	3838	1.95	ng/ul#	38
24) 2,4-Dimethylphenol	9.65	107	10490	2.31	ng/ul#	43
28) Naphthalene	10.51	128	49426	4.13	ng/ul#	76
32) Caprolactam	11.39	113	3715	3.34	ng/ul#	1
34) 2-Methylnaphthalene	12.13	142	16801	1.94	ng/ul	93
42) 2-Nitroaniline	13.42	65	3034	1.32	ng/ul#	9
44) Dimethylphthalate	13.79	163	42769	3.90	ng/ul#	96
47) Acenaphthylene	14.05	152	32137	2.40	ng/ul#	19
49) Acenaphthene	14.39	153	87020	9.72	ng/ul	95
52) 4-Nitrophenol	14.56	109	2657	1.39	ng/ul#	57
53) Dibenzofuran	14.64	168	15092	1.18	ng/ul	89
54) 2,4-Dinitrotoluene	14.69	165	5868	1.93	ng/ul#	77
58) Fluorene	15.38	166	111126	10.41	ng/ul#	95
64) N-Nitrosodiphenylamine	15.49	169	48959	5.59	ng/ul	91
69) Phenanthrene	17.12	178	462452	28.45	ng/ul	96
71) Anthracene	17.12	178	461594	27.66	ng/ul	97
76) Fluoranthene	19.14	202	167632	8.62	ng/ul	98

Data Path : Z:\HPCHEM1\BNA M\DATA\BM070615\
Data File : BM001875.D
Acq On : 07 Jul 2015 22:11
Operator : TP/UM
Sample : G2861-10RE
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Jul 08 05:35:36 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM061515.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 08 05:33:09 2015
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
79) Pyrene	19.50	202	259284	12.68	ng/ul	98
82) Benzo(a)anthracene	21.25	228	114165	5.30	ng/ul#	92
84) Chrysene	21.30	228	112607	5.52	ng/ul#	91
87) Benzo(b)fluoranthene	22.82	252	189092	8.46	ng/ul#	97
88) Benzo(k)fluoranthene	22.82	252	74935	3.47	ng/ul#	96
90) Benzo(a)pyrene	23.40	252	186860	8.60	ng/ul	95
91) Indeno(1,2,3-cd)pyrene	25.74	276	178877	6.94	ng/ul#	95
92) Dibenzo(a,h)anthracene	25.74	278	39106	1.80	ng/ul#	80
93) Benzo(g,h,i)perylene	26.44	276	168789	7.79	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM070615\
Data File : BM001875.D
Acq On : 07 Jul 2015 22:11
Operator : TP/UM
Sample : G2861-10RE
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Jul 08 05:35:36 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM061515.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 08 05:33:09 2015
Response via : Initial Calibration





#4

Benzaldehyde

Concen: 1.03 ng/ul

RT: 6.80 min Scan# 700

Delta R.T. -0.01 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11

Instrument :

BNA_M

ClientSampled :





Tot Ion: 77 Resp: 3476

Ion Ratio Lower Upper

77 100

105 95.4 79.3 118.9

106 30.1 72.8 109.2#



#14

Acetophenone

Concen: 1.63 ng/ul

RT: 8.49 min Scan# 987

Delta R.T. 0.01 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11




Tgt Ion: 105 Resp: 10762


Ion Ratio Lower Upper

105 100

77 125.9 64.2 96.4#

51 54.3 22.0 33.0#





#15

N-Nitroso-di-n-propylamine

Concen: 1.62 ng/ul

RT: 8.43 min Scan# 976

Delta R.T. -0.05 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11

Instrument :

BNA_M

ClientSampleId :

Tot Ion: 70 Resp: 5212





Ion Ratio Lower Upper

70 100

42 45.1 45.2 67.8#

101 0.0 7.8 11.6#

130 0.0 15.5 23.3#



#17

Hexachloroethane

Concen: 19.26 ng/ul

RT: 8.77 min Scan# 1034

Delta R.T. 0.04 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11




Tgt Ion: 117 Resp: 32745


Ion Ratio Lower Upper

117 100

201 0.0 92.1 138.1#

199 0.0 60.6 90.8#









#20
Nitrobenzene
Concen: 1.93 ng/ul
RT: 8.85 min Scan# 1048
Delta R.T. -0.02 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11




Instrument :
BNA_M
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
77	100		
123	17.8	34.0	51.0#
65	71.3	11.3	16.9#



#21
Isophorone
Concen: 4.27 ng/ul
RT: 9.37 min Scan# 1137
Delta R.T. -0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11

Tgt Ion	Ratio	Lower	Upper
82	100		
95	187.8	6.3	9.5#
138	11.5	14.9	22.3#









#23
2-Nitrophenol
Concen: 1.95 ng/ul
RT: 9.58 min Scan# 1172
Delta R.T. 0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11




Instrument :
BNA_M
ClientSampled :


Tot Ion:139 Resp: 3838
Ion Ratio Lower Upper
139 100
65 66.4 45.4 68.0
109 120.6 31.6 47.4#



#24
2,4-Dimethylphenol
Concen: 2.31 ng/ul
RT: 9.65 min Scan# 1184
Delta R.T. 0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11

Tgt Ion:107 Resp: 10490
Ion Ratio Lower Upper
107 100
121 24.8 37.4 56.2#
122 13.5 59.2 88.8#









#28
Naphthalene
Concen: 4.13 ng/ul
RT: 10.51 min Scan# 1330
Delta R.T. 0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11




Instrument :
BNA_M
ClientSampleId :


Tot Ion:128 Resp: 49426
Ion Ratio Lower Upper
128 100
129 24.0 8.7 13.1#
127 18.3 10.0 15.0#



#32
Caprolactam
Concen: 3.34 ng/ul
RT: 11.39 min Scan# 1479
Delta R.T. 0.00 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11

Tgt Ion:113 Resp: 3715
Ion Ratio Lower Upper
113 100
55 1818.0 123.3 184.9#
56 606.1 98.6 147.8#









#34
2-Methylnaphthalene
Concen: 1.94 ng/ul
RT: 12.13 min Scan# 1606
Delta R.T. 0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11




Instrument :
BNA_M
ClientSampled :


Tot Ion:142 Resp: 16801
Ion Ratio Lower Upper
142 100
141 96.5 71.6 107.4



#42
2-Nitroaniline
Concen: 1.32 ng/ul
RT: 13.42 min Scan# 1824
Delta R.T. 0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11

Tgt Ion: 65 Resp: 3034
Ion Ratio Lower Upper
65 100
92 153.1 52.1 78.1#
138 22.3 78.8 118.2#









#44
Dimethylphthalate
Concen: 3.90 ng/ul
RT: 13.79 min Scan# 1887
Delta R.T. 0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11




Instrument :
BNA_M
ClientSampleId :


Tot Ion:163 Resp: 42769
Ion Ratio Lower Upper
163 100
194 6.3 3.6 5.4#
164 11.0 8.0 12.0



#47
Acenaphthylene
Concen: 2.40 ng/ul
RT: 14.05 min Scan# 1932
Delta R.T. 0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11

Tgt Ion:152 Resp: 32137
Ion Ratio Lower Upper
152 100
151 38.2 15.9 23.9#
153 70.0 10.0 15.0#





#49

Acenaphthene

Concen: 9.72 ng/ul

RT: 14.39 min Scan# 1990

Delta R.T. 0.01 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11

Instrument :

BNA_M

ClientSampleId :





Tot Ion:153 Resp: 87020

Ion Ratio Lower Upper

153 100

152 46.5 36.6 55.0

154 84.0 72.6 108.8



#52

4-Nitrophenol

Concen: 1.39 ng/ul

RT: 14.56 min Scan# 2018

Delta R.T. 0.01 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11




Tgt Ion:109 Resp: 2657


Ion Ratio Lower Upper

109 100

139 133.7 81.7 122.5#

65 59.2 94.3 141.5#





#53

Dibenzofuran

Concen: 1.18 ng/ul

RT: 14.64 min Scan# 2032

Delta R.T. -0.08 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11

Instrument :

BNA_M





ClientSampled :

Tot Ion:168 Resp: 15092

Ion Ratio Lower Upper

168 100

139 30.8 30.1 45.1



#54

2,4-Dinitrotoluene

Concen: 1.93 ng/ul

RT: 14.69 min Scan# 2040

Delta R.T. -0.01 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11




Tgt Ion:165 Resp: 5868


Ion Ratio Lower Upper

165 100

63 43.0 35.4 53.2

182 130.8 2.1 3.1#





#58

Fluorene

Concen: 10.41 ng/ul

RT: 15.38 min Scan# 2158

Delta R.T. 0.01 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11

Instrument :

BNA_M

ClientSampleId :





Tot Ion:166 Resp: 111126

Ion Ratio Lower Upper

166 100

165 99.9 77.4 116.2

167 20.5 11.3 16.9#



#64

N-Nitrosodiphenylamine

Concen: 5.59 ng/ul

RT: 15.49 min Scan# 2176

Delta R.T. -0.10 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11




Tgt Ion:169 Resp: 48959


Ion Ratio Lower Upper

169 100

168 55.7 52.9 79.3

167 33.9 27.5 41.3





#69

Phenanthrene

Concen: 28.45 ng/ul

RT: 17.12 min Scan# 2453

Delta R.T. 0.00 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11

Instrument :

BNA_M

ClientSampleId :


Tot Ion:178 Resp: 462452

Ion Ratio Lower Upper

178 100

179 17.3 11.9 17.9




176 18.0 15.2 22.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 27.66 ng/ul

RT: 17.12 min Scan# 2453

Delta R.T. -0.09 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11


Tgt Ion:178 Resp: 461594

Ion Ratio Lower Upper

178 100

179 17.3 11.8 17.8



176 18.0 14.4 21.6




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E









#76
 Fluoranthene
 Concen: 8.62 ng/ul
 RT: 19.14 min Scan# 2797
 Delta R.T. 0.01 min
 Lab File: BM001875.D
 Acq: 07 Jul 2015 22:11




Instrument :
 BNA_M
 ClientSampleId :


Tot Ion:202 Resp: 167632
 Ion Ratio Lower Upper
 202 100
 101 8.3 6.1 9.1
 100 5.7 4.8 7.2



#79
 Pyrene
 Concen: 12.68 ng/ul
 RT: 19.50 min Scan# 2859
 Delta R.T. 0.01 min
 Lab File: BM001875.D
 Acq: 07 Jul 2015 22:11

Tgt Ion:202 Resp: 259284
 Ion Ratio Lower Upper
 202 100
 101 8.3 6.9 10.3
 100 6.6 6.2 9.2





#82

Benzo(a)anthracene

Concen: 5.30 ng/ul

RT: 21.25 min Scan# 3156

Delta R.T. 0.00 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11

Instrument :

BNA_M

ClientSampleId :


Tot Ion:228 Resp: 114165

Ion Ratio Lower Upper

228 100

229 24.1 15.4 23.0#




226 28.5 20.6 30.8



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 229.00 (228.70 to 229.70): E

Ion 226.00 (225.70 to 226.70): E



#84

Chrysene

Concen: 5.52 ng/ul

RT: 21.30 min Scan# 3165

Delta R.T. 0.00 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11


Tgt Ion:228 Resp: 112607

Ion Ratio Lower Upper

228 100

226 31.1 23.0 34.4



229 26.3 15.4 23.2#




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#87

Benzo(b)fluoranthene

Concen: 8.46 ng/ul

RT: 22.82 min Scan# 3423

Delta R.T. -0.01 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11

Instrument :

BNA_M

ClientSampled :


Tot Ion:252 Resp: 189092

Ion Ratio Lower Upper

252 100

253 23.1 17.8 26.6




125 8.8 5.3 7.9#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 3.47 ng/ul

RT: 22.82 min Scan# 3423

Delta R.T. -0.05 min

Lab File: BM001875.D

Acq: 07 Jul 2015 22:11


Tgt Ion:252 Resp: 74935

Ion Ratio Lower Upper

252 100

253 23.1 17.1 25.7



125 8.8 5.3 7.9#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#90
Benzo(a)pyrene
Concen: 8.60 ng/ul
RT: 23.40 min Scan# 3521
Delta R.T. -0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11




Instrument :
BNA_M
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
252	100		
253	24.5	17.4	26.0
125	8.8	6.0	9.0



#91
Indeno(1,2,3-cd)pyrene
Concen: 6.94 ng/ul
RT: 25.74 min Scan# 3920
Delta R.T. -0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11

Tgt Ion	Ratio	Lower	Upper
276	100		
138	13.5	14.7	22.1#
277	24.0	19.7	29.5









#92
Dibenzo(a,h)anthracene
Concen: 1.80 ng/ul
RT: 25.74 min Scan# 3920
Delta R.T. -0.02 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11

Instrument :
BNA_M
ClientSampleId :

Tot Ion:278 Resp: 39106
Ion Ratio Lower Upper
278 100
139 20.7 9.2 13.8#
279 32.5 19.1 28.7#



#93
Benzo(g,h,i)perylene
Concen: 7.79 ng/ul
RT: 26.44 min Scan# 4038
Delta R.T. -0.01 min
Lab File: BM001875.D
Acq: 07 Jul 2015 22:11

Tgt Ion:276 Resp: 168789
Ion Ratio Lower Upper
276 100
138 15.7 12.6 19.0
277 25.6 18.7 28.1

