

Data Path : Z:\HPCHEM1\BNA M\DATA\BM070716\
 Data File : BM006283.D
 Acq On : 06 Jul 2016 15:24
 Operator : UM/SJ
 Sample : SSTD0.837
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.837

Manual Integrations

umangi
 7/7/2016 6:03:42 PM

Quant Time: Jul 06 16:04:46 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM070716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 06 15:50:37 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	5303	0.40	ng/ul	0.00
4) Naphthalene-d8	10.43	136	20835	0.40	ng/ul	0.00
8) Acenaphthene-d10	14.30	164	11498	0.40	ng/ul	0.00
12) Phenanthrene-d10	17.04	188	25550	0.40	ng/ul	0.00
18) Chrysene-d12	21.24	240	19398	0.40	ng/ul	0.00
22) Perylene-d12	23.45	264	14414	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	3.17	96	8470	3.28	ng/uL	0.00
6) 2-Methylnaphthalene-d10	12.02	152	19634	0.75	ng/ul	0.00
16) Fluoranthene-d10	19.08	212	44783	0.83	ng/ul	0.00
Target Compounds				Ovalue		
3) 1,4-Dioxane	3.20	88	10510	2.93	ng/ul#	1
5) Naphthalene	10.48	128	37551	0.79	ng/ul	99
7) 2-Methylnaphthalene	12.09	142	26292	0.80	ng/ul	100
9) Acenaphthylene	14.01	152	30069	0.58	ng/ul	93
10) Acenaphthene	14.35	153	29608	0.81	ng/ul	90
11) Fluorene	15.34	166	34212	0.97	ng/ul#	82
13) Pentachlorophenol	16.71	266	2751	1.53	ng/ul	95
14) Phenanthrene	17.07	178	53944m	0.89	ng/ul	
15) Anthracene	17.18	178	51515	0.69	ng/ul	98
17) Fluoranthene	19.10	202	64744	0.85	ng/ul	96
19) Pyrene	19.46	202	61770	0.93	ng/ul	98
20) Benzo(a)anthracene	21.22	228	48377	1.07	ng/ul	97
21) Chrysene	21.28	228	47552	0.68	ng/ul	97
23) Benzo(b)fluoranthene	22.79	252	43076	1.07	ng/ul	96
24) Benzo(k)fluoranthene	22.83	252	36647	0.73	ng/ul	96
25) Benzo(a)pyrene	23.35	252	36831	0.82	ng/ul	95
26) Indeno(1,2,3-cd)pyrene	25.66	276	37858	0.73	ng/ul#	92
27) Dibenzo(a,h)anthracene	25.68	278	29989	0.75	ng/ul	97
28) Benzo(g,h,i)perylene	26.34	276	32278	0.71	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM070716\
Data File : BM006283.D
Acq On : 06 Jul 2016 15:24
Operator : UM/SJ
Sample : SSTD0.837
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.837

Manual Integrations

Quant Time: Jul 06 16:04:46 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM070716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 06 15:50:37 2016
Response via : Initial Calibration

umangi
7/7/2016 6:03:42 PM

