


Data Path : Z:\HPCHEM1\BNA M\DATA\BM070716\
 Data File : BM006278.D
 Acq On : 06 Jul 2016 10:59
 Operator : UM/SJ
 Sample : DFTPP45
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP45

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM070716.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 07 05:51:09 2016



AutoFind: Scans 2533, 2534, 2535; Background Corrected with Scan 2526

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.0	163696	PASS
68	69	0.00	2	1.5	2326	PASS
69	198	0.00	100	39.7	154805	PASS
70	69	0.00	2	0.5	837	PASS
127	198	10	80	46.9	182805	PASS
197	198	0.00	2	0.5	1941	PASS
198	198	100	100	100.0	389824	PASS
199	198	5	9	6.6	25904	PASS
275	198	10	60	25.2	98373	PASS
365	198	1	100	3.1	12078	PASS
441	443	0.01	100	83.0	48162	PASS
442	198	50	100	78.1	304512	PASS
443	442	15	24	19.1	58021	PASS

Scan 2526 (17.445 min): BM006278.D
DFTPP45

m/z	Abundance
40.00	912.0
43.00	330.0
43.90	1602.0
207.00	1153.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

m/z	Abundance
37.00	720.0
38.00	2119.0
39.00	13037.0
40.00	1250.0
40.90	586.0
44.00	1261.0
45.00	467.0
49.00	1681.0
50.00	45848.0
51.00	184832.0
52.00	8953.0
52.90	511.0
55.00	1083.0
56.00	4872.0
57.00	11734.0
58.00	541.0
61.00	2283.0
62.00	2445.0
63.00	7419.0
64.10	1243.0
65.00	3578.0
68.00	2955.0
69.00	170240.0
70.10	757.0
73.10	1249.0
74.00	17008.0
75.00	29072.0
76.00	9863.0
77.00	198272.0
78.00	13547.0
79.00	12713.0
80.00	9758.0
81.00	14247.0
82.00	3709.0
83.00	3341.0
83.90	385.0
85.00	2753.0
86.00	3980.0
87.00	1867.0
88.00	901.0
89.00	312.0
91.00	3206.0
92.00	3301.0
93.00	21848.0
94.00	1298.0
94.90	319.0
95.90	1071.0
97.10	444.0
98.00	16936.0
99.00	13459.0
99.90	1121.0
101.00	7920.0
101.90	398.0
103.00	1960.0
104.00	4894.0
105.00	4058.0
106.00	1375.0
107.00	62088.0
108.00	9641.0
109.00	1374.0
110.00	107880.0
111.00	15124.0
112.00	2659.0
113.00	776.0
116.00	3356.0
117.00	49192.0
118.00	3589.0
119.00	682.0
120.10	721.0
122.00	3861.0
123.00	6258.0
124.00	2189.0
125.00	2075.0
126.10	615.0
127.00	193152.0
128.00	14809.0
129.00	74616.0
130.00	6649.0
130.90	1488.0
132.00	678.0
133.00	375.0
133.90	1945.0
135.00	6060.0
136.00	2254.0
137.00	3593.0
138.00	699.0
139.00	337.0
140.00	960.0
141.00	9385.0
142.00	3328.0
143.00	1989.0
144.00	713.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

144.90	443.0
146.00	1766.0
147.00	5215.0
148.00	11415.0
149.00	2683.0
150.00	603.0
151.00	1317.0
151.80	589.0
153.00	2949.0
154.00	2384.0
155.00	5979.0
156.10	7793.0
157.00	1581.0
158.00	1502.0
159.00	1288.0
160.00	3269.0
161.00	5337.0
161.90	1221.0
163.90	379.0
165.00	3427.0
166.00	2850.0
167.00	20840.0
168.00	10539.0
169.00	1603.0
170.00	582.0
170.90	792.0
172.00	1698.0
172.90	1803.0
174.00	4852.0
175.00	7720.0
176.10	1997.0
177.00	3529.0
178.00	1089.0
179.00	14162.0
180.00	10535.0
181.00	4914.0
182.10	762.0
182.90	513.0
184.00	1292.0
185.00	7334.0
186.00	54360.0
187.00	15976.0
188.00	1829.0
189.00	4389.0
189.90	522.0
191.00	1952.0
192.00	4405.0
193.10	5268.0
193.90	1008.0
195.10	544.0
196.00	9780.0
197.00	2199.0
198.00	393152.0
199.00	27056.0
200.00	1969.0
201.40	1524.0
203.00	2823.0
204.00	14489.0
205.00	23440.0
206.00	100768.0
207.10	14878.0
208.00	3636.0
209.00	1255.0
210.00	1265.0
211.00	3602.0
211.70	1102.0
213.10	383.0
214.90	1048.0
216.00	2356.0
217.00	28984.0
218.00	3374.0
220.30	1042.0
221.00	13753.0
221.90	3887.0
223.00	5872.0
224.00	56128.0
225.00	14348.0
226.00	1786.0
227.00	25512.0
228.00	3550.0
229.00	4392.0
229.90	661.0
231.00	1955.0
232.00	473.0
233.90	1436.0
235.00	1895.0
236.00	1136.0
237.00	2147.0
239.00	1043.0
240.00	767.0
241.00	1416.0
242.00	3135.0
243.00	2567.0
244.00	41680.0
245.00	5195.0
246.00	8614.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

247.00	1630.0
247.90	404.0
249.00	1711.0
250.90	364.0
252.10	499.0
253.00	1065.0
254.00	1195.0
255.00	222720.0
256.00	33448.0
257.00	2780.0
258.00	12766.0
258.90	2052.0
259.90	363.0
261.00	376.0
265.00	5875.0
266.00	712.0
270.90	373.0
272.10	718.0
273.00	6426.0
274.00	17000.0
275.00	93704.0
276.00	12213.0
277.00	8362.0
278.00	1630.0
279.00	383.0
282.10	436.0
283.00	919.0
284.00	720.0
285.00	1355.0
286.10	317.0
291.90	401.0
293.00	1799.0
294.00	553.0
295.10	554.0
296.00	26272.0
297.00	3723.0
300.90	317.0
302.00	602.0
303.10	3076.0
304.00	882.0
308.00	306.0
312.90	362.0
314.00	1286.0
315.00	2711.0
316.00	1581.0
317.00	335.0
320.80	567.0
322.00	488.0
323.10	8558.0
324.00	1559.0
326.90	1629.0
328.00	1121.0
332.00	699.0
332.90	725.0
334.00	5946.0
335.00	1265.0
341.10	1180.0
346.00	1879.0
346.90	321.0
352.00	2125.0
353.10	1615.0
354.10	2281.0
355.00	532.0
365.00	10300.0
366.00	1441.0
371.10	438.0
372.10	3915.0
373.10	886.0
383.00	933.0
383.90	383.0
390.00	842.0
402.00	1668.0
403.10	2352.0
404.00	1025.0
421.10	2313.0
422.00	2028.0
423.10	12735.0
424.10	3581.0
425.00	432.0
441.00	41376.0
442.10	247360.0
443.10	47888.0
444.00	4195.0
445.10	348.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

m/z	Abundance
37.00	732.0
38.00	2040.0
39.10	13180.0
40.00	1135.0
41.00	736.0
44.00	1335.0
45.00	352.0
49.00	1135.0
50.00	47200.0
51.00	185856.0
52.00	9244.0
53.00	332.0
55.00	929.0
56.00	5743.0
57.00	12074.0
58.00	777.0
61.00	2024.0
62.00	2369.0
63.00	6870.0
64.00	1098.0
65.00	3763.0
66.00	345.0
68.00	2566.0
69.00	175424.0
70.00	963.0
72.90	1581.0
74.00	18264.0
75.00	27816.0
76.10	10780.0
77.00	200576.0
78.10	12996.0
79.00	13992.0
80.00	11274.0
81.00	14448.0
82.00	3796.0
83.00	3844.0
84.00	538.0
85.00	3273.0
86.00	4203.0
87.00	1911.0
88.10	703.0
89.00	356.0
91.00	3090.0
92.00	3747.0
93.00	22384.0
94.00	1890.0
94.90	380.0
96.00	914.0
96.90	387.0
98.00	17648.0
99.00	14079.0
100.00	1277.0
101.00	8613.0
101.90	362.0
103.00	2617.0
104.00	4763.0
105.10	4555.0
106.00	1666.0
107.00	66608.0
108.00	9592.0
109.00	1600.0
110.00	112488.0
111.00	17712.0
112.00	1968.0
113.00	609.0
116.00	3628.0
117.00	54560.0
118.00	3644.0
118.90	751.0
120.00	736.0
120.90	350.0
122.00	4525.0
123.00	5696.0
124.00	2504.0
125.00	2849.0
126.10	838.0
127.00	208128.0
128.00	15259.0
129.00	77008.0
130.00	6977.0
130.90	1250.0
131.90	778.0
132.90	535.0
134.00	2004.0
135.00	6409.0
136.00	2595.0
137.00	3177.0
137.90	513.0
138.80	368.0
139.90	917.0
141.00	10641.0
142.10	3417.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

143.00	2173.0
144.10	468.0
145.00	625.0
145.90	1667.0
147.00	5872.0
148.00	11976.0
149.00	2922.0
150.00	848.0
151.00	1345.0
152.00	639.0
152.90	3016.0
154.00	2774.0
155.00	6868.0
156.00	9790.0
157.10	2101.0
157.90	1878.0
159.00	1545.0
160.00	3300.0
161.00	4783.0
162.00	1207.0
164.10	378.0
165.00	3793.0
166.00	3233.0
167.00	22032.0
168.00	10415.0
169.00	1926.0
170.00	503.0
170.90	749.0
172.00	2112.0
172.90	3052.0
174.00	5415.0
175.00	8442.0
176.00	2730.0
177.00	3790.0
178.00	1487.0
179.00	16116.0
180.00	11326.0
181.00	4962.0
182.00	553.0
184.00	1507.0
185.00	8711.0
186.00	61808.0
187.00	17736.0
188.00	1925.0
189.00	4231.0
189.90	461.0
191.00	1923.0
192.00	5207.0
193.00	5351.0
194.10	962.0
195.00	647.0
196.00	11212.0
197.10	2128.0
198.00	438656.0
199.00	28904.0
200.00	2615.0
201.50	1635.0
203.00	2789.0
204.00	16118.0
205.00	27512.0
206.00	115928.0
207.00	16720.0
208.00	4107.0
209.00	1360.0
210.10	2093.0
211.00	4217.0
211.90	1008.0
213.10	315.0
215.00	1501.0
216.00	2718.0
217.00	31904.0
218.00	3965.0
220.30	628.0
221.00	14490.0
221.90	3785.0
223.00	6988.0
224.00	64088.0
225.00	14628.0
226.00	1626.0
227.00	27128.0
228.00	4509.0
229.00	6000.0
230.00	755.0
231.00	2116.0
232.10	482.0
233.00	319.0
234.00	1586.0
234.90	2094.0
236.00	1355.0
237.00	1986.0
238.00	353.0
238.90	1077.0
240.10	749.0
241.00	1431.0
242.00	3384.0
243.10	3632.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

244.10	47448.0
245.00	6211.0
246.00	10460.0
247.00	1836.0
247.90	442.0
248.90	1530.0
250.00	320.0
251.10	410.0
251.90	440.0
253.00	1330.0
254.10	1492.0
255.00	264960.0
256.00	37720.0
257.00	2592.0
258.00	15255.0
259.00	2346.0
260.00	418.0
260.90	385.0
264.00	339.0
265.00	5740.0
265.90	789.0
272.10	650.0
273.00	7071.0
274.00	20440.0
275.00	114328.0
276.00	14230.0
277.00	9347.0
278.00	1645.0
279.00	364.0
281.90	382.0
283.00	1154.0
284.00	728.0
285.00	1547.0
286.00	481.0
289.00	447.0
289.90	392.0
292.00	505.0
293.00	2519.0
293.90	671.0
295.00	627.0
296.00	32664.0
297.00	4494.0
298.00	440.0
301.10	351.0
301.90	593.0
303.10	4020.0
304.00	897.0
308.00	565.0
309.10	382.0
310.00	379.0
314.00	1764.0
315.00	3633.0
316.00	1781.0
317.10	317.0
320.90	949.0
322.00	535.0
323.00	10273.0
324.00	1779.0
327.00	2206.0
328.00	924.0
332.00	781.0
333.00	1277.0
334.10	6907.0
335.00	1259.0
341.00	1435.0
346.00	2252.0
352.00	3492.0
353.00	2066.0
354.00	2800.0
354.90	674.0
365.00	13888.0
366.00	1996.0
371.00	696.0
372.10	5232.0
373.00	1551.0
383.00	1441.0
383.90	448.0
390.00	652.0
391.00	415.0
402.00	2096.0
403.10	2661.0
404.10	1386.0
421.00	2654.0
422.10	2976.0
423.10	18264.0
424.10	3610.0
425.10	431.0
441.10	54248.0
442.10	349952.0
443.10	65184.0
444.10	6720.0
445.10	365.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

m/z	Abundance
37.00	686.0
38.00	1736.0
39.00	8483.0
40.00	1066.0
41.10	338.0
43.10	300.0
44.00	1138.0
49.00	1005.0
50.00	29432.0
51.00	120400.0
52.00	6137.0
53.00	480.0
54.90	785.0
56.00	3444.0
57.00	7684.0
58.00	414.0
61.00	1619.0
62.00	1733.0
63.10	4781.0
64.00	521.0
65.00	2478.0
66.90	316.0
68.00	1457.0
69.00	118752.0
70.00	792.0
73.00	1138.0
74.00	11701.0
75.00	18432.0
76.10	7263.0
77.00	133184.0
78.10	9073.0
79.00	8398.0
80.00	6955.0
81.00	9313.0
82.00	2793.0
83.00	2395.0
83.90	310.0
85.00	2186.0
85.90	2667.0
87.10	1120.0
87.90	519.0
91.00	1968.0
92.00	1945.0
93.00	16068.0
93.90	1021.0
94.90	319.0
96.00	678.0
98.00	13034.0
99.00	9411.0
99.90	684.0
101.00	5483.0
103.00	1747.0
104.00	2716.0
105.10	2903.0
105.90	1030.0
107.00	44056.0
108.00	6622.0
109.00	1044.0
110.00	75864.0
111.00	11749.0
112.00	1681.0
112.90	664.0
116.00	2294.0
117.00	38728.0
118.00	2236.0
119.00	521.0
120.00	860.0
122.00	2764.0
123.00	4042.0
124.00	1871.0
125.00	1513.0
126.00	632.0
127.00	147136.0
128.00	10667.0
129.00	57960.0
130.00	4713.0
131.00	917.0
132.00	449.0
134.00	1718.0
135.00	4502.0
135.90	1368.0
137.10	2229.0
137.90	567.0
139.90	715.0
141.00	7422.0
142.00	2079.0
143.00	1857.0
144.10	313.0
145.10	325.0
146.00	1690.0
147.00	4188.0
148.00	8421.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

149.00	1818.0
150.00	592.0
151.00	938.0
151.60	436.0
152.00	442.0
153.00	2567.0
154.00	1721.0
155.00	4047.0
156.00	6395.0
157.10	1026.0
158.00	1102.0
159.00	1293.0
160.00	2472.0
161.00	3623.0
161.90	1010.0
165.00	2771.0
166.00	2731.0
167.00	16848.0
168.00	8503.0
169.00	1576.0
169.90	547.0
171.00	477.0
171.90	1500.0
173.00	1953.0
174.00	3916.0
175.00	6898.0
176.10	1421.0
177.00	2815.0
177.90	946.0
179.00	11830.0
180.00	8953.0
181.00	3901.0
182.00	462.0
184.00	953.0
185.00	6013.0
186.00	45736.0
187.00	12801.0
188.00	1350.0
189.00	3197.0
191.10	1431.0
192.00	4027.0
193.00	4111.0
194.00	1023.0
195.00	829.0
196.00	10173.0
197.00	1496.0
198.00	337664.0
199.00	21752.0
200.00	1611.0
201.30	1463.0
203.00	2357.0
204.00	12308.0
205.00	21048.0
206.10	86832.0
207.00	11734.0
208.00	3189.0
209.00	1005.0
210.00	1455.0
211.00	3189.0
214.90	982.0
216.00	1758.0
217.00	24536.0
218.00	2803.0
219.10	351.0
220.20	393.0
221.00	10608.0
221.80	2888.0
223.00	5382.0
224.00	47480.0
225.00	11901.0
226.00	1274.0
227.00	20896.0
228.00	2798.0
229.00	4049.0
230.00	537.0
231.00	1830.0
232.90	380.0
234.00	1149.0
235.00	1580.0
236.00	862.0
237.10	1981.0
238.00	444.0
239.00	852.0
240.00	777.0
241.00	947.0
242.00	2528.0
243.00	3027.0
244.00	37560.0
245.10	4992.0
246.00	7283.0
246.90	1677.0
249.00	1243.0
249.90	316.0
251.00	481.0
253.00	904.0
254.10	1443.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

255.00	201792.0
256.00	29304.0
257.00	1947.0
258.00	12174.0
259.00	2022.0
259.90	471.0
265.00	5628.0
265.90	511.0
271.10	370.0
272.10	489.0
273.00	6093.0
274.00	16074.0
275.00	87088.0
276.00	11829.0
277.00	8660.0
278.00	1391.0
283.00	1091.0
284.00	803.0
285.00	1421.0
286.00	355.0
288.90	361.0
289.90	321.0
293.00	2112.0
293.90	485.0
295.00	593.0
296.00	27048.0
297.00	3736.0
302.00	573.0
303.00	3160.0
304.00	707.0
308.00	476.0
309.90	301.0
313.00	325.0
314.10	1209.0
315.00	3363.0
316.00	1456.0
317.00	410.0
320.90	1016.0
323.10	8680.0
324.00	1720.0
326.00	322.0
327.00	2085.0
328.00	756.0
332.00	600.0
333.00	943.0
334.00	5655.0
335.10	1565.0
341.00	928.0
342.00	306.0
346.00	1842.0
352.00	2537.0
353.10	1781.0
354.10	2743.0
355.10	321.0
358.90	304.0
365.00	12047.0
366.00	1945.0
371.00	689.0
372.10	4652.0
373.00	1173.0
383.00	1724.0
384.00	362.0
390.00	448.0
391.00	642.0
401.10	374.0
402.00	2567.0
403.00	2514.0
404.10	595.0
421.00	2451.0
422.00	2048.0
423.00	15772.0
424.10	3269.0
425.10	555.0
441.10	48864.0
442.10	316224.0
443.10	60992.0
444.00	5059.0
445.10	329.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45