

Data Path : Z:\HPCHEM1\BNA M\DATA\BM070816\
 Data File : BM006324.D
 Acq On : 07 Jul 2016 18:04
 Operator : UM/SJ
 Sample : H3811-07MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A4TV6MS

Manual Integrations

umangi
 7/8/2016 7:13:02 PM

Quant Time: Jul 08 04:27:50 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM070716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 08 03:28:13 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	4520	0.40	ng/ul	0.00
4) Naphthalene-d8	10.43	136	18121	0.40	ng/ul	0.00
8) Acenaphthene-d10	14.30	164	10798	0.40	ng/ul	0.00
12) Phenanthrene-d10	17.04	188	24187m	0.40	ng/ul	0.00
18) Chrysene-d12	21.24	240	19142	0.40	ng/ul	0.00
22) Perylene-d12	23.45	264	12811m	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	3.17	96	6984	3.66	ng/uL	0.00
6) 2-Methylnaphthalene-d10	12.02	152	6679	0.32	ng/ul	0.00
16) Fluoranthene-d10	19.07	212	16199	0.32	ng/ul	-0.01
Target Compounds				Ovalue		
5) Naphthalene	10.46	128	1373	0.03	ng/ul#	78
7) 2-Methylnaphthalene	12.09	142	739	0.03	ng/ul	98
9) Acenaphthylene	14.01	152	3595	0.10	ng/ul#	93
10) Acenaphthene	14.35	153	9694	0.27	ng/ul	92
11) Fluorene	15.34	166	1552	0.04	ng/ul#	82
13) Pentachlorophenol	16.70	266	1962	0.53	ng/ul	96
14) Phenanthrene	17.07	178	42921m	0.64	ng/ul	
15) Anthracene	17.16	178	7950	0.13	ng/ul#	88
17) Fluoranthene	19.11	202	94921	1.29	ng/ul	93
19) Pyrene	19.47	202	102252	1.28	ng/ul	95
20) Benzo(a)anthracene	21.22	228	33168	0.54	ng/ul#	93
21) Chrysene	21.27	228	39310	0.65	ng/ul	96
23) Benzo(b)fluoranthene	22.78	252	42190m	0.92	ng/ul	
24) Benzo(k)fluoranthene	22.83	252	10881m	0.25	ng/ul	
25) Benzo(a)pyrene	23.35	252	27237m	0.64	ng/ul	
26) Indeno(1,2,3-cd)pyrene	25.66	276	19467m	0.41	ng/ul	
28) Benzo(g,h,i)perylene	26.34	276	17990m	0.45	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM070816\
Data File : BM006324.D
Acq On : 07 Jul 2016 18:04
Operator : UM/SJ
Sample : H3811-07MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4TV6MS

Manual Integrations

umangi
7/8/2016 7:13:02 PM

Quant Time: Jul 08 04:27:50 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM070716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 08 03:28:13 2016
Response via : Initial Calibration

