

Data Path : Z:\HPCHEM1\BNA M\DATA\BM070816\
 Data File : BM006344.D
 Acq On : 08 Jul 2016 06:44
 Operator : UM/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.446

Quant Time: Jul 08 08:58:18 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM070716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 08 08:04:54 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2 SURR	1,4-Dioxane-d8	0.169	0.179	-5.9	86	0.00
3	1,4-Dioxane	0.219	0.226	-3.2	86	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
5	Naphthalene	0.933	0.996	-6.8	88	0.00
6 SURR	2-Methylnaphthalene-d10	0.468	0.529	-13.0	93	0.00
7	2-Methylnaphthalene	0.625	0.703	-12.5	92	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	94	-0.02
9	Acenaphthylene	1.347	1.378	-2.3	91	0.00
10 C	Acenaphthene	1.337	1.437	-7.5	93	0.00
11	Fluorene	1.500	1.607	-7.1	95	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
13	Pentachlorophenol	0.062	0.073	-17.7	122	0.00
14	Phenanthrene	1.106	1.191	-7.7	96	0.00
15	Anthracene	1.025	1.130	-10.2	97	0.00
16 SURR	Fluoranthene-d10	0.843	0.975	-15.7	103	0.00
17 C	Fluoranthene	1.220	1.399	-14.7	103	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
19	Pyrene	1.675	1.746	-4.2	105	0.00
20	Benzo(a)anthracene	1.295	1.462	-12.9	109	0.00
21	Chrysene	1.269	1.442	-13.6	110	0.00
22 I	Perylene-d12	1.000	1.000	0.0	108	0.00
23	Benzo(b)fluoranthene	1.432	1.610	-12.4	116	0.00
24	Benzo(k)fluoranthene	1.332	1.384	-3.9	99	0.00
25 C	Benzo(a)pyrene	1.333	1.452	-8.9	107	0.00
26	Indeno(1,2,3-cd)pyrene	1.479	1.700	-14.9	109	0.00
27	Dibenzo(a,h)anthracene	1.154	1.318	-14.2	109	0.00
28	Benzo(q,h,i)perylene	1.251	1.421	-13.6	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0