


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070821\
 Data File : BM030877.D
 Acq On : 08 Jul 2021 08:48
 Operator : CG/JU
 Sample : DFTPP757
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP755

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM062821.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 08 10:44:22 2021



AutoFind: Scans 2461, 2462, 2463; Background Corrected with Scan 2454

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.7	87949	PASS
68	69	0.00	2	1.7	1397	PASS
69	198	0.00	100	43.2	83053	PASS
70	69	0.00	2	0.5	401	PASS
127	198	10	80	51.2	98469	PASS
197	198	0.00	2	0.7	1266	PASS
198	198	100	100	100.0	192427	PASS
199	198	5	9	7.0	13389	PASS
275	198	10	60	29.2	56192	PASS
365	198	1	100	4.5	8605	PASS
441	198	0.01	100	11.1	21354	PASS
442	198	50	100	70.0	134664	PASS
443	442	15	24	19.1	25685	PASS

m/z	Abundance
39.00	378.0
43.00	172.0
44.00	681.0
73.00	667.0
95.00	174.0
147.00	174.0
190.00	223.0
207.00	1425.0
208.00	280.0
281.00	418.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755

m/z	Abundance
37.00	503.0
38.00	1422.0
39.10	7168.0
40.00	702.0
41.00	288.0
43.00	203.0
44.00	1107.0
44.90	239.0
49.10	841.0
50.00	25608.0
51.00	94576.0
52.00	5322.0
53.00	208.0
55.00	489.0
56.00	2489.0
57.00	6393.0
57.90	256.0
61.00	1310.0
62.00	1362.0
63.00	3884.0
63.90	548.0
65.00	1933.0
66.00	152.0
66.90	181.0
68.00	1420.0
69.00	88864.0
70.00	456.0
73.00	1187.0
74.00	9836.0
75.00	13955.0
76.00	4597.0
77.00	100280.0
78.00	7018.0
79.00	6552.0
80.00	5241.0
81.00	7270.0
82.00	1711.0
83.00	1562.0
84.00	278.0
85.00	1193.0
86.00	2278.0
87.00	967.0
88.00	234.0
88.90	193.0
91.00	1611.0
92.00	2159.0
93.00	11626.0
93.90	886.0
94.90	188.0
95.90	608.0
96.90	232.0
98.00	9675.0
99.00	7772.0
99.90	714.0
101.00	3949.0
101.80	217.0
102.90	1240.0
104.00	2607.0
105.00	2550.0
106.00	843.0
107.00	29688.0
108.00	4376.0
109.00	887.0
110.00	48752.0
111.00	8136.0
112.00	977.0
112.90	341.0
114.90	159.0
116.10	1510.0
117.00	29312.0
118.00	2003.0
118.90	269.0
120.00	384.0
121.00	206.0
122.00	2037.0
123.00	3086.0
123.90	1430.0
125.00	1108.0
126.10	382.0
127.00	98520.0
128.00	7489.0
129.00	38704.0
130.00	3580.0
130.90	649.0
131.90	398.0
133.00	273.0
133.90	1102.0
135.00	3129.0
136.00	1355.0
137.00	2066.0
137.90	278.0
138.90	236.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755

140.00	503.0
141.00	5075.0
142.00	1471.0
142.90	1084.0
143.90	185.0
144.90	246.0
146.00	1014.0
147.00	2461.0
148.00	6968.0
148.90	1534.0
150.00	402.0
151.00	562.0
151.70	290.0
152.00	264.0
153.00	1572.0
154.00	1410.0
155.00	2955.0
156.00	4026.0
156.90	770.0
157.90	1039.0
159.00	718.0
160.00	1709.0
161.00	2577.0
162.00	700.0
162.90	272.0
164.00	385.0
164.90	1986.0
166.00	1737.0
167.00	11661.0
168.00	8044.0
169.00	1316.0
170.00	365.0
170.90	652.0
171.90	1091.0
173.00	1279.0
174.00	2062.0
175.00	4084.0
176.00	1261.0
177.00	2008.0
178.00	627.0
179.00	8251.0
180.00	6140.0
181.00	2402.0
182.00	371.0
182.90	206.0
183.90	693.0
185.00	3718.0
186.00	27160.0
187.00	8290.0
188.00	1089.0
189.00	1860.0
190.00	431.0
191.00	989.0
192.00	2661.0
193.00	3119.0
194.00	634.0
195.00	293.0
196.00	5246.0
197.00	988.0
197.90	175680.0
198.90	12410.0
200.00	1044.0
201.40	898.0
203.00	1440.0
204.00	7929.0
205.00	12752.0
206.00	53760.0
207.00	8802.0
208.00	2470.0
209.00	949.0
210.20	874.0
211.00	2152.0
214.90	454.0
216.00	1472.0
217.00	15100.0
218.00	1920.0
218.90	264.0
221.00	7152.0
221.90	1639.0
223.00	3100.0
224.00	28624.0
225.00	7173.0
226.00	1162.0
227.00	12431.0
228.00	2019.0
229.00	2741.0
230.00	370.0
231.00	1055.0
231.90	242.0
232.80	207.0
234.00	705.0
234.90	856.0
236.00	880.0
237.00	1795.0
238.00	309.0
239.00	540.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755

240.00	450.0
241.00	882.0
242.00	1735.0
243.00	1642.0
244.00	21704.0
245.00	3007.0
246.00	4794.0
247.00	860.0
247.90	220.0
249.00	823.0
250.00	159.0
251.00	291.0
251.90	334.0
252.90	407.0
254.00	686.0
255.00	117024.0
256.00	17232.0
257.00	1848.0
258.00	7661.0
259.00	1064.0
259.80	156.0
260.90	159.0
264.00	187.0
265.00	3025.0
265.90	361.0
270.00	184.0
270.90	196.0
272.10	380.0
273.00	3259.0
274.00	8399.0
275.00	47704.0
276.00	6612.0
277.00	4333.0
278.00	1077.0
279.00	273.0
281.00	485.0
282.00	197.0
283.00	672.0
283.90	352.0
285.00	853.0
286.00	164.0
289.00	191.0
292.00	215.0
292.90	905.0
294.00	330.0
295.00	295.0
296.00	14458.0
297.00	2111.0
297.90	159.0
301.00	189.0
301.90	342.0
303.00	1595.0
304.10	388.0
307.90	288.0
314.00	691.0
314.90	1645.0
316.00	886.0
317.00	158.0
321.00	417.0
321.90	263.0
323.00	3883.0
324.00	846.0
326.90	855.0
328.00	427.0
332.00	262.0
332.90	421.0
334.00	2810.0
334.90	695.0
341.00	483.0
346.00	1074.0
347.00	332.0
352.00	1295.0
353.00	748.0
354.10	1313.0
355.00	282.0
365.00	6555.0
366.00	974.0
371.10	291.0
372.00	2459.0
373.00	552.0
382.90	690.0
390.10	292.0
391.00	183.0
402.00	725.0
402.90	1050.0
404.00	388.0
421.00	950.0
422.00	819.0
423.00	5530.0
424.00	2568.0
425.00	430.0
441.00	14077.0
442.10	92952.0
443.00	17960.0
444.10	1872.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755

m/z	Abundance
36.00	170.0
37.00	541.0
38.00	1384.0
39.00	7698.0
40.00	830.0
41.00	263.0
43.00	167.0
44.00	992.0
45.00	211.0
49.00	900.0
50.00	26784.0
51.00	103912.0
52.00	5057.0
53.00	234.0
55.00	524.0
56.00	3125.0
57.00	6571.0
57.90	373.0
61.00	1090.0
62.00	1447.0
63.00	4193.0
64.00	635.0
65.00	1951.0
68.00	1739.0
69.00	98584.0
70.00	499.0
73.00	1438.0
74.00	11177.0
75.00	16624.0
76.00	5946.0
77.00	115936.0
78.00	8268.0
79.00	7268.0
80.00	6106.0
81.00	8611.0
82.00	2019.0
83.00	1682.0
83.90	297.0
85.00	1262.0
86.00	2398.0
86.90	1032.0
87.90	479.0
89.00	243.0
91.00	1601.0
91.90	2160.0
93.00	14157.0
94.00	908.0
94.90	263.0
96.00	642.0
97.10	289.0
98.00	11243.0
99.00	9398.0
100.00	783.0
100.90	4297.0
102.10	167.0
103.00	1381.0
104.00	2858.0
105.00	2744.0
105.90	923.0
107.00	35656.0
108.00	5242.0
109.00	932.0
110.00	55744.0
111.00	9690.0
111.90	1177.0
112.90	483.0
115.00	183.0
116.00	2012.0
117.00	34400.0
118.00	2614.0
118.90	247.0
119.90	488.0
122.00	2332.0
123.00	3679.0
123.90	1740.0
124.90	1559.0
126.10	440.0
127.00	120264.0
128.00	8772.0
129.00	45992.0
130.00	4139.0
130.90	743.0
131.90	276.0
132.90	288.0
134.00	1737.0
135.00	3961.0
136.00	1521.0
137.00	2264.0
137.90	532.0
139.00	250.0
139.90	678.0
140.90	6099.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755

142.00	1985.0
142.90	1442.0
144.00	296.0
145.00	329.0
145.90	1105.0
147.00	3392.0
148.00	8812.0
149.00	1974.0
150.00	515.0
151.00	783.0
152.00	492.0
153.00	1996.0
154.00	1601.0
155.00	3466.0
156.00	5216.0
157.00	884.0
157.90	1181.0
158.90	892.0
160.00	2191.0
161.00	2933.0
161.90	857.0
162.90	307.0
163.90	237.0
165.00	2349.0
166.00	2072.0
167.00	14664.0
168.00	8562.0
169.00	1212.0
169.90	407.0
171.00	693.0
171.90	1229.0
173.00	1569.0
174.00	2918.0
175.00	5102.0
176.00	1506.0
177.00	2462.0
178.00	738.0
179.00	10668.0
180.00	7345.0
181.00	3423.0
181.90	537.0
183.00	291.0
184.00	798.0
185.00	4966.0
186.00	35752.0
187.00	10525.0
188.00	1340.0
188.90	2742.0
190.00	446.0
191.00	1435.0
192.00	3332.0
193.00	3698.0
194.00	838.0
195.10	415.0
196.00	6998.0
196.80	1021.0
197.00	1030.0
197.90	236736.0
198.90	16728.0
200.00	1428.0
201.50	1004.0
203.00	1985.0
204.00	10636.0
205.00	17656.0
206.00	68848.0
207.00	10680.0
208.00	2670.0
209.00	1132.0
210.00	1287.0
211.00	2406.0
212.00	341.0
213.00	274.0
215.00	766.0
216.00	1827.0
217.00	19496.0
218.00	2355.0
218.90	194.0
221.00	9187.0
221.90	2410.0
223.00	4791.0
224.00	40192.0
225.00	10434.0
226.00	1352.0
227.00	18320.0
228.00	2320.0
229.00	3718.0
229.90	441.0
231.00	1387.0
232.00	349.0
233.00	448.0
234.00	1296.0
234.90	1274.0
236.00	877.0
237.00	1959.0
237.90	323.0
239.00	824.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755

239.90	664.0
241.00	1223.0
242.00	2335.0
243.00	2043.0
244.00	29752.0
245.00	4224.0
246.00	6407.0
246.90	1387.0
248.00	262.0
249.00	1288.0
249.90	227.0
250.90	325.0
251.90	387.0
253.00	739.0
254.10	1003.0
255.00	161600.0
256.00	23776.0
257.00	1932.0
258.00	10583.0
259.00	1546.0
259.80	269.0
260.90	244.0
264.00	289.0
265.00	4462.0
266.00	708.0
270.80	296.0
272.00	484.0
273.00	4959.0
274.00	12169.0
275.00	69120.0
276.00	9528.0
277.00	6271.0
277.90	1201.0
279.00	319.0
281.00	442.0
281.90	201.0
283.00	777.0
284.00	412.0
285.00	1101.0
285.90	252.0
289.00	253.0
289.90	202.0
291.00	192.0
291.90	210.0
293.00	1438.0
294.00	525.0
295.00	410.0
296.00	22960.0
297.00	2904.0
298.00	293.0
301.00	364.0
302.00	440.0
303.00	2392.0
304.00	759.0
307.90	198.0
309.10	163.0
309.90	280.0
313.00	207.0
314.00	1081.0
315.00	2602.0
316.00	1370.0
317.00	213.0
320.90	548.0
322.10	419.0
323.00	6309.0
324.00	1184.0
325.00	155.0
327.00	1396.0
328.00	654.0
331.90	556.0
333.00	634.0
334.00	4009.0
335.00	1038.0
341.10	803.0
342.00	200.0
345.90	1644.0
347.00	379.0
352.00	1869.0
353.00	1356.0
354.00	1974.0
355.00	471.0
359.00	168.0
365.00	10527.0
365.90	1475.0
371.00	617.0
372.00	3342.0
373.10	695.0
383.00	911.0
384.00	283.0
389.90	420.0
390.90	273.0
401.00	299.0
402.00	1355.0
403.00	2110.0
404.00	674.0
404.90	173.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755

421.00	1450.0
421.90	1424.0
423.00	9953.0
424.00	3324.0
424.90	416.0
441.00	25512.0
442.00	161856.0
443.10	31112.0
444.10	2787.0
445.10	181.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755

m/z	Abundance
37.10	279.0
37.90	788.0
39.00	5027.0
40.00	666.0
41.00	291.0
42.90	184.0
44.00	785.0
44.90	197.0
49.00	594.0
50.00	17536.0
51.00	65360.0
52.10	3124.0
53.00	171.0
54.90	506.0
56.00	1980.0
57.00	4606.0
58.00	193.0
60.90	806.0
61.90	973.0
63.00	2341.0
64.00	400.0
65.00	1082.0
68.00	1033.0
69.00	61712.0
70.00	248.0
73.00	958.0
74.00	6949.0
75.00	10662.0
76.00	3531.0
77.00	72696.0
78.00	5067.0
79.00	4811.0
80.00	3767.0
81.00	5233.0
82.00	1114.0
82.90	1026.0
84.00	213.0
85.00	911.0
86.00	1426.0
86.90	611.0
87.90	377.0
91.00	1156.0
92.00	1462.0
93.00	8909.0
93.90	619.0
96.00	499.0
98.00	7371.0
99.00	5601.0
99.90	324.0
101.00	2870.0
101.90	186.0
102.90	970.0
104.00	1945.0
105.00	1767.0
106.00	632.0
107.00	22296.0
108.00	3487.0
109.00	584.0
110.00	36912.0
111.00	5348.0
111.90	728.0
113.00	271.0
115.90	1258.0
117.00	23064.0
118.00	1678.0
119.00	213.0
120.00	345.0
122.00	1405.0
123.00	2601.0
124.00	1197.0
125.00	924.0
126.00	395.0
127.00	76624.0
128.00	5669.0
129.00	30280.0
130.00	2913.0
130.90	610.0
131.90	253.0
133.00	212.0
133.90	891.0
135.00	2458.0
136.00	1034.0
137.00	1486.0
137.90	336.0
139.10	162.0
139.90	462.0
141.00	4289.0
142.00	1257.0
142.90	949.0
143.90	197.0
145.00	250.0
146.00	793.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755

147.00	2224.0
148.00	5792.0
149.00	1099.0
149.90	269.0
151.00	689.0
151.80	374.0
153.00	1290.0
153.90	1100.0
155.00	2221.0
156.00	3781.0
157.00	545.0
157.90	863.0
159.00	635.0
160.00	1251.0
161.00	1770.0
161.90	458.0
163.00	160.0
163.90	178.0
164.90	1786.0
166.00	1405.0
167.00	9922.0
168.00	5872.0
168.90	772.0
170.00	334.0
170.80	350.0
172.00	924.0
173.00	900.0
174.00	2079.0
175.00	3929.0
176.00	935.0
177.00	1772.0
178.00	601.0
179.00	7694.0
180.00	5126.0
181.00	2116.0
182.00	352.0
184.00	582.0
185.00	3381.0
186.00	24264.0
187.00	7108.0
188.00	772.0
189.00	1577.0
189.90	245.0
191.00	967.0
192.00	2358.0
193.00	2510.0
194.00	523.0
194.90	245.0
196.00	4702.0
196.70	589.0
197.10	758.0
197.90	164864.0
199.00	11030.0
199.90	885.0
201.40	660.0
202.90	1470.0
204.00	7632.0
205.00	12251.0
206.00	50064.0
207.00	7397.0
208.00	2023.0
209.00	777.0
210.00	1030.0
211.00	1852.0
211.70	465.0
214.90	585.0
216.00	1151.0
217.00	14262.0
218.00	1916.0
219.90	173.0
221.00	6579.0
222.00	2042.0
223.00	3262.0
224.00	29168.0
225.00	6517.0
226.00	869.0
227.00	12947.0
227.90	2096.0
229.00	2697.0
229.80	264.0
231.00	988.0
233.00	248.0
234.00	862.0
235.00	885.0
235.90	677.0
237.00	1311.0
238.00	222.0
239.00	550.0
239.90	542.0
240.90	765.0
242.00	1626.0
243.00	1720.0
244.00	21792.0
245.00	2829.0
246.00	4959.0
247.00	699.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755

247.90	214.0
249.00	1035.0
251.00	204.0
251.90	238.0
253.00	569.0
254.10	694.0
255.00	121032.0
256.00	16816.0
257.00	1482.0
258.00	7644.0
259.00	1248.0
260.00	208.0
260.90	170.0
263.90	255.0
265.00	3308.0
266.00	547.0
271.00	341.0
272.00	374.0
273.00	3656.0
274.00	9397.0
275.00	51752.0
276.00	6828.0
277.00	4781.0
278.00	887.0
278.90	186.0
280.90	424.0
281.90	245.0
283.00	754.0
284.00	366.0
285.00	874.0
288.90	203.0
289.90	220.0
292.00	264.0
292.90	1121.0
294.10	235.0
295.00	308.0
296.00	17752.0
297.00	2109.0
301.00	276.0
302.00	296.0
303.00	2053.0
303.90	476.0
307.90	267.0
309.10	166.0
310.00	154.0
312.90	172.0
314.00	730.0
315.00	1994.0
316.00	870.0
317.00	167.0
320.90	656.0
321.90	245.0
323.00	4770.0
324.00	1112.0
327.00	1113.0
327.90	492.0
331.90	451.0
333.00	575.0
334.00	3861.0
335.00	786.0
341.00	713.0
342.00	252.0
346.00	1269.0
346.90	227.0
352.00	1744.0
352.90	1026.0
354.00	1470.0
355.00	399.0
359.00	170.0
365.00	8734.0
365.90	986.0
371.00	386.0
372.00	3133.0
373.00	931.0
383.00	780.0
383.90	221.0
390.10	368.0
391.10	158.0
392.00	263.0
401.00	179.0
402.00	1051.0
403.00	1809.0
404.00	531.0
404.80	164.0
421.00	1395.0
422.00	1422.0
423.00	8783.0
424.00	2152.0
425.00	353.0
441.00	24472.0
442.00	149184.0
443.10	27984.0
444.00	2361.0

Instrument :
BNA_M
ClientSampleId :
DFTPP755