

Data Path : V:\HPCHEM1\BNA M\DATA\BM070916\
 Data File : BM006392.D
 Acq On : 09 Jul 2016 23:53
 Operator : UM/SJ
 Sample : H3914-16
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A4TS7

Quant Time: Jul 11 04:58:55 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM070216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 11 04:57:47 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	139944	20.00	ng/ul	0.00
18) Naphthalene-d8	10.41	136	580745	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.28	164	329574	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.03	188	748567	20.00	ng/ul	0.00
75) Chrysene-d12	21.23	240	870693	20.00	ng/ul	0.00
83) Perylene-d12	23.44	264	914252	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.15	96	2406	0.74	ng/uL	0.00
5) Phenol-d5	6.81	99	108712	8.30	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.97	67	238295	30.82	ng/ul	0.00
9) 2-Chlorophenol-d4	7.16	132	262167	26.14	ng/ul	0.00
13) 4-Methylphenol-d8	8.34	113	189755	17.89	ng/ul	0.00
19) Nitrobenzene-d5	8.79	128	154220	33.32	ng/ul	0.00
22) 2-Nitrophenol-d4	9.50	143	176499	33.65	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.04	165	301521	31.72	ng/ul	0.00
29) 4-Chloroaniline-d4	10.56	131	15974	1.62	ng/ul	0.00
43) Dimethylphthalate-d6	13.70	166	1083875	39.41	ng/ul	0.00
46) Acenaphthylene-d8	13.97	160	1233983	36.24	ng/ul	0.00
51) 4-Nitrophenol-d4	14.50	143	51806	9.94	ng/ul	0.00
57) Fluorene-d10	15.28	176	893138	37.67	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.41	200	180791	41.22	ng/ul	0.00
70) Anthracene-d10	17.13	188	1482693	41.51	ng/ul	0.00
76) Pyrene-d10	19.43	212	1705122	41.07	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.30	264	1826592	43.00	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : V:\HPCHEM1\BNA M\DATA\BM070916\
Data File : BM006392.D
Acq On : 09 Jul 2016 23:53
Operator : UM/SJ
Sample : H3914-16
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4TS7

Quant Time: Jul 11 04:58:55 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM070216.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 11 04:57:47 2016
Response via : Initial Calibration

