

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM071020\
 Data File : BM026736.D
 Acq On : 10 Jul 2020 16:24
 Operator : JU/CG
 Sample : L3217-14
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E3-41-BOILER-ROOM

Manual Integrations
 APPROVED

mohammad
 7/13/2020 1:02:52 PM

Quant Time: Jul 10 17:04:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM070820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 10 10:38:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.32	152	112190	20.00	ng	0.00
21) Naphthalene-d8	10.07	136	407667	20.00	ng	0.00
39) Acenaphthene-d10	13.98	164	223034	20.00	ng	0.00
64) Phenanthrene-d10	16.74	188	439979	20.00	ng	0.00
76) Chrysene-d12	20.97	240	449751	20.00	ng	0.00
86) Perylene-d12	23.03	264	480192	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.96	112	831597	124.24	ng	0.00
7) Phenol-d6	6.53	99	1024503	96.07	ng	0.00
23) Nitrobenzene-d5	8.47	82	882153	92.32	ng	0.00
42) 2,4,6-Tribromophenol	15.50	330	424612	130.69	ng	0.00
45) 2-Fluorobiphenyl	12.59	172	1636025	97.53	ng	0.00
79) Terphenyl-d14	19.41	244	2403885	105.19	ng	0.00
Target Compounds						
32) Benzoic acid	9.41	122	3128m	7.608	ng	98
50) Dimethylphthalate	13.46	163	299631	15.706	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM071020\
Data File : BM026736.D
Acq On : 10 Jul 2020 16:24
Operator : JU/CG
Sample : L3217-14
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E3-41-BOILER-ROOM

Manual Integrations
APPROVED

mohammad
7/13/2020 1:02:52 PM

Quant Time: Jul 10 17:04:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM070820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 10 10:38:33 2020
Response via : Initial Calibration

