


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM071117\
Data File : BM010787.D
Acq On : 11 Jul 2017 10:32
Operator : SJ/JU
Sample : DFTPP23
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP23

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
Title : SVOA CALIBRATION
Last Update : Wed Jul 12 16:49:56 2017



AutoFind: Scans 2454, 2455, 2456; Background Corrected with Scan 2449

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.4	26341	PASS
68	69	0.00	2	1.7	621	PASS
69	198	0.00	100	44.6	36290	PASS
70	69	0.00	2	0.5	165	PASS
127	198	10	80	45.1	36720	PASS
197	198	0.00	2	0.6	523	PASS
198	198	100	100	100.0	81378	PASS
199	198	5	9	7.3	5964	PASS
275	198	10	60	27.3	22210	PASS
365	198	1	100	5.1	4145	PASS
441	443	0.01	100	71.4	7651	PASS
442	198	50	100	64.7	52680	PASS
443	442	15	24	20.3	10720	PASS

m/z	Abundance
39.00	1284.0
50.00	5644.0
51.10	20072.0
52.10	1693.0
56.00	912.0
57.00	1504.0
60.90	382.0
62.20	573.0
62.90	1056.0
65.10	529.0
67.90	678.0
69.00	28136.0
72.90	472.0
74.00	3777.0
75.00	5296.0
76.00	2039.0
77.10	32632.0
78.00	2414.0
79.10	2432.0
80.00	1715.0
81.00	3058.0
81.90	543.0
82.80	432.0
85.00	510.0
86.00	856.0
86.80	513.0
90.90	691.0
92.00	844.0
93.00	5322.0
94.00	458.0
97.90	4087.0
99.00	3514.0
101.00	1382.0
102.90	382.0
104.00	866.0
105.00	1021.0
106.00	611.0
107.00	10276.0
107.90	1145.0
110.00	14520.0
111.00	2433.0
112.10	355.0
115.90	628.0
117.00	14047.0
118.00	1270.0
121.90	580.0
123.10	1131.0
123.90	787.0
124.80	466.0
125.10	506.0
126.20	382.0
127.00	25008.0
128.00	1882.0
128.90	13278.0
130.10	1047.0
133.90	502.0
134.90	836.0
136.10	391.0
136.80	791.0
141.10	1918.0
142.00	744.0
143.00	436.0
147.00	914.0
148.00	3228.0
149.00	1000.0
151.70	501.0
152.90	840.0
154.00	597.0
154.90	1149.0
156.00	1545.0
158.00	382.0
160.10	593.0
161.00	1555.0
162.00	421.0
165.00	934.0
165.80	486.0
167.00	3953.0
168.00	1806.0
169.00	380.0
171.90	477.0
173.00	334.0
174.00	883.0
175.00	1575.0
175.90	318.0
177.00	625.0
178.90	2831.0
180.10	1759.0
181.10	542.0
184.80	1178.0
186.00	8760.0
187.00	2419.0

Instrument :
BNA_M
ClientSampleId :
DFTPP23

188.10	326.0
188.90	562.0
190.90	833.0
192.00	918.0
193.00	854.0
195.90	1212.0
197.00	431.0
197.90	54328.0
198.90	3799.0
201.50	409.0
202.90	498.0
204.00	3428.0
205.00	5175.0
206.00	14880.0
207.00	2475.0
207.90	804.0
209.10	760.0
210.00	574.0
210.90	992.0
214.90	300.0
216.00	501.0
217.00	5004.0
217.80	671.0
221.10	1473.0
221.80	670.0
223.00	1144.0
224.00	10038.0
225.00	3088.0
226.10	383.0
227.00	4891.0
227.70	398.0
228.20	567.0
229.10	779.0
230.90	340.0
234.00	346.0
237.00	520.0
238.80	351.0
241.00	533.0
242.10	542.0
242.80	503.0
244.00	6388.0
244.90	902.0
246.00	1992.0
246.80	439.0
253.80	314.0
255.00	29288.0
255.90	4603.0
256.90	648.0
258.00	2821.0
259.10	399.0
264.90	1425.0
273.00	1093.0
274.00	2962.0
275.00	11687.0
276.10	1567.0
276.90	1447.0
277.90	350.0
292.90	614.0
296.00	4888.0
297.00	1127.0
314.90	350.0
315.20	357.0
322.90	733.0
324.00	319.0
327.00	615.0
333.90	533.0
352.00	490.0
365.00	1805.0
423.00	1035.0
424.10	806.0
441.00	4125.0
442.10	24680.0
443.10	5341.0
444.10	366.0

Instrument :
BNA_M
ClientSampleId :
DFTPP23

m/z	Abundance
38.10	485.0
39.00	2485.0
50.00	9912.0
51.10	31648.0
52.10	1670.0
56.10	907.0
57.00	3111.0
61.00	475.0
61.90	879.0
63.00	1910.0
65.00	686.0
68.10	797.0
69.00	44792.0
73.10	731.0
74.00	5410.0
75.00	7000.0
76.10	3607.0
77.10	55088.0
78.10	3524.0
79.00	3942.0
80.00	4273.0
81.00	3986.0
81.90	1015.0
82.90	1103.0
85.00	437.0
86.10	1692.0
86.90	686.0
87.90	336.0
90.90	1321.0
92.00	1391.0
93.00	10054.0
94.00	547.0
98.00	7651.0
99.00	4560.0
101.00	2184.0
102.90	632.0
103.90	1457.0
105.00	1530.0
106.10	381.0
107.00	18552.0
108.00	2387.0
109.10	615.0
110.00	26744.0
111.00	5780.0
111.90	776.0
116.00	917.0
117.00	25152.0
118.00	1450.0
122.00	1043.0
123.10	2302.0
123.90	989.0
125.00	468.0
127.00	44840.0
128.00	4165.0
129.00	20992.0
130.00	2221.0
131.10	457.0
131.90	398.0
133.90	633.0
135.10	1830.0
135.90	763.0
137.00	831.0
141.00	3488.0
142.00	1263.0
143.00	569.0
144.10	357.0
146.00	765.0
147.00	1462.0
148.00	5152.0
149.10	1071.0
151.10	547.0
151.70	368.0
152.90	1085.0
154.00	751.0
155.00	1369.0
156.10	2316.0
158.00	699.0
158.90	388.0
160.00	1435.0
161.00	2236.0
165.00	1352.0
165.90	1351.0
166.90	8183.0
168.00	4724.0
168.90	512.0
170.90	343.0
171.90	825.0
173.00	673.0
174.00	1436.0
175.10	2623.0
176.10	419.0

Instrument :
BNA_M
ClientSampleId :
DFTPP23

176.90	974.0
177.80	528.0
179.00	5071.0
180.00	3752.0
180.90	1389.0
184.00	537.0
185.00	3057.0
186.00	16856.0
187.00	5209.0
187.80	316.0
188.80	2006.0
189.80	363.0
191.00	712.0
191.90	1632.0
192.90	2542.0
194.20	355.0
194.60	309.0
196.10	2682.0
196.80	533.0
197.90	95056.0
199.00	7796.0
199.90	410.0
201.40	681.0
202.80	1090.0
203.90	4976.0
205.00	8909.0
206.00	32544.0
207.00	5096.0
207.90	1783.0
209.00	501.0
209.90	759.0
211.00	1394.0
211.60	373.0
215.00	1064.0
216.00	855.0
216.90	10122.0
218.00	1646.0
221.00	2580.0
221.90	1328.0
223.00	1754.0
224.00	18608.0
225.00	4316.0
225.90	445.0
227.00	11246.0
227.90	1439.0
229.00	1892.0
230.20	616.0
230.90	649.0
232.90	401.0
233.90	724.0
235.00	776.0
236.10	553.0
237.00	755.0
238.00	355.0
239.20	428.0
241.10	703.0
242.00	1010.0
243.00	1261.0
244.00	13957.0
244.90	1938.0
246.00	3427.0
246.90	705.0
248.90	347.0
253.10	325.0
255.00	67880.0
256.00	9719.0
257.00	1016.0
258.00	5378.0
258.90	1208.0
265.00	1968.0
272.90	2009.0
274.00	5715.0
275.00	26592.0
276.00	4003.0
276.90	3248.0
278.00	686.0
283.20	416.0
292.90	695.0
295.10	365.0
296.00	9950.0
297.10	1682.0
303.10	998.0
304.00	365.0
314.00	491.0
314.90	1394.0
316.00	567.0
321.00	373.0
323.10	2640.0
323.90	466.0
326.90	522.0
332.90	385.0
334.00	1833.0
335.00	653.0
345.90	510.0
352.00	864.0

Instrument :
BNA_M
ClientSampleId :
DFTPP23

353.00	577.0
354.00	723.0
364.90	4584.0
366.00	742.0
371.90	1223.0
373.00	599.0
383.00	330.0
402.10	504.0
420.80	412.0
421.90	404.0
423.00	4031.0
424.00	737.0
441.00	8383.0
442.00	55672.0
443.00	12878.0
444.00	1178.0

Instrument :
BNA_M
ClientSampleId :
DFTPP23

m/z	Abundance
38.10	537.0
39.00	1584.0
48.90	554.0
50.10	7732.0
51.10	27304.0
52.10	1473.0
55.90	923.0
57.00	1962.0
61.10	315.0
62.10	732.0
63.00	1501.0
65.10	647.0
68.10	1066.0
69.00	35944.0
70.00	496.0
73.00	640.0
74.00	4578.0
75.00	6582.0
76.00	2184.0
77.10	41616.0
78.10	2552.0
79.00	3302.0
80.00	3062.0
81.00	4075.0
82.00	860.0
83.00	865.0
85.10	502.0
85.90	1341.0
87.00	457.0
88.00	481.0
91.10	1017.0
92.00	730.0
93.00	7411.0
93.90	446.0
96.00	399.0
98.00	6210.0
99.00	4647.0
100.00	583.0
101.00	2267.0
102.90	424.0
103.90	1423.0
105.00	1136.0
105.90	476.0
107.00	16920.0
108.00	1994.0
109.10	625.0
110.00	20464.0
111.00	3960.0
111.90	453.0
115.90	867.0
117.00	20368.0
118.00	1422.0
121.90	1310.0
122.90	1787.0
124.00	657.0
124.90	675.0
127.00	40312.0
128.00	2612.0
129.00	20400.0
130.00	1690.0
131.00	818.0
132.80	354.0
134.10	732.0
135.00	2440.0
135.90	545.0
136.90	915.0
140.10	433.0
140.90	3048.0
142.00	1244.0
142.90	907.0
146.00	685.0
147.00	1961.0
148.00	4844.0
148.90	908.0
150.00	323.0
151.00	682.0
152.90	1340.0
153.90	896.0
154.90	1660.0
156.00	2128.0
157.20	599.0
158.00	950.0
159.90	970.0
161.00	1294.0
162.20	399.0
165.00	1214.0
166.10	1037.0
167.00	7236.0
167.90	3336.0
171.00	361.0
171.80	519.0

Instrument :
BNA_M
ClientSampleId :
DFTPP23

172.90	654.0
173.90	1539.0
175.00	2087.0
176.10	372.0
176.90	1535.0
177.90	731.0
178.90	4676.0
180.00	4636.0
181.00	1205.0
183.80	695.0
185.00	2303.0
186.00	17528.0
187.00	4401.0
188.10	687.0
189.00	1219.0
190.90	1138.0
191.90	1457.0
192.90	1970.0
196.00	2786.0
196.80	606.0
197.90	94752.0
198.90	6297.0
199.80	407.0
201.30	623.0
203.00	682.0
204.10	4119.0
205.00	8042.0
206.00	29528.0
207.10	4564.0
207.90	1549.0
209.10	528.0
210.00	341.0
210.90	1843.0
214.90	758.0
215.80	707.0
217.00	9693.0
218.00	1641.0
221.00	3224.0
221.70	1267.0
223.00	2015.0
224.00	16976.0
225.00	4384.0
225.80	397.0
227.00	10046.0
228.00	912.0
229.00	1281.0
231.00	831.0
234.00	716.0
234.80	604.0
235.90	525.0
237.10	499.0
238.90	474.0
240.00	403.0
241.00	418.0
241.90	737.0
243.00	826.0
244.10	13359.0
245.00	1921.0
246.00	3744.0
247.00	981.0
249.00	653.0
253.10	521.0
253.90	548.0
255.00	71536.0
256.00	7860.0
256.90	485.0
258.00	6000.0
258.80	744.0
259.20	576.0
264.90	2726.0
273.00	2487.0
274.00	6077.0
275.00	28352.0
276.00	4174.0
277.00	2773.0
277.80	510.0
283.10	683.0
292.90	736.0
296.00	10552.0
296.90	1529.0
303.00	1115.0
308.10	348.0
314.00	728.0
314.90	1471.0
315.90	713.0
323.00	2512.0
323.90	831.0
327.00	732.0
327.90	418.0
331.90	329.0
333.00	362.0
334.00	2106.0
334.90	583.0
341.10	543.0
345.90	563.0

Instrument :
BNA_M
ClientSampleId :
DFTPP23

351.90	738.0
353.10	463.0
354.00	1059.0
365.00	6046.0
365.90	870.0
372.00	1398.0
373.10	472.0
383.00	444.0
402.00	389.0
403.00	1005.0
422.10	746.0
423.00	3602.0
424.00	950.0
441.00	10447.0
442.10	77688.0
443.00	13941.0
444.00	1232.0

Instrument :
BNA_M
ClientSampleId :
DFTPP23

m/z	Abundance
206.90	1077.0

Instrument :
BNA_M
ClientSampleId :
DFTPP23