

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071124\
 Data File : BM046510.D
 Acq On : 11 Jul 2024 10:27
 Operator : MA/JU
 Sample : PB161755BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK755

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 07/12/2024
 Supervised By :mohammad ahmed 07/13/2024

Quant Time: Jul 12 04:19:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM071024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 12 04:18:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.514	152	1825	0.400	ng/ul	0.00
4) Naphthalene-d8	10.275	136	4453	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.158	164	3107	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.912	188	7018m	0.400	ng/ul	0.00
17) Chrysene-d12	21.117	240	5886	0.400	ng/ul #	0.00
23) Perylene-d12	23.262	264	6954m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.117	96	7058	4.707	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.875	152	1881	0.262	ng/ul	0.00
18) Fluoranthene-d10	18.950	212	4852	0.271	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071124\
Data File : BM046510.D
Acq On : 11 Jul 2024 10:27
Operator : MA/JU
Sample : PB161755BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK755

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 07/12/2024
Supervised By :mohammad ahmed 07/13/2024

Quant Time: Jul 12 04:19:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM071024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 12 04:18:37 2024
Response via : Initial Calibration

