



Data Path : Z:\HPCHEM1\BNA M\Data\BM071317\
Data File : BM010818.D
Acq On : 13 Jul 2017 18:33
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 14 09:18:36 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM071217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 12 14:21:16 2017
Response via : Initial Calibration



TIC: BM010818.D

(76) Benzidine

19.121min (-0.006) 66137.49ng

response 1598177

Ion	Exp%	Act%
184.00	100	100
185.00	14.10	14.88
183.00	11.10	11.88
0.00	0.00	0.00