

Data Path : Z:\HPCHEM1\BNA\_M\DATA\BM071317\  
 Data File : BM010824.D  
 Acq On : 13 Jul 2017 22:10  
 Operator : SJ/JU  
 Sample : PB100603BL  
 Misc :  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 PB100603BL

Quant Time: Jul 14 01:27:58 2017  
 Quant Method : Z:\HPCHEM1\BNA\_M\METHODS\8270-BM071217.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jul 12 14:21:16 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.45  | 152  | 243621   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.20 | 136  | 892192   | 20.00  | ng    | 0.00     |
| 38) Acenaphthene-d10        | 14.10 | 164  | 550931   | 20.00  | ng    | 0.00     |
| 63) Phenanthrene-d10        | 16.85 | 188  | 1253386  | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 21.06 | 240  | 1212014  | 20.00  | ng    | 0.00     |
| 86) Perylene-d12            | 23.19 | 264  | 1026284  | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.09  | 112  | 2000580  | 136.03 | ng    | 0.00     |
| 7) Phenol-d6                | 6.64  | 99   | 2634792  | 129.95 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 8.59  | 82   | 1999090  | 86.13  | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 15.60 | 330  | 1105366  | 131.29 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 12.71 | 172  | 4006728  | 90.47  | ng    | 0.00     |
| 78) Terphenyl-d14           | 19.51 | 244  | 5183937  | 82.72  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_M\DATA\BM071317\  
Data File : BM010824.D  
Acq On : 13 Jul 2017 22:10  
Operator : SJ/JU  
Sample : PB100603BL  
Misc :  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
PB100603BL

Quant Time: Jul 14 01:27:58 2017  
Quant Method : Z:\HPCHEM1\BNA\_M\METHODS\8270-BM071217.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jul 12 14:21:16 2017  
Response via : Initial Calibration

