

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM071415\
 Data File : BM001912.D
 Acq On : 14 Jul 2015 11:49
 Operator : TP/UM
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.448

Quant Time: Jul 14 18:24:13 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM071415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 14 04:27:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
3	Naphthalene	1.214	1.222	-0.7	106	0.00
4 SURR	2-Methylnaphthalene-d10	0.625	0.623	0.3	105	0.00
5	2-Methylnaphthalene	0.806	0.807	-0.1	107	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
7	Acenaphthylene	2.357	2.302	2.3	106	0.00
8 C	Acenaphthene	1.665	1.667	-0.1	109	0.00
9	Fluorene	1.915	1.941	-1.4	111	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
11	Pentachlorophenol	0.105	0.100	4.8	114	0.00
12	Phenanthrene	1.392	1.425	-2.4	111	0.00
13	Anthracene	1.339	1.351	-0.9	110	0.00
14 SURR	Fluoranthene-d10	1.085	1.096	-1.0	109	0.00
15 C	Fluoranthene	1.551	1.578	-1.7	111	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	108	0.01
17	Pyrene	1.816	1.819	-0.2	109	0.01
18	Benzo(a)anthracene	1.566	1.516	3.2	106	0.01
19	Chrysene	1.556	1.565	-0.6	110	0.01
20 I	Perylene-d12	1.000	1.000	0.0	106	0.01
21	Benzo(b)fluoranthene	1.835	1.688	8.0	101	0.00
22	Benzo(k)fluoranthene	1.524	1.666	-9.3	116	0.01
23 C	Benzo(a)pyrene	1.587	1.559	1.8	106	0.01
24	Indeno(1,2,3-cd)pyrene	1.702	1.696	0.4	106	0.00
25	Dibenzo(a,h)anthracene	1.357	1.345	0.9	105	0.01
26	Benzo(g,h,i)perylene	1.437	1.460	-1.6	107	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0