

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM071415\
 Data File : BM001917.D
 Acq On : 14 Jul 2015 16:43
 Operator : TP/UM
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.449

Quant Time: Jul 14 18:38:39 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM071415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 14 04:27:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
3	Naphthalene	1.214	1.224	-0.8	115	0.00
4 SURR	2-Methylnaphthalene-d10	0.625	0.633	-1.3	115	-0.01
5	2-Methylnaphthalene	0.806	0.815	-1.1	116	-0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
7	Acenaphthylene	2.357	2.355	0.1	116	0.00
8 C	Acenaphthene	1.665	1.665	0.0	116	0.00
9	Fluorene	1.915	1.935	-1.0	118	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
11	Pentachlorophenol	0.105	0.097	7.6	116	0.00
12	Phenanthrene	1.392	1.396	-0.3	113	0.00
13	Anthracene	1.339	1.341	-0.1	114	0.00
14 SURR	Fluoranthene-d10	1.085	1.096	-1.0	114	0.00
15 C	Fluoranthene	1.551	1.535	1.0	113	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
17	Pyrene	1.816	1.833	-0.9	112	0.00
18	Benzo(a)anthracene	1.566	1.549	1.1	110	0.00
19	Chrysene	1.556	1.564	-0.5	112	0.00
20 I	Perylene-d12	1.000	1.000	0.0	105	0.00
21	Benzo(b)fluoranthene	1.835	1.733	5.6	102	0.00
22	Benzo(k)fluoranthene	1.524	1.596	-4.7	109	0.00
23 C	Benzo(a)pyrene	1.587	1.546	2.6	104	0.01
24	Indeno(1,2,3-cd)pyrene	1.702	1.566	8.0	97	0.00
25	Dibenzo(a,h)anthracene	1.357	1.240	8.6	95	0.01
26	Benzo(g,h,i)perylene	1.437	1.329	7.5	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0