


Data Path : Z:\HPCHEM1\BNA M\DATA\BM071516\
 Data File : BM006442.D
 Acq On : 15 Jul 2016 14:13
 Operator : UM/SJ
 Sample : DFTPP51
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP51

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM071416.M
 Title : SVOA CALIBRATION
 Last Update : Fri Jul 15 23:20:22 2016



AutoFind: Scans 2529, 2530, 2531; Background Corrected with Scan 2524

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.1	54642	PASS
68	69	0.00	2	1.5	745	PASS
69	198	0.00	100	38.9	49384	PASS
70	69	0.00	2	0.5	255	PASS
127	198	10	80	46.3	58736	PASS
197	198	0.00	2	0.3	438	PASS
198	198	100	100	100.0	126912	PASS
199	198	5	9	6.9	8705	PASS
275	198	10	60	26.3	33344	PASS
365	198	1	100	3.3	4162	PASS
441	443	0.01	100	80.1	15628	PASS
442	198	50	100	79.5	100850	PASS
443	442	15	24	19.3	19500	PASS

Scan 2524 (17.433 min): BM006442.D
DFTPP51

m/z	Abundance
39.90	844.0
44.00	1048.0
48.90	404.0
73.00	367.0
207.00	769.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

m/z	Abundance
38.00	634.0
39.00	4657.0
39.90	856.0
44.00	814.0
48.90	793.0
50.00	13565.0
51.00	56416.0
52.10	2855.0
55.90	1770.0
57.00	3479.0
60.90	858.0
62.00	681.0
63.00	1854.0
65.00	998.0
68.00	730.0
69.00	48056.0
70.00	434.0
73.00	723.0
74.00	5145.0
75.00	8816.0
76.00	2577.0
77.00	57464.0
78.10	3762.0
79.00	3770.0
80.00	3367.0
81.00	4018.0
82.00	887.0
82.90	1073.0
85.00	692.0
86.00	1118.0
87.00	777.0
91.10	653.0
91.90	1010.0
93.00	6905.0
94.00	500.0
95.90	304.0
98.00	4621.0
99.00	4839.0
99.90	456.0
101.00	2131.0
102.90	575.0
104.00	1539.0
105.00	1444.0
107.00	17552.0
108.00	2420.0
109.10	419.0
110.00	28632.0
111.00	4427.0
112.00	491.0
115.90	860.0
117.00	13504.0
118.00	1195.0
121.90	1023.0
122.90	1578.0
123.90	716.0
125.00	699.0
127.00	53968.0
128.00	4348.0
129.00	22088.0
130.00	1942.0
131.00	341.0
134.00	759.0
135.00	1665.0
135.90	652.0
136.90	1569.0
139.90	317.0
141.00	2687.0
141.90	914.0
142.90	797.0
146.00	499.0
147.00	1445.0
148.00	3805.0
148.90	1095.0
150.90	363.0
153.00	961.0
154.00	796.0
155.00	1489.0
156.00	3136.0
157.00	604.0
157.80	460.0
158.90	488.0
160.00	1074.0
161.00	1253.0
161.80	446.0
165.00	883.0
166.00	800.0
167.00	5953.0
168.00	3871.0
168.90	520.0
171.80	439.0
172.90	879.0
174.00	1417.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

175.00	2357.0
176.00	502.0
176.90	1378.0
178.00	402.0
179.00	4495.0
180.00	3242.0
180.90	1314.0
185.00	1753.0
186.00	15322.0
187.00	4345.0
188.10	641.0
189.00	1081.0
191.00	792.0
191.90	1438.0
193.10	1635.0
196.00	2388.0
196.90	819.0
198.00	108496.0
198.90	7664.0
200.00	834.0
201.40	584.0
202.90	878.0
204.10	4487.0
205.00	7127.0
206.00	30048.0
207.10	3924.0
207.90	1341.0
209.00	443.0
210.00	550.0
211.10	1134.0
214.90	333.0
216.00	572.0
217.00	8066.0
218.00	1137.0
221.00	4202.0
221.80	813.0
222.90	1858.0
224.00	16334.0
225.00	4453.0
226.00	744.0
227.00	6619.0
227.90	941.0
229.00	1491.0
230.90	558.0
234.00	569.0
235.10	457.0
236.00	462.0
236.90	1137.0
241.00	415.0
241.90	557.0
243.00	964.0
244.00	11766.0
245.10	1608.0
246.00	2430.0
246.90	448.0
248.90	330.0
254.00	542.0
255.00	62672.0
256.00	9507.0
257.10	1033.0
258.00	3790.0
259.00	353.0
265.00	1622.0
273.00	1574.0
274.00	4837.0
275.00	26040.0
276.00	3595.0
277.00	1859.0
278.00	393.0
283.00	327.0
293.00	545.0
296.00	7971.0
296.90	1184.0
303.00	851.0
314.10	435.0
315.00	879.0
316.00	508.0
323.10	2446.0
324.00	666.0
326.80	345.0
334.00	1547.0
335.00	390.0
341.00	305.0
346.00	354.0
352.00	545.0
353.00	534.0
354.10	739.0
365.00	3360.0
366.00	635.0
372.10	855.0
402.00	472.0
403.00	676.0
403.90	397.0
421.00	531.0
423.00	3844.0
424.10	1319.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

441.10	10838.0
442.10	67920.0
443.00	13262.0
444.00	1766.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

m/z	Abundance
38.00	598.0
39.00	4232.0
40.00	986.0
41.00	392.0
44.00	859.0
49.00	733.0
50.00	16520.0
51.00	63624.0
52.00	3241.0
54.90	489.0
56.00	1871.0
57.00	4442.0
61.00	786.0
62.10	1032.0
63.00	3117.0
65.00	1025.0
68.00	940.0
69.00	57808.0
70.00	333.0
73.00	992.0
74.00	5938.0
75.00	8697.0
76.00	3406.0
77.00	64320.0
78.00	4537.0
79.00	4528.0
79.90	3311.0
81.00	4659.0
82.00	1007.0
83.00	1019.0
84.90	892.0
86.00	1342.0
87.00	758.0
91.00	1059.0
91.90	1403.0
93.00	8279.0
93.90	609.0
96.00	424.0
98.00	6073.0
99.00	5658.0
100.00	507.0
101.00	2687.0
103.10	889.0
104.00	1534.0
105.00	1463.0
105.90	447.0
107.00	20960.0
108.00	3274.0
108.90	679.0
110.00	36704.0
111.00	5660.0
112.00	752.0
116.00	1117.0
117.00	18864.0
117.90	1310.0
119.90	355.0
122.00	1361.0
123.00	1923.0
124.00	1019.0
125.00	831.0
127.00	68552.0
128.00	5194.0
129.00	25776.0
129.90	2027.0
130.90	641.0
133.90	1224.0
135.00	2449.0
136.00	791.0
136.90	1121.0
139.90	498.0
141.00	3388.0
142.00	1267.0
143.00	776.0
146.00	515.0
147.00	1962.0
148.00	4761.0
149.00	1109.0
150.90	413.0
152.90	985.0
153.90	877.0
154.90	2132.0
156.00	2849.0
157.10	697.0
158.00	600.0
159.00	606.0
159.90	1404.0
161.00	1912.0
161.80	584.0
165.00	1401.0
166.00	1070.0
167.00	6470.0
168.00	5047.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

169.00	833.0
169.90	354.0
171.00	429.0
171.90	438.0
172.90	1060.0
174.10	2054.0
175.00	3433.0
176.00	1167.0
176.90	1012.0
178.00	568.0
179.00	5659.0
180.00	4309.0
181.00	2227.0
182.00	374.0
184.00	447.0
185.10	3007.0
186.00	20840.0
187.00	6641.0
188.00	605.0
189.00	1160.0
190.90	645.0
192.00	1774.0
193.00	2356.0
194.00	523.0
196.00	4495.0
197.10	582.0
198.00	151552.0
199.00	10870.0
199.90	1038.0
201.30	566.0
201.80	332.0
203.00	1101.0
204.00	6282.0
205.00	10071.0
206.00	39160.0
207.00	5744.0
208.00	1732.0
208.90	659.0
210.00	616.0
211.10	1450.0
214.90	492.0
215.90	936.0
216.90	11065.0
218.00	1314.0
221.00	4782.0
222.00	1154.0
223.00	2626.0
224.00	23008.0
225.00	6211.0
226.00	718.0
227.00	9517.0
228.00	1704.0
229.00	1763.0
231.00	757.0
233.90	577.0
235.00	490.0
237.00	1019.0
238.90	395.0
240.00	337.0
240.90	491.0
242.00	1002.0
243.00	1110.0
244.00	18144.0
245.00	2837.0
246.00	3855.0
246.90	587.0
249.00	563.0
252.90	444.0
254.00	528.0
255.00	90392.0
256.00	13667.0
257.00	1334.0
258.00	4978.0
259.00	960.0
265.00	1733.0
266.00	305.0
273.00	3071.0
274.00	7292.0
275.00	39672.0
276.00	5187.0
276.90	3917.0
278.10	645.0
282.90	310.0
284.00	309.0
285.10	746.0
293.00	496.0
295.10	415.0
296.00	11443.0
297.10	1691.0
303.00	1291.0
304.00	412.0
310.00	337.0
313.90	414.0
314.10	410.0
315.00	1252.0
316.10	800.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

321.00	312.0
323.00	3465.0
323.90	579.0
327.00	602.0
328.00	316.0
332.00	300.0
333.00	462.0
334.00	2511.0
335.00	645.0
341.10	378.0
345.90	694.0
352.00	1006.0
353.00	773.0
354.10	1179.0
365.00	4845.0
366.00	664.0
372.00	1860.0
373.00	428.0
383.10	442.0
402.00	888.0
403.00	988.0
404.00	437.0
421.00	1156.0
422.10	1064.0
423.00	6148.0
424.00	1672.0
425.10	357.0
441.00	18448.0
442.10	117384.0
443.10	22504.0
444.00	2599.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

m/z	Abundance
38.00	599.0
39.00	3053.0
40.00	908.0
44.00	979.0
48.90	690.0
50.00	10564.0
51.00	43888.0
52.00	2217.0
56.00	1264.0
57.00	3050.0
61.00	826.0
62.00	538.0
63.00	1540.0
65.00	770.0
68.00	566.0
69.00	42288.0
73.00	391.0
74.00	4083.0
75.00	7298.0
76.10	2495.0
77.00	49480.0
78.00	2886.0
78.90	3369.0
80.00	2273.0
81.00	3668.0
81.90	805.0
83.00	791.0
85.00	829.0
85.90	1002.0
86.90	615.0
90.90	744.0
91.90	965.0
93.00	5056.0
96.00	410.0
98.00	4713.0
98.90	3357.0
99.90	385.0
101.00	1926.0
102.90	628.0
103.90	1155.0
105.00	1162.0
105.90	357.0
107.00	15151.0
108.00	2137.0
109.00	404.0
110.00	26728.0
111.00	3772.0
112.00	354.0
116.10	888.0
117.00	13081.0
118.00	687.0
122.00	877.0
123.00	1608.0
124.00	535.0
124.90	881.0
127.00	53688.0
128.00	3478.0
129.00	20224.0
130.00	1599.0
134.00	640.0
135.00	1764.0
135.90	840.0
137.00	722.0
137.80	313.0
141.00	2673.0
141.90	807.0
142.90	658.0
146.00	435.0
147.00	1360.0
148.00	3455.0
148.90	800.0
150.90	438.0
151.90	369.0
152.90	1020.0
154.00	701.0
155.00	1505.0
156.00	2254.0
157.90	643.0
159.00	360.0
160.00	869.0
161.00	1300.0
162.00	368.0
165.00	1123.0
165.90	768.0
166.10	768.0
167.00	5054.0
168.00	3443.0
169.10	502.0
171.90	465.0
173.00	751.0
174.00	1300.0
175.10	2436.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

175.90	511.0
177.00	937.0
177.80	360.0
179.00	4428.0
180.00	2961.0
181.00	1525.0
182.00	333.0
184.00	303.0
185.00	2124.0
186.00	16520.0
187.00	4818.0
188.00	551.0
188.90	1262.0
191.00	822.0
191.90	1412.0
193.00	2014.0
196.10	3066.0
197.00	732.0
198.00	120688.0
199.00	7583.0
199.90	615.0
201.70	490.0
203.00	940.0
204.00	4881.0
205.00	7473.0
206.10	32152.0
207.00	4807.0
208.00	1028.0
209.00	309.0
210.10	648.0
211.00	1331.0
215.00	370.0
216.00	674.0
217.00	8836.0
218.00	1056.0
221.00	4598.0
221.90	1089.0
223.00	2383.0
224.00	20016.0
225.00	4100.0
226.00	384.0
227.00	7896.0
228.00	1346.0
229.00	1432.0
230.00	353.0
231.00	749.0
233.90	651.0
235.00	656.0
236.00	364.0
236.90	737.0
238.90	424.0
240.00	388.0
241.00	512.0
242.10	862.0
243.00	1047.0
244.00	13518.0
245.00	1735.0
246.00	3733.0
246.90	552.0
249.00	451.0
253.00	466.0
254.10	558.0
255.00	77200.0
256.00	10470.0
256.90	727.0
258.00	4439.0
258.90	529.0
264.90	1468.0
272.00	363.0
273.00	2123.0
274.00	6131.0
275.00	34320.0
276.00	4035.0
277.00	2569.0
277.90	597.0
283.00	332.0
285.10	599.0
293.00	760.0
296.00	8868.0
297.00	1223.0
303.00	1205.0
314.10	568.0
314.90	1008.0
316.10	857.0
321.00	300.0
323.10	3853.0
324.00	630.0
327.00	888.0
328.00	401.0
332.90	335.0
334.00	2056.0
335.00	655.0
341.00	365.0
345.80	636.0
352.00	800.0
353.10	748.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

354.10	952.0
365.00	4282.0
365.90	526.0
372.00	1732.0
373.10	542.0
382.90	367.0
402.00	637.0
403.00	1330.0
421.10	919.0
422.00	733.0
423.00	5317.0
424.10	1732.0
441.00	17600.0
442.10	117248.0
443.10	22736.0
444.10	2132.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51