

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071521\
 Data File : BM030988.D
 Acq On : 15 Jul 2021 21:03
 Operator : CG/JU
 Sample : M2958-03DL 5X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBK28DL

Quant Time: Jul 16 03:34:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM071321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 15 16:14:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.656	152	1201	0.40	ng/ul	0.00
4) Naphthalene-d8	10.419	136	4234	0.40	ng/ul	# 0.00
9) Acenaphthene-d10	14.280	164	3435	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.030	188	6578	0.40	ng/ul	# 0.00
17) Chrysene-d12	21.214	240	5781	0.40	ng/ul	0.00
23) Perylene-d12	23.396	264	5700	0.40	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.230	96	636	0.51	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.011	152	514	0.07	ng/ul	0.00
18) Fluoranthene-d10	19.058	212	1625	0.08	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.262	88	30	0.02	ng/ul#	60
5) Naphthalene	10.469	128	11526	0.91	ng/ul	97
7) 2-Methylnaphthalene	12.088	142	9097	1.02	ng/ul	99
8) 1-Methylnaphthalene	12.308	142	7263	0.83	ng/ul	100
11) Acenaphthene	14.345	153	580	0.05	ng/ul	92
12) Fluorene	15.334	166	1681	0.12	ng/ul#	96
14) Pentachlorophenol	16.676	266	115	0.05	ng/ul	91
15) Phenanthrene	17.072	178	1618	0.08	ng/ul#	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071521\
Data File : BM030988.D
Acq On : 15 Jul 2021 21:03
Operator : CG/JU
Sample : M2958-03DL 5X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBK28DL

Quant Time: Jul 16 03:34:03 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM071321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 16:14:04 2021
Response via : Initial Calibration

