

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM071616\
 Data File : BM006488.D
 Acq On : 16 Jul 2016 21:15
 Operator : UM/SJ
 Sample : H3985-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BDJ40

Quant Time: Jul 18 07:15:38 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM071416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 15 23:20:22 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	144595	20.00	ng/ul	0.00
18) Naphthalene-d8	10.39	136	642777	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.27	164	368356	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.02	188	844278	20.00	ng/ul	0.00
75) Chrysene-d12	21.22	240	1015924	20.00	ng/ul	-0.01
83) Perylene-d12	23.41	264	1074767	20.00	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.14	96	10324	3.01	ng/uL	0.00
5) Phenol-d5	6.79	99	452870	38.22	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.96	67	294488	40.97	ng/ul	0.00
9) 2-Chlorophenol-d4	7.15	132	361967	37.70	ng/ul	0.00
13) 4-Methylphenol-d8	8.32	113	369013	39.19	ng/ul	0.00
19) Nitrobenzene-d5	8.77	128	179253	36.80	ng/ul	0.00
22) 2-Nitrophenol-d4	9.49	143	199511	36.22	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.02	165	346551	33.35	ng/ul	0.00
29) 4-Chloroaniline-d4	10.53	131	475788	43.80	ng/ul	0.00
43) Dimethylphthalate-d6	13.68	166	1133217	37.59	ng/ul	0.00
46) Acenaphthylene-d8	13.96	160	1397656	37.50	ng/ul	0.00
51) 4-Nitrophenol-d4	14.49	143	175209	33.45	ng/ul	0.00
57) Fluorene-d10	15.26	176	982206	36.65	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.40	200	141393	28.95	ng/ul	0.00
70) Anthracene-d10	17.12	188	1525486	38.23	ng/ul	0.00
76) Pyrene-d10	19.42	212	1736783	37.71	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.27	264	1922685	39.12	ng/ul	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Phenol	6.82	94	13459	1.08	ng/ul#	91
44) Dimethylphthalate	13.73	163	589796	19.28	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM071616\
Data File : BM006488.D
Acq On : 16 Jul 2016 21:15
Operator : UM/SJ
Sample : H3985-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BDJ40

Quant Time: Jul 18 07:15:38 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM071416.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 15 23:20:22 2016
Response via : Initial Calibration

