


Data Path : Z:\HPCHEM1\BNA M\DATA\BM071816\
Data File : BM006515.D
Acq On : 18 Jul 2016 09:58
Operator : UM/SJ
Sample : DFTPP53
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP53

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM071416.M
Title : SVOA CALIBRATION
Last Update : Tue Jul 19 02:05:48 2016



AutoFind: Scans 2528, 2529, 2530; Background Corrected with Scan 2523

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.8	69213	PASS
68	69	0.00	2	1.6	1035	PASS
69	198	0.00	100	40.6	64120	PASS
70	69	0.00	2	0.6	393	PASS
127	198	10	80	47.4	74949	PASS
197	198	0.00	2	0.4	657	PASS
198	198	100	100	100.0	158016	PASS
199	198	5	9	6.8	10723	PASS
275	198	10	60	26.2	41442	PASS
365	198	1	100	3.2	5047	PASS
441	443	0.01	100	81.6	19729	PASS
442	198	50	100	79.5	125701	PASS
443	442	15	24	19.2	24185	PASS

Scan 2523 (17.427 min): BM006515.D
DFTPP53

m/z	Abundance
39.90	837.0
41.10	345.0
42.90	375.0
44.00	1497.0
49.00	402.0
55.00	443.0
73.00	1033.0
84.00	430.0
91.00	368.0
96.00	357.0
167.00	314.0
207.00	1723.0
281.00	314.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

m/z	Abundance
38.00	800.0
39.10	4960.0
40.00	908.0
41.00	597.0
42.90	412.0
44.00	1681.0
48.90	773.0
50.00	16664.0
51.00	67728.0
52.00	3180.0
52.90	375.0
55.00	615.0
56.00	2142.0
57.00	4969.0
61.00	986.0
62.00	1064.0
63.00	2302.0
64.00	308.0
65.10	1206.0
68.00	856.0
69.00	61976.0
70.00	492.0
73.00	1245.0
74.10	6105.0
75.00	9298.0
76.00	3133.0
77.00	68344.0
78.10	5059.0
79.00	4956.0
80.00	3253.0
81.00	5039.0
82.00	1264.0
83.00	1371.0
83.90	434.0
85.00	800.0
86.00	1497.0
87.00	603.0
87.90	312.0
91.00	1599.0
92.00	1310.0
93.00	7017.0
93.90	784.0
94.90	457.0
96.00	502.0
98.10	5593.0
99.00	5057.0
99.90	547.0
101.00	2912.0
103.00	754.0
104.00	1580.0
105.00	1620.0
105.90	482.0
107.00	20920.0
108.00	3193.0
109.00	685.0
110.00	37328.0
111.00	5620.0
111.90	872.0
113.00	349.0
116.00	1227.0
117.00	16784.0
118.00	1446.0
119.00	446.0
119.90	462.0
122.00	1223.0
123.00	2245.0
123.90	768.0
124.90	1012.0
127.00	68184.0
128.00	4996.0
129.00	25752.0
130.00	2253.0
131.00	669.0
131.90	375.0
134.00	896.0
135.00	2516.0
135.90	1058.0
137.00	974.0
140.00	467.0
141.00	3333.0
142.00	1113.0
143.10	1079.0
145.90	458.0
146.90	2181.0
148.00	3931.0
149.00	959.0
150.10	351.0
151.30	339.0
152.90	1170.0
153.90	984.0
155.00	1872.0
156.10	2850.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

157.00	572.0
158.00	427.0
159.00	528.0
159.90	1211.0
161.00	1949.0
161.90	548.0
165.10	1200.0
166.00	992.0
167.00	6056.0
168.00	4189.0
169.00	585.0
171.00	310.0
172.10	412.0
173.00	650.0
173.90	1398.0
175.00	3018.0
175.90	934.0
177.00	1027.0
177.90	387.0
179.00	4597.0
180.00	3654.0
181.00	1456.0
183.90	407.0
185.00	2425.0
186.00	18704.0
187.00	5311.0
187.90	764.0
189.00	1240.0
191.00	654.0
192.00	1460.0
193.00	1740.0
194.10	391.0
195.20	360.0
196.00	3150.0
197.10	817.0
198.00	131136.0
199.00	9224.0
200.00	557.0
202.90	971.0
204.00	5227.0
205.10	8143.0
206.00	34384.0
207.00	5783.0
208.10	1415.0
209.00	775.0
210.00	569.0
211.00	1491.0
211.90	384.0
214.90	446.0
215.90	797.0
217.00	8714.0
217.90	1533.0
221.00	5568.0
221.90	1509.0
222.90	2084.0
224.10	18808.0
225.00	4524.0
226.00	763.0
227.00	8360.0
228.00	1172.0
229.00	1832.0
231.10	705.0
234.00	747.0
235.00	807.0
236.10	423.0
237.00	882.0
238.90	395.0
239.90	323.0
241.00	613.0
242.10	964.0
243.00	1095.0
244.10	14852.0
245.10	2155.0
246.00	2582.0
246.90	408.0
248.90	462.0
252.90	407.0
254.00	398.0
255.00	73792.0
256.00	10991.0
257.00	826.0
258.00	4383.0
259.10	628.0
265.00	1660.0
273.00	2307.0
274.00	5511.0
275.00	31936.0
276.00	4775.0
277.00	2219.0
278.00	726.0
281.10	310.0
284.90	443.0
293.00	671.0
293.90	393.0
296.00	8984.0
297.00	1377.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

303.00	825.0
313.90	388.0
315.00	949.0
316.00	612.0
323.10	2671.0
324.10	445.0
327.00	687.0
328.00	369.0
334.00	1956.0
335.00	431.0
341.10	508.0
345.90	445.0
351.90	671.0
353.10	354.0
354.00	1026.0
365.00	3363.0
366.00	577.0
370.80	312.0
372.00	1372.0
382.90	446.0
402.00	726.0
403.00	940.0
404.00	385.0
421.00	620.0
422.00	396.0
423.00	4056.0
424.00	1353.0
441.10	12324.0
442.10	78864.0
443.10	16379.0
444.00	1871.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

m/z	Abundance
37.00	424.0
38.00	866.0
39.00	5675.0
40.00	678.0
41.00	537.0
43.00	390.0
44.00	1773.0
49.00	707.0
50.00	19792.0
51.00	79088.0
52.00	4196.0
55.00	736.0
56.00	2534.0
57.00	6356.0
61.00	1297.0
62.00	956.0
63.00	3059.0
64.00	653.0
65.10	1600.0
66.90	350.0
67.90	1135.0
69.00	73432.0
70.10	334.0
73.00	894.0
74.00	8103.0
75.00	12590.0
76.10	4196.0
77.00	84864.0
78.00	6001.0
79.00	5643.0
80.00	4244.0
81.00	6852.0
82.00	1708.0
83.00	1837.0
84.00	362.0
85.00	1427.0
86.00	1994.0
87.00	673.0
87.90	377.0
91.00	1430.0
92.00	1750.0
93.00	10014.0
93.90	837.0
95.90	512.0
97.10	420.0
98.00	7737.0
99.00	6386.0
100.00	462.0
101.00	3549.0
101.80	318.0
103.00	1213.0
104.00	2276.0
105.00	2360.0
106.10	557.0
107.00	28128.0
108.00	4347.0
109.10	849.0
110.00	46504.0
111.00	6893.0
112.00	900.0
115.90	1303.0
117.00	20864.0
118.00	1866.0
119.00	390.0
120.00	398.0
122.00	1782.0
123.00	2760.0
124.00	1351.0
124.90	1187.0
127.00	88256.0
128.00	6806.0
129.00	33872.0
130.00	3122.0
131.00	642.0
132.00	309.0
134.00	1096.0
135.00	2624.0
136.00	918.0
137.00	1522.0
138.00	362.0
139.90	327.0
141.00	4044.0
142.00	1459.0
143.00	1006.0
144.00	442.0
145.00	378.0
146.00	488.0
146.90	2836.0
148.00	5769.0
149.00	1294.0
150.00	368.0
151.30	676.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

151.80	399.0
153.00	1345.0
154.00	1222.0
155.10	2231.0
156.00	3639.0
157.00	906.0
158.00	748.0
159.00	784.0
159.90	1502.0
161.00	2177.0
162.00	680.0
163.90	535.0
164.90	1695.0
166.00	1672.0
167.00	8709.0
168.00	4691.0
168.90	705.0
170.10	317.0
171.00	355.0
172.00	1111.0
172.90	1335.0
174.00	1738.0
175.00	3755.0
176.10	953.0
177.00	1877.0
177.90	801.0
179.00	6904.0
180.00	5534.0
181.00	2167.0
182.00	600.0
183.90	621.0
185.00	3416.0
186.00	26576.0
187.00	7175.0
188.00	739.0
189.00	1785.0
191.00	1210.0
192.00	2130.0
193.00	2400.0
194.00	601.0
194.90	343.0
196.00	4492.0
196.80	1017.0
198.00	184000.0
199.00	12693.0
200.00	875.0
201.60	655.0
203.00	1500.0
204.00	7058.0
205.00	11817.0
206.00	47832.0
207.00	6732.0
208.00	1812.0
209.00	857.0
210.00	887.0
211.00	1957.0
211.80	590.0
214.90	799.0
216.00	1189.0
217.00	13512.0
218.00	1777.0
221.00	7107.0
222.00	1862.0
223.00	3181.0
224.10	27080.0
225.00	6703.0
226.10	857.0
227.00	11216.0
228.00	1650.0
229.00	2495.0
231.00	1103.0
233.90	771.0
235.00	1049.0
235.80	554.0
237.00	1460.0
239.00	437.0
241.00	895.0
242.00	1520.0
243.00	1602.0
244.00	20536.0
245.00	2622.0
246.00	4338.0
247.00	1082.0
249.00	1052.0
252.90	527.0
254.00	740.0
255.00	110528.0
256.00	17104.0
257.00	1567.0
258.00	6650.0
259.00	1301.0
265.00	2528.0
265.90	428.0
272.10	339.0
273.00	3394.0
274.00	8132.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

275.00	48824.0
276.00	6386.0
277.00	3627.0
278.00	873.0
281.00	304.0
282.90	468.0
284.00	357.0
284.90	972.0
293.00	974.0
293.90	307.0
295.00	408.0
296.00	13672.0
297.00	1786.0
303.00	1848.0
304.00	420.0
310.00	349.0
314.00	555.0
314.90	1670.0
316.00	849.0
321.00	414.0
323.00	4560.0
324.10	931.0
327.00	675.0
328.00	477.0
332.00	453.0
332.80	362.0
334.10	3093.0
335.00	542.0
341.00	635.0
346.00	971.0
352.00	981.0
353.00	973.0
354.00	1146.0
365.00	5968.0
366.00	876.0
370.90	343.0
372.00	2122.0
373.00	509.0
383.00	481.0
390.00	443.0
402.00	1162.0
403.10	1186.0
404.00	580.0
421.00	1182.0
422.00	805.0
423.10	7063.0
424.00	1886.0
441.10	22624.0
442.10	145408.0
443.10	28424.0
444.00	2897.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

m/z	Abundance
38.10	868.0
39.00	4663.0
40.00	1062.0
41.00	709.0
43.00	576.0
43.90	1502.0
49.00	770.0
50.00	14883.0
51.00	60824.0
52.00	3287.0
55.00	696.0
56.00	2000.0
57.00	3998.0
57.90	413.0
61.00	677.0
62.00	1073.0
63.00	2287.0
64.00	375.0
65.00	1131.0
66.90	342.0
68.00	1114.0
69.00	56952.0
70.00	354.0
73.10	999.0
74.00	6516.0
75.00	10038.0
76.00	3173.0
77.00	67744.0
78.00	3983.0
79.00	4440.0
80.00	3244.0
81.00	5306.0
82.00	1165.0
83.00	1219.0
83.90	327.0
85.00	861.0
86.00	1648.0
86.90	848.0
87.90	368.0
91.00	1484.0
92.00	1356.0
93.00	7716.0
93.90	744.0
95.00	562.0
95.90	468.0
98.00	5416.0
99.00	4631.0
99.90	561.0
101.00	2336.0
103.00	912.0
104.00	1526.0
105.00	1527.0
106.10	539.0
107.00	21744.0
108.00	3043.0
109.00	569.0
110.00	36984.0
111.00	5597.0
111.90	579.0
115.90	1000.0
117.00	16888.0
118.00	1247.0
118.90	325.0
120.00	463.0
122.00	1267.0
122.90	1858.0
124.00	661.0
125.00	995.0
127.00	68408.0
128.00	5404.0
129.00	27176.0
130.00	2159.0
130.90	519.0
132.00	397.0
133.00	469.0
134.00	817.0
135.00	2247.0
136.00	1068.0
137.10	1405.0
138.00	400.0
140.00	341.0
141.00	3390.0
142.00	1116.0
143.00	796.0
145.00	375.0
146.00	510.0
147.00	2062.0
148.00	4254.0
148.90	1089.0
149.90	459.0
151.00	424.0
151.60	425.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

153.00	1255.0
154.00	882.0
155.00	2428.0
156.00	3606.0
157.00	513.0
157.90	593.0
159.00	617.0
159.90	975.0
161.00	1815.0
162.00	644.0
165.00	1646.0
166.10	981.0
167.00	6666.0
168.00	4575.0
168.90	767.0
171.00	301.0
171.90	759.0
173.00	867.0
174.00	1337.0
175.00	3550.0
176.00	809.0
176.90	1658.0
178.00	606.0
179.00	5050.0
180.00	4711.0
181.00	1893.0
182.10	344.0
183.90	597.0
185.00	2960.0
186.00	22920.0
187.00	6333.0
188.10	905.0
189.00	1748.0
191.00	737.0
192.00	2164.0
193.00	2640.0
193.90	602.0
196.00	4074.0
196.90	955.0
198.00	158912.0
199.00	10252.0
200.00	766.0
201.30	768.0
203.00	973.0
204.00	6008.0
205.00	9210.0
206.10	42128.0
207.00	6149.0
208.00	1824.0
209.00	508.0
209.90	571.0
211.00	1798.0
215.00	513.0
216.00	1040.0
217.00	12206.0
218.00	1679.0
221.00	5533.0
221.90	1756.0
223.00	2745.0
224.00	25776.0
225.00	5968.0
226.00	756.0
227.00	10284.0
228.00	1360.0
229.00	1905.0
231.00	846.0
234.00	881.0
234.90	779.0
236.00	388.0
237.10	912.0
239.00	488.0
241.00	486.0
242.00	1150.0
243.00	1419.0
244.10	17432.0
245.00	2378.0
246.00	3753.0
247.00	836.0
249.00	940.0
252.90	475.0
254.10	619.0
255.00	99760.0
256.00	14842.0
257.00	1388.0
258.00	5437.0
259.00	1046.0
265.00	2075.0
273.00	2577.0
274.00	8043.0
275.00	43568.0
276.00	5373.0
277.00	3633.0
278.00	668.0
283.00	314.0
284.10	399.0
285.00	706.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

293.00	1010.0
296.00	11546.0
297.00	1530.0
302.00	354.0
303.00	1384.0
303.90	305.0
314.00	754.0
314.90	946.0
316.00	1231.0
320.90	402.0
323.10	3846.0
324.00	901.0
326.90	842.0
328.00	416.0
333.00	452.0
334.10	2978.0
335.00	658.0
341.00	695.0
346.00	613.0
352.00	1280.0
353.00	811.0
354.00	1596.0
355.10	384.0
365.00	5812.0
365.90	562.0
371.00	345.0
372.10	1679.0
373.00	547.0
383.00	442.0
389.90	330.0
391.10	338.0
402.00	782.0
403.00	1337.0
404.00	632.0
421.00	906.0
422.00	954.0
423.00	8466.0
424.00	1737.0
441.10	24240.0
442.10	152832.0
443.10	27752.0
444.00	2234.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53