

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071824\
Data File : BM046712.D
Acq On : 18 Jul 2024 21:22
Operator : MA/JU
Sample : P3246-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-CTW2-BL-02

Quant Time: Jul 19 01:15:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM071724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 18 06:01:46 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.487	152	89281	20.000	ng	0.00
21) Naphthalene-d8	10.245	136	309007	20.000	ng	0.00
39) Acenaphthene-d10	14.133	164	194599	20.000	ng	0.00
64) Phenanthrene-d10	16.886	188	370297	20.000	ng	0.00
76) Chrysene-d12	21.109	240	319992	20.000	ng	-0.01
86) Perylene-d12	23.874	264	415689	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.134	112	582776	132.582	ng	0.00
7) Phenol-d6	6.687	99	724290	129.403	ng	0.00
23) Nitrobenzene-d5	8.628	82	504396	84.749	ng	0.00
42) 2,4,6-Tribromophenol	15.633	330	342198	113.371	ng	0.00
45) 2-Fluorobiphenyl	12.745	172	1240545	90.047	ng	0.00
79) Terphenyl-d14	19.539	244	1567524	88.587	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071824\
Data File : BM046712.D
Acq On : 18 Jul 2024 21:22
Operator : MA/JU
Sample : P3246-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-CTW2-BL-02

Quant Time: Jul 19 01:15:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM071724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 18 06:01:46 2024
Response via : Initial Calibration

