

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071922\
 Data File : BM035872.D
 Acq On : 19 Jul 2022 19:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 20 02:46:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM071822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 18 17:08:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.533	0.523	1.9	116	0.00
3	Pyridine	1.417	1.471	-3.8	121	0.00
4	n-Nitrosodimethylamine	0.592	0.626	-5.7	123	0.00
5 S	2-Fluorophenol	1.290	1.297	-0.5	117	0.00
6	Aniline	1.782	1.896	-6.4	124	0.00
7 S	Phenol-d6	1.622	1.691	-4.3	122	0.00
8	2-Chlorophenol	1.333	1.356	-1.7	119	0.00
9	Benzaldehyde	0.911	0.883	3.1	128	0.00
10 C	Phenol	1.634	1.706	-4.4	123	0.00
11	bis(2-Chloroethyl)ether	1.318	1.376	-4.4	124	0.00
12	1,3-Dichlorobenzene	1.487	1.497	-0.7	118	0.00
13 C	1,4-Dichlorobenzene	1.514	1.529	-1.0	119	0.00
14	1,2-Dichlorobenzene	1.464	1.477	-0.9	119	0.00
15	Benzyl Alcohol	1.069	1.144	-7.0	125	0.00
16	2,2'-oxybis(1-Chloropropane	1.716	1.836	-7.0	125	0.00
17	2-Methylphenol	1.061	1.120	-5.6	124	0.00
18	Hexachloroethane	0.545	0.552	-1.3	118	0.00
19 P	n-Nitroso-di-n-propylamine	0.941	1.036	-10.1	129	0.00
20	3+4-Methylphenols	1.434	1.525	-6.3	125	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
22	Acetophenone	0.494	0.515	-4.3	128	0.00
23 S	Nitrobenzene-d5	0.374	0.387	-3.5	126	0.00
24	Nitrobenzene	0.351	0.362	-3.1	126	0.00
25	Isophorone	0.585	0.618	-5.6	130	0.00
26 C	2-Nitrophenol	0.158	0.159	-0.6	122	0.00
27	2,4-Dimethylphenol	0.250	0.257	-2.8	126	0.00
28	bis(2-Chloroethoxy)methane	0.407	0.427	-4.9	129	0.00
29 C	2,4-Dichlorophenol	0.277	0.284	-2.5	124	0.00
30	1,2,4-Trichlorobenzene	0.319	0.320	-0.3	123	0.00
31	Naphthalene	1.016	1.042	-2.6	126	0.00
32	Benzoic acid	0.189	0.186	1.6	119	0.02
33	4-Chloroaniline	0.407	0.426	-4.7	128	0.00
34 C	Hexachlorobutadiene	0.183	0.184	-0.5	123	0.00
35	Caprolactam	0.085	0.085	0.0	126	0.00
36 C	4-Chloro-3-methylphenol	0.284	0.299	-5.3	130	0.00
37	2-Methylnaphthalene	0.661	0.697	-5.4	130	0.00
38	1-Methylnaphthalene	0.643	0.672	-4.5	129	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.00
40	1,2,4,5-Tetrachlorobenzene	0.571	0.581	-1.8	127	0.00
41 P	Hexachlorocyclopentadiene	0.308	0.314	-1.9	125	0.00
42 S	2,4,6-Tribromophenol	0.224	0.224	0.0	128	0.00
43 C	2,4,6-Trichlorophenol	0.378	0.388	-2.6	127	0.00
44	2,4,5-Trichlorophenol	0.398	0.407	-2.3	128	0.00
45 S	2-Fluorobiphenyl	1.415	1.473	-4.1	130	0.00
46	1,1'-Biphenyl	1.521	1.580	-3.9	130	0.00
47	2-Chloronaphthalene	1.169	1.201	-2.7	129	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071922\
 Data File : BM035872.D
 Acq On : 19 Jul 2022 19:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 20 02:46:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM071822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 18 17:08:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.322	0.338	-5.0	131	0.00
49	Acenaphthylene	1.801	1.864	-3.5	130	0.00
50	Dimethylphthalate	1.316	1.334	-1.4	130	0.00
51	2,6-Dinitrotoluene	0.275	0.282	-2.5	128	0.00
52 C	Acenaphthene	1.079	1.109	-2.8	130	0.00
53	3-Nitroaniline	0.331	0.343	-3.6	130	0.00
54 P	2,4-Dinitrophenol	0.136	0.129	5.1	126	0.00
55	Dibenzofuran	1.728	1.777	-2.8	131	0.00
56 P	4-Nitrophenol	0.260	0.257	1.2	125	0.00
57	2,4-Dinitrotoluene	0.360	0.370	-2.8	129	0.00
58	Fluorene	1.357	1.396	-2.9	131	0.00
59	2,3,4,6-Tetrachlorophenol	0.320	0.324	-1.3	129	0.00
60	Diethylphthalate	1.304	1.310	-0.5	129	0.00
61	4-Chlorophenyl-phenylether	0.660	0.683	-3.5	132	0.00
62	4-Nitroaniline	0.339	0.349	-2.9	128	0.00
63	Azobenzene	1.339	1.407	-5.1	133	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
65	4,6-Dinitro-2-methylphenol	0.100	0.097	3.0	121	0.00
66 c	n-Nitrosodiphenylamine	0.587	0.619	-5.5	130	0.00
67	4-Bromophenyl-phenylether	0.201	0.209	-4.0	129	0.00
68	Hexachlorobenzene	0.233	0.242	-3.9	129	0.00
69	Atrazine	0.154	0.164	-6.5	125	0.00
70 C	Pentachlorophenol	0.130	0.135	-3.8	128	0.00
71	Phenanthrene	1.052	1.078	-2.5	128	0.00
72	Anthracene	1.023	1.055	-3.1	128	0.00
73	Carbazole	1.037	1.058	-2.0	127	0.00
74	Di-n-butylphthalate	1.056	1.064	-0.8	127	0.00
75 C	Fluoranthene	1.144	1.132	1.0	125	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
77	Benzydine	0.351	0.331	5.7	108	0.00
78	Pyrene	1.308	1.393	-6.5	125	0.00
79 S	Terphenyl-d14	1.028	1.104	-7.4	124	0.00
80	Butylbenzylphthalate	0.467	0.469	-0.4	120	0.00
81	Benzo(a)anthracene	1.266	1.291	-2.0	120	0.00
82	3,3'-Dichlorobenzidine	0.297	0.295	0.7	117	0.00
83	Chrysene	1.208	1.235	-2.2	120	0.00
84	Bis(2-ethylhexyl)phthalate	0.658	0.662	-0.6	119	0.00
85 c	Di-n-octyl phthalate	1.044	1.020	2.3	116	0.00
86 I	Perylene-d12	1.000	1.000	0.0	118	0.00
87	Indeno(1,2,3-cd)pyrene	1.461	1.508	-3.2	117	0.00
88	Benzo(b)fluoranthene	1.191	1.211	-1.7	114	0.00
89	Benzo(k)fluoranthene	1.232	1.286	-4.4	121	0.00
90 C	Benzo(a)pyrene	1.013	1.037	-2.4	116	0.00
91	Dibenzo(a,h)anthracene	1.214	1.250	-3.0	116	0.00
92	Benzo(g,h,i)perylene	1.198	1.226	-2.3	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071922\
Data File : BM035872.D
Acq On : 19 Jul 2022 19:56
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jul 20 02:46:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM071822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 18 17:08:15 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0