

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM072115\
 Data File : BM002001.D
 Acq On : 22 Jul 2015 00:22
 Operator : TP/UM
 Sample : SSTD08064
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD08064

Quant Time: Jul 22 03:18:03 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072115.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 22 00:57:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	76643	20.00	ng/ul	0.00
18) Naphthalene-d8	10.41	136	332363	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.29	164	206433	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.03	188	457848	20.00	ng/ul	0.00
77) Chrysene-d12	21.23	240	576015	20.00	ng/ul	0.00
85) Perylene-d12	23.44	264	363989	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.09	96	63008	39.64	ng/uL	0.00
5) Phenol-d5	6.80	99	510120	81.47	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.96	67	260217	75.01	ng/ul	0.00
9) 2-Chlorophenol-d4	7.16	132	429281	87.02	ng/ul	0.00
13) 4-Methylphenol-d8	8.35	113	432365	85.93	ng/ul	0.01
19) Nitrobenzene-d5	8.79	128	209242	86.85	ng/ul	0.00
22) 2-Nitrophenol-d4	9.50	143	244240	91.49	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.04	165	475753	87.47	ng/ul	0.00
29) 4-Chloroaniline-d4	10.56	131	446533	79.13	ng/ul	0.00
43) Dimethylphthalate-d6	13.70	166	1322672	80.62	ng/ul	0.01
46) Acenaphthylene-d8	13.98	160	1643714	82.60	ng/ul	0.00
51) 4-Nitrophenol-d4	14.52	143	280822	101.97	ng/ul	0.02
57) Fluorene-d10	15.29	176	1439905	100.36	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.42	200	336582	117.15	ng/ul	0.01
70) Anthracene-d10	17.14	188	1762210	83.49	ng/ul	0.01
78) Pyrene-d10	19.44	212	1889462	74.74	ng/ul	0.01
89) Benzo(a)pyrene-d12	23.30	264	1564159	85.16	ng/ul	0.01

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.12	88	65438	38.55	ng/uL	95
4) Benzaldehyde	6.77	77	214613	71.45	ng/ul	94
6) Phenol	6.83	94	524309	81.37	ng/ul	99
8) Bis(2-Chloroethyl)ether	7.06	93	381109	80.62	ng/ul	99
10) 2-Chlorophenol	7.19	128	435031	86.53	ng/ul	98
11) 2-Methylphenol	8.07	108	412475	84.93	ng/ul	98
12) 2,2'-oxybis(1-Chloropropan	8.16	45	369853	69.22	ng/ul	100
14) Acetophenone	8.45	105	668412	83.96	ng/ul	99
15) N-Nitroso-di-n-propylamine	8.45	70	324397	80.75	ng/ul	98
16) 4-Methylphenol	8.42	108	451415	85.04	ng/ul	95
17) Hexachloroethane	8.69	117	164459	82.53	ng/ul	95
20) Nitrobenzene	8.83	77	473441	78.68	ng/ul	98
21) Isophorone	9.36	82	876273	80.41	ng/ul	99
23) 2-Nitrophenol	9.54	139	262161	90.14	ng/ul	98
24) 2,4-Dimethylphenol	9.60	107	508665	81.62	ng/ul	100
25) Bis(2-Chloroethoxy)methane	9.84	93	530545	77.10	ng/ul	100
27) 2,4-Dichlorophenol	10.07	162	457859	86.06	ng/ul	98
28) Naphthalene	10.46	128	1369295	82.68	ng/ul	99
30) 4-Chloroaniline	10.58	127	467816	80.03	ng/ul	99
31) Hexachlorobutadiene	10.74	225	320009	82.78	ng/ul	98
32) Caprolactam	11.39	113	145668	60.38	ng/ul	97
33) 4-Chloro-3-methylphenol	11.73	107	471623	81.76	ng/ul	97
34) 2-Methylnaphthalene	12.09	142	1027989	83.21	ng/ul	100

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM072115\
 Data File : BM002001.D
 Acq On : 22 Jul 2015 00:22
 Operator : TP/UM
 Sample : SSTD08064
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD08064

Quant Time: Jul 22 03:18:03 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072115.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 22 00:57:39 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 1,2,4,5-Tetrachlorobenzene	12.46	216	618455	79.42	ng/ul	99
37) Hexachlorocyclopentadiene	12.43	237	372624	85.88	ng/ul	99
38) 2,4,6-Trichlorophenol	12.71	196	392204	84.31	ng/ul	99
39) 2,4,5-Trichlorophenol	12.79	196	427631	86.28	ng/ul	98
40) 1,1'-Biphenyl	13.12	154	1367453	82.01	ng/ul	99
41) 2-Chloronaphthalene	13.16	162	1074424	82.82	ng/ul	99
42) 2-Nitroaniline	13.37	65	275025	78.69	ng/ul	97
44) Dimethylphthalate	13.75	163	1340212	79.74	ng/ul	100
45) 2,6-Dinitrotoluene	13.88	165	291248	86.10	ng/ul	98
47) Acenaphthylene	14.01	152	1678103	82.53	ng/ul	99
48) 3-Nitroaniline	14.21	138	243328	79.76	ng/ul	95
49) Acenaphthene	14.35	153	1102023	79.90	ng/ul	99
50) 2,4-Dinitrophenol	14.42	184	189295	93.59	ng/ul	92
52) 4-Nitrophenol	14.53	109	235891	93.67	ng/ul	99
53) Dibenzofuran	14.69	168	1980863	101.15	ng/ul	99
54) 2,4-Dinitrotoluene	14.67	165	529992	107.38	ng/ul	97
55) 2,3,4,6-Tetrachlorophenol	14.92	232	464169	105.69	ng/ul	97
56) Diethylphthalate	15.12	149	1685145	101.41	ng/ul	99
58) Fluorene	15.34	166	1718704	103.96	ng/ul	98
59) 4-Chlorophenyl-phenylether	15.33	204	921416	102.62	ng/ul	96
60) 4-Nitroaniline	15.39	138	364999	114.13	ng/ul	91
63) 4,6-Dinitro-2-methylphenol	15.44	198	352841	115.21	ng/ul#	98
64) N-Nitrosodiphenylamine	15.56	169	1456971	106.68	ng/ul	99
65) 4-Bromophenyl-phenylether	16.23	248	438532	84.77	ng/ul	98
66) Hexachlorobenzene	16.34	284	482744	84.39	ng/ul	99
67) Atrazine	16.51	200	416958	80.02	ng/ul	100
68) Pentachlorophenol	16.69	266	300885	87.72	ng/ul	98
69) Phenanthrene	17.08	178	2023310	82.16	ng/ul	99
71) Anthracene	17.17	178	2093626	81.23	ng/ul	99
72) 1,2,3,4-Tetrachlorobenzene	13.07	216	638471	90.54	ng/uL	100
73) Pentachlorobenzene	14.60	250	709435	102.85	ng/uL	99
74) Carbazole	17.44	167	1823198	84.39	ng/ul	100
75) Di-n-butylphthalate	18.00	149	2207500	88.97	ng/ul	100
76) Fluoranthene	19.10	202	2364707	85.89	ng/ul	99
79) Pyrene	19.47	202	2453199	74.00	ng/ul	99
80) Butylbenzylphthalate	20.37	149	1221118	95.41	ng/ul	99
81) 3,3'-Dichlorobenzidine	21.15	252	844277	77.78	ng/ul	100
82) Benzo(a)anthracene	21.22	228	2826970	82.28	ng/ul	99
83) Bis(2-ethylhexyl)phthalate	21.14	149	1766510	92.36	ng/ul	99
84) Chrysene	21.27	228	2554372	77.21	ng/ul	99
86) Di-n-octyl phthalate	22.01	149	2816633	117.04	ng/ul	100
87) Benzo(b)fluoranthene	22.78	252	1909949	78.05	ng/ul	99
88) Benzo(k)fluoranthene	22.82	252	1869552	81.29	ng/ul	99
90) Benzo(a)pyrene	23.35	252	1769579	84.38	ng/ul	99
91) Indeno(1,2,3-cd)pyrene	25.68	276	1881810	83.62	ng/ul	98
92) Dibenzo(a,h)anthracene	25.69	278	1610264	84.48	ng/ul	98
93) Benzo(g,h,i)perylene	26.36	276	1545184	83.02	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM072115\
 Data File : BM002001.D
 Acq On : 22 Jul 2015 00:22
 Operator : TP/UM
 Sample : SSTD08064
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampled :
 SSTD08064

Quant Time: Jul 22 03:18:03 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072115.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 22 00:57:39 2015
 Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 38.55 ng/uL

RT: 3.12 min Scan# 74

Delta R.T. 0.00 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064


Tgt Ion: 88 Resp: 65438

Ion Ratio Lower Upper

88 100

43 24.8 22.5 33.7




58 64.4 48.5 72.7



Abundance Ion 88.00 (87.70 to 88.70): BM001999.D (-70) (-)

Ion 43.00 (42.70 to 43.70): BM001999.D (-70) (-)

Ion 58.00 (57.70 to 58.70): BM001999.D (-70) (-)



#4

Benzaldehyde

Concen: 71.45 ng/uL

RT: 6.77 min Scan# 694

Delta R.T. 0.00 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22


Tgt Ion: 77 Resp: 214613

Ion Ratio Lower Upper

77 100

105 103.3 79.6 119.4



106 98.0 72.5 108.7




Abundance Ion 77.00 (76.70 to 77.70): BM001999.D (-688) (-)

Ion 105.00 (104.70 to 105.70): BM001999.D (-688) (-)

Ion 106.00 (105.70 to 106.70): BM001999.D (-688) (-)





#6

Phenol

Concen: 81.37 ng/ul

RT: 6.83 min Scan# 704

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064


Tgt Ion: 94 Resp: 524309

Ion Ratio Lower Upper

94 100

65 26.6 21.8 32.6

66 38.1 30.6 46.0






Abundance Ion 94.00 (93.70 to 94.70): BM001999.D (-696) (-)

Ion 65.00 (64.70 to 65.70): BM001999.D (-696) (-)

Ion 66.00 (65.70 to 66.70): BM001999.D (-696) (-)

6.83

Time-->



#8

Bis(2-Chloroethyl)ether

Concen: 80.62 ng/ul

RT: 7.06 min Scan# 743

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22


Tgt Ion: 93 Resp: 381109

Ion Ratio Lower Upper

93 100

63 67.0 53.8 80.8

95 32.7 26.9 40.3





Abundance Ion 93.00 (92.70 to 93.70): BM001999.D (-736) (-)


Ion 63.00 (62.70 to 63.70): BM001999.D (-736) (-)

Ion 95.00 (94.70 to 95.70): BM001999.D (-736) (-)

7.06

Time-->









#10
2-Chlorophenol
Concen: 86.53 ng/ul
RT: 7.19 min Scan# 765
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Instrument :
BNA_M
ClientSampleId :
SSTD08064

Tgt Ion:128 Resp: 435031
Ion Ratio Lower Upper
128 100
64 40.4 33.2 49.8
130 32.6 25.4 38.2




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 64.00 (63.70 to 64.70): BM0
Ion 130.00 (129.70 to 130.70): E






#11
2-Methylphenol
Concen: 84.93 ng/ul
RT: 8.07 min Scan# 915
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion:108 Resp: 412475
Ion Ratio Lower Upper
108 100
107 88.1 72.3 108.5



Abundance Ion 108.00 (107.70 to 108.70): E
Ion 107.00 (106.70 to 107.70): E









#12
2,2'-oxybis(1-chloropropane)
Concen: 69.22 ng/ul
RT: 8.16 min Scan# 931
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22




Instrument :
BNA_M
ClientSampleId :
SSTD08064


Tgt Ion	Ratio	Lower	Upper
45	100		
77	20.4	16.5	24.7
79	16.3	13.0	19.4



#14
Acetophenone
Concen: 83.96 ng/ul
RT: 8.45 min Scan# 980
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion	Ratio	Lower	Upper
105	100		
77	74.8	59.6	89.4
51	21.5	17.6	26.4









#15
N-Nitroso-di-n-propylamine
Concen: 80.75 ng/ul
RT: 8.45 min Scan# 980
Delta R.T. 0.02 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22




Instrument :
BNA_M
ClientSampleId :
SSTD08064


Tgt Ion:	70	Resp:	324397
Ion	Ratio	Lower	Upper
70	100		
42	49.6	38.2	57.2
101	10.5	8.1	12.1
130	24.3	18.5	27.7



#16
4-Methylphenol
Concen: 85.04 ng/ul
RT: 8.42 min Scan# 974
Delta R.T. 0.02 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion:	108	Resp:	451415
Ion	Ratio	Lower	Upper
108	100		
107	107.5	90.0	135.0





#17

Hexachloroethane

Concen: 82.53 ng/ul

RT: 8.69 min Scan# 1021

Delta R.T. 0.00 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064


Tgt Ion:117 Resp: 164459

Ion Ratio Lower Upper

117 100

201 121.9 92.4 138.6




199 76.9 59.6 89.4



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 78.68 ng/ul

RT: 8.83 min Scan# 1045

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22


Tgt Ion: 77 Resp: 473441

Ion Ratio Lower Upper

77 100

123 48.8 37.4 56.2



65 13.6 11.0 16.4




Abundance Ion 77.00 (76.70 to 77.70): BM0

Ion 123.00 (122.70 to 123.70): E

Ion 65.00 (64.70 to 65.70): BM0





#21

Isophorone

Concen: 80.41 ng/ul

RT: 9.36 min Scan# 1134

Delta R.T. 0.02 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064





Tgt Ion: 82 Resp: 876273

Ion Ratio Lower Upper

82 100

95 8.0 6.3 9.5

138 19.3 15.1 22.7



#23

2-Nitrophenol

Concen: 90.14 ng/ul

RT: 9.54 min Scan# 1165

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22




Tgt Ion: 139 Resp: 262161


Ion Ratio Lower Upper

139 100

65 46.0 37.7 56.5

109 31.4 25.8 38.8









#24
 2,4-Dimethylphenol
 Concen: 81.62 ng/ul
 RT: 9.60 min Scan# 1175
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22




Instrument :
 BNA_M
 ClientSampleId :
 SSTD08064


Tgt Ion	Ratio	Lower	Upper
107	100		
121	48.9	39.0	58.4
122	83.7	66.6	99.8



#25
 Bis(2-Chloroethoxy)methane
 Concen: 77.10 ng/ul
 RT: 9.84 min Scan# 1216
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Tgt Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.3	39.5
123	11.9	9.4	14.0









#27
2,4-Dichlorophenol
Concen: 86.06 ng/ul
RT: 10.07 min Scan# 1255
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22




Instrument :
BNA_M
ClientSampleId :
SSTD08064


Tgt Ion	Ratio	Lower	Upper
162	100		
164	65.9	53.1	79.7
98	33.4	28.2	42.4



#28
Naphthalene
Concen: 82.68 ng/ul
RT: 10.46 min Scan# 1322
Delta R.T. 0.00 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.4	12.6
127	13.1	10.3	15.5









#30
4-Chloroaniline
Concen: 80.03 ng/ul
RT: 10.58 min Scan# 1342
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22




Instrument :
BNA_M
ClientSampleId :
SSTD08064


Tgt Ion:127 Resp: 467816
Ion Ratio Lower Upper
127 100
129 31.6 25.6 38.4



#31
Hexachlorobutadiene
Concen: 82.78 ng/ul
RT: 10.74 min Scan# 1369
Delta R.T. 0.00 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion:225 Resp: 320009
Ion Ratio Lower Upper
225 100
223 62.2 48.0 72.0
227 63.3 49.9 74.9





#32

Caprolactam

Concen: 60.38 ng/ul

RT: 11.39 min Scan# 1480

Delta R.T. 0.06 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064


Tgt Ion:113 Resp: 145668

Ion Ratio Lower Upper

113 100

55 118.8 97.8 146.6




56 98.6 81.8 122.6



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BM0

Ion 56.00 (55.70 to 56.70): BM0



#33

4-Chloro-3-methylphenol

Concen: 81.76 ng/ul

RT: 11.73 min Scan# 1537

Delta R.T. 0.02 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22


Tgt Ion:107 Resp: 471623

Ion Ratio Lower Upper

107 100

144 25.9 20.6 30.8



142 81.9 68.3 102.5




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#34
2-Methylnaphthalene
Concen: 83.21 ng/ul
RT: 12.09 min Scan# 1598
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22




Instrument :
BNA_M
ClientSampleId :
SSTD08064


Tgt Ion:142 Resp: 1027989
Ion Ratio Lower Upper
142 100
141 87.7 70.0 105.0



#36
1,2,4,5-Tetrachlorobenzene
Concen: 79.42 ng/ul
RT: 12.46 min Scan# 1662
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion:216 Resp: 618455
Ion Ratio Lower Upper
216 100
214 77.6 61.8 92.6
179 20.0 16.0 24.0
108 13.3 11.3 16.9





#37

Hexachlorocyclopentadiene

Concen: 85.88 ng/ul

RT: 12.43 min Scan# 1657

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064


Tgt Ion:237 Resp: 372624

Ion Ratio Lower Upper

237 100

235 62.3 50.3 75.5




272 12.6 9.1 13.7



Abundance Ion 237.00 (236.70 to 237.70): E

Ion 235.00 (234.70 to 235.70): E

Ion 272.00 (271.70 to 272.70): E



#38

2,4,6-Trichlorophenol

Concen: 84.31 ng/ul

RT: 12.71 min Scan# 1704

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22


Tgt Ion:196 Resp: 392204

Ion Ratio Lower Upper

196 100

198 96.7 77.8 116.6



200 31.1 25.8 38.6




Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E









#39
 2,4,5-Trichlorophenol
 Concen: 86.28 ng/ul
 RT: 12.79 min Scan# 1717
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22




Instrument :
 BNA_M
 ClientSampleId :
 SSTD08064


Tgt Ion:196 Resp: 427631
 Ion Ratio Lower Upper
 196 100
 198 96.9 78.9 118.3
 200 30.7 25.9 38.9



#40
 1,1'-Biphenyl
 Concen: 82.01 ng/ul
 RT: 13.12 min Scan# 1773
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Tgt Ion:154 Resp: 1367453
 Ion Ratio Lower Upper
 154 100
 153 40.1 31.2 46.8
 76 10.9 8.8 13.2









#41
 2-Chloronaphthalene
 Concen: 82.82 ng/ul
 RT: 13.16 min Scan# 1780
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22




Instrument :
 BNA_M
 ClientSampleId :
 SST08064


Tgt Ion: 162 Resp: 1074424
 Ion Ratio Lower Upper
 162 100
 164 33.3 26.7 40.1
 127 37.5 30.8 46.2



#42
 2-Nitroaniline
 Concen: 78.69 ng/ul
 RT: 13.37 min Scan# 1817
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Tgt Ion: 65 Resp: 275025
 Ion Ratio Lower Upper
 65 100
 92 74.1 58.2 87.2
 138 122.9 95.0 142.6





#44

Dimethylphthalate

Concen: 79.74 ng/ul

RT: 13.75 min Scan# 1881

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064


Tgt Ion:163 Resp: 1340212

Ion Ratio Lower Upper

163 100

194 4.3 3.4 5.0




164 10.1 8.2 12.4



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#45

2,6-Dinitrotoluene

Concen: 86.10 ng/ul

RT: 13.88 min Scan# 1903

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22


Tgt Ion:165 Resp: 291248

Ion Ratio Lower Upper

165 100

89 52.8 43.2 64.8



121 19.6 17.2 25.8




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 89.00 (88.70 to 89.70): BM0

Ion 121.00 (120.70 to 121.70): E





#47

Acenaphthylene

Concen: 82.53 ng/ul

RT: 14.01 min Scan# 1925

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064


Tgt Ion:152 Resp: 1678103

Ion Ratio Lower Upper




152 100

151 19.6 15.2 22.8

153 13.1 10.4 15.6



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E



#48

3-Nitroaniline

Concen: 79.76 ng/ul

RT: 14.21 min Scan# 1959

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22


Tgt Ion:138 Resp: 243328

Ion Ratio Lower Upper



138 100


108 10.9 9.3 13.9

92 108.0 91.0 136.4



Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BM





#49

Acenaphthene

Concen: 79.90 ng/ul

RT: 14.35 min Scan# 1983

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064


Tgt Ion:153 Resp: 1102023

Ion Ratio Lower Upper

153 100

152 45.6 37.4 56.2




154 90.5 73.3 109.9



Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E



#50

2,4-Dinitrophenol

Concen: 93.59 ng/ul

RT: 14.42 min Scan# 1995

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22


Tgt Ion:184 Resp: 189295

Ion Ratio Lower Upper

184 100

63 50.0 45.0 67.6



154 60.1 44.2 66.4




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BM0

Ion 154.00 (153.70 to 154.70): E





#52

4-Nitrophenol

Concen: 93.67 ng/ul

RT: 14.53 min Scan# 2013

Delta R.T. 0.02 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064


Tgt Ion:109 Resp: 235891

Ion Ratio Lower Upper

109 100

139 126.8 101.9 152.9




65 119.5 93.8 140.8



Abundance Ion 109.00 (108.70 to 109.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 65.00 (64.70 to 65.70): BM



#53

Dibenzofuran

Concen: 101.15 ng/ul

RT: 14.69 min Scan# 2041

Delta R.T. 0.01 min

Lab File: BM002001.D


Acq: 22 Jul 2015 00:22

Tgt Ion:168 Resp: 1980863

Ion Ratio Lower Upper



168 100


139 35.6 28.2 42.4



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E









#54
 2,4-Dinitrotoluene
 Concen: 107.38 ng/ul
 RT: 14.67 min Scan# 2038
 Delta R.T. 0.02 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22




Instrument :
 BNA_M
 ClientSampleId :
 SST08064


Tgt Ion	Ratio	Lower	Upper
165	100		
63	39.9	30.3	45.5
182	3.2	2.3	3.5



#55
 2,3,4,6-Tetrachlorophenol
 Concen: 105.69 ng/ul
 RT: 14.92 min Scan# 2079
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Tgt Ion	Ratio	Lower	Upper
232	100		
131	40.5	30.4	45.6
130	2.1	1.8	2.6
166	28.9	22.1	33.1









#56
Diethylphthalate
Concen: 101.41 ng/ul
RT: 15.12 min Scan# 2114
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Instrument :
BNA_M
ClientSampleId :
SSTD08064

Tgt Ion:149 Resp: 1685145
Ion Ratio Lower Upper
149 100
177 23.2 17.8 26.6
150 12.8 10.3 15.5




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#58
Fluorene
Concen: 103.96 ng/ul
RT: 15.34 min Scan# 2151
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion:166 Resp: 1718704
Ion Ratio Lower Upper
166 100
165 95.2 77.8 116.6
167 13.7 10.6 16.0



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E









#59
4-Chlorophenyl-phenylether
Concen: 102.62 ng/ul
RT: 15.33 min Scan# 2150
Delta R.T. 0.00 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Instrument :
BNA_M
ClientSampleId :
SSTD08064

Tgt Ion:204 Resp: 921416
Ion Ratio Lower Upper
204 100
206 33.1 26.8 40.2
141 56.0 41.6 62.4




Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E






#60
4-Nitroaniline
Concen: 114.13 ng/ul
RT: 15.39 min Scan# 2159
Delta R.T. 0.02 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion:138 Resp: 364999
Ion Ratio Lower Upper
138 100
92 48.3 44.6 67.0
108 81.5 71.3 106.9



Abundance Ion 138.00 (137.70 to 138.70): E
Ion 92.00 (91.70 to 92.70): BM0
Ion 108.00 (107.70 to 108.70): E









#63
 4,6-Dinitro-2-methylphenol
 Concen: 115.21 ng/ul
 RT: 15.44 min Scan# 2168
 Delta R.T. 0.02 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Instrument :
 BNA_M
 ClientSampleId :
 SSTD08064

Tgt Ion:198 Resp: 352841
 Ion Ratio Lower Upper
 198 100
 182 3.2 3.6 5.4#
 77 21.4 17.7 26.5




Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 77.00 (76.70 to 77.70): BM






#64
 N-Nitrosodiphenylamine
 Concen: 106.68 ng/ul
 RT: 15.56 min Scan# 2188
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Tgt Ion:169 Resp: 1456971
 Ion Ratio Lower Upper
 169 100
 168 64.8 52.3 78.5
 167 34.3 27.0 40.6



Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E









#65
 4-Bromophenyl-phenylether
 Concen: 84.77 ng/ul
 RT: 16.23 min Scan# 2302
 Delta R.T. 0.00 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Instrument :
 BNA_M
 ClientSampleId :
 SSTD08064

Tgt Ion:248 Resp: 438532
 Ion Ratio Lower Upper
 248 100
 250 98.5 79.6 119.4
 141 69.9 53.5 80.3




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E






#66
 Hexachlorobenzene
 Concen: 84.39 ng/ul
 RT: 16.34 min Scan# 2321
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Tgt Ion:284 Resp: 482744
 Ion Ratio Lower Upper
 284 100
 142 30.4 24.7 37.1
 249 32.5 25.2 37.8



Abundance Ion 284.00 (283.70 to 284.70): E
 Ion 142.00 (141.70 to 142.70): E
 Ion 249.00 (248.70 to 249.70): E









#67
 Atrazine
 Concen: 80.02 ng/ul
 RT: 16.51 min Scan# 2350
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22




Instrument :
 BNA_M
 ClientSampleId :
 SSTD08064


Tgt Ion	Ratio	Lower	Upper
200	100		
173	25.9	20.9	31.3
215	49.9	39.8	59.6



#68
 Pentachlorophenol
 Concen: 87.72 ng/ul
 RT: 16.69 min Scan# 2380
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.0	48.2	72.2
268	63.5	50.2	75.2





#69

Phenanthrene

Concen: 82.16 ng/ul

RT: 17.08 min Scan# 2447

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22

Instrument :

BNA_M

ClientSampleId :

SSTD08064


Tgt Ion:178 Resp: 2023310

Ion Ratio Lower Upper

178 100

179 15.3 12.1 18.1




176 18.1 14.1 21.1



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 81.23 ng/ul

RT: 17.17 min Scan# 2463

Delta R.T. 0.01 min

Lab File: BM002001.D

Acq: 22 Jul 2015 00:22


Tgt Ion:178 Resp: 2093626

Ion Ratio Lower Upper

178 100

179 15.3 12.6 19.0



176 17.8 14.1 21.1




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E









#72
 1,2,3,4-Tetrachlorobenzene
 Concen: 90.54 ng/uL
 RT: 13.07 min Scan# 1766
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22




Instrument :
 BNA_M
 ClientSampleId :
 SSTD08064


Tgt Ion:216 Resp: 638471
 Ion Ratio Lower Upper
 216 100
 214 78.5 62.8 94.2
 218 47.6 38.5 57.7



#73
 Pentachlorobenzene
 Concen: 102.85 ng/uL
 RT: 14.60 min Scan# 2026
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Tgt Ion:250 Resp: 709435
 Ion Ratio Lower Upper
 250 100
 248 62.9 50.8 76.2
 252 63.6 51.2 76.8









#74
 Carbazole
 Concen: 84.39 ng/ul
 RT: 17.44 min Scan# 2509
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22




Instrument :
 BNA_M
 ClientSampleId :
 SST08064


Tgt Ion:167 Resp: 1823198
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.0 25.4
 139 11.9 9.8 14.6



#75
 Di-n-butylphthalate
 Concen: 88.97 ng/ul
 RT: 18.00 min Scan# 2604
 Delta R.T. 0.01 min
 Lab File: BM002001.D
 Acq: 22 Jul 2015 00:22

Tgt Ion:149 Resp: 2207500
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.3 10.9
 104 5.0 4.2 6.2









#76
Fluoranthene
Concen: 85.89 ng/ul
RT: 19.10 min Scan# 2790
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Instrument :
BNA_M
ClientSampleId :
SSTD08064

Tgt Ion:202 Resp: 2364707
Ion Ratio Lower Upper
202 100
101 9.3 7.0 10.6
100 7.3 5.6 8.4




Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 100.00 (99.70 to 100.70): B






#79
Pyrene
Concen: 74.00 ng/ul
RT: 19.47 min Scan# 2853
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion:202 Resp: 2453199
Ion Ratio Lower Upper
202 100
101 10.2 7.8 11.8
100 8.6 6.5 9.7



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 100.00 (99.70 to 100.70): B









#80
Butylbenzylphthalate
Concen: 95.41 ng/ul
RT: 20.37 min Scan# 3006
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22




Instrument :
BNA_M
ClientSampleId :
SSTD08064


Tgt Ion:149 Resp: 1221118
Ion Ratio Lower Upper
149 100
91 65.3 52.0 78.0
206 23.7 19.3 28.9



#81
3,3'-Dichlorobenzidine
Concen: 77.78 ng/ul
RT: 21.15 min Scan# 3139
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion:252 Resp: 844277
Ion Ratio Lower Upper
252 100
254 64.6 51.8 77.6
126 9.5 7.9 11.9









#82
Benzo(a)anthracene
Concen: 82.28 ng/ul
RT: 21.22 min Scan# 3150
Delta R.T. 0.01 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Instrument :
BNA_M
ClientSampleId :
SSTD08064

Tgt Ion:228 Resp: 2826970
Ion Ratio Lower Upper
228 100
229 19.7 15.8 23.6
226 25.9 20.2 30.4



#83
Bis(2-ethylhexyl)phthalate
Concen: 92.36 ng/ul
RT: 21.14 min Scan# 3137
Delta R.T. 0.00 min
Lab File: BM002001.D
Acq: 22 Jul 2015 00:22

Tgt Ion:149 Resp: 1766510
Ion Ratio Lower Upper
149 100
167 29.0 22.9 34.3
279 4.8 3.4 5.2

