

Data Path : Z:\HPCHEM1\BNA M\Data\BM072116\
 Data File : BM006583.D
 Acq On : 21 Jul 2016 09:01
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02057

Quant Time: Jul 21 15:41:47 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM072016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 21 01:09:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2 S	1,4-Dioxane-d8	0.495	0.456	7.9	111	0.00
3 S	Phenol-d5	1.702	1.768	-3.9	125	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.054	1.086	-3.0	122	0.00
5 S	2-Chlorophenol-d4	1.360	1.371	-0.8	123	0.00
6 S	4-Methylphenol-d8	1.320	1.348	-2.1	122	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
8 S	Nitrobenzene-d5	0.151	0.152	-0.7	124	0.00
9 S	2-Nitrophenol-d4	0.161	0.163	-1.2	124	0.00
10 S	2,4-Dichlorophenol-d3	0.304	0.304	0.0	122	0.00
11	Naphthalene	1.031	1.039	-0.8	124	0.00
12 S	4-Chloroaniline-d4	0.335	0.411	-22.7#	122	0.00
13	2-Methylnaphthalene	0.757	0.751	0.8	121	0.00
14	1-Methylnaphthalene	0.722	0.716	0.8	121	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
16 S	Dimethylphthalate-d6	1.568	1.571	-0.2	119	0.00
17 S	Acenaphthylene-d8	2.002	2.053	-2.5	121	0.00
18	Acenaphthylene	2.144	2.182	-1.8	120	0.00
19 C	Acenaphthene	1.378	1.395	-1.2	119	0.00
20 S	4-Nitrophenol-d4	0.272	0.271	0.4	120	0.00
21 S	Fluorene-d10	1.402	1.412	-0.7	119	0.00
22	Fluorene	1.552	1.583	-2.0	118	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.104	0.101	2.9	114	0.00
25	Phenanthrene	1.125	1.153	-2.5	118	0.00
26 S	Anthracene-d10	0.950	0.967	-1.8	115	0.00
27	Anthracene	1.158	1.201	-3.7	118	0.00
28 C	Fluoranthene	1.321	1.369	-3.6	117	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
30 S	Pyrene-d10	0.911	0.930	-2.1	116	0.00
31	Pyrene	1.211	1.246	-2.9	116	0.00
32	Benzo(a)anthracene	1.212	1.234	-1.8	113	0.00
33	Chrysene	1.118	1.135	-1.5	113	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.199	1.207	-0.7	112	0.00
36	Benzo(k)fluoranthene	1.159	1.192	-2.8	113	0.00
37 S	Benzo(a)pyrene-d12	0.920	0.936	-1.7	113	0.00
38 C	Benzo(a)pyrene	1.167	1.179	-1.0	111	0.00
39	Indeno(1,2,3-cd)pyrene	1.351	1.361	-0.7	111	0.00
40	Dibenzo(a,h)anthracene	1.120	1.140	-1.8	111	0.00
41	Benzo(g,h,i)perylene	1.158	1.157	0.1	111	0.00

Data Path : Z:\HPCHEM1\BNA M\Data\BM072116\
Data File : BM006583.D
Acq On : 21 Jul 2016 09:01
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02057

Quant Time: Jul 21 15:41:47 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM072016.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jul 21 01:09:30 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0