

Data File : BM010942.D
Acq On : 21 Jul 2017 18:59
Operator : SJ/JU
Sample : I4280-03MSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD

Quant Time: Jul 21 23:25:08 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM071217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 12 14:21:16 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.43	152	299158	20.00	ng	-0.02
21) Naphthalene-d8	10.19	136	1154117	20.00	ng	-0.02
38) Acenaphthene-d10	14.08	164	799669	20.00	ng	-0.02
63) Phenanthrene-d10	16.83	188	1815923	20.00	ng	-0.02
75) Chrysene-d12	21.06	240	1486632	20.00	ng	-0.01
86) Perylene-d12	23.18	264	1337153	20.00	ng	-0.02

System Monitoring Compounds						
5) 2-Fluorophenol	5.07	112	2375256	131.52	ng	-0.01
7) Phenol-d6	6.63	99	3052263	122.60	ng	-0.01
23) Nitrobenzene-d5	8.57	82	2688237	89.53	ng	-0.02
41) 2,4,6-Tribromophenol	15.59	330	1705939	139.60	ng	-0.01
44) 2-Fluorobiphenyl	12.69	172	5531344	86.05	ng	-0.02
78) Terphenyl-d14	19.50	244	7267013	94.54	ng	-0.01

Target Compounds						Qvalue
2) 1,4-Dioxane	3.09	88	267782	33.60	ng	# 100
3) Pyridine	3.46	79	725594	33.31	ng	# 88
4) n-Nitrosodimethylamine	3.38	42	420793	37.80	ng	# 78
6) Aniline	6.77	93	853838	27.38	ng	94
8) 2-Chlorophenol	7.00	128	806550	45.11	ng	87
9) Benzaldehyde	6.59	77	352180	20.37	ng	86
10) Phenol	6.66	94	1067980	40.57	ng	91
11) bis(2-Chloroethyl)ether	6.87	93	905325	44.66	ng	94
12) 1,3-Dichlorobenzene	7.32	146	886917	40.14	ng	# 84
13) 1,4-Dichlorobenzene	7.46	146	912498	40.02	ng	# 90
14) 1,2-Dichlorobenzene	7.77	146	871367	40.10	ng	# 91
15) Benzyl Alcohol	7.67	79	1145347	48.58	ng	# 85
16) 2,2'-oxybis(1-Chloropropan	7.96	45	809065	46.03	ng	78
17) 2-Methylphenol	7.87	107	787719	46.27	ng	# 80
18) Hexachloroethane	8.48	117	391906	40.97	ng	# 82
19) n-Nitroso-di-n-propylamine	8.23	70	895256	45.39	ng	# 91
20) 3+4-Methylphenols	8.20	107	1068954	45.19	ng	86
22) Acetophenone	8.25	105	1549485	44.59	ng	# 96
24) Nitrobenzene	8.62	77	1376651	44.74	ng	91
25) Isophorone	9.14	82	2340174	47.58	ng	# 86
26) 2-Nitrophenol	9.32	139	469412	45.84	ng	# 33
27) 2,4-Dimethylphenol	9.39	122	843574	46.16	ng	# 72
28) bis(2-Chloroethoxy)methane	9.63	93	1297957	47.07	ng	97
29) 2,4-Dichlorophenol	9.85	162	945973	45.82	ng	91
30) 1,2,4-Trichlorobenzene	10.05	180	1068816	42.35	ng	99
31) Naphthalene	10.23	128	2617942	44.61	ng	99
32) Benzoic acid	9.54	122	582604	40.63	ng	# 66
33) 4-Chloroaniline	10.36	127	290182	12.27	ng	# 80
34) Hexachlorobutadiene	10.52	225	797533	40.73	ng	91
35) Caprolactam	11.15	113	226795	41.26	ng	85
36) 4-Chloro-3-methylphenol	11.49	107	1169258	46.82	ng	# 75
37) 2-Methylnaphthalene	11.86	142	2045469	48.53	ng	# 89
39) 1,2,4,5-Tetrachlorobenzene	12.24	216	1434804	41.16	ng	# 100
40) Hexachlorocyclopentadiene	12.22	237	1841943	92.18	ng	99

Data File : BM010942.D
Acq On : 21 Jul 2017 18:59
Operator : SJ/JU
Sample : I4280-03MSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD

Quant Time: Jul 21 23:25:08 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM071217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 12 14:21:16 2017
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	12.49	196	917591	42.75	ng	97
43) 2,4,5-Trichlorophenol	12.56	196	990754	46.75	ng #	91
45) 1,1'-Biphenyl	12.90	154	2842079	40.79	ng	97
46) 2-Chloronaphthalene	12.94	162	2235239	43.76	ng #	93
47) 2-Nitroaniline	13.16	65	828427	49.33	ng #	76
48) Acenaphthylene	13.80	152	3406111	44.83	ng	99
49) Dimethylphthalate	13.56	163	3035162	44.61	ng #	99
50) 2,6-Dinitrotoluene	13.67	165	638032	47.81	ng #	63
51) Acenaphthene	14.14	154	2076939	44.74	ng	97
52) 3-Nitroaniline	14.00	138	401091	33.06	ng #	44
53) 2,4-Dinitrophenol	14.21	184	872188	94.55	ng #	88
54) Dibenzofuran	14.49	168	3563119	47.45	ng #	90
55) 4-Nitrophenol	14.33	139	786465	74.65	ng #	1
56) 2,4-Dinitrotoluene	14.46	165	873986	47.20	ng	92
57) Fluorene	15.14	166	2876969	44.57	ng	99
58) 2,3,4,6-Tetrachlorophenol	14.72	232	966277	49.08	ng #	100
59) Diethylphthalate	14.94	149	3079324	45.83	ng	100
60) 4-Chlorophenyl-phenylether	15.14	204	1736083	43.78	ng	95
61) 4-Nitroaniline	15.17	138	514276	40.56	ng #	1
62) Azobenzene	15.43	77	3483054	51.05	ng	89
64) 4,6-Dinitro-2-methylphenol	15.23	198	550883	46.01	ng	93
65) n-Nitrosodiphenylamine	15.36	169	2518734	48.47	ng	99
66) 4-Bromophenyl-phenylether	16.04	248	1138231	49.46	ng #	91
67) Hexachlorobenzene	16.14	284	1180666	44.87	ng #	88
68) Atrazine	16.33	200	978116	45.45	ng	96
69) Pentachlorophenol	16.49	266	1467954	87.02	ng	96
70) Phenanthrene	16.88	178	4524756	47.90	ng	99
71) Anthracene	16.97	178	4503817	48.03	ng	99
72) Carbazole	17.24	167	3513798	42.55	ng	99
73) Di-n-butylphthalate	17.83	149	4718629	48.29	ng #	97
74) Fluoranthene	18.91	202	4899667	41.88	ng	98
76) Benzidine	19.11	184	314258	7.90	ng	99
77) Pyrene	19.27	202	4895167	56.80	ng	99
79) Butylbenzylphthalate	20.21	149	1696432	56.22	ng #	80
80) Benzo(a)anthracene	21.04	228	4041619	46.92	ng	100
81) 3,3'-Dichlorobenzidine	20.99	252	1103984	32.59	ng #	97
82) Chrysene	21.09	228	3785144	46.14	ng	99
83) Bis(2-ethylhexyl)phthalate	21.00	149	2277121	51.83	ng #	98
84) Di-n-octyl phthalate	21.85	149	3461589	48.08	ng	100
85) Indeno(1,2,3-cd)pyrene	25.28	276	4207565	44.73	ng #	100
87) Benzo(b)fluoranthene	22.55	252	3880291	45.68	ng #	93
88) Benzo(k)fluoranthene	22.60	252	3646331	44.95	ng #	94
89) Benzo(a)pyrene	23.09	252	3635185	47.02	ng #	96
90) Dibenzo(a,h)anthracene	25.29	278	3425828	45.66	ng #	92
91) Benzo(g,h,i)perylene	25.91	276	3544774	48.06	ng #	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

ALS Vial : 6 Sample Multiplier: 1

CL-01-071817-AMSD

Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.43 min Scan# 772
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD




Tgt Ion:152 Resp: 299158
Ion Ratio Lower Upper
152 100
150 142.3 119.5 179.3
115 67.7 43.0 64.4#




#2

1,4-Dioxane
Concen: 33.60 ng
RT: 3.09 min Scan# 35
Delta R.T. 0.00 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion: 88 Resp: 267782
Ion Ratio Lower Upper
88 100
58 63.7 0.0 0.0#
43 37.3 0.0 0.0#





#3

Pyridine

Concen: 33.31 ng

RT: 3.46 min Scan# 97

Delta R.T. -0.01 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion: 79 Resp: 725594

Ion Ratio Lower Upper

79 100

52 50.7 51.1 76.7#




51 31.6 26.1 39.1



Abundance Ion 79.00 (78.70 to 79.70): BM010796.D (-93) (-)

Ion 52.00 (51.70 to 52.70): BM010796.D (-93) (-)

Ion 51.00 (50.70 to 51.70): BM010796.D (-93) (-)



#4

n-Nitrosodimethylamine

Concen: 37.80 ng

RT: 3.38 min Scan# 83

Delta R.T. -0.01 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion: 42 Resp: 420793

Ion Ratio Lower Upper

42 100

74 120.5 119.3 178.9



44 5.4 5.4 8.2#




Abundance Ion 42.00 (41.70 to 42.70): BM010796.D (-78) (-)

Ion 74.00 (73.70 to 74.70): BM010796.D (-78) (-)

Ion 44.00 (43.70 to 44.70): BM010796.D (-78) (-)





#5

2-Fluorophenol

Concen: 131.52 ng

RT: 5.07 min Scan# 371

Delta R.T. -0.01 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion: 112 Resp: 2375256

Ion Ratio Lower Upper

112 100

64 65.3 38.6 57.8#




63 40.9 19.3 28.9#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BM0

Ion 63.00 (62.70 to 63.70): BM0



#6

Aniline

Concen: 27.38 ng

RT: 6.77 min Scan# 661

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion: 93 Resp: 853838

Ion Ratio Lower Upper

93 100

66 42.9 30.8 46.2



65 22.2 16.0 24.0






Abundance Ion 93.00 (92.70 to 93.70): BM0

Ion 66.00 (65.70 to 66.70): BM0

Ion 65.00 (64.70 to 65.70): BM0





#7

Phenol-d6

Concen: 122.60 ng

RT: 6.63 min Scan# 636

Delta R.T. -0.01 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD





Tgt Ion: 99 Resp: 3052263

Ion Ratio Lower Upper

99 100

42 20.7 11.0 16.4#

71 52.5 23.4 35.2#



#8

2-Chlorophenol

Concen: 45.11 ng

RT: 7.00 min Scan# 700

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion: 128 Resp: 806550


Ion Ratio Lower Upper

128 100

130 30.8 12.0 52.0

64 58.2 24.9 64.9





#9

Benzaldehyde

Concen: 20.37 ng

RT: 6.59 min Scan# 629

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD





Tgt Ion: 77 Resp: 352180

Ion Ratio Lower Upper

77 100

105 82.6 78.8 118.8

106 83.2 73.7 113.7



#10

Phenol

Concen: 40.57 ng

RT: 6.66 min Scan# 641

Delta R.T. -0.01 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59




Tgt Ion: 94 Resp: 1067980


Ion Ratio Lower Upper

94 100

65 34.5 18.8 58.8

66 52.5 39.9 79.9









#11
bis(2-Chloroethyl)ether
Concen: 44.66 ng
RT: 6.87 min Scan# 678
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion	Ratio	Lower	Upper
93	100		
63	75.6	49.5	89.5
95	31.2	13.5	53.5



#12
1,3-Dichlorobenzene
Concen: 40.14 ng
RT: 7.32 min Scan# 753
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion	Ratio	Lower	Upper
146	100		
148	65.5	50.9	76.3
75	51.8	21.4	32.2





#13

1,4-Dichlorobenzene

Concen: 40.02 ng

RT: 7.46 min Scan# 778

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion:146 Resp: 912498

Ion Ratio Lower Upper

146 100

148 67.9 51.9 77.9




111 49.3 29.2 43.8#



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 40.10 ng

RT: 7.77 min Scan# 830

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion:146 Resp: 871367

Ion Ratio Lower Upper

146 100

148 66.3 52.3 78.5



111 52.5 30.6 45.8#




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E









#15
Benzyl Alcohol
Concen: 48.58 ng
RT: 7.67 min Scan# 814
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion: 79 Resp: 1145347
Ion Ratio Lower Upper
79 100
108 58.0 65.0 97.4#
77 67.9 53.0 79.6



#16
2,2'-oxybis(1-chloropropane)
Concen: 46.03 ng
RT: 7.96 min Scan# 863
Delta R.T. -0.01 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion: 45 Resp: 809065
Ion Ratio Lower Upper
45 100
77 25.9 0.0 35.2
79 18.4 0.0 32.0





#17

2-Methylphenol

Concen: 46.27 ng

RT: 7.87 min Scan# 848

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD

Tgt Ion:	107	Resp:	787719
Ion	Ratio	Lower	Upper
107	100		
108	108.3	89.8	134.8
77	67.7	32.6	48.8#
79	67.4	32.8	49.2#






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BM0

Ion 79.00 (78.70 to 79.70): BM0



#18

Hexachloroethane

Concen: 40.97 ng


RT: 8.48 min Scan# 951

Delta R.T. -0.03 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59



Tgt Ion:	117	Resp:	391906
Ion	Ratio	Lower	Upper
117	100		
119	92.1	79.2	118.8
201	116.3	69.8	104.6#






Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 45.39 ng

RT: 8.23 min Scan# 909

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD

Tgt Ion: 70 Resp: 895256





Ion Ratio Lower Upper

70 100

42 48.5 44.5 66.7

101 8.1 7.4 11.2

130 14.4 15.3 22.9#



#20

3+4-Methylphenols

Concen: 45.19 ng

RT: 8.20 min Scan# 904

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Tgt Ion: 107 Resp: 1068954


Ion Ratio Lower Upper


107 100

108 89.0 73.2 113.2

77 48.5 10.7 50.7

79 42.1 9.6 49.6





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 10.19 min Scan# 1241

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD

Tgt Ion:136 Resp: 1154117





Ion Ratio Lower Upper

136 100

137 11.9 8.7 13.1

54 7.8 4.6 6.8#

68 6.2 3.7 5.5#



#22

Acetophenone

Concen: 44.59 ng

RT: 8.25 min Scan# 911

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Tgt Ion:105 Resp: 1549485




Ion Ratio Lower Upper


105 100

71 5.8 0.6 1.0#

51 29.2 21.6 32.4

120 18.3 16.1 24.1









#23
 Nitrobenzene-d5
 Concen: 89.53 ng
 RT: 8.57 min Scan# 967
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59




Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD


Tgt Ion: 82 Resp: 2688237
 Ion Ratio Lower Upper
 82 100
 128 30.2 32.8 49.2#
 54 46.3 40.6 60.8



#24
 Nitrobenzene
 Concen: 44.74 ng
 RT: 8.62 min Scan# 974
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion: 77 Resp: 1376651
 Ion Ratio Lower Upper
 77 100
 123 34.1 32.6 49.0
 65 15.1 10.9 16.3





#25

Isophorone

Concen: 47.58 ng

RT: 9.14 min Scan# 1063

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion: 82 Resp: 2340174

Ion Ratio Lower Upper

82 100

95 9.5 5.8 8.6#




138 12.9 16.3 24.5#



Abundance Ion 82.00 (81.70 to 82.70): BM010796.D (-1056) (-)

Ion 95.00 (94.70 to 95.70): BM010796.D (-1056) (-)

Ion 138.00 (137.70 to 138.70): BM010796.D (-1056) (-)



#26

2-Nitrophenol

Concen: 45.84 ng

RT: 9.32 min Scan# 1093

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion: 139 Resp: 469412

Ion Ratio Lower Upper

139 100

109 62.3 20.1 30.1#



65 77.8 31.5 47.3#




Abundance Ion 139.00 (138.70 to 139.70): BM010796.D (-1089) (-)

Ion 109.00 (108.70 to 109.70): BM010796.D (-1089) (-)

Ion 65.00 (64.70 to 65.70): BM010796.D (-1089) (-)





#27

2,4-Dimethylphenol

Concen: 46.16 ng

RT: 9.39 min Scan# 1106

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion:122 Resp: 843574

Ion Ratio Lower Upper

122 100

107 150.1 84.7 127.1#




121 57.0 46.6 69.8



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 47.07 ng

RT: 9.63 min Scan# 1146

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion: 93 Resp: 1297957

Ion Ratio Lower Upper

93 100

95 33.2 25.5 38.3



123 9.6 8.6 13.0




Abundance Ion 93.00 (92.70 to 93.70): BM0

Ion 95.00 (94.70 to 95.70): BM0

Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 45.82 ng
 RT: 9.85 min Scan# 1183
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD

Tgt Ion	Ratio	Lower	Upper
162	100		
164	68.1	42.8	82.8
98	42.4	14.5	54.5




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BM0






#30
 1,2,4-Trichlorobenzene
 Concen: 42.35 ng
 RT: 10.05 min Scan# 1218
 Delta R.T. -0.03 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion	Ratio	Lower	Upper
180	100		
182	97.8	77.6	116.4
145	28.7	23.4	35.2



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E









#31
Naphthalene
Concen: 44.61 ng
RT: 10.23 min Scan# 1249
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion:128 Resp: 2617942
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 13.7 10.4 15.6



#32
Benzoic acid
Concen: 40.63 ng
RT: 9.54 min Scan# 1131
Delta R.T. -0.01 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:122 Resp: 582604
Ion Ratio Lower Upper
122 100
105 147.1 99.9 139.9#
77 139.0 73.7 113.7#









#33
4-Chloroaniline
Concen: 12.27 ng
RT: 10.36 min Scan# 1270
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion:	127	Resp:	290182
Ion	Ratio	Lower	Upper
127	100		
129	32.3	25.3	37.9
65	44.5	17.6	26.4
92	24.8	13.1	19.7



#34
Hexachlorobutadiene
Concen: 40.73 ng
RT: 10.52 min Scan# 1298
Delta R.T. -0.03 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:	225	Resp:	797533
Ion	Ratio	Lower	Upper
225	100		
223	65.5	47.9	71.9
227	70.3	50.1	75.1





#35

Caprolactam

Concen: 41.26 ng

RT: 11.15 min Scan# 1405

Delta R.T. -0.01 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion:113 Resp: 226795

Ion Ratio Lower Upper

113 100

55 146.5 145.9 185.9




56 137.4 101.0 141.0



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BM0

Ion 56.00 (55.70 to 56.70): BM0



#36

4-Chloro-3-methylphenol

Concen: 46.82 ng

RT: 11.49 min Scan# 1463

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion:107 Resp: 1169258

Ion Ratio Lower Upper

107 100

144 20.3 23.3 34.9#



142 64.6 72.6 109.0#




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 48.53 ng
RT: 11.86 min Scan# 1525
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion:142 Resp: 2045469
Ion Ratio Lower Upper
142 100
141 88.1 71.9 107.9
115 53.0 25.4 38.0#



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.08 min Scan# 1903
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:164 Resp: 799669
Ion Ratio Lower Upper
164 100
162 95.5 82.4 123.6
160 44.9 35.8 53.6









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 41.16 ng
 RT: 12.24 min Scan# 1590
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD

Tgt Ion:	216	Resp:	1434804
Ion	Ratio	Lower	Upper
216	100		
214	77.9	0.0	0.0#
179	20.1	0.0	0.0#
108	22.1	0.0	0.0#




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 92.18 ng
 RT: 12.22 min Scan# 1587
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion:	237	Resp:	1841943
Ion	Ratio	Lower	Upper
237	100		
235	61.5	42.0	82.0
272	11.7	0.0	33.4



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E









#41
2,4,6-Tribromophenol
Concen: 139.60 ng
RT: 15.59 min Scan# 2159
Delta R.T. -0.01 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion:330 Resp: 1705939
Ion Ratio Lower Upper
330 100
332 96.8 0.0 0.0#
141 39.7 0.0 0.0#



#42
2,4,6-Trichlorophenol
Concen: 42.75 ng
RT: 12.49 min Scan# 1633
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:196 Resp: 917591
Ion Ratio Lower Upper
196 100
198 99.4 76.3 114.5
200 32.3 25.6 38.4









#43
 2,4,5-Trichlorophenol
 Concen: 46.75 ng
 RT: 12.56 min Scan# 1645
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59




Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD


Tgt Ion	Ratio	Lower	Upper
196	100		
198	99.9	79.4	119.0
97	56.5	26.8	40.2
132	24.4	18.6	28.0



#44
 2-Fluorobiphenyl
 Concen: 86.05 ng
 RT: 12.69 min Scan# 1667
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.5	28.5	42.7
170	23.3	18.8	28.2









#45
 1,1'-Biphenyl
 Concen: 40.79 ng
 RT: 12.90 min Scan# 1703
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD

Tgt Ion:154 Resp: 2842079
 Ion Ratio Lower Upper
 154 100
 153 39.4 20.7 60.7
 76 13.6 0.0 30.9




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BM






#46
 2-Chloronaphthalene
 Concen: 43.76 ng
 RT: 12.94 min Scan# 1709
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion:162 Resp: 2235239
 Ion Ratio Lower Upper
 162 100
 127 41.6 26.9 40.3#
 164 34.0 26.8 40.2



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 49.33 ng
RT: 13.16 min Scan# 1746
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion: 65 Resp: 828427
Ion Ratio Lower Upper
65 100
92 68.4 59.1 88.7
138 78.6 93.9 140.9#



#48
Acenaphthylene
Concen: 44.83 ng
RT: 13.80 min Scan# 1855
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion: 152 Resp: 3406111
Ion Ratio Lower Upper
152 100
151 20.4 15.9 23.9
153 13.9 10.6 16.0









#49
Dimethylphthalate
Concen: 44.61 ng
RT: 13.56 min Scan# 1814
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion:163 Resp: 3035162
Ion Ratio Lower Upper
163 100
194 4.3 2.7 4.1#
164 10.2 8.2 12.2



#50
2,6-Dinitrotoluene
Concen: 47.81 ng
RT: 13.67 min Scan# 1833
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:165 Resp: 638032
Ion Ratio Lower Upper
165 100
63 79.6 44.1 66.1#
89 84.4 44.3 66.5#





#51

Acenaphthene

Concen: 44.74 ng

RT: 14.14 min Scan# 1914

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion:154 Resp: 2076939

Ion Ratio Lower Upper

154 100

153 116.6 90.0 135.0




152 54.1 42.6 63.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 33.06 ng

RT: 14.00 min Scan# 1889

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion:138 Resp: 401091

Ion Ratio Lower Upper

138 100

108 13.2 7.3 10.9#



92 154.3 76.6 114.8#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BM









#53
 2,4-Dinitrophenol
 Concen: 94.55 ng
 RT: 14.21 min Scan# 1925
 Delta R.T. -0.01 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59




Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD


Tgt Ion	Ratio	Lower	Upper
184	100		
63	83.3	69.2	103.8
154	85.6	53.3	79.9#



#54
 Dibenzofuran
 Concen: 47.45 ng
 RT: 14.49 min Scan# 1972
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion	Ratio	Lower	Upper
168	100		
139	42.0	27.3	40.9#
169	13.4	10.6	16.0









#55
4-Nitrophenol
Concen: 74.65 ng
RT: 14.33 min Scan# 1945
Delta R.T. -0.01 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion:139 Resp: 786465
Ion Ratio Lower Upper
139 100
109 149.4 37.7 77.7#
65 143.8 49.9 89.9#



#56
2,4-Dinitrotoluene
Concen: 47.20 ng
RT: 14.46 min Scan# 1968
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:165 Resp: 873986
Ion Ratio Lower Upper
165 100
63 60.8 52.4 78.6
89 98.9 72.4 108.6
182 2.9 2.0 3.0









#57
 Fluorene
 Concen: 44.57 ng
 RT: 15.14 min Scan# 2083
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59




Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD


Tgt Ion:166 Resp: 2876969
 Ion Ratio Lower Upper
 166 100
 165 97.8 79.0 118.4
 167 13.3 10.3 15.5



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 49.08 ng
 RT: 14.72 min Scan# 2011
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion:232 Resp: 966277
 Ion Ratio Lower Upper
 232 100
 131 40.8 0.0 0.0#
 130 2.9 0.0 0.0#
 166 28.8 0.0 0.0#









#59
Diethylphthalate
Concen: 45.83 ng
RT: 14.94 min Scan# 2049
Delta R.T. -0.01 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion:149 Resp: 3079324
Ion Ratio Lower Upper
149 100
177 22.7 18.3 27.5
150 12.1 9.8 14.8



#60
4-Chlorophenyl-phenylether
Concen: 43.78 ng
RT: 15.14 min Scan# 2083
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:204 Resp: 1736083
Ion Ratio Lower Upper
204 100
206 32.3 26.6 39.8
141 61.3 45.2 67.8









#61
4-Nitroaniline
Concen: 40.56 ng
RT: 15.17 min Scan# 2089
Delta R.T. -0.01 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion: 138 Resp: 514276
Ion Ratio Lower Upper
138 100
92 73.5 16.7 56.7#
108 164.2 37.6 77.6#



#62
Azobenzene
Concen: 51.05 ng
RT: 15.43 min Scan# 2133
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion: 77 Resp: 3483054
Ion Ratio Lower Upper
77 100
182 21.4 6.5 46.5
105 16.8 3.8 43.8
51 30.8 5.5 45.5









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 16.83 min Scan# 2371
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion:188 Resp: 1815923
Ion Ratio Lower Upper
188 100
94 8.0 8.7 13.1#
80 9.2 9.3 13.9#



#64
4,6-Dinitro-2-methylphenol
Concen: 46.01 ng
RT: 15.23 min Scan# 2099
Delta R.T. -0.01 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:198 Resp: 550883
Ion Ratio Lower Upper
198 100
51 43.7 28.8 68.8
105 45.3 30.5 70.5









#65
 n-Nitrosodiphenylamine
 Concen: 48.47 ng
 RT: 15.36 min Scan# 2121
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59




Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD


Tgt Ion:169 Resp: 2518734
 Ion Ratio Lower Upper
 169 100
 168 67.5 53.9 80.9
 167 35.4 28.9 43.3



#66
 4-Bromophenyl-phenylether
 Concen: 49.46 ng
 RT: 16.04 min Scan# 2236
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion:248 Resp: 1138231
 Ion Ratio Lower Upper
 248 100
 250 98.7 77.4 116.0
 141 72.3 45.2 67.8#









#67
Hexachlorobenzene
Concen: 44.87 ng
RT: 16.14 min Scan# 2254
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD

Tgt Ion:284 Resp: 1180666
Ion Ratio Lower Upper
284 100
142 33.3 20.4 30.6#
249 34.6 23.8 35.6




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 45.45 ng
RT: 16.33 min Scan# 2285
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:200 Resp: 978116
Ion Ratio Lower Upper
200 100
173 24.0 4.8 44.8
215 50.2 26.9 66.9



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 87.02 ng
 RT: 16.49 min Scan# 2313
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59




Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD


Tgt Ion:266 Resp: 1467954
 Ion Ratio Lower Upper
 266 100
 268 67.1 52.0 78.0
 264 64.7 49.0 73.4



#70
 Phenanthrene
 Concen: 47.90 ng
 RT: 16.88 min Scan# 2379
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion:178 Resp: 4524756
 Ion Ratio Lower Upper
 178 100
 176 19.3 15.2 22.8
 179 15.3 12.1 18.1









#71
 Anthracene
 Concen: 48.03 ng
 RT: 16.97 min Scan# 2394
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59




Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD


Tgt Ion:178 Resp: 4503817
 Ion Ratio Lower Upper
 178 100
 176 18.2 14.6 22.0
 179 15.1 12.6 19.0



#72
 Carbazole
 Concen: 42.55 ng
 RT: 17.24 min Scan# 2441
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion:167 Resp: 3513798
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.2 25.8
 139 13.3 10.8 16.2









#73
Di-n-butylphthalate
Concen: 48.29 ng
RT: 17.83 min Scan# 2541
Delta R.T. -0.01 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion:149 Resp: 4718629
Ion Ratio Lower Upper
149 100
150 9.1 7.5 11.3
104 7.5 3.7 5.5#



#74
Fluoranthene
Concen: 41.88 ng
RT: 18.91 min Scan# 2724
Delta R.T. -0.02 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:202 Resp: 4899667
Ion Ratio Lower Upper
202 100
101 6.8 0.0 28.7
203 17.2 0.0 37.2





#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.06 min Scan# 3089

Delta R.T. -0.01 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion:240 Resp: 1486632

Ion Ratio Lower Upper

240 100

120 5.8 8.2 12.2#




236 25.0 20.0 30.0



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 7.90 ng

RT: 19.11 min Scan# 2758

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion:184 Resp: 314258

Ion Ratio Lower Upper

184 100

185 13.7 11.3 16.9



183 11.3 8.9 13.3




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 56.80 ng

RT: 19.27 min Scan# 2786

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion:202 Resp: 4895167

Ion Ratio Lower Upper

202 100

200 20.7 16.9 25.3




203 17.3 14.1 21.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 94.54 ng

RT: 19.50 min Scan# 2824

Delta R.T. -0.01 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion:244 Resp: 7267013

Ion Ratio Lower Upper

244 100

212 8.1 4.9 7.3#



122 6.7 6.2 9.4




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#79
Butylbenzylphthalate
Concen: 56.22 ng
RT: 20.21 min Scan# 2945
Delta R.T. -0.01 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59




Instrument :
BNA_M
ClientSampleId :
CL-01-071817-AMSD


Tgt Ion:149 Resp: 1696432
Ion Ratio Lower Upper
149 100
91 80.5 50.1 75.1#
206 25.7 16.2 24.2#



#80
Benzo(a)anthracene
Concen: 46.92 ng
RT: 21.04 min Scan# 3086
Delta R.T. -0.01 min
Lab File: BM010942.D
Acq: 21 Jul 2017 18:59

Tgt Ion:228 Resp: 4041619
Ion Ratio Lower Upper
228 100
226 27.4 21.7 32.5
229 20.0 16.0 24.0









#81
 3,3'-Dichlorobenzidine
 Concen: 32.59 ng
 RT: 20.99 min Scan# 3077
 Delta R.T. -0.01 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59




Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD


Tgt Ion:252 Resp: 1103984
 Ion Ratio Lower Upper
 252 100
 254 63.9 51.9 77.9
 126 7.5 9.8 14.8#



#82
 Chrysene
 Concen: 46.14 ng
 RT: 21.09 min Scan# 3095
 Delta R.T. -0.01 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion:228 Resp: 3785144
 Ion Ratio Lower Upper
 228 100
 226 29.5 24.1 36.1
 229 19.3 15.5 23.3









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 51.83 ng
 RT: 21.00 min Scan# 3079
 Delta R.T. -0.01 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59




Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD


Tgt Ion:149 Resp: 2277121
 Ion Ratio Lower Upper
 149 100
 167 28.7 22.4 33.6
 279 5.4 3.4 5.0#



#84
 Di-n-octyl phthalate
 Concen: 48.08 ng
 RT: 21.85 min Scan# 3224
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion:149 Resp: 3461589
 Ion Ratio Lower Upper
 149 100
 167 1.6 1.2 1.8
 43 9.1 7.3 10.9









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 44.73 ng
 RT: 25.28 min Scan# 3807
 Delta R.T. -0.03 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59




Instrument :
 BNA_M
 ClientSampleId :
 CL-01-071817-AMSD


Tgt Ion: 276 Resp: 4207565
 Ion Ratio Lower Upper
 276 100
 138 12.1 0.0 0.0#
 277 25.7 0.0 0.0#



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 23.18 min Scan# 3450
 Delta R.T. -0.02 min
 Lab File: BM010942.D
 Acq: 21 Jul 2017 18:59

Tgt Ion: 264 Resp: 1337153
 Ion Ratio Lower Upper
 264 100
 260 22.7 18.6 27.8
 265 22.4 17.3 25.9





#87

Benzo(b)fluoranthene

Concen: 45.68 ng

RT: 22.55 min Scan# 3343

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion:252 Resp: 3880291

Ion Ratio Lower Upper

252 100

253 22.3 18.0 27.0




125 5.2 9.9 14.9#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 44.95 ng

RT: 22.60 min Scan# 3351

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59


Tgt Ion:252 Resp: 3646331

Ion Ratio Lower Upper

252 100

253 22.1 17.2 25.8



125 4.6 8.1 12.1#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 47.02 ng

RT: 23.09 min Scan# 3435

Delta R.T. -0.02 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :


BNA_M

ClientSampleId :

CL-01-071817-AMSD

Tgt Ion:252 Resp: 3635185




Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.6	26.4
125	5.4	7.4	11.2#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 45.66 ng

RT: 25.29 min Scan# 3809


Delta R.T. -0.03 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Tgt Ion:278 Resp: 3425828



Ion	Ratio	Lower	Upper
278	100		
139	8.7	13.4	20.0#
279	23.0	19.0	28.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 48.06 ng

RT: 25.91 min Scan# 3915

Delta R.T. -0.04 min

Lab File: BM010942.D

Acq: 21 Jul 2017 18:59

Instrument :

BNA_M

ClientSampleId :

CL-01-071817-AMSD


Tgt Ion:276 Resp: 3544774

Ion Ratio Lower Upper

276 100

277 23.3 18.5 27.7

138 10.3 19.0 28.6#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

