

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM072120\
 Data File : BM026798.D
 Acq On : 21 Jul 2020 09:05
 Operator : JU/CG
 Sample : PB130302BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK02

Quant Time: Jul 21 10:19:32 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM063020MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 21 10:10:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.30	152	54803	20.00	ng/ul	0.00
18) Naphthalene-d8	10.05	136	214863	20.00	ng/ul	0.00
36) Acenaphthene-d10	13.97	164	135479	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.72	188	292690	20.00	ng/ul	0.00
78) Chrysene-d12	20.95	240	287326	20.00	ng/ul	0.00
86) Perylene-d12	23.01	264	305077	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	2.90	96	11501	6.31	ng/uL	0.00
5) Phenol-d5	6.52	99	166341	33.14	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.66	67	110373	32.52	ng/ul	0.00
9) 2-Chlorophenol-d4	6.84	132	131590	34.12	ng/ul	0.00
13) 4-Methylphenol-d8	8.03	113	131500	32.76	ng/ul	0.00
19) Nitrobenzene-d5	8.45	128	55377	32.59	ng/ul	0.00
22) 2-Nitrophenol-d4	9.16	143	60465	33.37	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.70	165	112049	31.58	ng/ul	0.00
29) 4-Chloroaniline-d4	10.21	131	150191	40.20	ng/ul	0.00
44) Dimethylphthalate-d6	13.40	166	362529	33.54	ng/ul	0.00
47) Acenaphthylene-d8	13.65	160	407914	31.24	ng/ul	0.00
52) 4-Nitrophenol-d4	14.24	143	31374	20.12	ng/ul	0.01
58) Fluorene-d10	14.97	176	305778	32.31	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.13	200	42867	23.44	ng/ul	0.00
71) Anthracene-d10	16.82	188	460608	31.96	ng/ul	0.00
79) Pyrene-d10	19.14	212	501923	33.61	ng/ul	0.00
90) Benzo(a)pyrene-d12	22.88	264	547374	33.25	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM072120\
Data File : BM026798.D
Acq On : 21 Jul 2020 09:05
Operator : JU/CG
Sample : PB130302BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK02

Quant Time: Jul 21 10:19:32 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM063020MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 21 10:10:59 2020
Response via : Initial Calibration

