

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM072122\  
 Data File : BM035899.D  
 Acq On : 21 Jul 2022 21:24  
 Operator : CG/JU  
 Sample : N3786-11  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 0718-WC-TP-WC3EA

Quant Time: Jul 22 01:09:19 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM071822.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jul 18 17:08:15 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|--------|------|----------|--------|-------|----------|
| -----                       |        |      |          |        |       |          |
| Internal Standards          |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.892  | 152  | 341655   | 20.000 | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.698 | 136  | 1458030  | 20.000 | ng    | #-0.01   |
| 39) Acenaphthene-d10        | 14.527 | 164  | 812863   | 20.000 | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.268 | 188  | 1488579  | 20.000 | ng    | 0.00     |
| 76) Chrysene-d12            | 21.439 | 240  | 1154965  | 20.000 | ng    | -0.01    |
| 86) Perylene-d12            | 23.815 | 264  | 1143408  | 20.000 | ng    | -0.01    |
| System Monitoring Compounds |        |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.451  | 112  | 774493   | 35.144 | ng    | 0.00     |
| 7) Phenol-d6                | 7.057  | 99   | 991270   | 35.770 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.057  | 82   | 590337   | 21.655 | ng    | -0.01    |
| 42) 2,4,6-Tribromophenol    | 16.009 | 330  | 264983   | 29.144 | ng    | -0.01    |
| 45) 2-Fluorobiphenyl        | 13.157 | 172  | 1316837  | 22.899 | ng    | -0.01    |
| 79) Terphenyl-d14           | 19.892 | 244  | 1482147  | 24.968 | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM072122\  
Data File : BM035899.D  
Acq On : 21 Jul 2022 21:24  
Operator : CG/JU  
Sample : N3786-11  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
0718-WC-TP-WC3EA

Quant Time: Jul 22 01:09:19 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM071822.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jul 18 17:08:15 2022  
Response via : Initial Calibration

