


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM072216\
Data File : BM006617.D
Acq On : 22 Jul 2016 09:54
Operator : UM/SJ
Sample : DFTPP57
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP57

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072016.M
Title : SVOA CALIBRATION
Last Update : Mon Jul 25 15:53:06 2016



AutoFind: Scans 2528, 2529, 2530; Background Corrected with Scan 2523

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.2	74978	PASS
68	69	0.00	2	1.6	1130	PASS
69	198	0.00	100	39.8	69008	PASS
70	69	0.00	2	0.6	437	PASS
127	198	10	80	47.3	81954	PASS
197	198	0.00	2	0.5	929	PASS
198	198	100	100	100.0	173392	PASS
199	198	5	9	6.5	11292	PASS
275	198	10	60	24.8	42962	PASS
365	198	1	100	3.0	5179	PASS
441	443	0.01	100	80.4	20301	PASS
442	198	50	100	75.7	131336	PASS
443	442	15	24	19.2	25254	PASS

m/z	Abundance
38.00	658.0
39.00	4829.0
39.90	952.0
41.00	404.0
44.00	1130.0
49.00	692.0
50.00	16227.0
51.00	67096.0
52.00	3794.0
55.10	570.0
56.00	1924.0
57.00	3922.0
61.00	824.0
62.00	1025.0
63.00	2376.0
65.00	1193.0
68.10	878.0
69.00	59104.0
70.00	447.0
73.00	833.0
74.00	5818.0
75.00	9757.0
76.00	3462.0
77.00	68120.0
78.00	5533.0
79.00	4278.0
80.00	3199.0
81.00	4791.0
82.00	1154.0
83.00	1536.0
85.00	824.0
85.90	1328.0
87.00	784.0
91.00	1377.0
92.00	1044.0
93.00	7381.0
94.00	727.0
96.00	343.0
98.00	5220.0
98.90	4354.0
99.90	349.0
101.00	2825.0
102.90	709.0
104.00	1480.0
105.00	1706.0
106.00	374.0
107.00	21864.0
108.00	3163.0
109.00	589.0
110.00	35552.0
111.00	5803.0
111.90	737.0
115.90	1084.0
117.00	16120.0
118.00	1164.0
119.90	306.0
121.90	1220.0
123.00	1690.0
123.90	883.0
125.00	704.0
127.00	65728.0
128.00	5724.0
129.00	25224.0
129.90	1960.0
130.90	518.0
134.10	781.0
135.00	1819.0
135.90	685.0
137.00	1411.0
137.80	317.0
140.00	478.0
141.00	3206.0
142.00	1038.0
143.00	741.0
146.10	533.0
147.00	1956.0
147.90	3747.0
149.00	897.0
150.90	479.0
153.10	996.0
153.90	530.0
155.00	1797.0
156.00	2772.0
157.90	534.0
158.90	377.0
159.90	1173.0
161.00	2142.0
161.90	532.0
164.90	1265.0
166.00	1171.0
167.00	6235.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

168.00	3477.0
169.00	714.0
171.00	381.0
172.00	596.0
173.00	932.0
174.00	1688.0
175.00	2682.0
175.90	578.0
177.00	969.0
177.90	474.0
179.00	5020.0
180.00	3625.0
181.00	1496.0
183.90	369.0
185.00	2537.0
186.00	17880.0
187.00	5352.0
188.00	695.0
188.90	1027.0
190.10	347.0
190.90	423.0
191.90	1295.0
193.00	1934.0
193.90	488.0
196.00	2914.0
197.10	935.0
198.00	130672.0
199.00	9538.0
200.00	1186.0
201.40	651.0
202.90	967.0
204.00	4718.0
205.00	7709.0
206.00	32432.0
207.00	5383.0
208.00	1343.0
209.00	470.0
210.10	562.0
211.00	1332.0
211.90	544.0
215.00	459.0
215.90	718.0
217.00	8401.0
218.00	986.0
221.00	4437.0
221.60	1432.0
223.00	1764.0
224.00	18576.0
225.00	4643.0
226.10	487.0
227.00	7776.0
228.00	991.0
229.00	1826.0
231.00	596.0
234.10	355.0
235.00	662.0
235.90	355.0
237.10	678.0
240.80	400.0
242.00	1011.0
243.00	983.0
244.10	12576.0
245.00	1772.0
246.00	2640.0
246.90	614.0
248.90	469.0
252.90	413.0
254.00	458.0
255.00	71352.0
256.00	11179.0
257.00	933.0
258.00	4039.0
259.10	665.0
265.00	1551.0
271.90	308.0
272.90	1946.0
274.00	5076.0
275.00	30504.0
276.00	4567.0
277.00	2295.0
283.10	374.0
285.00	485.0
292.90	446.0
296.00	8192.0
297.00	1329.0
303.10	1005.0
314.00	419.0
315.00	1008.0
315.90	524.0
323.00	2398.0
324.00	572.0
326.80	398.0
334.10	1304.0
335.00	446.0
340.90	381.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

345.80	345.0
346.00	340.0
352.10	788.0
353.00	428.0
354.10	850.0
365.00	3085.0
365.90	490.0
372.00	1379.0
382.90	345.0
402.00	415.0
403.10	706.0
421.00	527.0
423.10	3880.0
424.10	1082.0
441.00	11017.0
442.10	77144.0
443.10	15579.0
444.00	1352.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
36.90	411.0
38.00	1029.0
39.10	6614.0
40.00	1053.0
41.00	559.0
43.10	333.0
44.00	1263.0
49.00	889.0
50.00	21744.0
51.00	86216.0
52.00	4723.0
53.10	475.0
55.00	783.0
56.00	2717.0
57.00	5587.0
58.00	316.0
61.00	958.0
62.00	1249.0
63.00	3626.0
64.10	444.0
65.00	2246.0
67.00	435.0
67.90	1546.0
69.00	80176.0
69.90	398.0
73.00	971.0
74.00	8063.0
75.00	12278.0
76.00	4704.0
77.00	93344.0
78.10	6653.0
79.00	5682.0
80.00	4850.0
81.00	6257.0
82.00	1716.0
83.00	1660.0
83.90	536.0
84.90	1391.0
86.00	1986.0
87.00	1147.0
91.00	1869.0
91.90	1296.0
93.00	10898.0
93.90	611.0
95.10	520.0
96.00	525.0
96.90	320.0
98.00	8138.0
99.00	6503.0
100.00	412.0
101.00	3658.0
102.90	1320.0
104.00	2061.0
105.00	2188.0
106.00	949.0
107.00	30392.0
108.00	4667.0
109.00	975.0
110.00	50864.0
111.00	7867.0
112.00	959.0
116.00	1559.0
117.00	24304.0
118.00	1758.0
119.00	528.0
122.00	1361.0
123.00	3030.0
123.90	1179.0
125.00	1008.0
126.10	306.0
127.00	97544.0
128.00	7232.0
129.00	35752.0
130.00	3019.0
130.90	693.0
133.90	954.0
135.00	3024.0
136.10	1439.0
137.00	1536.0
140.00	499.0
141.00	4312.0
142.10	1501.0
142.90	992.0
145.90	777.0
147.00	2610.0
148.00	6540.0
149.00	1388.0
150.10	532.0
151.00	628.0
152.00	462.0
153.00	1643.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

153.90	1046.0
155.00	2775.0
156.00	4458.0
157.00	678.0
157.90	1145.0
158.90	596.0
159.90	1870.0
161.10	2343.0
162.10	680.0
164.90	1387.0
166.00	1717.0
167.00	9357.0
168.00	5042.0
169.10	821.0
170.90	346.0
172.00	777.0
173.00	1228.0
173.90	1668.0
175.00	3571.0
176.00	1042.0
177.00	1909.0
178.00	591.0
179.00	7015.0
180.00	5290.0
181.00	2575.0
182.00	335.0
184.00	681.0
185.00	3241.0
186.00	26496.0
187.00	8724.0
188.00	1002.0
189.00	1646.0
191.00	997.0
192.00	2783.0
193.00	2847.0
194.00	830.0
196.00	4812.0
197.10	1031.0
198.00	202240.0
199.00	12656.0
199.90	1044.0
201.30	898.0
203.00	1195.0
204.00	7856.0
205.00	11975.0
206.00	51872.0
207.00	7410.0
208.00	1774.0
209.00	581.0
210.00	882.0
211.10	1815.0
212.00	387.0
215.00	455.0
216.00	1036.0
217.00	13984.0
218.00	1987.0
219.00	441.0
219.90	300.0
221.00	7538.0
221.90	1689.0
223.00	3349.0
224.00	28336.0
225.00	7084.0
226.00	955.0
227.00	12226.0
228.00	1243.0
229.00	2413.0
231.00	1015.0
234.00	619.0
235.00	584.0
236.00	649.0
237.00	972.0
239.00	494.0
241.00	597.0
242.00	1683.0
243.00	1652.0
244.00	21416.0
245.10	3302.0
246.00	3958.0
247.00	839.0
249.00	936.0
253.00	731.0
254.10	890.0
255.00	114456.0
256.00	16848.0
257.00	1296.0
258.00	6350.0
259.00	1006.0
265.00	3324.0
266.00	381.0
272.10	364.0
273.00	3336.0
274.00	9137.0
275.00	49856.0
276.00	6155.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

277.00	4758.0
278.00	640.0
283.00	447.0
284.00	331.0
285.00	1000.0
292.90	1080.0
294.00	341.0
295.00	320.0
296.00	14819.0
297.00	2144.0
302.00	392.0
303.00	1671.0
304.00	418.0
308.00	412.0
314.00	680.0
315.00	1641.0
316.00	899.0
321.00	373.0
323.00	4125.0
324.00	860.0
327.00	769.0
332.00	330.0
332.90	365.0
334.00	2792.0
335.10	938.0
341.10	456.0
345.90	732.0
352.00	1573.0
353.00	891.0
354.00	1751.0
365.00	6124.0
365.90	902.0
372.00	2916.0
373.00	521.0
383.00	666.0
402.00	771.0
403.10	1262.0
404.00	470.0
421.00	1164.0
422.00	832.0
423.00	7835.0
424.10	1909.0
441.10	23744.0
442.10	146496.0
443.10	28416.0
444.10	2406.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
38.00	774.0
39.10	5546.0
40.00	1102.0
41.00	582.0
44.00	1359.0
48.90	756.0
50.00	18624.0
51.00	71624.0
52.10	3371.0
55.00	744.0
56.00	1952.0
57.00	5156.0
60.90	1049.0
62.00	876.0
63.00	2929.0
64.00	532.0
65.00	1325.0
67.10	303.0
67.90	968.0
69.00	68936.0
70.00	468.0
73.10	989.0
74.00	6617.0
75.00	10272.0
76.10	3798.0
77.00	80024.0
78.10	5204.0
79.00	5386.0
80.00	3711.0
81.00	5665.0
81.90	1546.0
83.00	1337.0
83.90	438.0
85.00	1136.0
86.00	2030.0
87.00	966.0
91.00	1484.0
92.00	1350.0
93.00	8767.0
94.00	455.0
95.00	383.0
96.00	476.0
98.00	6116.0
99.00	5380.0
99.90	444.0
101.00	2909.0
103.00	974.0
104.00	1951.0
105.00	1925.0
105.90	685.0
107.00	25176.0
108.00	3805.0
109.00	773.0
110.00	44312.0
111.00	6474.0
112.00	734.0
116.00	1417.0
117.00	19600.0
118.00	1503.0
119.90	416.0
122.00	1372.0
123.00	2497.0
124.00	997.0
125.00	1332.0
126.00	366.0
127.00	82592.0
128.00	6042.0
129.00	30448.0
130.00	3011.0
130.90	597.0
132.10	349.0
133.90	1116.0
135.00	2376.0
136.00	931.0
136.90	1366.0
140.00	428.0
141.00	3671.0
141.90	1243.0
143.00	945.0
143.90	311.0
144.90	387.0
146.00	848.0
147.00	2404.0
148.00	4977.0
149.00	846.0
149.90	475.0
151.00	461.0
153.00	1422.0
153.90	1142.0
155.10	2221.0
156.00	3980.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

157.00	432.0
158.00	641.0
158.90	576.0
159.90	1191.0
161.00	2223.0
162.10	726.0
165.00	1953.0
166.00	1382.0
167.00	9072.0
168.00	4048.0
169.00	497.0
170.00	307.0
170.90	605.0
172.00	743.0
173.00	1087.0
174.00	2165.0
175.00	3416.0
176.00	975.0
177.00	1235.0
178.00	628.0
179.00	6307.0
180.00	4468.0
181.00	2370.0
182.90	419.0
184.00	639.0
185.00	3440.0
186.00	26048.0
187.00	7559.0
188.00	679.0
189.00	1796.0
191.00	757.0
192.00	2325.0
193.00	2252.0
194.00	679.0
194.90	600.0
196.00	4448.0
197.00	823.0
198.00	187264.0
198.90	11682.0
199.90	1334.0
201.40	723.0
203.00	1152.0
204.00	7290.0
205.00	10862.0
206.00	47624.0
207.10	6992.0
208.00	1714.0
209.00	581.0
210.10	656.0
211.00	1963.0
211.90	322.0
215.10	778.0
215.90	863.0
217.00	13181.0
218.00	1564.0
221.00	6023.0
222.00	1591.0
223.00	3076.0
224.00	29272.0
225.00	7196.0
226.10	781.0
227.00	11614.0
228.00	1665.0
229.00	2375.0
230.00	436.0
231.00	863.0
233.90	681.0
235.00	839.0
235.90	573.0
237.00	1089.0
239.00	398.0
240.00	550.0
241.00	650.0
242.00	1701.0
243.10	1755.0
244.10	21512.0
245.10	2562.0
246.00	4728.0
247.00	986.0
248.90	769.0
253.00	516.0
254.10	891.0
255.00	115040.0
256.00	16440.0
257.10	1446.0
258.00	6472.0
259.00	824.0
265.00	2216.0
273.00	3203.0
274.00	8474.0
275.00	48528.0
276.00	6710.0
277.00	3675.0
278.00	979.0
283.10	507.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

285.00	691.0
293.00	814.0
295.00	382.0
296.00	14110.0
297.00	1689.0
301.90	345.0
303.10	1669.0
303.90	312.0
308.10	370.0
314.00	628.0
315.00	1516.0
316.00	911.0
321.00	404.0
323.10	4149.0
324.00	752.0
326.90	872.0
328.00	444.0
333.00	395.0
334.00	3104.0
335.10	750.0
341.00	629.0
346.00	937.0
352.00	1368.0
353.00	1128.0
354.00	2143.0
355.00	413.0
365.00	6330.0
366.00	721.0
371.00	342.0
372.10	2591.0
373.00	819.0
383.00	403.0
390.00	302.0
402.10	950.0
403.10	1185.0
404.00	415.0
421.00	1059.0
422.00	1310.0
423.00	8540.0
424.10	2027.0
441.10	26144.0
442.10	170368.0
443.10	31768.0
444.10	2732.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

Scan 2523 (17.427 min): BM006617.D
DFTPP57

m/z	Abundance
39.90	788.0
41.00	471.0
43.10	479.0
44.00	1765.0
48.80	407.0
55.00	420.0
68.90	397.0
73.00	488.0
91.10	341.0
207.10	1018.0
281.00	343.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57