

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072222\
 Data File : BM035906.D
 Acq On : 22 Jul 2022 15:01
 Operator : CG/JU
 Sample : N3786-11
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 0718-WC-TP-WC3EA

Quant Time: Jul 22 15:33:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM071822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 22 14:25:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.886	152	343673	20.000	ng	0.00
21) Naphthalene-d8	10.692	136	1374165	20.000	ng	0.00
39) Acenaphthene-d10	14.521	164	739114	20.000	ng	0.00
64) Phenanthrene-d10	17.262	188	1359673	20.000	ng	0.00
76) Chrysene-d12	21.439	240	1158989	20.000	ng	0.00
86) Perylene-d12	23.809	264	1159317	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.445	112	783419	35.340	ng	0.00
7) Phenol-d6	7.051	99	965634	34.640	ng	0.00
23) Nitrobenzene-d5	9.051	82	749329	29.164	ng	0.00
42) 2,4,6-Tribromophenol	16.004	330	254805	30.820	ng	0.00
45) 2-Fluorobiphenyl	13.151	172	1578476	30.187	ng	0.00
79) Terphenyl-d14	19.886	244	1950002	32.736	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072222\
Data File : BM035906.D
Acq On : 22 Jul 2022 15:01
Operator : CG/JU
Sample : N3786-11
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
0718-WC-TP-WC3EA

Quant Time: Jul 22 15:33:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM071822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 22 14:25:12 2022
Response via : Initial Calibration

