

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072522\
 Data File : BM035910.D
 Acq On : 25 Jul 2022 11:39
 Operator : CG/JU
 Sample : SSTD01055
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010006

Quant Time: Jul 25 13:22:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 25 13:20:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.886	152	379688	20.000	ng/u1	0.00
20) Naphthalene-d8	10.692	136	1636200	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.521	164	970234	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.262	188	1951130	20.000	ng/u1	0.00
79) Chrysene-d12	21.438	240	1882884	20.000	ng/u1	0.00
88) Perylene-d12	23.809	264	1891346	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	38374	4.034	ng/uL	0.00
4) Pyridine-d5	3.698	84	243811	10.408	ng/u1	0.00
7) Phenol-d5	7.045	99	271642	9.335	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.216	67	190667	10.537	ng/u1	0.00
11) 2-Chlorophenol-d4	7.416	132	233955	9.282	ng/u1	0.00
15) 4-Methylphenol-d8	8.598	113	221698	9.101	ng/u1	0.00
21) Nitrobenzene-d5	9.051	128	114902	9.434	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	106131	8.081	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.310	165	201518	9.037	ng/u1	0.00
31) 4-Chloroaniline-d4	10.833	131	337132	9.609	ng/u1	0.00
46) Dimethylphthalate-d6	13.939	166	590790	8.952	ng/u1	0.00
49) Acenaphthylene-d8	14.215	160	748051	9.147	ng/u1	0.00
54) 4-Nitrophenol-d4	14.715	143	108506	10.703	ng/u1	0.00
60) Fluorene-d10	15.509	176	539414	9.440	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.627	200	72534	6.236	ng/u1	0.00
73) Anthracene-d10	17.356	188	795478	9.070	ng/u1	0.00
81) Pyrene-d10	19.644	212	900999	9.733	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.650	264	840969	9.219	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.316	88	39776	4.125	ng/uL	96
5) Pyridine	3.722	79	262399	10.620	ng/u1	90
6) Benzaldehyde	7.028	77	174278	13.982	ng/u1	94
8) Phenol	7.075	94	305040	9.824	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.310	93	252511	10.412	ng/u1	99
12) 2-Chlorophenol	7.445	128	243594	9.332	ng/u1	98
13) 2-Methylphenol	8.333	108	222076	9.655	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.416	45	334836	9.263	ng/u1	98
16) Acetophenone	8.716	105	370180	10.023	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.698	70	183314	10.519	ng/u1	95
18) 4-Methylphenol	8.663	108	242163	9.783	ng/u1	96
19) Hexachloroethane	8.969	117	101127	9.688	ng/u1	90
22) Nitrobenzene	9.092	77	272107	10.525	ng/u1	98
23) Isophorone	9.622	82	497751	10.179	ng/u1	97
25) 2-Nitrophenol	9.804	139	121696	8.513	ng/u1	97
26) 2,4-Dimethylphenol	9.869	107	258873	9.557	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.110	93	320510	10.107	ng/u1	99
29) 2,4-Dichlorophenol	10.339	162	211411	9.344	ng/u1	98
30) Naphthalene	10.739	128	799475	9.564	ng/u1	99
32) 4-Chloroaniline	10.857	127	339946	9.733	ng/u1	98
33) Hexachlorobutadiene	11.027	225	133012	9.343	ng/u1	96
34) Caprolactam	11.621	113	62999	8.270	ng/u1	86
35) 4-Chloro-3-methylphenol	11.974	107	218549	9.214	ng/u1	98
36) 2-Methylnaphthalene	12.351	142	512394	9.427	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072522\
 Data File : BM035910.D
 Acq On : 25 Jul 2022 11:39
 Operator : CG/JU
 Sample : SSTD01055
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010006

Quant Time: Jul 25 13:22:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 25 13:20:56 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.568	142	522332	9.513	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.716	216	247314	9.103	ng/ul	98
40) Hexachlorocyclopentadiene	12.698	237	144645	14.478	ng/ul	99
41) 2,4,6-Trichlorophenol	12.957	196	151241	8.925	ng/ul	98
42) 2,4,5-Trichlorophenol	13.027	196	165582	9.418	ng/ul	99
43) 1,1'-Biphenyl	13.363	154	684934	9.440	ng/ul	99
44) 2-Chloronaphthalene	13.398	162	520708	9.391	ng/ul	100
45) 2-Nitroaniline	13.604	65	135187	9.457	ng/ul	93
47) Dimethylphthalate	13.986	163	599328	9.004	ng/ul	100
48) 2,6-Dinitrotoluene	14.104	165	108335	8.233	ng/ul	93
50) Acenaphthylene	14.245	152	849842	9.419	ng/ul	100
51) 3-Nitroaniline	14.427	138	129547	9.608	ng/ul	99
52) Acenaphthene	14.586	153	554481	9.264	ng/ul	98
53) 2,4-Dinitrophenol	14.627	184	47044	6.487	ng/ul	96
55) 4-Nitrophenol	14.727	109	88836	13.483	ng/ul	89
56) Dibenzofuran	14.915	168	773886	9.447	ng/ul	97
57) 2,4-Dinitrotoluene	14.880	165	157454	8.241	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.139	232	127009	8.746	ng/ul	90
59) Diethylphthalate	15.345	149	593006	8.793	ng/ul	99
61) Fluorene	15.568	166	626338	9.415	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.562	204	298112	9.548	ng/ul	98
63) 4-Nitroaniline	15.586	138	132402	10.993	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.639	198	74777	6.460	ng/ul	95
67) N-Nitrosodiphenylamine	15.774	169	514535	9.010	ng/ul	99
68) 4-Bromophenyl-phenylether	16.456	248	169710	9.070	ng/ul	97
69) Hexachlorobenzene	16.562	284	206926	9.559	ng/ul	98
70) Atrazine	16.727	200	167244	8.294	ng/ul	98
71) Pentachlorophenol	16.909	266	109056	9.556	ng/ul	98
72) Phenanthrene	17.303	178	951987	9.062	ng/ul	99
74) Anthracene	17.392	178	961845	8.948	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.321	216	266782	8.958	ng/ul	99
76) Pentachlorobenzene	14.833	250	245056	9.254	ng/ul	99
77) Carbazole	17.662	167	884111	9.114	ng/ul	99
78) Di-n-butylphthalate	18.233	149	965268	7.182	ng/ul	99
80) Fluoranthene	19.315	202	1067365	9.479	ng/ul	98
82) Pyrene	19.674	202	1131454	9.600	ng/ul	98
83) Butylbenzylphthalate	20.580	149	385803	7.614	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.356	252	326108	9.496	ng/ul	97
85) Benzo(a)anthracene	21.421	228	1076726	9.471	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.356	149	586227	7.796	ng/ul	97
87) Chrysene	21.474	228	1055899	9.317	ng/ul	99
89) Di-n-octyl phthalate	22.274	149	948368	7.640	ng/ul	100
90) Benzo(b)fluoranthene	23.080	252	1120029	9.970	ng/ul	98
91) Benzo(k)fluoranthene	23.132	252	1027148	9.506	ng/ul	97
93) Benzo(a)pyrene	23.697	252	1060329	9.415	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.268	276	1170956	8.881	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.285	278	1004210	8.901	ng/ul	97
96) Benzo(g,h,i)perylene	27.026	276	977167	8.647	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072522\
Data File : BM035910.D
Acq On : 25 Jul 2022 11:39
Operator : CG/JU
Sample : SST01055
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD010006

Quant Time: Jul 25 13:22:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 25 13:20:56 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.125 ng/uL
RT: 3.316 min Scan# 39
Delta R.T. 0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion: 88 Resp: 39776
Ion Ratio Lower Upper
88 100
43 30.8 27.8 41.8
58 74.5 61.8 92.8



#5
Pyridine
Concen: 10.620 ng/uL
RT: 3.722 min Scan# 108
Delta R.T. 0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion: 79 Resp: 262399
Ion Ratio Lower Upper
79 100
52 74.5 67.8 101.6
51 35.4 31.4 47.0









#6
Benzaldehyde
Concen: 13.982 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion: 77 Resp: 174278
Ion Ratio Lower Upper
77 100
105 98.9 84.6 127.0
106 94.6 80.3 120.5



#8
Phenol
Concen: 9.824 ng/ul
RT: 7.075 min Scan# 678
Delta R.T. -0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion: 94 Resp: 305040
Ion Ratio Lower Upper
94 100
65 26.6 21.3 31.9
66 33.7 27.4 41.0









#10
Bis(2-Chloroethyl)ether
Concen: 10.412 ng/ul
RT: 7.310 min Scan# 718
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion: 93 Resp: 252511
Ion Ratio Lower Upper
93 100
63 79.0 64.0 96.0
95 32.7 25.8 38.6



#12
2-Chlorophenol
Concen: 9.332 ng/ul
RT: 7.445 min Scan# 741
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:128 Resp: 243594
Ion Ratio Lower Upper
128 100
64 46.8 36.9 55.3
130 33.5 25.6 38.4





#13

2-Methylphenol

Concen: 9.655 ng/ul

RT: 8.333 min Scan# 892

Delta R.T. 0.000 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39

Instrument :

BNA_M

ClientSampleId :





SSTD010006

Tgt Ion:108 Resp: 222076

Ion Ratio Lower Upper

108 100

107 88.2 75.0 112.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 9.263 ng/ul

RT: 8.416 min Scan# 906

Delta R.T. 0.000 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39




Tgt Ion: 45 Resp: 334836


Ion Ratio Lower Upper

45 100

77 14.2 10.6 15.8

79 10.5 8.3 12.5









#16
Acetophenone
Concen: 10.023 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:105 Resp: 370180
Ion Ratio Lower Upper
105 100
77 76.5 59.8 89.6
51 26.8 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 10.519 ng/ul
RT: 8.698 min Scan# 954
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion: 70 Resp: 183314
Ion Ratio Lower Upper
70 100
42 57.5 49.6 74.4
101 8.7 8.1 12.1
130 20.7 18.4 27.6









#18
4-Methylphenol
Concen: 9.783 ng/ul
RT: 8.663 min Scan# 948
Delta R.T. -0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:108 Resp: 242163
Ion Ratio Lower Upper
108 100
107 108.4 90.3 135.5



#19
Hexachloroethane
Concen: 9.688 ng/ul
RT: 8.969 min Scan# 1000
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:117 Resp: 101127
Ion Ratio Lower Upper
117 100
201 91.7 81.8 122.6
199 56.6 51.6 77.4









#22
Nitrobenzene
Concen: 10.525 ng/ul
RT: 9.092 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion: 77 Resp: 272107
Ion Ratio Lower Upper
77 100
123 45.1 37.4 56.0
65 12.6 10.8 16.2



#23
Isophorone
Concen: 10.179 ng/ul
RT: 9.622 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion: 82 Resp: 497751
Ion Ratio Lower Upper
82 100
95 6.4 5.5 8.3
138 17.7 15.3 22.9









#25
2-Nitrophenol
Concen: 8.513 ng/ul
RT: 9.804 min Scan# 1142
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:139 Resp: 121696
Ion Ratio Lower Upper
139 100
65 51.5 39.8 59.6
109 29.4 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 9.557 ng/ul
RT: 9.869 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:107 Resp: 258873
Ion Ratio Lower Upper
107 100
121 49.2 42.7 64.1
122 88.0 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 10.107 ng/ul

RT: 10.110 min Scan# 1194

Delta R.T. 0.000 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion: 93 Resp: 320510

Ion Ratio Lower Upper

93 100

95 31.7 26.1 39.1

123 11.7 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 9.344 ng/ul

RT: 10.339 min Scan# 1233

Delta R.T. 0.000 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39




Tgt Ion: 162 Resp: 211411


Ion Ratio Lower Upper

162 100

164 63.3 51.8 77.6

98 35.6 29.5 44.3









#30
Naphthalene
Concen: 9.564 ng/ul
RT: 10.739 min Scan# 1302
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:128 Resp: 799475
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 12.7 10.6 16.0



#32
4-Chloroaniline
Concen: 9.733 ng/ul
RT: 10.857 min Scan# 1321
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:127 Resp: 339946
Ion Ratio Lower Upper
127 100
129 32.9 25.3 37.9









#33
Hexachlorobutadiene
Concen: 9.343 ng/u1
RT: 11.027 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:225 Resp: 133012
Ion Ratio Lower Upper
225 100
223 65.4 50.0 75.0
227 65.7 49.9 74.9



#34
Caprolactam
Concen: 8.270 ng/u1
RT: 11.621 min Scan# 1451
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:113 Resp: 62999
Ion Ratio Lower Upper
113 100
55 164.0 150.2 225.4
56 115.6 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 9.214 ng/ul
RT: 11.974 min Scan# 1511
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:107 Resp: 218549
Ion Ratio Lower Upper
107 100
144 26.5 20.3 30.5
142 82.3 64.4 96.6



#36
2-Methylnaphthalene
Concen: 9.427 ng/ul
RT: 12.351 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:142 Resp: 512394
Ion Ratio Lower Upper
142 100
141 88.0 70.7 106.1





#37

1-Methylnaphthalene

Concen: 9.513 ng/ul

RT: 12.568 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:142 Resp: 522332

Ion Ratio Lower Upper

142 100

141 91.0 75.2 112.8

116 3.7 2.6 3.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 9.103 ng/ul

RT: 12.716 min Scan# 1637

Delta R.T. 0.000 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39

Tgt Ion:216 Resp: 247314




Ion Ratio Lower Upper


216 100

214 78.9 62.4 93.6

179 20.8 17.3 25.9

108 19.5 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 14.478 ng/ul
RT: 12.698 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:237 Resp: 144645
Ion Ratio Lower Upper
237 100
235 60.7 48.4 72.6
272 13.5 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 8.925 ng/ul
RT: 12.957 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:196 Resp: 151241
Ion Ratio Lower Upper
196 100
198 95.3 77.8 116.6
200 30.4 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 9.418 ng/ul
RT: 13.027 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:196 Resp: 165582
Ion Ratio Lower Upper
196 100
198 97.4 76.8 115.2
200 30.8 24.9 37.3



#43
1,1'-Biphenyl
Concen: 9.440 ng/ul
RT: 13.363 min Scan# 1747
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:154 Resp: 684934
Ion Ratio Lower Upper
154 100
153 39.6 32.4 48.6
76 12.9 10.1 15.1









#44
2-Chloronaphthalene
Concen: 9.391 ng/ul
RT: 13.398 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:162 Resp: 520708
Ion Ratio Lower Upper
162 100
164 32.2 25.4 38.0
127 39.6 31.6 47.4



#45
2-Nitroaniline
Concen: 9.457 ng/ul
RT: 13.604 min Scan# 1788
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion: 65 Resp: 135187
Ion Ratio Lower Upper
65 100
92 71.3 54.5 81.7
138 105.3 91.2 136.8









#47
Dimethylphthalate
Concen: 9.004 ng/ul
RT: 13.986 min Scan# 1853
Delta R.T. -0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:163 Resp: 599328
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 9.8 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 8.233 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:165 Resp: 108335
Ion Ratio Lower Upper
165 100
89 59.4 42.5 63.7
121 22.3 17.2 25.8









#50
Acenaphthylene
Concen: 9.419 ng/ul
RT: 14.245 min Scan# 1897
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:152 Resp: 849842
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1
153 13.1 10.6 15.8



#51
3-Nitroaniline
Concen: 9.608 ng/ul
RT: 14.427 min Scan# 1928
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:138 Resp: 129547
Ion Ratio Lower Upper
138 100
108 11.4 8.1 12.1
92 113.8 90.4 135.6









#52
Acenaphthene
Concen: 9.264 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:153 Resp: 554481
Ion Ratio Lower Upper
153 100
152 48.7 37.1 55.7
154 87.8 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 6.487 ng/ul
RT: 14.627 min Scan# 1962
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:184 Resp: 47044
Ion Ratio Lower Upper
184 100
63 71.5 54.2 81.4
154 64.8 49.3 73.9









#55
4-Nitrophenol
Concen: 13.483 ng/ul
RT: 14.727 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:109 Resp: 88836
Ion Ratio Lower Upper
109 100
139 144.7 124.9 187.3
65 135.6 122.2 183.4



#56
Dibenzofuran
Concen: 9.447 ng/ul
RT: 14.915 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:168 Resp: 773886
Ion Ratio Lower Upper
168 100
139 39.3 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 8.241 ng/ul
RT: 14.880 min Scan# 2049
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:165 Resp: 157454
Ion Ratio Lower Upper
165 100
63 48.2 35.9 53.9
182 2.6 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 8.746 ng/ul
RT: 15.139 min Scan# 2049
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:232 Resp: 127009
Ion Ratio Lower Upper
232 100
131 54.2 36.7 55.1
130 2.4 1.7 2.5
166 33.5 24.4 36.6









#59
Diethylphthalate
Concen: 8.793 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:149 Resp: 593006
Ion Ratio Lower Upper
149 100
177 21.0 16.8 25.2
150 12.2 10.2 15.4



#61
Fluorene
Concen: 9.415 ng/ul
RT: 15.568 min Scan# 2122
Delta R.T. -0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:166 Resp: 626338
Ion Ratio Lower Upper
166 100
165 99.5 77.2 115.8
167 13.1 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 9.548 ng/ul
RT: 15.562 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:204 Resp: 298112
Ion Ratio Lower Upper
204 100
206 32.4 25.9 38.9
141 60.6 50.6 75.8



#63
4-Nitroaniline
Concen: 10.993 ng/ul
RT: 15.586 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:138 Resp: 132402
Ion Ratio Lower Upper
138 100
92 49.1 39.2 58.8
108 71.2 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 6.460 ng/ul
RT: 15.639 min Scan# 2134
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:198 Resp: 74777
Ion Ratio Lower Upper
198 100
182 6.1 4.2 6.4
77 27.4 19.9 29.9



#67
N-Nitrosodiphenylamine
Concen: 9.010 ng/ul
RT: 15.774 min Scan# 2157
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:169 Resp: 514535
Ion Ratio Lower Upper
169 100
168 67.8 53.8 80.6
167 37.2 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 9.070 ng/ul
RT: 16.456 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:248 Resp: 169710
Ion Ratio Lower Upper
248 100
250 98.9 77.3 115.9
141 77.1 59.8 89.8



#69
Hexachlorobenzene
Concen: 9.559 ng/ul
RT: 16.562 min Scan# 2291
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:284 Resp: 206926
Ion Ratio Lower Upper
284 100
142 35.8 28.6 42.8
249 28.6 24.6 37.0





#70

Atrazine

Concen: 8.294 ng/ul

RT: 16.727 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:200 Resp: 167244

Ion Ratio Lower Upper

200 100

173 27.9 21.3 31.9

215 47.4 39.0 58.6



#71

Pentachlorophenol

Concen: 9.556 ng/ul

RT: 16.909 min Scan# 2350

Delta R.T. 0.000 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39




Tgt Ion:266 Resp: 109056


Ion Ratio Lower Upper

266 100

264 60.7 49.5 74.3

268 65.1 51.0 76.4









#72
Phenanthrene
Concen: 9.062 ng/ul
RT: 17.303 min Scan# 2417
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:178 Resp: 951987
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 19.1 14.9 22.3



#74
Anthracene
Concen: 8.948 ng/ul
RT: 17.392 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:178 Resp: 961845
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 18.8 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 8.958 ng/uL
RT: 13.321 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:216 Resp: 266782
Ion Ratio Lower Upper
216 100
214 77.1 60.9 91.3
218 46.8 37.5 56.3



#76
Pentachlorobenzene
Concen: 9.254 ng/uL
RT: 14.833 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:250 Resp: 245056
Ion Ratio Lower Upper
250 100
248 62.0 50.4 75.6
252 62.9 51.1 76.7









#77
 Carbazole
 Concen: 9.114 ng/ul
 RT: 17.662 min Scan# 2479
 Delta R.T. -0.006 min
 Lab File: BM035910.D
 Acq: 25 Jul 2022 11:39




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010006


Tgt Ion:167 Resp: 884111
 Ion Ratio Lower Upper
 167 100
 166 22.2 17.4 26.0
 139 13.1 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 7.182 ng/ul
 RT: 18.233 min Scan# 2575
 Delta R.T. -0.006 min
 Lab File: BM035910.D
 Acq: 25 Jul 2022 11:39

Tgt Ion:149 Resp: 965268
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.3 10.9
 104 5.0 3.7 5.5









#80
Fluoranthene
Concen: 9.479 ng/u1
RT: 19.315 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:202 Resp: 1067365
Ion Ratio Lower Upper
202 100
101 12.2 10.1 15.1
100 8.8 7.8 11.6



#82
Pyrene
Concen: 9.600 ng/u1
RT: 19.674 min Scan# 2820
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:202 Resp: 1131454
Ion Ratio Lower Upper
202 100
101 14.8 11.9 17.9
100 11.5 10.2 15.4









#83
Butylbenzylphthalate
Concen: 7.614 ng/ul
RT: 20.580 min Scan# 2974
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:149 Resp: 385803
Ion Ratio Lower Upper
149 100
91 74.4 53.2 79.8
206 18.6 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 9.496 ng/ul
RT: 21.356 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:252 Resp: 326108
Ion Ratio Lower Upper
252 100
254 63.4 52.4 78.6
126 12.2 10.8 16.2









#85
Benzo(a)anthracene
Concen: 9.471 ng/ul
RT: 21.421 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:228 Resp: 1076726
Ion Ratio Lower Upper
228 100
229 19.1 15.6 23.4
226 26.4 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 7.796 ng/ul
RT: 21.356 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:149 Resp: 586227
Ion Ratio Lower Upper
149 100
167 25.5 21.5 32.3
279 3.5 3.4 5.2





#87

Chrysene

Concen: 9.317 ng/ul

RT: 21.474 min Scan# 3126

Delta R.T. 0.000 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:228 Resp: 1055899

Ion Ratio Lower Upper

228 100

226 30.1 23.7 35.5

229 19.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 7.640 ng/ul




RT: 22.274 min Scan# 3262


Delta R.T. -0.006 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39

Tgt Ion:149 Resp: 948368









#90
Benzo(b)fluoranthene
Concen: 9.970 ng/ul
RT: 23.080 min Scan# 311
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:252 Resp: 1120029
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 11.6 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 9.506 ng/ul
RT: 23.132 min Scan# 3408
Delta R.T. 0.000 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:252 Resp: 1027148
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 10.8 10.6 15.8





#93

Benzo(a)pyrene

Concen: 9.415 ng/ul

RT: 23.697 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:252 Resp: 1060329

Ion Ratio Lower Upper

252 100

253 22.3 17.3 25.9

125 12.6 11.8 17.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 8.881 ng/ul

RT: 26.268 min Scan# 3941

Delta R.T. -0.012 min

Lab File: BM035910.D

Acq: 25 Jul 2022 11:39




Tgt Ion:276 Resp: 1170956


Ion Ratio Lower Upper

276 100

138 25.1 25.6 38.4#

277 25.0 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 8.901 ng/ul
RT: 26.285 min Scan# 3946
Delta R.T. -0.012 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Instrument :
BNA_M
ClientSampleId :
SSTD010006

Tgt Ion:278 Resp: 1004210
Ion Ratio Lower Upper
278 100
139 17.6 16.2 24.2
279 24.0 18.7 28.1



#96
Benzo(g,h,i)perylene
Concen: 8.647 ng/ul
RT: 27.026 min Scan# 4070
Delta R.T. -0.006 min
Lab File: BM035910.D
Acq: 25 Jul 2022 11:39

Tgt Ion:276 Resp: 977167
Ion Ratio Lower Upper
276 100
138 23.7 21.3 31.9
277 23.7 18.6 28.0

