

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072522\
 Data File : BM035911.D
 Acq On : 25 Jul 2022 12:16
 Operator : CG/JU
 Sample : SSTD02056
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020007

Quant Time: Jul 25 13:23:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 25 13:20:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.886	152	345168	20.000	ng/u1	0.00
20) Naphthalene-d8	10.692	136	1536309	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.521	164	931575	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.262	188	1944537	20.000	ng/u1	0.00
79) Chrysene-d12	21.438	240	1985133	20.000	ng/u1	0.00
88) Perylene-d12	23.809	264	1992449	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	75598	8.743	ng/uL	0.00
4) Pyridine-d5	3.698	84	515587	24.210	ng/u1	0.00
7) Phenol-d5	7.051	99	592637	22.403	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	405464	24.647	ng/u1	0.00
11) 2-Chlorophenol-d4	7.416	132	511691	22.332	ng/u1	0.00
15) 4-Methylphenol-d8	8.598	113	482466	21.787	ng/u1	0.00
21) Nitrobenzene-d5	9.051	128	256821	22.457	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	248784	20.175	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.310	165	451726	21.576	ng/u1	0.00
31) 4-Chloroaniline-d4	10.833	131	799681	24.274	ng/u1	0.00
46) Dimethylphthalate-d6	13.939	166	1360484	21.470	ng/u1	0.00
49) Acenaphthylene-d8	14.215	160	1728202	22.008	ng/u1	0.00
54) 4-Nitrophenol-d4	14.721	143	277136	28.471	ng/u1	0.00
60) Fluorene-d10	15.509	176	1235355	22.515	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.627	200	196994	16.993	ng/u1	0.00
73) Anthracene-d10	17.362	188	1883634	21.551	ng/u1	0.00
81) Pyrene-d10	19.650	212	2209476	22.639	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.656	264	2104761	21.902	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.310	88	79429	9.061	ng/uL	97
5) Pyridine	3.716	79	546362	24.324	ng/u1	92
6) Benzaldehyde	7.028	77	125793	11.102	ng/u1	92
8) Phenol	7.075	94	656210	23.247	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.316	93	530118	24.045	ng/u1	96
12) 2-Chlorophenol	7.445	128	519855	21.907	ng/u1	100
13) 2-Methylphenol	8.333	108	488315	23.354	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.416	45	707492	21.529	ng/u1	98
16) Acetophenone	8.716	105	803241	23.923	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.704	70	411956	26.004	ng/u1	94
18) 4-Methylphenol	8.663	108	531761	23.631	ng/u1	98
19) Hexachloroethane	8.969	117	210994	22.235	ng/u1	91
22) Nitrobenzene	9.098	77	591870	24.382	ng/u1	99
23) Isophorone	9.622	82	1127959	24.567	ng/u1	98
25) 2-Nitrophenol	9.810	139	279009	20.787	ng/u1	98
26) 2,4-Dimethylphenol	9.869	107	566662	22.279	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.110	93	702169	23.583	ng/u1	99
29) 2,4-Dichlorophenol	10.339	162	471725	22.205	ng/u1	99
30) Naphthalene	10.745	128	1729575	22.036	ng/u1	99
32) 4-Chloroaniline	10.857	127	797951	24.331	ng/u1	99
33) Hexachlorobutadiene	11.027	225	284210	21.262	ng/u1	98
34) Caprolactam	11.627	113	155806	21.782	ng/u1	93
35) 4-Chloro-3-methylphenol	11.980	107	502273	22.554	ng/u1	97
36) 2-Methylnaphthalene	12.351	142	1134713	22.234	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072522\
 Data File : BM035911.D
 Acq On : 25 Jul 2022 12:16
 Operator : CG/JU
 Sample : SST02056
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST020007


Quant Time: Jul 25 13:23:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 25 13:20:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.574	142	1144705	22.205	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.716	216	541572	20.761	ng/ul	97
40) Hexachlorocyclopentadiene	12.698	237	334164	34.836	ng/ul	96
41) 2,4,6-Trichlorophenol	12.957	196	355098	21.824	ng/ul	98
42) 2,4,5-Trichlorophenol	13.027	196	375068	22.217	ng/ul	100
43) 1,1'-Biphenyl	13.363	154	1499891	21.530	ng/ul	100
44) 2-Chloronaphthalene	13.404	162	1152255	21.643	ng/ul	99
45) 2-Nitroaniline	13.610	65	330634	24.088	ng/ul	99
47) Dimethylphthalate	13.986	163	1371017	21.453	ng/ul	99
48) 2,6-Dinitrotoluene	14.104	165	276338	21.872	ng/ul	92
50) Acenaphthylene	14.245	152	1924235	22.211	ng/ul	100
51) 3-Nitroaniline	14.427	138	230835	17.831	ng/ul	96
52) Acenaphthene	14.586	153	1224842	21.313	ng/ul	99
53) 2,4-Dinitrophenol	14.633	184	134754	19.354	ng/ul	98
55) 4-Nitrophenol	14.733	109	212908	33.655	ng/ul	87
56) Dibenzofuran	14.921	168	1726884	21.956	ng/ul	99
57) 2,4-Dinitrotoluene	14.886	165	399598	21.781	ng/ul#	98
58) 2,3,4,6-Tetrachlorophenol	15.145	232	306261	21.965	ng/ul	96
59) Diethylphthalate	15.345	149	1398780	21.601	ng/ul	99
61) Fluorene	15.568	166	1409172	22.061	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.562	204	668525	22.299	ng/ul	99
63) 4-Nitroaniline	15.586	138	218408	18.887	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.645	198	199717	17.312	ng/ul#	98
67) N-Nitrosodiphenylamine	15.780	169	1178582	20.708	ng/ul	99
68) 4-Bromophenyl-phenylether	16.456	248	388557	20.837	ng/ul	98
69) Hexachlorobenzene	16.562	284	475151	22.024	ng/ul	97
70) Atrazine	16.727	200	407212	20.263	ng/ul	98
71) Pentachlorophenol	16.909	266	280110	24.628	ng/ul	99
72) Phenanthrene	17.303	178	2207982	21.088	ng/ul	99
74) Anthracene	17.392	178	2245412	20.959	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.321	216	564602	19.021	ng/ul	97
76) Pentachlorobenzene	14.833	250	545508	20.670	ng/ul	99
77) Carbazole	17.668	167	2104186	21.764	ng/ul	99
78) Di-n-butylphthalate	18.239	149	2366127	17.665	ng/ul	99
80) Fluoranthene	19.315	202	2604486	21.938	ng/ul	99
82) Pyrene	19.680	202	2717201	21.867	ng/ul	96
83) Butylbenzylphthalate	20.580	149	1004139	18.797	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.356	252	750759	20.735	ng/ul	99
85) Benzo(a)anthracene	21.421	228	2654088	22.143	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.356	149	1502582	18.954	ng/ul	98
87) Chrysene	21.474	228	2577573	21.572	ng/ul	99
89) Di-n-octyl phthalate	22.280	149	2400465	18.357	ng/ul	100
90) Benzo(b)fluoranthene	23.085	252	2700256	22.817	ng/ul	98
91) Benzo(k)fluoranthene	23.132	252	2618650	23.005	ng/ul	98
93) Benzo(a)pyrene	23.703	252	2635706	22.215	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.279	276	2908609	20.940	ng/ul	94
95) Dibenzo(a,h)anthracene	26.297	278	2484826	20.907	ng/ul	96
96) Benzo(g,h,i)perylene	27.032	276	2421082	20.337	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 25 13:23:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 25 13:20:56 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.061 ng/uL
RT: 3.310 min Scan# 38
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion: 88 Resp: 79429
Ion Ratio Lower Upper
88 100
43 31.0 27.8 41.8
58 76.0 61.8 92.8



#5
Pyridine
Concen: 24.324 ng/uL
RT: 3.716 min Scan# 107
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion: 79 Resp: 546362
Ion Ratio Lower Upper
79 100
52 76.4 67.8 101.6
51 35.4 31.4 47.0









#6
Benzaldehyde
Concen: 11.102 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion: 77 Resp: 125793
Ion Ratio Lower Upper
77 100
105 97.4 84.6 127.0
106 93.3 80.3 120.5



#8
Phenol
Concen: 23.247 ng/ul
RT: 7.075 min Scan# 678
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion: 94 Resp: 656210
Ion Ratio Lower Upper
94 100
65 26.8 21.3 31.9
66 34.6 27.4 41.0









#10
Bis(2-Chloroethyl)ether
Concen: 24.045 ng/ul
RT: 7.316 min Scan# 719
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion: 93 Resp: 530118
Ion Ratio Lower Upper
93 100
63 76.0 64.0 96.0
95 33.1 25.8 38.6



#12
2-Chlorophenol
Concen: 21.907 ng/ul
RT: 7.445 min Scan# 741
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:128 Resp: 519855
Ion Ratio Lower Upper
128 100
64 46.4 36.9 55.3
130 31.9 25.6 38.4





#13

2-Methylphenol

Concen: 23.354 ng/ul

RT: 8.333 min Scan# 892

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16

Instrument :

BNA_M

ClientSampleId :





SSTD020007

Tgt Ion:108 Resp: 488315

Ion Ratio Lower Upper

108 100

107 88.4 75.0 112.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.529 ng/ul

RT: 8.416 min Scan# 906

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16




Tgt Ion: 45 Resp: 707492


Ion Ratio Lower Upper

45 100

77 14.0 10.6 15.8

79 10.8 8.3 12.5









#16
Acetophenone
Concen: 23.923 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:105 Resp: 803241
Ion Ratio Lower Upper
105 100
77 77.3 59.8 89.6
51 26.5 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 26.004 ng/ul
RT: 8.704 min Scan# 955
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion: 70 Resp: 411956
Ion Ratio Lower Upper
70 100
42 56.5 49.6 74.4
101 9.9 8.1 12.1
130 20.7 18.4 27.6









#18
4-Methylphenol
Concen: 23.631 ng/ul
RT: 8.663 min Scan# 948
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:108 Resp: 531761
Ion Ratio Lower Upper
108 100
107 111.2 90.3 135.5



#19
Hexachloroethane
Concen: 22.235 ng/ul
RT: 8.969 min Scan# 1000
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:117 Resp: 210994
Ion Ratio Lower Upper
117 100
201 94.0 81.8 122.6
199 56.2 51.6 77.4









#22
Nitrobenzene
Concen: 24.382 ng/ul
RT: 9.098 min Scan# 1022
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion: 77 Resp: 591870
Ion Ratio Lower Upper
77 100
123 45.6 37.4 56.0
65 13.2 10.8 16.2



#23
Isophorone
Concen: 24.567 ng/ul
RT: 9.622 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion: 82 Resp: 1127959
Ion Ratio Lower Upper
82 100
95 6.7 5.5 8.3
138 17.8 15.3 22.9









#25
 2-Nitrophenol
 Concen: 20.787 ng/ul
 RT: 9.810 min Scan# 1143
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020007


Tgt Ion:	139	Resp:	279009
Ion Ratio	Lower	Upper	
139	100		
65	48.1	39.8	59.6
109	28.8	22.2	33.4



#26
 2,4-Dimethylphenol
 Concen: 22.279 ng/ul
 RT: 9.869 min Scan# 1153
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16

Tgt Ion:	107	Resp:	566662
Ion Ratio	Lower	Upper	
107	100		
121	50.1	42.7	64.1
122	87.7	71.5	107.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 23.583 ng/ul
 RT: 10.110 min Scan# 1194
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020007


Tgt Ion: 93 Resp: 702169
 Ion Ratio Lower Upper
 93 100
 95 31.8 26.1 39.1
 123 11.6 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 22.205 ng/ul
 RT: 10.339 min Scan# 1233
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16

Tgt Ion: 162 Resp: 471725
 Ion Ratio Lower Upper
 162 100
 164 65.4 51.8 77.6
 98 35.9 29.5 44.3









#30
Naphthalene
Concen: 22.036 ng/ul
RT: 10.745 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:128 Resp: 1729575
Ion Ratio Lower Upper
128 100
129 10.5 8.8 13.2
127 12.8 10.6 16.0



#32
4-Chloroaniline
Concen: 24.331 ng/ul
RT: 10.857 min Scan# 1321
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:127 Resp: 797951
Ion Ratio Lower Upper
127 100
129 32.3 25.3 37.9









#33
Hexachlorobutadiene
Concen: 21.262 ng/u1
RT: 11.027 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:225 Resp: 284210
Ion Ratio Lower Upper
225 100
223 62.4 50.0 75.0
227 64.6 49.9 74.9



#34
Caprolactam
Concen: 21.782 ng/u1
RT: 11.627 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:113 Resp: 155806
Ion Ratio Lower Upper
113 100
55 172.8 150.2 225.4
56 129.3 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 22.554 ng/ul
RT: 11.980 min Scan# 1512
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:107 Resp: 502273
Ion Ratio Lower Upper
107 100
144 26.4 20.3 30.5
142 83.9 64.4 96.6



#36
2-Methylnaphthalene
Concen: 22.234 ng/ul
RT: 12.351 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:142 Resp: 1134713
Ion Ratio Lower Upper
142 100
141 88.8 70.7 106.1





#37

1-Methylnaphthalene

Concen: 22.205 ng/ul

RT: 12.574 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16

Instrument :

BNA_M

ClientSampleId :

SSTD020007





Tgt Ion:142 Resp: 1144705

Ion Ratio Lower Upper

142 100

141 92.3 75.2 112.8

116 3.9 2.6 3.8#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.761 ng/ul

RT: 12.716 min Scan# 1637

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16

Tgt Ion:216 Resp: 541572




Ion Ratio Lower Upper


216 100

214 80.8 62.4 93.6

179 21.7 17.3 25.9

108 19.2 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 34.836 ng/ul
RT: 12.698 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:237 Resp: 334164
Ion Ratio Lower Upper
237 100
235 63.2 48.4 72.6
272 13.2 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 21.824 ng/ul
RT: 12.957 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:196 Resp: 355098
Ion Ratio Lower Upper
196 100
198 94.4 77.8 116.6
200 30.6 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 22.217 ng/ul
RT: 13.027 min Scan# 1690
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:196 Resp: 375068
Ion Ratio Lower Upper
196 100
198 95.7 76.8 115.2
200 31.2 24.9 37.3



#43
1,1'-Biphenyl
Concen: 21.530 ng/ul
RT: 13.363 min Scan# 1747
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:154 Resp: 1499891
Ion Ratio Lower Upper
154 100
153 40.4 32.4 48.6
76 13.2 10.1 15.1









#44
2-Chloronaphthalene
Concen: 21.643 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:162 Resp: 1152255
Ion Ratio Lower Upper
162 100
164 32.8 25.4 38.0
127 39.1 31.6 47.4



#45
2-Nitroaniline
Concen: 24.088 ng/ul
RT: 13.610 min Scan# 1789
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion: 65 Resp: 330634
Ion Ratio Lower Upper
65 100
92 70.7 54.5 81.7
138 113.9 91.2 136.8









#47
 Dimethylphthalate
 Concen: 21.453 ng/ul
 RT: 13.986 min Scan# 1853
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020007


Tgt Ion:163 Resp: 1371017
 Ion Ratio Lower Upper
 163 100
 194 3.8 3.2 4.8
 164 10.3 7.9 11.9



#48
 2,6-Dinitrotoluene
 Concen: 21.872 ng/ul
 RT: 14.104 min Scan# 1873
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16

Tgt Ion:165 Resp: 276338
 Ion Ratio Lower Upper
 165 100
 89 60.2 42.5 63.7
 121 23.3 17.2 25.8









#50
Acenaphthylene
Concen: 22.211 ng/ul
RT: 14.245 min Scan# 1897
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:152 Resp: 1924235
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1
153 13.1 10.6 15.8



#51
3-Nitroaniline
Concen: 17.831 ng/ul
RT: 14.427 min Scan# 1928
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:138 Resp: 230835
Ion Ratio Lower Upper
138 100
108 11.0 8.1 12.1
92 117.7 90.4 135.6









#52
Acenaphthene
Concen: 21.313 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:153 Resp: 1224842
Ion Ratio Lower Upper
153 100
152 48.6 37.1 55.7
154 88.6 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 19.354 ng/ul
RT: 14.633 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:184 Resp: 134754
Ion Ratio Lower Upper
184 100
63 64.9 54.2 81.4
154 62.1 49.3 73.9









#55
4-Nitrophenol
Concen: 33.655 ng/ul
RT: 14.733 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:109 Resp: 212908
Ion Ratio Lower Upper
109 100
139 140.9 124.9 187.3
65 133.9 122.2 183.4



#56
Dibenzofuran
Concen: 21.956 ng/ul
RT: 14.921 min Scan# 1212
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:168 Resp: 1726884
Ion Ratio Lower Upper
168 100
139 38.0 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 21.781 ng/ul
RT: 14.886 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:165 Resp: 399598
Ion Ratio Lower Upper
165 100
63 43.4 35.9 53.9
182 2.3 2.5 3.7#



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.965 ng/ul
RT: 15.145 min Scan# 2050
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:232 Resp: 306261
Ion Ratio Lower Upper
232 100
131 49.6 36.7 55.1
130 2.5 1.7 2.5
166 31.8 24.4 36.6





#59

Diethylphthalate

Concen: 21.601 ng/ul

RT: 15.345 min Scan# 2084

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16

Instrument :

BNA_M

ClientSampleId :

SSTD020007





Tgt Ion:149 Resp: 1398780

Ion Ratio Lower Upper

149 100

177 21.4 16.8 25.2

150 12.3 10.2 15.4



#61

Fluorene

Concen: 22.061 ng/ul

RT: 15.568 min Scan# 2122

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16




Tgt Ion:166 Resp: 1409172


Ion Ratio Lower Upper

166 100

165 98.6 77.2 115.8

167 13.6 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 22.299 ng/ul
RT: 15.562 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:204 Resp: 668525
Ion Ratio Lower Upper
204 100
206 32.5 25.9 38.9
141 62.0 50.6 75.8



#63
4-Nitroaniline
Concen: 18.887 ng/ul
RT: 15.586 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:138 Resp: 218408
Ion Ratio Lower Upper
138 100
92 50.5 39.2 58.8
108 74.6 51.6 77.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.312 ng/ul
 RT: 15.645 min Scan# 2135
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16




Instrument :
 BNA_M
 ClientSampleId :
 SST020007


Tgt Ion:198 Resp: 199717
 Ion Ratio Lower Upper
 198 100
 182 4.0 4.2 6.4#
 77 25.4 19.9 29.9



#67
 N-Nitrosodiphenylamine
 Concen: 20.708 ng/ul
 RT: 15.780 min Scan# 2158
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16

Tgt Ion:169 Resp: 1178582
 Ion Ratio Lower Upper
 169 100
 168 67.6 53.8 80.6
 167 37.4 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 20.837 ng/ul
RT: 16.456 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:248 Resp: 388557
Ion Ratio Lower Upper
248 100
250 98.4 77.3 115.9
141 76.5 59.8 89.8



#69
Hexachlorobenzene
Concen: 22.024 ng/ul
RT: 16.562 min Scan# 2291
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:284 Resp: 475151
Ion Ratio Lower Upper
284 100
142 38.0 28.6 42.8
249 30.1 24.6 37.0









#70
Atrazine
Concen: 20.263 ng/ul
RT: 16.727 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:200 Resp: 407212
Ion Ratio Lower Upper
200 100
173 27.3 21.3 31.9
215 47.5 39.0 58.6



#71
Pentachlorophenol
Concen: 24.628 ng/ul
RT: 16.909 min Scan# 2350
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:266 Resp: 280110
Ion Ratio Lower Upper
266 100
264 62.9 49.5 74.3
268 64.3 51.0 76.4





#72

Phenanthrene

Concen: 21.088 ng/u1

RT: 17.303 min Scan# 2417

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16

Instrument :

BNA_M

ClientSampleId :

SSTD020007





Tgt Ion:178 Resp: 2207982

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.2

176 19.7 14.9 22.3



#74

Anthracene

Concen: 20.959 ng/u1

RT: 17.392 min Scan# 2432

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16




Tgt Ion:178 Resp: 2245412


Ion Ratio Lower Upper

178 100

179 15.8 12.2 18.2

176 19.1 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.021 ng/uL
RT: 13.321 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:216 Resp: 564602
Ion Ratio Lower Upper
216 100
214 80.2 60.9 91.3
218 47.4 37.5 56.3



#76
Pentachlorobenzene
Concen: 20.670 ng/uL
RT: 14.833 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:250 Resp: 545508
Ion Ratio Lower Upper
250 100
248 62.2 50.4 75.6
252 64.1 51.1 76.7









#77
 Carbazole
 Concen: 21.764 ng/ul
 RT: 17.668 min Scan# 2479
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020007

Tgt Ion:167 Resp: 2104186




Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.4	26.0
139	13.4	10.1	15.1




#78
 Di-n-butylphthalate
 Concen: 17.665 ng/ul
 RT: 18.239 min Scan# 2576
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16

Tgt Ion:149 Resp: 2366127

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.3	10.9
104	5.0	3.7	5.5





#80

Fluoranthene

Concen: 21.938 ng/ul

RT: 19.315 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16

Instrument :

BNA_M

ClientSampleId :

SSTD020007





Tgt Ion:202 Resp: 2604486

Ion Ratio Lower Upper

202 100

101 12.3 10.1 15.1

100 9.2 7.8 11.6



#82

Pyrene

Concen: 21.867 ng/ul

RT: 19.680 min Scan# 2821

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16




Tgt Ion:202 Resp: 2717201


Ion Ratio Lower Upper

202 100

101 13.7 11.9 17.9

100 11.0 10.2 15.4









#83
Butylbenzylphthalate
Concen: 18.797 ng/ul
RT: 20.580 min Scan# 2974
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:149 Resp: 1004139
Ion Ratio Lower Upper
149 100
91 73.5 53.2 79.8
206 18.7 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 20.735 ng/ul
RT: 21.356 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:252 Resp: 750759
Ion Ratio Lower Upper
252 100
254 64.7 52.4 78.6
126 12.4 10.8 16.2









#85
Benzo(a)anthracene
Concen: 22.143 ng/ul
RT: 21.421 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:228 Resp: 2654088
Ion Ratio Lower Upper
228 100
229 19.9 15.6 23.4
226 27.1 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.954 ng/ul
RT: 21.356 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:149 Resp: 1502582
Ion Ratio Lower Upper
149 100
167 26.1 21.5 32.3
279 3.5 3.4 5.2





#87

Chrysene

Concen: 21.572 ng/ul

RT: 21.474 min Scan# 3126

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16

Instrument :

BNA_M

ClientSampleId :

SSTD020007





Tgt Ion:228 Resp: 2577573

Ion Ratio Lower Upper

228 100

226 30.3 23.7 35.5

229 19.5 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 18.357 ng/ul




RT: 22.280 min Scan# 3263


Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16

Tgt Ion:149 Resp: 2400465









#90
Benzo(b)fluoranthene
Concen: 22.817 ng/ul
RT: 23.085 min Scan# 3400
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16




Instrument :
BNA_M
ClientSampleId :
SSTD020007


Tgt Ion:252 Resp: 2700256
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 11.2 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 23.005 ng/ul
RT: 23.132 min Scan# 3408
Delta R.T. 0.000 min
Lab File: BM035911.D
Acq: 25 Jul 2022 12:16

Tgt Ion:252 Resp: 2618650
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 11.3 10.6 15.8





#93

Benzo(a)pyrene

Concen: 22.215 ng/ul

RT: 23.703 min Scan# 3505

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16

Instrument :

BNA_M

ClientSampleId :

SSTD020007





Tgt Ion:252 Resp: 2635706

Ion Ratio Lower Upper

252 100

253 21.8 17.3 25.9

125 12.6 11.8 17.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.940 ng/ul

RT: 26.279 min Scan# 3943

Delta R.T. 0.000 min

Lab File: BM035911.D

Acq: 25 Jul 2022 12:16




Tgt Ion:276 Resp: 2908609


Ion Ratio Lower Upper

276 100

138 26.4 25.6 38.4

277 24.7 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 20.907 ng/ul
 RT: 26.297 min Scan# 3946
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020007

Tgt Ion:278 Resp: 2484826

Ion	Ratio	Lower	Upper
278	100		
139	16.7	16.2	24.2
279	24.0	18.7	28.1



#96
 Benzo(g,h,i)perylene
 Concen: 20.337 ng/ul
 RT: 27.032 min Scan# 4071
 Delta R.T. 0.000 min
 Lab File: BM035911.D
 Acq: 25 Jul 2022 12:16

Tgt Ion:276 Resp: 2421082

Ion	Ratio	Lower	Upper
276	100		
138	24.3	21.3	31.9
277	23.9	18.6	28.0

