

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072522\
 Data File : BM035914.D
 Acq On : 25 Jul 2022 14:09
 Operator : CG/JU
 Sample : SSTD16059
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160010

Quant Time: Jul 25 14:43:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 25 13:20:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	415700	20.000	ng/u1	0.00
20) Naphthalene-d8	10.698	136	1782920	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.527	164	1027695	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.268	188	1965657	20.000	ng/u1	0.00
79) Chrysene-d12	21.445	240	1434256	20.000	ng/u1	0.00
88) Perylene-d12	23.815	264	1236683	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	706003	67.796	ng/uL	0.00
4) Pyridine-d5	3.699	84	4878244	190.198	ng/u1	0.00
7) Phenol-d5	7.069	99	5700787	178.937	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	7.234	67	3588336	181.119	ng/u1	0.01
11) 2-Chlorophenol-d4	7.428	132	4867968	176.408	ng/u1	0.01
15) 4-Methylphenol-d8	8.622	113	4537485	170.133	ng/u1	0.02
21) Nitrobenzene-d5	9.069	128	2414633	181.939	ng/u1	0.02
24) 2-Nitrophenol-d4	9.787	143	2548425	178.075	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.328	165	4272088	175.822	ng/u1	0.02
31) 4-Chloroaniline-d4	10.851	131	6446174	168.608	ng/u1	0.02
46) Dimethylphthalate-d6	13.957	166	11186195	160.019	ng/u1	0.02
49) Acenaphthylene-d8	14.227	160	14131386	163.127	ng/u1	0.01
54) 4-Nitrophenol-d4	14.763	143	2319708	216.019	ng/u1	0.04
60) Fluorene-d10	15.527	176	9871890	163.095	ng/u1	0.02
65) 4,6-Dinitro-2-methylph...	15.657	200	1974147	168.460	ng/u1	0.03
73) Anthracene-d10	17.374	188	13752501	155.651	ng/u1	0.01
81) Pyrene-d10	19.662	212	13880007	196.844	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.674	264	10663993	178.782	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.316	88	729419	69.091	ng/uL	92
5) Pyridine	3.716	79	4984640	184.267	ng/u1	92
6) Benzaldehyde	7.028	77	1620062	118.718	ng/u1	95
8) Phenol	7.099	94	6067166	178.471	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.328	93	4757551	179.180	ng/u1	96
12) 2-Chlorophenol	7.463	128	4903214	171.566	ng/u1	96
13) 2-Methylphenol	8.345	108	4572514	181.582	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.434	45	6197174	156.581	ng/u1	97
16) Acetophenone	8.734	105	6838703	169.122	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.751	70	3424573	179.492	ng/u1	92
18) 4-Methylphenol	8.704	108	4790431	176.763	ng/u1	95
19) Hexachloroethane	8.975	117	2024478	177.142	ng/u1	94
22) Nitrobenzene	9.116	77	5082453	180.409	ng/u1	96
23) Isophorone	9.657	82	9809904	184.106	ng/u1	99
25) 2-Nitrophenol	9.822	139	2683345	172.268	ng/u1	97
26) 2,4-Dimethylphenol	9.887	107	5037371	170.658	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.128	93	6046967	175.002	ng/u1	99
29) 2,4-Dichlorophenol	10.357	162	4306205	174.664	ng/u1	98
30) Naphthalene	10.751	128	14034861	154.084	ng/u1	97
32) 4-Chloroaniline	10.875	127	6345456	166.719	ng/u1	97
33) Hexachlorobutadiene	11.034	225	2682621	172.931	ng/u1	98
34) Caprolactam	11.716	113	770267	92.790	ng/u1	91
35) 4-Chloro-3-methylphenol	12.010	107	4496869	173.994	ng/u1	95
36) 2-Methylnaphthalene	12.363	142	9702626	163.821	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072522\
 Data File : BM035914.D
 Acq On : 25 Jul 2022 14:09
 Operator : CG/JU
 Sample : SSTD16059
 Misc :
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD160010


Quant Time: Jul 25 14:43:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 25 13:20:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.581	142	9644517	161.205	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.728	216	4936219	171.530	ng/ul	99
40) Hexachlorocyclopentadiene	12.704	237	3493195	330.098	ng/ul	96
41) 2,4,6-Trichlorophenol	12.969	196	3350687	186.670	ng/ul	99
42) 2,4,5-Trichlorophenol	13.045	196	3554270	190.848	ng/ul	100
43) 1,1'-Biphenyl	13.375	154	12034576	156.593	ng/ul	96
44) 2-Chloronaphthalene	13.416	162	9518400	162.060	ng/ul	97
45) 2-Nitroaniline	13.628	65	2909158	192.123	ng/ul	96
47) Dimethylphthalate	14.004	163	11035008	156.523	ng/ul	97
48) 2,6-Dinitrotoluene	14.127	165	2553007	183.166	ng/ul	97
50) Acenaphthylene	14.257	152	13779999	144.185	ng/ul#	90
51) 3-Nitroaniline	14.457	138	2515663	176.149	ng/ul	94
52) Acenaphthene	14.598	153	9890662	156.008	ng/ul	97
53) 2,4-Dinitrophenol	14.657	184	1533483	199.647	ng/ul	95
55) 4-Nitrophenol	14.780	109	1728819	247.723	ng/ul#	83
56) Dibenzofuran	14.933	168	12825255	147.811	ng/ul#	87
57) 2,4-Dinitrotoluene	14.910	165	3486635	172.274	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.157	232	2851944	185.410	ng/ul	96
59) Diethylphthalate	15.369	149	11322045	158.489	ng/ul	97
61) Fluorene	15.586	166	10322764	146.492	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.574	204	5193105	157.019	ng/ul	96
63) 4-Nitroaniline	15.627	138	1841260	144.332	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.674	198	1946686	166.928	ng/ul#	97
67) N-Nitrosodiphenylamine	15.792	169	9383504	163.097	ng/ul	98
68) 4-Bromophenyl-phenylether	16.468	248	3448022	182.919	ng/ul	97
69) Hexachlorobenzene	16.574	284	3965814	181.849	ng/ul	99
70) Atrazine	16.751	200	3188455	156.950	ng/ul	99
71) Pentachlorophenol	16.921	266	2505197	217.898	ng/ul	99
72) Phenanthrene	17.310	178	14669178	138.600	ng/ul	93
74) Anthracene	17.404	178	14038181	129.628	ng/ul	93
75) 1,2,3,4-Tetrachloroben...	13.333	216	5166704	172.196	ng/ul	97
76) Pentachlorobenzene	14.851	250	4707077	176.442	ng/ul	98
77) Carbazole	17.680	167	14045202	143.711	ng/ul#	89
78) Di-n-butylphthalate	18.239	149	13953482	103.056	ng/ul#	87
80) Fluoranthene	19.321	202	14748811	171.946	ng/ul	96
82) Pyrene	19.686	202	13913349	154.974	ng/ul	99
83) Butylbenzylphthalate	20.592	149	6642452	172.101	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.368	252	3493612	133.550	ng/ul#	98
85) Benzo(a)anthracene	21.433	228	13159305	151.954	ng/ul#	86
86) Bis(2-ethylhexyl)phtha...	21.362	149	8592557	150.019	ng/ul	95
87) Chrysene	21.486	228	12531317	145.158	ng/ul#	83
89) Di-n-octyl phthalate	22.280	149	12631520	155.625	ng/ul	100
90) Benzo(b)fluoranthene	23.150	252	12222433	166.397	ng/ul	95
91) Benzo(k)fluoranthene	23.150	252	12261749	173.550	ng/ul	95
93) Benzo(a)pyrene	23.868	252	117540	1.596	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.315	276	15018984	174.202	ng/ul	93
95) Dibenzo(a,h)anthracene	26.333	278	12613112	170.983	ng/ul	95
96) Benzo(g,h,i)perylene	27.080	276	12604500	170.585	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 25 14:43:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 25 13:20:56 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 69.091 ng/uL
RT: 3.316 min Scan# 38
Delta R.T. 0.006 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion: 88 Resp: 729419
Ion Ratio Lower Upper
88 100
43 28.0 27.8 41.8
58 72.2 61.8 92.8



#5
Pyridine
Concen: 184.267 ng/uL
RT: 3.716 min Scan# 107
Delta R.T. 0.000 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion: 79 Resp: 4984640
Ion Ratio Lower Upper
79 100
52 77.6 67.8 101.6
51 34.6 31.4 47.0









#6
Benzaldehyde
Concen: 118.718 ng/ul
RT: 7.028 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion: 77 Resp: 1620062
Ion Ratio Lower Upper
77 100
105 99.2 84.6 127.0
106 96.0 80.3 120.5



#8
Phenol
Concen: 178.471 ng/ul
RT: 7.099 min Scan# 682
Delta R.T. 0.024 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion: 94 Resp: 6067166
Ion Ratio Lower Upper
94 100
65 27.7 21.3 31.9
66 35.2 27.4 41.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 179.180 ng/ul
 RT: 7.328 min Scan# 719
 Delta R.T. 0.012 min
 Lab File: BM035914.D
 Acq: 25 Jul 2022 14:09




Instrument :
 BNA_M
 ClientSampleId :
 SSTD160010


Tgt Ion: 93 Resp: 4757551
 Ion Ratio Lower Upper
 93 100
 63 75.4 64.0 96.0
 95 33.0 25.8 38.6



#12
 2-Chlorophenol
 Concen: 171.566 ng/ul
 RT: 7.463 min Scan# 744
 Delta R.T. 0.018 min
 Lab File: BM035914.D
 Acq: 25 Jul 2022 14:09

Tgt Ion:128 Resp: 4903214
 Ion Ratio Lower Upper
 128 100
 64 42.4 36.9 55.3
 130 32.9 25.6 38.4









#13
2-Methylphenol
Concen: 181.582 ng/ul
RT: 8.345 min Scan# 892
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:108 Resp: 4572514
Ion Ratio Lower Upper
108 100
107 88.2 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 156.581 ng/ul
RT: 8.434 min Scan# 909
Delta R.T. 0.018 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion: 45 Resp: 6197174
Ion Ratio Lower Upper
45 100
77 14.8 10.6 15.8
79 11.5 8.3 12.5









#16
Acetophenone
Concen: 169.122 ng/ul
RT: 8.734 min Scan# 90
Delta R.T. 0.018 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:105 Resp: 6838703
Ion Ratio Lower Upper
105 100
77 74.2 59.8 89.6
51 24.7 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 179.492 ng/ul
RT: 8.751 min Scan# 963
Delta R.T. 0.047 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion: 70 Resp: 3424573
Ion Ratio Lower Upper
70 100
42 52.9 49.6 74.4
101 10.0 8.1 12.1
130 23.1 18.4 27.6









#18
4-Methylphenol
Concen: 176.763 ng/ul
RT: 8.704 min Scan# 91
Delta R.T. 0.041 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:108 Resp: 4790431
Ion Ratio Lower Upper
108 100
107 107.3 90.3 135.5



#19
Hexachloroethane
Concen: 177.142 ng/ul
RT: 8.975 min Scan# 1001
Delta R.T. 0.006 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:117 Resp: 2024478
Ion Ratio Lower Upper
117 100
201 95.8 81.8 122.6
199 60.5 51.6 77.4









#22
Nitrobenzene
Concen: 180.409 ng/ul
RT: 9.116 min Scan# 1025
Delta R.T. 0.018 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion: 77 Resp: 5082453
Ion Ratio Lower Upper
77 100
123 49.7 37.4 56.0
65 13.6 10.8 16.2



#23
Isophorone
Concen: 184.106 ng/ul
RT: 9.657 min Scan# 1117
Delta R.T. 0.036 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion: 82 Resp: 9809904
Ion Ratio Lower Upper
82 100
95 7.1 5.5 8.3
138 18.8 15.3 22.9









#25
2-Nitrophenol
Concen: 172.268 ng/ul
RT: 9.822 min Scan# 1145
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:139 Resp: 2683345
Ion Ratio Lower Upper
139 100
65 47.1 39.8 59.6
109 28.6 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 170.658 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. 0.018 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:107 Resp: 5037371
Ion Ratio Lower Upper
107 100
121 51.4 42.7 64.1
122 88.8 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 175.002 ng/ul

RT: 10.128 min Scan# 1194

Delta R.T. 0.018 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Instrument :





BNA_M

ClientSampleId :

SSTD160010

Tgt Ion: 93 Resp: 6046967

Ion	Ratio	Lower	Upper
93	100		
95	33.3	26.1	39.1
123	12.0	9.4	14.0



#29

2,4-Dichlorophenol

Concen: 174.664 ng/ul

RT: 10.357 min Scan# 1236




Delta R.T. 0.018 min


Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Tgt Ion: 162 Resp: 4306205

Ion	Ratio	Lower	Upper
162	100		
164	65.7	51.8	77.6
98	35.6	29.5	44.3









#30
Naphthalene
Concen: 154.084 ng/ul
RT: 10.751 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:128 Resp:14034861
Ion Ratio Lower Upper
128 100
129 12.3 8.8 13.2
127 14.3 10.6 16.0



#32
4-Chloroaniline
Concen: 166.719 ng/ul
RT: 10.875 min Scan# 1324
Delta R.T. 0.018 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:127 Resp: 6345456
Ion Ratio Lower Upper
127 100
129 33.2 25.3 37.9





#33

Hexachlorobutadiene

Concen: 172.931 ng/ul

RT: 11.034 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Instrument :

BNA_M

ClientSampleId :

SSTD160010





Tgt Ion:225 Resp: 2682621

Ion Ratio Lower Upper

225 100

223 63.5 50.0 75.0

227 64.8 49.9 74.9



#34

Caprolactam

Concen: 92.790 ng/ul

RT: 11.716 min Scan# 1467

Delta R.T. 0.088 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09




Tgt Ion:113 Resp: 770267


Ion Ratio Lower Upper

113 100

55 168.8 150.2 225.4

56 123.3 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 173.994 ng/ul
RT: 12.010 min Scan# 1517
Delta R.T. 0.030 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:107 Resp: 4496869
Ion Ratio Lower Upper
107 100
144 27.5 20.3 30.5
142 84.8 64.4 96.6



#36
2-Methylnaphthalene
Concen: 163.821 ng/ul
RT: 12.363 min Scan# 1577
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:142 Resp: 9702626
Ion Ratio Lower Upper
142 100
141 88.4 70.7 106.1





#37

1-Methylnaphthalene

Concen: 161.205 ng/ul

RT: 12.581 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Instrument :

BNA_M

ClientSampleId :

SSTD160010





Tgt Ion:142 Resp: 9644517

Ion Ratio Lower Upper

142 100

141 91.9 75.2 112.8

116 4.0 2.6 3.8#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 171.530 ng/ul

RT: 12.728 min Scan# 1639

Delta R.T. 0.012 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Tgt Ion:216 Resp: 4936219




Ion Ratio Lower Upper


216 100

214 79.7 62.4 93.6

179 21.9 17.3 25.9

108 17.6 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 330.098 ng/ul
RT: 12.704 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:237 Resp: 3493195
Ion Ratio Lower Upper
237 100
235 63.8 48.4 72.6
272 13.4 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 186.670 ng/ul
RT: 12.969 min Scan# 1680
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:196 Resp: 3350687
Ion Ratio Lower Upper
196 100
198 96.0 77.8 116.6
200 30.8 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 190.848 ng/ul
RT: 13.045 min Scan# 1
Delta R.T. 0.018 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:196 Resp: 3554270
Ion Ratio Lower Upper
196 100
198 95.9 76.8 115.2
200 31.3 24.9 37.3



#43
1,1'-Biphenyl
Concen: 156.593 ng/ul
RT: 13.375 min Scan# 1749
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:154 Resp:12034576
Ion Ratio Lower Upper
154 100
153 43.5 32.4 48.6
76 12.0 10.1 15.1





#44

2-Chloronaphthalene

Concen: 162.060 ng/ul

RT: 13.416 min Scan# 1

Delta R.T. 0.012 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Instrument :

BNA_M

ClientSampleId :

SSTD160010





Tgt Ion:162 Resp: 9518400

Ion Ratio Lower Upper

162 100

164 34.9 25.4 38.0

127 40.5 31.6 47.4



#45

2-Nitroaniline

Concen: 192.123 ng/ul

RT: 13.628 min Scan# 1792

Delta R.T. 0.018 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09




Tgt Ion: 65 Resp: 2909158


Ion Ratio Lower Upper

65 100

92 72.3 54.5 81.7

138 117.2 91.2 136.8









#47
Dimethylphthalate
Concen: 156.523 ng/ul
RT: 14.004 min Scan# 1856
Delta R.T. 0.018 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:163 Resp:11035008
Ion Ratio Lower Upper
163 100
194 4.2 3.2 4.8
164 11.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 183.166 ng/ul
RT: 14.127 min Scan# 1877
Delta R.T. 0.024 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:165 Resp: 2553007
Ion Ratio Lower Upper
165 100
89 56.4 42.5 63.7
121 21.5 17.2 25.8









#50
Acenaphthylene
Concen: 144.185 ng/ul
RT: 14.257 min Scan# 1899
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:152 Resp:13779999
Ion Ratio Lower Upper
152 100
151 24.9 16.1 24.1#
153 16.8 10.6 15.8#



#51
3-Nitroaniline
Concen: 176.149 ng/ul
RT: 14.457 min Scan# 1933
Delta R.T. 0.030 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:138 Resp: 2515663
Ion Ratio Lower Upper
138 100
108 9.6 8.1 12.1
92 106.3 90.4 135.6









#52
Acenaphthene
Concen: 156.008 ng/ul
RT: 14.598 min Scan# 1957
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:153 Resp: 9890662
Ion Ratio Lower Upper
153 100
152 50.0 37.1 55.7
154 90.9 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 199.647 ng/ul
RT: 14.657 min Scan# 1967
Delta R.T. 0.024 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:184 Resp: 1533483
Ion Ratio Lower Upper
184 100
63 61.3 54.2 81.4
154 62.1 49.3 73.9









#55
4-Nitrophenol
Concen: 247.723 ng/ul
RT: 14.780 min Scan# 1980
Delta R.T. 0.047 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:109 Resp: 1728819
Ion Ratio Lower Upper
109 100
139 143.9 124.9 187.3
65 120.4 122.2 183.4#



#56
Dibenzofuran
Concen: 147.811 ng/ul
RT: 14.933 min Scan# 2014
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:168 Resp:12825255
Ion Ratio Lower Upper
168 100
139 45.5 30.1 45.1#





#57

2,4-Dinitrotoluene

Concen: 172.274 ng/ul

RT: 14.910 min Scan# 2006

Delta R.T. 0.024 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Instrument :

BNA_M

ClientSampleId :

SSTD160010





Tgt Ion:165 Resp: 3486635

Ion Ratio Lower Upper

165 100

63 41.5 35.9 53.9

182 2.6 2.5 3.7



#58

2,3,4,6-Tetrachlorophenol

Concen: 185.410 ng/ul

RT: 15.157 min Scan# 2052

Delta R.T. 0.012 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Tgt Ion:232 Resp: 2851944




Ion Ratio Lower Upper


232 100

131 48.6 36.7 55.1

130 2.5 1.7 2.5

166 32.3 24.4 36.6









#59
Diethylphthalate
Concen: 158.489 ng/ul
RT: 15.369 min Scan# 2088
Delta R.T. 0.024 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:149 Resp:11322045
Ion Ratio Lower Upper
149 100
177 23.2 16.8 25.2
150 13.3 10.2 15.4



#61
Fluorene
Concen: 146.492 ng/ul
RT: 15.586 min Scan# 2125
Delta R.T. 0.018 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:166 Resp:10322764
Ion Ratio Lower Upper
166 100
165 97.6 77.2 115.8
167 14.1 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 157.019 ng/ul
RT: 15.574 min Scan# 2121
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:204 Resp: 5193105
Ion Ratio Lower Upper
204 100
206 33.5 25.9 38.9
141 59.6 50.6 75.8



#63
4-Nitroaniline
Concen: 144.332 ng/ul
RT: 15.627 min Scan# 2132
Delta R.T. 0.041 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:138 Resp: 1841260
Ion Ratio Lower Upper
138 100
92 48.0 39.2 58.8
108 76.6 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 166.928 ng/ul
RT: 15.674 min Scan# 2135
Delta R.T. 0.030 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:198 Resp: 1946686
Ion Ratio Lower Upper
198 100
182 3.2 4.2 6.4#
77 23.5 19.9 29.9



#67
N-Nitrosodiphenylamine
Concen: 163.097 ng/ul
RT: 15.792 min Scan# 2160
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:169 Resp: 9383504
Ion Ratio Lower Upper
169 100
168 68.4 53.8 80.6
167 38.3 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 182.919 ng/ul
RT: 16.468 min Scan# 21
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:248 Resp: 3448022
Ion Ratio Lower Upper
248 100
250 97.4 77.3 115.9
141 69.3 59.8 89.8



#69
Hexachlorobenzene
Concen: 181.849 ng/ul
RT: 16.574 min Scan# 2293
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:284 Resp: 3965814
Ion Ratio Lower Upper
284 100
142 34.9 28.6 42.8
249 30.9 24.6 37.0









#70
Atrazine
Concen: 156.950 ng/ul
RT: 16.751 min Scan# 21
Delta R.T. 0.024 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:200 Resp: 3188455
Ion Ratio Lower Upper
200 100
173 27.9 21.3 31.9
215 48.8 39.0 58.6



#71
Pentachlorophenol
Concen: 217.898 ng/ul
RT: 16.921 min Scan# 2352
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:266 Resp: 2505197
Ion Ratio Lower Upper
266 100
264 63.5 49.5 74.3
268 64.3 51.0 76.4





#72

Phenanthrene

Concen: 138.600 ng/ul

RT: 17.310 min Scan# 2417

Delta R.T. 0.006 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Instrument :

BNA_M

ClientSampleId :

SSTD160010





Tgt Ion:178 Resp:14669178

Ion Ratio Lower Upper

178 100

179 17.8 12.2 18.2

176 22.0 14.9 22.3



#74

Anthracene

Concen: 129.628 ng/ul

RT: 17.404 min Scan# 2434

Delta R.T. 0.012 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09




Tgt Ion:178 Resp:14038181


Ion Ratio Lower Upper

178 100

179 17.8 12.2 18.2

176 21.1 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 172.196 ng/uL
RT: 13.333 min Scan# 1
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:216 Resp: 5166704
Ion Ratio Lower Upper
216 100
214 79.2 60.9 91.3
218 48.6 37.5 56.3



#76
Pentachlorobenzene
Concen: 176.442 ng/uL
RT: 14.851 min Scan# 2000
Delta R.T. 0.018 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:250 Resp: 4707077
Ion Ratio Lower Upper
250 100
248 64.4 50.4 75.6
252 65.3 51.1 76.7





#77

Carbazole

Concen: 143.711 ng/ul

RT: 17.680 min Scan# 2479

Delta R.T. 0.012 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Instrument :

BNA_M

ClientSampleId :

SSTD160010





Tgt Ion:167 Resp:14045202

Ion Ratio Lower Upper

167 100

166 27.0 17.4 26.0#

139 16.6 10.1 15.1#



#78

Di-n-butylphthalate

Concen: 103.056 ng/ul

RT: 18.239 min Scan# 2576

Delta R.T. 0.000 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09




Tgt Ion:149 Resp:13953482


Ion Ratio Lower Upper

149 100

150 14.7 7.3 10.9#

104 7.6 3.7 5.5#









#80
Fluoranthene
Concen: 171.946 ng/uI
RT: 19.321 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:202 Resp:14748811
Ion Ratio Lower Upper
202 100
101 14.4 10.1 15.1
100 10.8 7.8 11.6



#82
Pyrene
Concen: 154.974 ng/uI
RT: 19.686 min Scan# 2822
Delta R.T. 0.006 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:202 Resp:13913349
Ion Ratio Lower Upper
202 100
101 15.9 11.9 17.9
100 12.9 10.2 15.4









#83
Butylbenzylphthalate
Concen: 172.101 ng/ul
RT: 20.592 min Scan# 2976
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:149 Resp: 6642452
Ion Ratio Lower Upper
149 100
91 64.2 53.2 79.8
206 21.8 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 133.550 ng/ul
RT: 21.368 min Scan# 3108
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:252 Resp: 3493612
Ion Ratio Lower Upper
252 100
254 64.8 52.4 78.6
126 10.0 10.8 16.2#









#85
Benzo(a)anthracene
Concen: 151.954 ng/ul
RT: 21.433 min Scan# 3117
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:228 Resp:13159305
Ion Ratio Lower Upper
228 100
229 25.1 15.6 23.4#
226 35.0 21.7 32.5#



#86
Bis(2-ethylhexyl)phthalate
Concen: 150.019 ng/ul
RT: 21.362 min Scan# 3107
Delta R.T. 0.006 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:149 Resp: 8592557
Ion Ratio Lower Upper
149 100
167 30.1 21.5 32.3
279 4.5 3.4 5.2









#87
Chrysene
Concen: 145.158 ng/ul
RT: 21.486 min Scan# 3126
Delta R.T. 0.012 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09




Instrument :
BNA_M
ClientSampleId :
SSTD160010


Tgt Ion:228 Resp:12531317
Ion Ratio Lower Upper
228 100
226 39.6 23.7 35.5#
229 25.6 15.5 23.3#



#89
Di-n-octyl phthalate
Concen: 155.625 ng/ul
RT: 22.280 min Scan# 3263
Delta R.T. 0.000 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:149 Resp:12631520





#90

Benzo(b)fluoranthene

Concen: 166.397 ng/ul

RT: 23.150 min Scan# 3411

Delta R.T. 0.065 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Instrument :

BNA_M

ClientSampleId :

SSTD160010





Tgt Ion:252 Resp:12222433

Ion Ratio Lower Upper

252 100

253 23.9 17.5 26.3

125 10.9 10.7 16.1



#91

Benzo(k)fluoranthene

Concen: 173.550 ng/ul

RT: 23.150 min Scan# 3411

Delta R.T. 0.018 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09




Tgt Ion:252 Resp:12261749


Ion Ratio Lower Upper

252 100

253 23.9 17.4 26.0

125 10.9 10.6 15.8





#93

Benzo(a)pyrene

Concen: 1.596 ng/ul

RT: 23.868 min Scan# 3

Delta R.T. 0.165 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09

Instrument :

BNA_M

ClientSampleId :

SSTD160010





Tgt Ion:252 Resp: 117540

Ion Ratio Lower Upper

252 100

253 22.2 17.3 25.9

125 15.3 11.8 17.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 174.202 ng/ul

RT: 26.315 min Scan# 3949

Delta R.T. 0.036 min

Lab File: BM035914.D

Acq: 25 Jul 2022 14:09




Tgt Ion:276 Resp:15018984


Ion Ratio Lower Upper

276 100

138 25.8 25.6 38.4

277 25.8 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 170.983 ng/ul
RT: 26.333 min Scan# 3946
Delta R.T. 0.036 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Instrument :
BNA_M
ClientSampleId :
SSTD160010

Tgt Ion:278 Resp:12613112
Ion Ratio Lower Upper
278 100
139 16.5 16.2 24.2
279 24.9 18.7 28.1



#96
Benzo(g,h,i)perylene
Concen: 170.585 ng/ul
RT: 27.080 min Scan# 4079
Delta R.T. 0.047 min
Lab File: BM035914.D
Acq: 25 Jul 2022 14:09

Tgt Ion:276 Resp:12604500
Ion Ratio Lower Upper
276 100
138 23.0 21.3 31.9
277 24.6 18.6 28.0

