

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072522\
 Data File : BM035915.D
 Acq On : 25 Jul 2022 15:27
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV011

Quant Time: Jul 25 17:35:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 25 14:55:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	339716	20.000	ng/u1	0.00
20) Naphthalene-d8	10.692	136	1531526	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.521	164	954966	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.262	188	1977488	20.000	ng/u1	0.00
79) Chrysene-d12	21.439	240	1995180	20.000	ng/u1	0.00
88) Perylene-d12	23.809	264	2007562	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	73331	8.123	ng/uL	0.00
4) Pyridine-d5	3.699	84	495261	20.025	ng/u1	0.00
7) Phenol-d5	7.051	99	578756	20.316	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.216	67	396890	20.898	ng/u1	0.00
11) 2-Chlorophenol-d4	7.416	132	499294	21.115	ng/u1	0.00
15) 4-Methylphenol-d8	8.598	113	481899	20.880	ng/u1	0.00
21) Nitrobenzene-d5	9.051	128	258130	21.270	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	252820	21.382	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.310	165	456242	21.335	ng/u1	0.00
31) 4-Chloroaniline-d4	10.833	131	809056	22.349	ng/u1	0.00
46) Dimethylphthalate-d6	13.939	166	1428364	21.958	ng/u1	0.00
49) Acenaphthylene-d8	14.216	160	1770137	21.328	ng/u1	0.00
54) 4-Nitrophenol-d4	14.721	143	281887	21.102	ng/u1	0.00
60) Fluorene-d10	15.510	176	1263746	21.499	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.627	200	204939	19.280	ng/u1	0.00
73) Anthracene-d10	17.362	188	1929629	21.583	ng/u1	0.00
81) Pyrene-d10	19.651	212	2224027	20.113	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.656	264	2115224	20.753	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.310	88	76424	8.169	ng/uL	93
5) Pyridine	3.716	79	520710	20.063	ng/u1	90
6) Benzaldehyde	7.022	77	239642	20.042	ng/u1	92
8) Phenol	7.075	94	641720	20.584	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.310	93	524274	20.917	ng/u1	97
12) 2-Chlorophenol	7.445	128	514864	21.225	ng/u1	100
13) 2-Methylphenol	8.334	108	485529	20.847	ng/u1	93
14) 2,2'-oxybis(1-Chloropr...	8.428	45	705248	21.317	ng/u1	96
16) Acetophenone	8.716	105	809571	21.848	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.704	70	411714	22.281	ng/u1	94
18) 4-Methylphenol	8.663	108	525955	21.051	ng/u1	98
19) Hexachloroethane	8.969	117	212390	21.221	ng/u1	91
22) Nitrobenzene	9.092	77	589205	21.216	ng/u1	97
23) Isophorone	9.622	82	1146900	21.877	ng/u1	97
25) 2-Nitrophenol	9.804	139	284663	21.701	ng/u1	100
26) 2,4-Dimethylphenol	9.869	107	572220	21.349	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.110	93	707512	21.399	ng/u1	100
29) 2,4-Dichlorophenol	10.339	162	475924	21.442	ng/u1	98
30) Naphthalene	10.745	128	1714694	21.211	ng/u1	99
32) 4-Chloroaniline	10.857	127	802411	22.313	ng/u1	100
33) Hexachlorobutadiene	11.028	225	290893	21.230	ng/u1	99
34) Caprolactam	11.628	113	156918	21.392	ng/u1	90
35) 4-Chloro-3-methylphenol	11.980	107	507540	21.991	ng/u1	98
36) 2-Methylnaphthalene	12.351	142	1141214	21.578	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072522\
 Data File : BM035915.D
 Acq On : 25 Jul 2022 15:27
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SICV011


Quant Time: Jul 25 17:35:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 25 14:55:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.569	142	1154178	21.650	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.716	216	555496	20.435	ng/ul	99
40) Hexachlorocyclopentadiene	12.698	237	351828	19.433	ng/ul	98
41) 2,4,6-Trichlorophenol	12.957	196	368239	21.168	ng/ul	97
42) 2,4,5-Trichlorophenol	13.027	196	392542	21.148	ng/ul	99
43) 1,1'-Biphenyl	13.363	154	1526921	20.774	ng/ul	99
44) 2-Chloronaphthalene	13.404	162	1167165	20.691	ng/ul	98
45) 2-Nitroaniline	13.610	65	336741	21.653	ng/ul	98
47) Dimethylphthalate	13.986	163	1441379	22.029	ng/ul	100
48) 2,6-Dinitrotoluene	14.104	165	289687	22.300	ng/ul	94
50) Acenaphthylene	14.245	152	1959526	21.343	ng/ul	100
51) 3-Nitroaniline	14.427	138	232200	16.174	ng/ul	99
52) Acenaphthene	14.586	153	1244313	20.890	ng/ul	98
53) 2,4-Dinitrophenol	14.633	184	140632	19.066	ng/ul	97
55) 4-Nitrophenol	14.733	109	215316	21.075	ng/ul	85
56) Dibenzofuran	14.921	168	1763975	21.404	ng/ul	99
57) 2,4-Dinitrotoluene	14.886	165	416991	22.772	ng/ul#	96
58) 2,3,4,6-Tetrachlorophenol	15.145	232	317510	21.771	ng/ul	97
59) Diethylphthalate	15.345	149	1471340	22.231	ng/ul	99
61) Fluorene	15.568	166	1452077	21.569	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.563	204	690325	21.401	ng/ul	98
63) 4-Nitroaniline	15.586	138	216429	16.375	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.645	198	212306	19.960	ng/ul	98
67) N-Nitrosodiphenylamine	15.774	169	1203881	20.708	ng/ul	99
68) 4-Bromophenyl-phenylether	16.457	248	410544	21.177	ng/ul	98
69) Hexachlorobenzene	16.563	284	500872	21.204	ng/ul	98
70) Atrazine	16.727	200	425320	21.093	ng/ul	98
71) Pentachlorophenol	16.910	266	291007	20.166	ng/ul	97
72) Phenanthrene	17.304	178	2248044	21.210	ng/ul	98
74) Anthracene	17.392	178	2278799	21.427	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.322	216	585216	19.030	ng/ul	98
76) Pentachlorobenzene	14.833	250	567173	20.347	ng/ul	99
77) Carbazole	17.662	167	2118264	21.565	ng/ul	99
78) Di-n-butylphthalate	18.233	149	2504401	22.598	ng/ul	99
80) Fluoranthene	19.315	202	2629623	19.904	ng/ul	98
82) Pyrene	19.680	202	2695561	19.795	ng/ul	97
83) Butylbenzylphthalate	20.586	149	1042770	21.055	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.362	252	769004	20.887	ng/ul	99
85) Benzo(a)anthracene	21.421	228	2668007	21.246	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.362	149	1572302	21.971	ng/ul	99
87) Chrysene	21.474	228	2585103	21.098	ng/ul	99
89) Di-n-octyl phthalate	22.280	149	2538175	19.578	ng/ul	100
90) Benzo(b)fluoranthene	23.086	252	2707921	20.211	ng/ul	97
91) Benzo(k)fluoranthene	23.133	252	2634493	20.721	ng/ul	97
93) Benzo(a)pyrene	23.703	252	2667909	20.906	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.274	276	2975632	21.876	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.297	278	2572274	22.042	ng/ul	97
96) Benzo(g,h,i)perylene	27.032	276	2468532	21.899	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 25 17:35:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 25 14:55:39 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.169 ng/uL
RT: 3.310 min Scan# 38
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion: 88 Resp: 76424
Ion Ratio Lower Upper
88 100
43 29.1 27.8 41.8
58 72.2 61.8 92.8



#5
Pyridine
Concen: 20.063 ng/uL
RT: 3.716 min Scan# 107
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion: 79 Resp: 520710
Ion Ratio Lower Upper
79 100
52 74.2 67.8 101.6
51 34.4 31.4 47.0









#6
Benzaldehyde
Concen: 20.042 ng/ul
RT: 7.022 min Scan# 669
Delta R.T. -0.006 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion: 77 Resp: 239642
Ion Ratio Lower Upper
77 100
105 97.3 84.6 127.0
106 93.6 80.3 120.5



#8
Phenol
Concen: 20.584 ng/ul
RT: 7.075 min Scan# 678
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion: 94 Resp: 641720
Ion Ratio Lower Upper
94 100
65 27.4 21.3 31.9
66 34.6 27.4 41.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.917 ng/ul
RT: 7.310 min Scan# 719
Delta R.T. -0.006 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion: 93 Resp: 524274
Ion Ratio Lower Upper
93 100
63 76.7 64.0 96.0
95 32.8 25.8 38.6



#12
2-Chlorophenol
Concen: 21.225 ng/ul
RT: 7.445 min Scan# 741
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion: 128 Resp: 514864
Ion Ratio Lower Upper
128 100
64 46.4 36.9 55.3
130 31.9 25.6 38.4





#13

2-Methylphenol

Concen: 20.847 ng/ul

RT: 8.334 min Scan# 892

Delta R.T. 0.000 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27

Instrument :

BNA_M

ClientSampleId :





SICV011

Tgt Ion:108 Resp: 485529

Ion Ratio Lower Upper

108 100

107 87.4 75.0 112.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.317 ng/ul

RT: 8.428 min Scan# 908

Delta R.T. 0.012 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27




Tgt Ion: 45 Resp: 705248


Ion Ratio Lower Upper

45 100

77 14.8 10.6 15.8

79 11.5 8.3 12.5









#16
Acetophenone
Concen: 21.848 ng/ul
RT: 8.716 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:105 Resp: 809571
Ion Ratio Lower Upper
105 100
77 74.4 59.8 89.6
51 26.0 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 22.281 ng/ul
RT: 8.704 min Scan# 955
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion: 70 Resp: 411714
Ion Ratio Lower Upper
70 100
42 56.0 49.6 74.4
101 9.8 8.1 12.1
130 22.2 18.4 27.6









#18
4-Methylphenol
Concen: 21.051 ng/ul
RT: 8.663 min Scan# 948
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:108 Resp: 525955
Ion Ratio Lower Upper
108 100
107 110.8 90.3 135.5



#19
Hexachloroethane
Concen: 21.221 ng/ul
RT: 8.969 min Scan# 1000
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:117 Resp: 212390
Ion Ratio Lower Upper
117 100
201 92.1 81.8 122.6
199 59.6 51.6 77.4









#22
Nitrobenzene
Concen: 21.216 ng/ul
RT: 9.092 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion: 77 Resp: 589205
Ion Ratio Lower Upper
77 100
123 44.4 37.4 56.0
65 12.8 10.8 16.2



#23
Isophorone
Concen: 21.877 ng/ul
RT: 9.622 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion: 82 Resp: 1146900
Ion Ratio Lower Upper
82 100
95 6.4 5.5 8.3
138 17.8 15.3 22.9









#25
2-Nitrophenol
Concen: 21.701 ng/ul
RT: 9.804 min Scan# 1142
Delta R.T. -0.006 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:139 Resp: 284663
Ion Ratio Lower Upper
139 100
65 49.6 39.8 59.6
109 28.3 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 21.349 ng/ul
RT: 9.869 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:107 Resp: 572220
Ion Ratio Lower Upper
107 100
121 50.7 42.7 64.1
122 87.0 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 21.399 ng/ul

RT: 10.110 min Scan# 1194

Delta R.T. 0.000 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV011





Tgt Ion: 93 Resp: 707512

Ion Ratio Lower Upper

93 100

95 32.6 26.1 39.1

123 11.9 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 21.442 ng/ul

RT: 10.339 min Scan# 1233

Delta R.T. 0.000 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27




Tgt Ion:162 Resp: 475924


Ion Ratio Lower Upper

162 100

164 65.3 51.8 77.6

98 34.7 29.5 44.3









#30
Naphthalene
Concen: 21.211 ng/ul
RT: 10.745 min Scan# 1302
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:128 Resp: 1714694
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 12.8 10.6 16.0



#32
4-Chloroaniline
Concen: 22.313 ng/ul
RT: 10.857 min Scan# 1321
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:127 Resp: 802411
Ion Ratio Lower Upper
127 100
129 31.7 25.3 37.9









#33
Hexachlorobutadiene
Concen: 21.230 ng/u1
RT: 11.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:225 Resp: 290893
Ion Ratio Lower Upper
225 100
223 61.6 50.0 75.0
227 62.2 49.9 74.9



#34
Caprolactam
Concen: 21.392 ng/u1
RT: 11.628 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:113 Resp: 156918
Ion Ratio Lower Upper
113 100
55 169.6 150.2 225.4
56 120.1 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 21.991 ng/ul
RT: 11.980 min Scan# 1512
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:107 Resp: 507540
Ion Ratio Lower Upper
107 100
144 27.3 20.3 30.5
142 82.4 64.4 96.6



#36
2-Methylnaphthalene
Concen: 21.578 ng/ul
RT: 12.351 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:142 Resp: 1141214
Ion Ratio Lower Upper
142 100
141 89.4 70.7 106.1





#37

1-Methylnaphthalene

Concen: 21.650 ng/ul

RT: 12.569 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV011





Tgt Ion:142 Resp: 1154178

Ion Ratio Lower Upper

142 100

141 93.4 75.2 112.8

116 3.8 2.6 3.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.435 ng/ul

RT: 12.716 min Scan# 1637

Delta R.T. 0.000 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27

Tgt Ion:216 Resp: 555496




Ion Ratio Lower Upper


216 100

214 78.4 62.4 93.6

179 21.4 17.3 25.9

108 18.4 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 19.433 ng/ul
RT: 12.698 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:237 Resp: 351828
Ion Ratio Lower Upper
237 100
235 62.1 48.4 72.6
272 12.9 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 21.168 ng/ul
RT: 12.957 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:196 Resp: 368239
Ion Ratio Lower Upper
196 100
198 93.6 77.8 116.6
200 30.1 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 21.148 ng/ul
RT: 13.027 min Scan# 1690
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:196 Resp: 392542
Ion Ratio Lower Upper
196 100
198 94.9 76.8 115.2
200 30.5 24.9 37.3



#43
1,1'-Biphenyl
Concen: 20.774 ng/ul
RT: 13.363 min Scan# 1747
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:154 Resp: 1526921
Ion Ratio Lower Upper
154 100
153 40.0 32.4 48.6
76 13.2 10.1 15.1









#44
2-Chloronaphthalene
Concen: 20.691 ng/ul
RT: 13.404 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:162 Resp: 1167165
Ion Ratio Lower Upper
162 100
164 32.6 25.4 38.0
127 38.5 31.6 47.4



#45
2-Nitroaniline
Concen: 21.653 ng/ul
RT: 13.610 min Scan# 1789
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion: 65 Resp: 336741
Ion Ratio Lower Upper
65 100
92 71.1 54.5 81.7
138 114.7 91.2 136.8









#47
Dimethylphthalate
Concen: 22.029 ng/ul
RT: 13.986 min Scan# 1853
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:163 Resp: 1441379
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.1 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 22.300 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:165 Resp: 289687
Ion Ratio Lower Upper
165 100
89 59.2 42.5 63.7
121 21.5 17.2 25.8









#50
Acenaphthylene
Concen: 21.343 ng/ul
RT: 14.245 min Scan# 1897
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:152 Resp: 1959526
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1
153 13.2 10.6 15.8



#51
3-Nitroaniline
Concen: 16.174 ng/ul
RT: 14.427 min Scan# 1928
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:138 Resp: 232200
Ion Ratio Lower Upper
138 100
108 10.3 8.1 12.1
92 112.1 90.4 135.6









#52
Acenaphthene
Concen: 20.890 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:153 Resp: 1244313
Ion Ratio Lower Upper
153 100
152 48.4 37.1 55.7
154 88.2 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 19.066 ng/ul
RT: 14.633 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:184 Resp: 140632
Ion Ratio Lower Upper
184 100
63 65.9 54.2 81.4
154 63.6 49.3 73.9









#55
4-Nitrophenol
Concen: 21.075 ng/ul
RT: 14.733 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:109 Resp: 215316
Ion Ratio Lower Upper
109 100
139 140.6 124.9 187.3
65 128.8 122.2 183.4



#56
Dibenzofuran
Concen: 21.404 ng/ul
RT: 14.921 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:168 Resp: 1763975
Ion Ratio Lower Upper
168 100
139 38.0 30.1 45.1









#57
 2,4-Dinitrotoluene
 Concen: 22.772 ng/ul
 RT: 14.886 min Scan# 2006
 Delta R.T. 0.000 min
 Lab File: BM035915.D
 Acq: 25 Jul 2022 15:27




Instrument :
 BNA_M
 ClientSampleId :
 SICV011


Tgt Ion:165 Resp: 416991
 Ion Ratio Lower Upper
 165 100
 63 42.0 35.9 53.9
 182 2.4 2.5 3.7#



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 21.771 ng/ul
 RT: 15.145 min Scan# 2050
 Delta R.T. 0.000 min
 Lab File: BM035915.D
 Acq: 25 Jul 2022 15:27

Tgt Ion:232 Resp: 317510
 Ion Ratio Lower Upper
 232 100
 131 48.3 36.7 55.1
 130 2.5 1.7 2.5
 166 32.1 24.4 36.6









#59
Diethylphthalate
Concen: 22.231 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:149 Resp: 1471340
Ion Ratio Lower Upper
149 100
177 20.9 16.8 25.2
150 12.2 10.2 15.4



#61
Fluorene
Concen: 21.569 ng/ul
RT: 15.568 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:166 Resp: 1452077
Ion Ratio Lower Upper
166 100
165 96.9 77.2 115.8
167 13.4 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 21.401 ng/ul
RT: 15.563 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:204 Resp: 690325
Ion Ratio Lower Upper
204 100
206 33.0 25.9 38.9
141 61.0 50.6 75.8



#63
4-Nitroaniline
Concen: 16.375 ng/ul
RT: 15.586 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:138 Resp: 216429
Ion Ratio Lower Upper
138 100
92 49.1 39.2 58.8
108 75.3 51.6 77.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.960 ng/ul
 RT: 15.645 min Scan# 2135
 Delta R.T. 0.000 min
 Lab File: BM035915.D
 Acq: 25 Jul 2022 15:27




Instrument :
 BNA_M
 ClientSampleId :
 SICV011


Tgt Ion:198 Resp: 212306
 Ion Ratio Lower Upper
 198 100
 182 4.5 4.2 6.4
 77 24.1 19.9 29.9



#67
 N-Nitrosodiphenylamine
 Concen: 20.708 ng/ul
 RT: 15.774 min Scan# 2157
 Delta R.T. -0.006 min
 Lab File: BM035915.D
 Acq: 25 Jul 2022 15:27

Tgt Ion:169 Resp: 1203881
 Ion Ratio Lower Upper
 169 100
 168 67.6 53.8 80.6
 167 36.7 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 21.177 ng/ul
RT: 16.457 min Scan# 211
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:248 Resp: 410544
Ion Ratio Lower Upper
248 100
250 99.2 77.3 115.9
141 75.1 59.8 89.8



#69
Hexachlorobenzene
Concen: 21.204 ng/ul
RT: 16.563 min Scan# 2291
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:284 Resp: 500872
Ion Ratio Lower Upper
284 100
142 36.4 28.6 42.8
249 29.5 24.6 37.0









#70
Atrazine
Concen: 21.093 ng/ul
RT: 16.727 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:200 Resp: 425320
Ion Ratio Lower Upper
200 100
173 27.5 21.3 31.9
215 47.1 39.0 58.6



#71
Pentachlorophenol
Concen: 20.166 ng/ul
RT: 16.910 min Scan# 2350
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:266 Resp: 291007
Ion Ratio Lower Upper
266 100
264 63.5 49.5 74.3
268 66.3 51.0 76.4









#72
Phenanthrene
Concen: 21.210 ng/ul
RT: 17.304 min Scan# 2417
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:178 Resp: 2248044
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 19.7 14.9 22.3



#74
Anthracene
Concen: 21.427 ng/ul
RT: 17.392 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:178 Resp: 2278799
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 18.9 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.030 ng/uL
RT: 13.322 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:216 Resp: 585216
Ion Ratio Lower Upper
216 100
214 78.1 60.9 91.3
218 47.4 37.5 56.3



#76
Pentachlorobenzene
Concen: 20.347 ng/uL
RT: 14.833 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:250 Resp: 567173
Ion Ratio Lower Upper
250 100
248 62.4 50.4 75.6
252 63.2 51.1 76.7





#77

Carbazole

Concen: 21.565 ng/ul

RT: 17.662 min Scan# 24

Delta R.T. -0.006 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV011





Tgt Ion:167 Resp: 2118264

Ion Ratio Lower Upper

167 100

166 21.4 17.4 26.0

139 13.6 10.1 15.1



#78

Di-n-butylphthalate

Concen: 22.598 ng/ul

RT: 18.233 min Scan# 2575

Delta R.T. -0.006 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27




Tgt Ion:149 Resp: 2504401


Ion Ratio Lower Upper

149 100

150 9.3 7.3 10.9

104 5.1 3.7 5.5









#80
Fluoranthene
Concen: 19.904 ng/u1
RT: 19.315 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:202 Resp: 2629623
Ion Ratio Lower Upper
202 100
101 12.0 10.1 15.1
100 9.1 7.8 11.6



#82
Pyrene
Concen: 19.795 ng/u1
RT: 19.680 min Scan# 2821
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:202 Resp: 2695561
Ion Ratio Lower Upper
202 100
101 13.7 11.9 17.9
100 11.2 10.2 15.4









#83
Butylbenzylphthalate
Concen: 21.055 ng/ul
RT: 20.586 min Scan# 2974
Delta R.T. 0.006 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:149 Resp: 1042770
Ion Ratio Lower Upper
149 100
91 68.1 53.2 79.8
206 19.7 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 20.887 ng/ul
RT: 21.362 min Scan# 3107
Delta R.T. 0.006 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:252 Resp: 769004
Ion Ratio Lower Upper
252 100
254 65.6 52.4 78.6
126 11.3 10.8 16.2





#85

Benzo(a)anthracene

Concen: 21.246 ng/ul

RT: 21.421 min Scan# 3117

Delta R.T. 0.000 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV011





Tgt Ion:228 Resp: 2668007

Ion Ratio Lower Upper

228 100

229 19.9 15.6 23.4

226 26.9 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.971 ng/ul

RT: 21.362 min Scan# 3107

Delta R.T. 0.006 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27




Tgt Ion:149 Resp: 1572302


Ion Ratio Lower Upper

149 100

167 26.7 21.5 32.3

279 3.9 3.4 5.2





#87

Chrysene

Concen: 21.098 ng/ul

RT: 21.474 min Scan# 3126

Delta R.T. 0.000 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV011





Tgt Ion:228 Resp: 2585103

Ion Ratio Lower Upper

228 100

226 30.0 23.7 35.5

229 19.9 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 19.578 ng/ul




RT: 22.280 min Scan# 3263


Delta R.T. 0.000 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27

Tgt Ion:149 Resp: 2538175









#90
Benzo(b)fluoranthene
Concen: 20.211 ng/ul
RT: 23.086 min Scan# 3400
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV011


Tgt Ion:252 Resp: 2707921
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 11.0 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 20.721 ng/ul
RT: 23.133 min Scan# 3408
Delta R.T. 0.000 min
Lab File: BM035915.D
Acq: 25 Jul 2022 15:27

Tgt Ion:252 Resp: 2634493
Ion Ratio Lower Upper
252 100
253 22.3 17.4 26.0
125 11.0 10.6 15.8





#93

Benzo(a)pyrene

Concen: 20.906 ng/ul

RT: 23.703 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV011





Tgt Ion:252 Resp: 2667909

Ion Ratio Lower Upper

252 100

253 22.1 17.3 25.9

125 12.4 11.8 17.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.876 ng/ul

RT: 26.274 min Scan# 3942

Delta R.T. -0.006 min

Lab File: BM035915.D

Acq: 25 Jul 2022 15:27




Tgt Ion:276 Resp: 2975632


Ion Ratio Lower Upper

276 100

138 25.0 25.6 38.4#

277 23.9 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 22.042 ng/ul
 RT: 26.297 min Scan# 3946
 Delta R.T. 0.000 min
 Lab File: BM035915.D
 Acq: 25 Jul 2022 15:27

Instrument :
 BNA_M
 ClientSampleId :
 SICV011

Tgt Ion:278 Resp: 2572274

Ion	Ratio	Lower	Upper
278	100		
139	17.3	16.2	24.2
279	24.0	18.7	28.1



#96
 Benzo(g,h,i)perylene
 Concen: 21.899 ng/ul
 RT: 27.032 min Scan# 4071
 Delta R.T. 0.000 min
 Lab File: BM035915.D
 Acq: 25 Jul 2022 15:27

Tgt Ion:276 Resp: 2468532

Ion	Ratio	Lower	Upper
276	100		
138	23.6	21.3	31.9
277	23.4	18.6	28.0

