

Data Path : Z:\HPCHEM1\BNA M\DATA\BM072617\
 Data File : BM010971.D
 Acq On : 26 Jul 2017 15:04
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM072617

Quant Time: Jul 26 15:46:47 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 26 15:23:06 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	239155	20.00	ng	0.00
21) Naphthalene-d8	10.15	136	957666	20.00	ng	0.00
38) Acenaphthene-d10	14.05	164	651413	20.00	ng	0.00
63) Phenanthrene-d10	16.81	188	1604676	20.00	ng	0.00
75) Chrysene-d12	21.03	240	1774157	20.00	ng	0.00
86) Perylene-d12	23.14	264	1540092	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.03	112	1155722	81.69	ng	0.00
7) Phenol-d6	6.59	99	1673061	83.23	ng	0.00
23) Nitrobenzene-d5	8.54	82	2134184	83.64	ng	0.00
41) 2,4,6-Tribromophenol	15.56	330	840805	80.53	ng	0.00
44) 2-Fluorobiphenyl	12.66	172	4185862	80.59	ng	0.00
78) Terphenyl-d14	19.47	244	7628227	85.17	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.03	88	241475	38.015	ng	# 100
3) Pyridine	3.41	79	705638	40.669	ng	# 88
4) n-Nitrosodimethylamine	3.33	42	368196	41.642	ng	# 73
6) Aniline	6.74	93	1030623	40.681	ng	# 93
8) 2-Chlorophenol	6.96	128	593420	41.315	ng	# 81
9) Benzaldehyde	6.55	77	524936	40.362	ng	# 80
10) Phenol	6.62	94	880484	40.328	ng	# 96
11) bis(2-Chloroethyl)ether	6.84	93	692527	40.709	ng	# 93
12) 1,3-Dichlorobenzene	7.28	146	706620	40.367	ng	# 85
13) 1,4-Dichlorobenzene	7.43	146	720603	40.161	ng	# 91
14) 1,2-Dichlorobenzene	7.74	146	698602	40.722	ng	# 92
15) Benzyl Alcohol	7.64	79	772155	40.043	ng	# 84
16) 2,2'-oxybis(1-Chloropropan	7.93	45	616377	40.265	ng	# 76
17) 2-Methylphenol	7.85	107	573931	41.131	ng	# 78
18) Hexachloroethane	8.45	117	308371	39.409	ng	# 73
19) n-Nitroso-di-n-propylamine	8.20	70	681409	39.892	ng	# 93
20) 3+4-Methylphenols	8.17	107	784040	40.421	ng	# 86
22) Acetophenone	8.21	105	1182690	41.199	ng	# 95
24) Nitrobenzene	8.58	77	1083106	40.963	ng	# 90
25) Isophorone	9.10	82	1737572	40.842	ng	# 86
26) 2-Nitrophenol	9.28	139	364283	43.299	ng	# 25
27) 2,4-Dimethylphenol	9.36	122	598596	40.417	ng	# 69
28) bis(2-Chloroethoxy)methane	9.59	93	973231	41.015	ng	# 97
29) 2,4-Dichlorophenol	9.81	162	679912	40.854	ng	# 95
30) 1,2,4-Trichlorobenzene	10.02	180	851678	41.389	ng	# 99
31) Naphthalene	10.20	128	1974917	40.993	ng	# 100
32) Benzoic acid	9.51	122	487290	40.933	ng	# 70
33) 4-Chloroaniline	10.32	127	784471	40.818	ng	# 76
34) Hexachlorobutadiene	10.49	225	660889	40.777	ng	# 98
35) Caprolactam	11.12	113	187597	39.748	ng	# 93
36) 4-Chloro-3-methylphenol	11.46	107	878077	40.861	ng	# 77
37) 2-Methylnaphthalene	11.83	142	1448812	40.937	ng	# 87
39) 1,2,4,5-Tetrachlorobenzene	12.21	216	1151997	41.162	ng	# 100
40) Hexachlorocyclopentadiene	12.19	237	679221	41.650	ng	# 98

Data Path : Z:\HPCHEM1\BNA M\DATA\BM072617\
 Data File : BM010971.D
 Acq On : 26 Jul 2017 15:04
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM072617

Quant Time: Jul 26 15:46:47 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 26 15:23:06 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)


42) 2,4,6-Trichlorophenol	12.46	196	702728	41.485	nq	98
43) 2,4,5-Trichlorophenol	12.53	196	689299	39.509	nq	# 88
45) 1,1'-Biphenyl	12.87	154	2273324	40.360	nq	98
46) 2-Chloronaphthalene	12.91	162	1664604	40.110	nq	92
47) 2-Nitroaniline	13.13	65	606059	41.279	nq	# 69
48) Acenaphthylene	13.77	152	2509951	40.526	nq	100
49) Dimethylphthalate	13.53	163	2237557	40.153	nq	# 98
50) 2,6-Dinitrotoluene	13.64	165	470763	41.776	nq	# 61
51) Acenaphthene	14.12	154	1542731	40.229	nq	98
52) 3-Nitroaniline	13.97	138	411943	42.126	nq	# 44
53) 2,4-Dinitrophenol	14.18	184	324141	40.474	nq	# 88
54) Dibenzofuran	14.46	168	2470566	39.760	nq	# 90
55) 4-Nitrophenol	14.29	139	351400	40.901	nq	# 1
56) 2,4-Dinitrotoluene	14.44	165	648803	41.012	nq	# 96
57) Fluorene	15.11	166	2161592	39.955	nq	99
58) 2,3,4,6-Tetrachlorophenol	14.69	232	650491	39.778	nq	# 100
59) Diethylphthalate	14.91	149	2283645	40.127	nq	99
60) 4-Chlorophenyl-phenylether	15.12	204	1292836	39.580	nq	99
61) 4-Nitroaniline	15.14	138	431114	41.506	nq	# 1
62) Azobenzene	15.41	77	2427367	39.853	nq	89
64) 4,6-Dinitro-2-methylphenol	15.20	198	459736	42.320	nq	92
65) n-Nitrosodiphenylamine	15.33	169	1868684	40.691	nq	99
66) 4-Bromophenyl-phenylether	16.01	248	855253	41.459	nq	# 88
67) Hexachlorobenzene	16.12	284	949496	40.732	nq	# 88
68) Atrazine	16.30	200	782903	42.541	nq	95
69) Pentachlorophenol	16.46	266	625902	42.293	nq	98
70) Phenanthrene	16.85	178	3452834	41.093	nq	99
71) Anthracene	16.94	178	3473255	41.448	nq	99
72) Carbazole	17.22	167	3029638	41.341	nq	99
73) Di-n-butylphthalate	17.80	149	3788325	42.449	nq	# 97
74) Fluoranthene	18.88	202	4268411	40.993	nq	98
76) Benzidine	19.08	184	1759976	41.467	nq	99
77) Pyrene	19.24	202	4396328	41.704	nq	99
79) Butylbenzylphthalate	20.19	149	1614664	43.073	nq	# 71
80) Benzo(a)anthracene	21.01	228	4192441	41.385	nq	99
81) 3,3'-Dichlorobenzidine	20.96	252	1689207	42.505	nq	# 98
82) Chrysene	21.06	228	3936880	40.486	nq	100
83) Bis(2-ethylhexyl)phthalate	20.97	149	2366686	42.916	nq	# 99
84) Di-n-octyl phthalate	21.82	149	3836996	42.869	nq	100
85) Indeno(1,2,3-cd)pyrene	25.22	276	4127979	40.989	nq	# 100
87) Benzo(b)fluoranthene	22.52	252	4007002	40.538	nq	# 93
88) Benzo(k)fluoranthene	22.56	252	3867542	40.823	nq	# 94
89) Benzo(a)pyrene	23.05	252	3646665	40.771	nq	# 97
90) Dibenzo(a,h)anthracene	25.23	278	3356972	40.487	nq	# 92
91) Benzo(g,h,i)perylene	25.85	276	3276675	40.245	nq	# 86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\HPCHEM1\BNA M\DATA\BM072617\
Data File : BM010971.D
Acq On    : 26 Jul 2017   15:04
Operator  : SJ/JU
Sample    : SSTDICV040
Misc      :
ALS Vial  : 9      Sample Multiplier: 1
```

Instrument :
BNA_M
ClientSampleId :
ICVBM072617

Quant Time: Jul 26 15:46:47 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 26 15:23:06 2017
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.39 min Scan# 766
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Instrument :
BNA_M
ClientSampleId :
ICVBM072617




Tot Ion	Ratio	Lower	Upper
152	100		
150	150.3	119.5	179.3
115	66.0	43.0	64.4




#2

1,4-Dioxane
Concen: 38.015 ng
RT: 3.03 min Scan# 25
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion	Ratio	Lower	Upper
88	100		
58	61.5	0.0	0.0
43	30.1	0.0	0.0





#3

Pvridine

Concen: 40.669 ng

RT: 3.41 min Scan# 89

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion: 79 Resp: 705638

Ion Ratio Lower Upper

79 100

52 49.5 51.1 76.7#

51 32.4 26.1 39.1




Abundance Ion 79.00 (78.70 to 79.70): BM010967.D (-84) (-)

Ion 52.00 (51.70 to 52.70): BM010967.D (-84) (-)

Ion 51.00 (50.70 to 51.70): BM010967.D (-84) (-)


3.41

Time-->



Abundance Ion 79.00 (78.70 to 79.70): BM010971.D (-38) (-)

Time-->



#4

n-Nitrosodimethylamine

Concen: 41.642 ng

RT: 3.33 min Scan# 75

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion: 42 Resp: 368196

Ion Ratio Lower Upper

42 100

74 113.5 119.3 178.9#

44 3.7 5.4 8.2#




Abundance Ion 42.00 (41.70 to 42.70): BM010967.D (-70) (-)

Ion 74.00 (73.70 to 74.70): BM010967.D (-70) (-)

Ion 44.00 (43.70 to 44.70): BM010967.D (-70) (-)


3.33

Time-->



Abundance Ion 42.00 (41.70 to 42.70): BM010971.D (-58) (-)

Time-->



#5

2-Fluorophenol

Concen: 81.691 ng

RT: 5.03 min Scan# 364

Delta R.T. 0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion:112 Resp: 1155722

Ion Ratio Lower Upper

112 100

64 67.1 38.6 57.8#




63 43.6 19.3 28.9#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BM0

Ion 63.00 (62.70 to 63.70): BM0



#6

Aniline

Concen: 40.681 ng

RT: 6.74 min Scan# 655

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion: 93 Resp: 1030623

Ion Ratio Lower Upper

93 100

66 44.0 30.8 46.2



65 20.9 16.0 24.0




Abundance Ion 93.00 (92.70 to 93.70): BM0

Ion 66.00 (65.70 to 66.70): BM0

Ion 65.00 (64.70 to 65.70): BM0





#7

Phenol-d6

Concen: 83.233 ng

RT: 6.59 min Scan# 630

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion: 99 Resp: 1673061

Ion Ratio Lower Upper

99 100

42 22.0 11.0 16.4#




71 53.2 23.4 35.2#



Abundance Ion 99.00 (98.70 to 99.70): BM010971.D (-613) (-)

Ion 42.00 (41.70 to 42.70): BM010971.D (-613) (-)

Ion 71.00 (70.70 to 71.70): BM010971.D (-613) (-)



#8

2-Chlorophenol

Concen: 41.315 ng

RT: 6.96 min Scan# 693

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion: 128 Resp: 593420

Ion Ratio Lower Upper

128 100

130 29.2 12.0 52.0



64 63.3 24.9 64.9




Abundance Ion 128.00 (127.70 to 128.70): BM010971.D (-676) (-)

Ion 130.00 (129.70 to 130.70): BM010971.D (-676) (-)

Ion 64.00 (63.70 to 64.70): BM010971.D (-676) (-)





#9

Benzaldehyde

Concen: 40.362 ng

RT: 6.55 min Scan# 623

Delta R.T. 0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion: 77 Resp: 524936

Ion Ratio Lower Upper

77 100

105 78.4 78.8 118.8#




106 75.6 73.7 113.7



Abundance Ion 77.00 (76.70 to 77.70): BM010971.D (-606) (-)

Ion 105.00 (104.70 to 105.70): BM010971.D (-606) (-)

Ion 106.00 (105.70 to 106.70): BM010971.D (-606) (-)



#10

Phenol

Concen: 40.328 ng

RT: 6.62 min Scan# 634

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion: 94 Resp: 880484

Ion Ratio Lower Upper

94 100

65 35.8 18.8 58.8



66 57.8 39.9 79.9




Abundance Ion 94.00 (93.70 to 94.70): BM010971.D (-617) (-)

Ion 65.00 (64.70 to 65.70): BM010971.D (-617) (-)

Ion 66.00 (65.70 to 66.70): BM010971.D (-617) (-)





#11

bis(2-Chloroethyl)ether

Concen: 40.709 ng

RT: 6.84 min Scan# 672

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617





Tot Ion: 93 Resp: 692527

Ion Ratio Lower Upper

93 100

63 76.1 49.5 89.5

95 31.6 13.5 53.5



#12

1,3-Dichlorobenzene

Concen: 40.367 ng

RT: 7.28 min Scan# 747

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04




Tgt Ion: 146 Resp: 706620


Ion Ratio Lower Upper

146 100

148 65.0 50.9 76.3

75 51.5 21.4 32.2#









#13
 1,4-Dichlorobenzene
 Concen: 40.161 ng
 RT: 7.43 min Scan# 772
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04




Instrument :
 BNA_M
 ClientSampleId :
 ICVBM072617


Tot Ion	Ratio	Lower	Upper
146	100		
148	64.8	51.9	77.9
111	50.4	29.2	43.8



#14
 1,2-Dichlorobenzene
 Concen: 40.722 ng
 RT: 7.74 min Scan# 825
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04

Tgt Ion	Ratio	Lower	Upper
146	100		
148	63.7	52.3	78.5
111	49.6	30.6	45.8





#15

Benzyl Alcohol

Concen: 40.043 ng

RT: 7.64 min Scan# 808

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion: 79 Resp: 772155

Ion Ratio Lower Upper

79 100

108 56.8 65.0 97.4#




77 67.7 53.0 79.6



Abundance Ion 79.00 (78.70 to 79.70): BM010971.D (-791) (-)

Ion 108.00 (107.70 to 108.70): BM010971.D (-800) (-)

Ion 77.00 (76.70 to 77.70): BM010971.D (-770) (-)



#16

2,2'-oxybis(1-chloropropane)

Concen: 40.265 ng

RT: 7.93 min Scan# 858

Delta R.T. 0.01 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion: 45 Resp: 616377

Ion Ratio Lower Upper

45 100

77 26.4 0.0 35.2



79 20.1 0.0 32.0




Abundance Ion 45.00 (44.70 to 45.70): BM010971.D (-840) (-)

Ion 77.00 (76.70 to 77.70): BM010971.D (-770) (-)

Ion 79.00 (78.70 to 79.70): BM010971.D (-791) (-)





#17

2-Methylphenol

Concen: 41.131 ng

RT: 7.85 min Scan# 843

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617

Tot Ion:107 Resp: 573931


Ion Ratio Lower Upper

107 100

108 106.4 89.8 134.8

77 69.6 32.6 48.8#

79 67.1 32.8 49.2#






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BM

Ion 79.00 (78.70 to 79.70): BM



#18

Hexachloroethane

Concen: 39.409 ng

RT: 8.45 min Scan# 946

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion:117 Resp: 308371

Ion Ratio Lower Upper

117 100

119 103.9 79.2 118.8



201 134.9 69.8 104.6#




Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 39.892 ng

RT: 8.20 min Scan# 904





Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :
BNA_M
ClientSampleId :
ICVBM072617

Tot Ion:	70	Resp:	681409
Ion	Ratio	Lower	Upper
70	100		
42	50.0	44.5	66.7
101	8.6	7.4	11.2
130	14.8	15.3	22.9#



#20

3+4-Methylphenols

Concen: 40.421 ng




RT: 8.17 min Scan# 898


Delta R.T. -0.01 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Tgt Ion:	107	Resp:	784040
Ion	Ratio	Lower	Upper
107	100		
108	89.2	73.2	113.2
77	51.2	10.7	50.7#
79	41.5	9.6	49.6





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.15 min Scan# 1235

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617

Tot Ion:136 Resp: 957666


Ion Ratio Lower Upper

136 100

137 11.7 8.7 13.1

54 7.8 4.6 6.8#

68 5.9 3.7 5.5#




Abundance Ion 136.00 (135.70 to 136.70): E



Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BM

Ion 68.00 (67.70 to 68.70): BM



Time-->



#22

Acetophenone

Concen: 41.199 ng

RT: 8.21 min Scan# 905

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Tgt Ion:105 Resp: 1182690


Ion Ratio Lower Upper

105 100

71 7.5 0.6 1.0#

51 29.7 21.6 32.4

120 18.1 16.1 24.1




Abundance Ion 105.00 (104.70 to 105.70): E


Ion 71.00 (70.70 to 71.70): BM


Ion 51.00 (50.70 to 51.70): BM

Ion 120.00 (119.70 to 120.70): E



Time-->









#23
 Nitrobenzene-d5
 Concen: 83.640 ng
 RT: 8.54 min Scan# 961
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04




Instrument :
 BNA_M
 ClientSampleId :
 ICVBM072617


Tot Ion: 82 Resp: 2134184
 Ion Ratio Lower Upper
 82 100
 128 28.0 32.8 49.2#
 54 48.0 40.6 60.8



#24
 Nitrobenzene
 Concen: 40.963 ng
 RT: 8.58 min Scan# 968
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04

Tgt Ion: 77 Resp: 1083106
 Ion Ratio Lower Upper
 77 100
 123 32.8 32.6 49.0
 65 14.2 10.9 16.3





#25

Isophorone

Concen: 40.842 ng

RT: 9.10 min Scan# 1057

Delta R.T. -0.01 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617





Tot Ion: 82 Resp: 1737572

Ion Ratio Lower Upper

82 100

95 9.0 5.8 8.6#

138 12.4 16.3 24.5#



#26

2-Nitrophenol

Concen: 43.299 ng

RT: 9.28 min Scan# 1087

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04




Tgt Ion: 139 Resp: 364283


Ion Ratio Lower Upper

139 100

109 63.1 20.1 30.1#

65 85.1 31.5 47.3#









#27
 2,4-Dimethylphenol
 Concen: 40.417 ng
 RT: 9.36 min Scan# 1100
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04




Instrument :
 BNA_M
 ClientSampleId :
 ICVBM072617


Tot Ion	Ratio	Lower	Upper
122	100		
107	153.9	84.7	127.1#
121	59.5	46.6	69.8



#28
 bis(2-Chloroethoxy)methane
 Concen: 41.015 ng
 RT: 9.59 min Scan# 1140
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04

Tgt Ion	Ratio	Lower	Upper
93	100		
95	30.8	25.5	38.3
123	8.7	8.6	13.0









#29
2,4-Dichlorophenol
Concen: 40.854 ng
RT: 9.81 min Scan# 1177
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion	Ratio	Lower	Upper
162	100		
164	64.2	42.8	82.8
98	40.7	14.5	54.5



#30
1,2,4-Trichlorobenzene
Concen: 41.389 ng
RT: 10.02 min Scan# 1213
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion	Ratio	Lower	Upper
180	100		
182	97.1	77.6	116.4
145	27.7	23.4	35.2









#31
Naphthalene
Concen: 40.993 ng
RT: 10.20 min Scan# 1244
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion:128 Resp: 1974917
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.2 10.4 15.6



#32
Benzoic acid
Concen: 40.933 ng
RT: 9.51 min Scan# 1126
Delta R.T. -0.01 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion:122 Resp: 487290
Ion Ratio Lower Upper
122 100
105 142.0 99.9 139.9#
77 135.8 73.7 113.7#









#33
4-Chloroaniline
Concen: 40.818 ng
RT: 10.32 min Scan# 1264
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion	Ratio	Lower	Upper
127	100		
129	31.4	25.3	37.9
65	47.3	17.6	26.4
92	28.8	13.1	19.7



#34
Hexachlorobutadiene
Concen: 40.777 ng
RT: 10.49 min Scan# 1293
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion	Ratio	Lower	Upper
225	100		
223	58.5	47.9	71.9
227	64.3	50.1	75.1





#35

Caprolactam

Concen: 39.748 ng

RT: 11.12 min Scan# 1399

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion:113 Resp: 187597

Ion Ratio Lower Upper

113 100

55 157.7 145.9 185.9




56 129.0 101.0 141.0



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BM0

Ion 56.00 (55.70 to 56.70): BM0



#36

4-Chloro-3-methylphenol

Concen: 40.861 ng

RT: 11.46 min Scan# 1458

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion:107 Resp: 878077

Ion Ratio Lower Upper

107 100

144 22.4 23.3 34.9#



142 65.2 72.6 109.0#




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 40.937 ng
RT: 11.83 min Scan# 1520
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion:142 Resp: 1448812
Ion Ratio Lower Upper
142 100
141 86.4 71.9 107.9
115 52.9 25.4 38.0#



#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.05 min Scan# 1898
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion:164 Resp: 651413
Ion Ratio Lower Upper
164 100
162 100.2 82.4 123.6
160 47.8 35.8 53.6





#39

1,2,4,5-Tetrachlorobenzene

Concen: 41.162 ng

RT: 12.21 min Scan# 1585

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617

Tot Ion:216 Resp: 1151997


Ion	Ratio	Lower	Upper
216	100		
214	77.6	0.0	0.0#
179	20.3	0.0	0.0#
108	20.2	0.0	0.0#

216 100

214 77.6 0.0 0.0#

179 20.3 0.0 0.0#

108 20.2 0.0 0.0#






Abundance

Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 41.650 ng

RT: 12.19 min Scan# 1582

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion:237 Resp: 679221

Ion	Ratio	Lower	Upper
237	100		
235	64.1	42.0	82.0
272	12.9	0.0	33.4

237 100

235 64.1 42.0 82.0

272 12.9 0.0 33.4






Abundance

Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 80.531 ng
 RT: 15.56 min Scan# 2154
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM072617

Tot Ion	Ratio	Lower	Upper
330	100		
332	97.1	0.0	0.0#
141	40.9	0.0	0.0#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#42
 2,4,6-Trichlorophenol
 Concen: 41.485 ng
 RT: 12.46 min Scan# 1628
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04

Tgt Ion	Ratio	Lower	Upper
196	100		
198	96.4	76.3	114.5
200	30.2	25.6	38.4



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 39.509 ng
 RT: 12.53 min Scan# 1640
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04




Instrument :
 BNA_M
 ClientSampleId :
 ICVBM072617


Tot Ion	Ratio	Lower	Upper
196	100		
198	96.1	79.4	119.0
97	61.1	26.8	40.2#
132	23.1	18.6	28.0



#44
 2-Fluorobiphenyl
 Concen: 80.589 ng
 RT: 12.66 min Scan# 1662
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.8	28.5	42.7
170	23.8	18.8	28.2









#45
 1,1'-Biphenyl
 Concen: 40.360 ng
 RT: 12.87 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04




Instrument :
 BNA_M
 ClientSampleId :
 ICVBM072617


Tot Ion:154 Resp: 2273324
 Ion Ratio Lower Upper
 154 100
 153 40.4 20.7 60.7
 76 13.6 0.0 30.9



#46
 2-Chloronaphthalene
 Concen: 40.110 ng
 RT: 12.91 min Scan# 1704
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04

Tgt Ion:162 Resp: 1664604
 Ion Ratio Lower Upper
 162 100
 127 40.2 26.9 40.3
 164 31.2 26.8 40.2









#47
2-Nitroaniline
Concen: 41.279 ng
RT: 13.13 min Scan# 1741
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion	Ratio	Lower	Upper
65	100		
92	61.9	59.1	88.7
138	71.8	93.9	140.9#



#48
Acenaphthylene
Concen: 40.526 ng
RT: 13.77 min Scan# 1850
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.9	23.9
153	13.1	10.6	16.0





#49

Dimethylphthalate

Concen: 40.153 ng

RT: 13.53 min Scan# 1809

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion:163 Resp: 2237557

Ion Ratio Lower Upper

163 100

194 4.5 2.7 4.1#




164 9.7 8.2 12.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 41.776 ng

RT: 13.64 min Scan# 1828

Delta R.T. -0.01 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion:165 Resp: 470763

Ion Ratio Lower Upper

165 100

63 81.8 44.1 66.1#



89 85.1 44.3 66.5#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BM0

Ion 89.00 (88.70 to 89.70): BM0





#51

Acenaphthene

Concen: 40.229 ng

RT: 14.12 min Scan# 1909

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion:154 Resp: 1542731

Ion Ratio Lower Upper

154 100

153 114.7 90.0 135.0




152 53.6 42.6 63.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 42.126 ng

RT: 13.97 min Scan# 1884

Delta R.T. -0.01 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion:138 Resp: 411943

Ion Ratio Lower Upper

138 100

108 14.9 7.3 10.9#



92 154.3 76.6 114.8#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BM









#53
2,4-Dinitrophenol
Concen: 40.474 ng
RT: 14.18 min Scan# 1920
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion:184 Resp: 324141
Ion Ratio Lower Upper
184 100
63 81.4 69.2 103.8
154 83.0 53.3 79.9#



#54
Dibenzofuran
Concen: 39.760 ng
RT: 14.46 min Scan# 1967
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion:168 Resp: 2470566
Ion Ratio Lower Upper
168 100
139 41.9 27.3 40.9#
169 13.3 10.6 16.0









#55
4-Nitrophenol
Concen: 40.901 ng
RT: 14.29 min Scan# 1939
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion:139 Resp: 351400
Ion Ratio Lower Upper
139 100
109 151.9 37.7 77.7#
65 145.9 49.9 89.9#



#56
2,4-Dinitrotoluene
Concen: 41.012 ng
RT: 14.44 min Scan# 1964
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion:165 Resp: 648803
Ion Ratio Lower Upper
165 100
63 63.9 52.4 78.6
89 96.1 72.4 108.6
182 3.4 2.0 3.0#





#57

Fluorene

Concen: 39.955 ng

RT: 15.11 min Scan# 2078

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion:166 Resp: 2161592

Ion Ratio Lower Upper

166 100

165 99.3 79.0 118.4




167 13.4 10.3 15.5



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 39.778 ng

RT: 14.69 min Scan# 2007

Delta R.T. 0.01 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Tgt Ion:232 Resp: 650491


Ion Ratio Lower Upper

232 100

131 42.6 0.0 0.0#

130 2.8 0.0 0.0#

166 28.5 0.0 0.0#






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E









#59
Diethylphthalate
Concen: 40.127 ng
RT: 14.91 min Scan# 2044
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion:149 Resp: 2283645
Ion Ratio Lower Upper
149 100
177 22.6 18.3 27.5
150 12.7 9.8 14.8



#60
4-Chlorophenyl-phenylether
Concen: 39.580 ng
RT: 15.12 min Scan# 2079
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion:204 Resp: 1292836
Ion Ratio Lower Upper
204 100
206 32.8 26.6 39.8
141 57.0 45.2 67.8









#61
4-Nitroaniline
Concen: 41.506 ng
RT: 15.14 min Scan# 2084
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion:138 Resp: 431114
Ion Ratio Lower Upper
138 100
92 80.3 16.7 56.7#
108 170.0 37.6 77.6#



#62
Azobenzene
Concen: 39.853 ng
RT: 15.41 min Scan# 2129
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion: 77 Resp: 2427367
Ion Ratio Lower Upper
77 100
182 23.0 6.5 46.5
105 16.4 3.8 43.8
51 31.3 5.5 45.5









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.81 min Scan# 2367
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion:188 Resp: 1604676
Ion Ratio Lower Upper
188 100
94 7.4 8.7 13.1#
80 9.2 9.3 13.9#



#64
4,6-Dinitro-2-methylphenol
Concen: 42.320 ng
RT: 15.20 min Scan# 2094
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion:198 Resp: 459736
Ion Ratio Lower Upper
198 100
51 45.6 28.8 68.8
105 42.1 30.5 70.5









#65
n-Nitrosodiphenylamine
Concen: 40.691 ng
RT: 15.33 min Scan# 2116
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617

Tot Ion:169 Resp: 1868684
Ion Ratio Lower Upper
169 100
168 67.2 53.9 80.9
167 36.7 28.9 43.3



#66
4-Bromophenyl-phenylether
Concen: 41.459 ng
RT: 16.01 min Scan# 2231
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion:248 Resp: 855253
Ion Ratio Lower Upper
248 100
250 94.2 77.4 116.0
141 77.5 45.2 67.8#





#67

Hexachlorobenzene

Concen: 40.732 ng

RT: 16.12 min Scan# 2249

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion:284 Resp: 949496

Ion Ratio Lower Upper

284 100

142 34.0 20.4 30.6#




249 33.7 23.8 35.6



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 42.541 ng

RT: 16.30 min Scan# 2280

Delta R.T. -0.01 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion:200 Resp: 782903

Ion Ratio Lower Upper

200 100

173 23.4 4.8 44.8



215 51.5 26.9 66.9




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 42.293 ng
 RT: 16.46 min Scan# 2308
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04




Instrument :
 BNA_M
 ClientSampleId :
 ICVBM072617


Tot Ion	Ratio	Lower	Upper
266	100		
268	61.9	52.0	78.0
264	60.8	49.0	73.4



#70
 Phenanthrene
 Concen: 41.093 ng
 RT: 16.85 min Scan# 2374
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.2	22.8
179	14.6	12.1	18.1





#71

Anthracene

Concen: 41.448 ng

RT: 16.94 min Scan# 2389

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617





Tot Ion:178 Resp: 3473255

Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.6	22.0
179	15.7	12.6	19.0

178 100

176 18.6 14.6 22.0

179 15.7 12.6 19.0



#72

Carbazole

Concen: 41.341 ng

RT: 17.22 min Scan# 2437

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04




Tgt Ion:167 Resp: 3029638


Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.2	25.8
139	12.8	10.8	16.2

167 100

166 21.7 17.2 25.8

139 12.8 10.8 16.2





#73

Di-n-butylphthalate

Concen: 42.449 ng

RT: 17.80 min Scan# 2536

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :


BNA_M

ClientSampleId :

ICVBM072617

Tot Ion:149 Resp: 3788325




Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.5	11.3
104	7.6	3.7	5.5#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 40.993 ng

RT: 18.88 min Scan# 2719


Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Tgt Ion:202 Resp: 4268411



Ion	Ratio	Lower	Upper
202	100		
101	7.8	0.0	28.7
203	17.8	0.0	37.2




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.000 ng

RT: 21.03 min Scan# 3084

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :


BNA_M

ClientSampleId :

ICVBM072617

Tot Ion:240 Resp: 1774157



Ion	Ratio	Lower	Upper
240	100		
120	5.8	8.2	12.2#
236	26.2	20.0	30.0




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



Time--> 20.95 21.00 21.05 21.10



#76

Benzidine

Concen: 41.467 ng

RT: 19.08 min Scan# 2753


Delta R.T. -0.01 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Tgt Ion:184 Resp: 1759976



Ion	Ratio	Lower	Upper
184	100		
185	13.8	11.3	16.9
183	11.8	8.9	13.3




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



Time--> 19.00 19.10



#77

Pvrene

Concen: 41.704 ng

RT: 19.24 min Scan# 2781

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion:202 Resp: 4396328

Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.9	25.3
203	17.5	14.1	21.1

202 100

200 20.6 16.9 25.3




203 17.5 14.1 21.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 85.170 ng

RT: 19.47 min Scan# 2819

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion:244 Resp: 7628227

Ion	Ratio	Lower	Upper
244	100		
212	7.8	4.9	7.3#
122	6.9	6.2	9.4

244 100

212 7.8 4.9 7.3#



122 6.9 6.2 9.4




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#79
Butylbenzylphthalate
Concen: 43.073 ng
RT: 20.19 min Scan# 2941
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04




Instrument :
BNA_M
ClientSampleId :
ICVBM072617


Tot Ion:149 Resp: 1614664
Ion Ratio Lower Upper
149 100
91 88.6 50.1 75.1#
206 27.3 16.2 24.2#



#80
Benzo(a)anthracene
Concen: 41.385 ng
RT: 21.01 min Scan# 3081
Delta R.T. -0.00 min
Lab File: BM010971.D
Acq: 26 Jul 2017 15:04

Tgt Ion:228 Resp: 4192441
Ion Ratio Lower Upper
228 100
226 26.2 21.7 32.5
229 19.7 16.0 24.0









#81
 3,3'-Dichlorobenzidine
 Concen: 42.505 ng
 RT: 20.96 min Scan# 3072
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM072617

Tot Ion:252 Resp: 1689207




Ion	Ratio	Lower	Upper
252	100		
254	64.4	51.9	77.9
126	8.2	9.8	14.8#




#82
 Chrysene
 Concen: 40.486 ng
 RT: 21.06 min Scan# 3090
 Delta R.T. -0.00 min
 Lab File: BM010971.D
 Acq: 26 Jul 2017 15:04

Tgt Ion:228 Resp: 3936880

Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.1	36.1
229	19.3	15.5	23.3





#83

Bis(2-ethylhexyl)phthalate

Concen: 42.916 ng

RT: 20.97 min Scan# 3075

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :


BNA_M

ClientSampleId :

ICVBM072617

Tot Ion:149 Resp: 2366686

Ion	Ratio	Lower	Upper
149	100		
167	27.8	22.4	33.6
279	5.6	3.4	5.0






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#84

Di-n-octyl phthalate

Concen: 42.869 ng

RT: 21.82 min Scan# 3219


Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Tgt Ion:149 Resp: 3836996

Ion	Ratio	Lower	Upper
149	100		
167	1.7	1.2	1.8
43	9.1	7.3	10.9






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BM0





#85

Indeno(1,2,3-cd)pyrene

Concen: 40.989 ng

RT: 25.22 min Scan# 3797

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion:276 Resp: 4127979

Ion	Ratio	Lower	Upper
276	100		
138	11.7	0.0	0.0#
277	25.7	0.0	0.0#

276 100

138 11.7 0.0 0.0#

277 25.7 0.0 0.0#






Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 23.14 min Scan# 3443

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion:264 Resp: 1540092

Ion	Ratio	Lower	Upper
264	100		
260	22.7	18.6	27.8
265	21.5	17.3	25.9

264 100

260 22.7 18.6 27.8

265 21.5 17.3 25.9






Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E





#87

Benzo(b)fluoranthene

Concen: 40.538 ng

RT: 22.52 min Scan# 3337

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617


Tot Ion:252 Resp: 4007002

Ion Ratio Lower Upper

252 100

253 22.7 18.0 27.0




125 5.3 9.9 14.9#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 40.823 ng

RT: 22.56 min Scan# 3345

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04


Tgt Ion:252 Resp: 3867542

Ion Ratio Lower Upper

252 100

253 21.9 17.2 25.8



125 3.9 8.1 12.1#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 40.771 ng

RT: 23.05 min Scan# 3428

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :


BNA_M

ClientSampleId :

ICVBM072617

Tot Ion:252 Resp: 3646665

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	5.2	7.4	11.2#






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 40.487 ng

RT: 25.23 min Scan# 3799


Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Tgt Ion:278 Resp: 3356972

Ion	Ratio	Lower	Upper
278	100		
139	8.0	13.4	20.0#
279	23.7	19.0	28.4








Abundance

Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(a,h,i)perylene

Concen: 40.245 ng

RT: 25.85 min Scan# 3904

Delta R.T. -0.00 min

Lab File: BM010971.D

Acq: 26 Jul 2017 15:04

Instrument :

BNA_M

ClientSampleId :

ICVBM072617

Tot Ion:276 Resp: 3276675

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

276	100		
-----	-----	--	--

277	22.6	18.5	27.7
-----	------	------	------

138	10.8	19.0	28.6
-----	------	------	------

