


Data Path : Z:\HPCHEM1\BNA_M\Data\BM072715\
Data File : BM002056.D
Acq On : 27 Jul 2015 11:26
Operator : UM/HP
Sample : DFTPP81
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP81

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072715.M
Title : SVOA CALIBRATION
Last Update : Wed Jul 29 01:50:38 2015



AutoFind: Scans 2509, 2510, 2511; Background Corrected with Scan 2504

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.3	28570	PASS
68	69	0.00	2	1.8	637	PASS
69	198	0.00	100	33.9	35482	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	41.7	43618	PASS
197	198	0.00	2	0.6	650	PASS
198	198	100	100	100.0	104637	PASS
199	198	5	9	6.9	7199	PASS
275	198	10	60	28.5	29805	PASS
365	198	1	100	3.6	3755	PASS
441	443	0.01	100	72.6	13424	PASS
442	198	50	100	89.7	93869	PASS
443	442	15	24	19.7	18479	PASS

Scan 2504 (17.415 min): BM002056.D
DFTPP81

m/z	Abundance
44.00	885.0
207.10	489.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

m/z	Abundance
38.00	420.0
39.00	2709.0
40.00	328.0
44.00	862.0
49.10	335.0
50.10	7752.0
51.10	30184.0
52.00	1616.0
56.00	1159.0
57.00	2185.0
61.10	460.0
62.00	485.0
63.00	1406.0
65.00	646.0
68.00	596.0
69.00	36920.0
73.00	525.0
74.00	4091.0
75.00	6016.0
76.10	1945.0
77.10	40056.0
78.00	2852.0
79.00	2940.0
80.00	2342.0
81.00	3201.0
82.00	690.0
83.00	812.0
85.00	688.0
86.00	737.0
86.90	382.0
90.90	754.0
91.90	930.0
93.00	5029.0
93.90	306.0
96.00	310.0
98.00	3856.0
99.00	3227.0
101.00	1715.0
103.00	573.0
104.00	980.0
105.10	1122.0
106.00	310.0
107.00	13200.0
108.00	2236.0
109.10	417.0
110.00	23240.0
110.90	3204.0
111.90	492.0
116.00	752.0
117.00	10479.0
118.00	703.0
122.00	865.0
123.00	1313.0
124.00	583.0
125.00	512.0
127.00	42200.0
128.00	3594.0
129.00	17408.0
130.00	1499.0
134.00	551.0
135.00	1492.0
135.90	444.0
137.00	871.0
141.00	2143.0
142.00	835.0
143.00	529.0
146.00	503.0
146.90	1231.0
148.00	3059.0
148.90	770.0
151.00	350.0
152.90	830.0
153.90	826.0
155.00	1493.0
156.00	2140.0
158.90	305.0
160.00	759.0
161.00	1194.0
161.90	316.0
164.90	823.0
165.90	756.0
167.00	4823.0
167.90	2472.0
168.90	418.0
172.00	410.0
172.90	625.0
174.00	940.0
175.00	1968.0
175.80	493.0
177.00	955.0
178.90	3780.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

179.90	2478.0
181.00	931.0
181.90	328.0
185.10	1569.0
186.00	13383.0
187.00	4093.0
187.90	468.0
189.00	926.0
191.00	433.0
191.90	1024.0
192.90	1289.0
193.90	406.0
196.00	2830.0
197.10	573.0
197.90	95840.0
198.90	6611.0
199.90	643.0
201.40	450.0
203.10	757.0
204.00	3632.0
205.00	5898.0
206.00	26592.0
207.00	4106.0
208.00	1214.0
209.00	390.0
210.00	405.0
211.00	874.0
216.00	463.0
217.00	7090.0
218.00	993.0
221.00	4345.0
221.80	1065.0
223.00	1467.0
224.00	15269.0
225.00	3916.0
226.00	447.0
227.00	6539.0
227.90	755.0
228.90	1407.0
231.00	503.0
233.90	463.0
235.00	367.0
236.00	304.0
237.00	558.0
241.10	310.0
242.00	761.0
243.00	923.0
244.00	10689.0
245.00	1709.0
245.90	2384.0
247.00	488.0
248.90	441.0
254.10	423.0
255.00	57344.0
256.00	8322.0
257.10	911.0
258.00	3495.0
258.90	651.0
265.00	1375.0
273.00	1600.0
274.00	4115.0
275.00	25568.0
276.00	3727.0
277.00	1862.0
278.10	464.0
282.90	331.0
285.00	367.0
293.00	315.0
296.00	7163.0
297.00	955.0
303.00	681.0
313.80	361.0
314.90	848.0
316.00	349.0
323.00	2300.0
324.00	418.0
326.90	425.0
334.00	1399.0
345.90	484.0
351.90	483.0
352.90	472.0
354.00	661.0
365.00	2925.0
366.00	596.0
372.00	1215.0
373.10	348.0
402.00	334.0
403.00	588.0
421.00	461.0
422.00	507.0
423.00	3410.0
424.10	973.0
441.10	9143.0
442.10	63832.0
443.10	12766.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

444.00

1419.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

m/z	Abundance
38.00	501.0
39.10	2660.0
40.00	351.0
44.00	764.0
49.10	318.0
50.00	9494.0
51.00	35336.0
52.10	1871.0
56.00	1062.0
57.00	2429.0
61.00	566.0
61.90	472.0
63.00	1550.0
65.00	856.0
68.10	913.0
69.00	43952.0
73.00	594.0
74.00	5042.0
75.00	7355.0
76.00	2409.0
77.10	47608.0
78.00	3495.0
79.00	3629.0
80.00	2442.0
81.00	3425.0
82.00	895.0
83.10	772.0
85.00	779.0
86.00	1162.0
87.00	443.0
90.90	775.0
92.00	915.0
93.00	6087.0
94.00	467.0
98.00	5090.0
99.00	3731.0
100.00	388.0
101.00	1902.0
102.90	660.0
103.90	1413.0
105.00	1284.0
106.10	407.0
107.00	15575.0
108.00	2356.0
108.90	469.0
110.00	28480.0
111.00	3867.0
111.90	468.0
116.00	1044.0
117.00	13062.0
117.90	829.0
121.90	931.0
123.00	1927.0
123.90	836.0
124.90	725.0
126.10	431.0
127.00	54160.0
128.00	3909.0
129.00	22688.0
130.00	1594.0
131.00	353.0
133.90	700.0
135.00	1899.0
135.90	713.0
137.00	949.0
137.90	326.0
140.00	402.0
140.90	3271.0
142.00	1132.0
143.00	646.0
145.90	598.0
147.00	1529.0
147.90	3297.0
149.00	797.0
151.80	370.0
152.90	1005.0
154.00	727.0
155.00	1655.0
156.00	2267.0
157.10	474.0
157.90	593.0
159.00	408.0
160.00	1112.0
160.90	1328.0
161.90	427.0
164.90	1173.0
166.10	965.0
167.00	5824.0
168.00	3388.0
168.90	345.0
171.90	600.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

173.00	655.0
174.00	1431.0
175.00	2330.0
176.00	928.0
176.90	1318.0
177.90	380.0
179.00	4969.0
180.00	2977.0
181.00	1629.0
184.00	428.0
185.00	2543.0
186.00	18488.0
187.00	4944.0
187.90	501.0
189.00	1370.0
191.00	518.0
191.90	1560.0
193.00	1806.0
194.00	426.0
196.00	3880.0
197.00	884.0
197.90	127184.0
198.90	9039.0
199.90	526.0
201.20	672.0
203.00	1095.0
204.00	5014.0
205.00	8764.0
206.00	35400.0
207.00	4802.0
207.90	1276.0
208.90	492.0
209.90	513.0
211.00	1271.0
214.90	351.0
215.90	609.0
216.90	9120.0
218.00	1289.0
221.10	5130.0
221.70	1404.0
223.00	2254.0
224.00	20272.0
225.00	5154.0
226.00	459.0
227.00	9150.0
227.90	1175.0
229.00	1722.0
231.00	588.0
233.90	587.0
234.90	554.0
235.90	406.0
237.00	628.0
239.00	376.0
239.90	308.0
240.90	412.0
242.00	1171.0
243.00	912.0
244.00	14941.0
245.00	2288.0
246.00	3450.0
246.90	569.0
249.00	643.0
253.00	332.0
254.10	494.0
255.00	82056.0
256.00	11608.0
257.00	1111.0
258.00	4550.0
259.00	911.0
265.00	1996.0
273.00	2232.0
274.00	5645.0
275.00	35688.0
276.00	4871.0
276.90	3214.0
278.00	724.0
283.10	366.0
285.10	626.0
293.00	854.0
296.00	10229.0
297.00	1475.0
302.90	1093.0
304.00	327.0
314.00	360.0
314.90	899.0
316.00	764.0
323.00	2689.0
324.00	595.0
327.00	565.0
328.00	401.0
333.20	336.0
334.00	1922.0
335.00	561.0
341.10	430.0
346.00	686.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

352.00	846.0
353.10	684.0
354.00	1157.0
365.00	4337.0
365.90	523.0
372.00	1699.0
373.00	425.0
383.00	403.0
401.90	482.0
403.00	1114.0
404.00	326.0
421.00	864.0
422.00	829.0
423.00	6150.0
424.10	1446.0
441.00	16384.0
442.10	109688.0
443.10	21864.0
444.10	2133.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

m/z	Abundance
39.10	1829.0
39.90	337.0
43.90	648.0
50.10	5496.0
51.10	20192.0
52.10	855.0
56.00	736.0
57.00	1606.0
60.90	359.0
62.00	425.0
63.00	1030.0
65.10	482.0
68.00	403.0
69.00	25576.0
74.00	2773.0
75.00	4236.0
76.00	1662.0
77.00	28088.0
78.00	1910.0
79.00	2313.0
80.00	1753.0
81.00	2398.0
81.90	510.0
83.00	450.0
84.90	458.0
85.90	492.0
91.00	567.0
92.00	615.0
93.00	3599.0
98.00	2850.0
99.00	2140.0
101.00	1223.0
102.90	447.0
104.00	931.0
104.90	788.0
106.00	346.0
107.00	9875.0
108.00	1561.0
109.20	303.0
110.00	17040.0
111.00	2670.0
116.00	543.0
117.00	8624.0
117.90	522.0
122.00	795.0
123.00	1077.0
123.90	535.0
125.00	469.0
127.00	34496.0
128.00	2879.0
129.00	13897.0
130.00	1371.0
133.90	351.0
134.90	1319.0
135.90	362.0
137.10	740.0
141.00	2136.0
142.00	706.0
142.90	531.0
145.90	354.0
146.90	1192.0
148.00	2107.0
148.90	479.0
151.00	326.0
153.00	535.0
154.00	427.0
155.00	1165.0
156.00	1823.0
156.90	473.0
158.00	343.0
159.00	307.0
160.00	830.0
161.00	962.0
164.90	856.0
165.90	639.0
167.00	4090.0
168.00	2248.0
169.10	336.0
171.90	449.0
172.90	505.0
173.90	682.0
175.00	1580.0
175.90	522.0
176.90	741.0
177.80	312.0
178.90	3446.0
180.00	2357.0
181.00	1032.0
185.00	1444.0
186.00	12883.0
187.00	3727.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

188.00	391.0
189.00	933.0
191.00	447.0
191.90	1223.0
193.00	982.0
196.00	2590.0
196.70	527.0
197.00	493.0
198.00	90888.0
199.00	5948.0
199.90	501.0
201.20	431.0
202.90	811.0
204.00	3607.0
205.10	6106.0
206.00	25984.0
207.00	3769.0
208.00	933.0
209.00	375.0
210.00	421.0
210.80	976.0
214.90	321.0
215.90	493.0
217.00	7681.0
217.90	909.0
221.00	4040.0
221.80	890.0
223.00	1395.0
224.00	14775.0
225.00	3881.0
226.00	365.0
227.00	6888.0
227.90	871.0
228.90	1398.0
231.00	634.0
233.80	336.0
235.10	374.0
236.00	354.0
236.90	701.0
241.00	403.0
242.00	710.0
243.00	778.0
244.00	10867.0
245.00	1512.0
246.00	2194.0
247.10	457.0
248.90	577.0
254.10	523.0
255.00	61984.0
256.00	8799.0
257.00	607.0
258.00	4068.0
258.90	448.0
264.90	1550.0
273.00	1650.0
274.00	4807.0
275.00	28160.0
276.00	3790.0
277.00	2372.0
278.00	409.0
285.00	434.0
292.90	443.0
296.00	8764.0
296.90	1231.0
302.90	970.0
314.00	404.0
314.90	1078.0
316.00	659.0
323.10	2451.0
324.00	490.0
326.90	489.0
334.10	1783.0
335.00	450.0
341.10	385.0
346.00	528.0
352.00	832.0
353.00	562.0
354.00	928.0
365.00	4005.0
366.00	519.0
372.10	1547.0
382.90	510.0
401.90	550.0
403.00	694.0
404.10	382.0
421.00	697.0
422.00	723.0
423.00	5322.0
424.10	1312.0
441.10	14747.0
442.00	108088.0
443.10	20808.0
444.00	1617.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81