



Data Path : Z:\HPCHEM1\BNA M\DATA\BM072717\  
Data File : BM011000.D  
Acq On : 27 Jul 2017 08:49  
Operator : SJ/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 28 04:54:39 2017  
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jul 26 17:29:06 2017  
Response via : Initial Calibration



TIC: BM011000.D

(69) Pentachlorophenol (C)

16.457min (-0.006) 0.00ng

response 368194

| Ion    | Exp%  | Act%  |
|--------|-------|-------|
| 265.70 | 100   | 100   |
| 268.00 | 65.00 | 62.35 |
| 264.00 | 61.20 | 64.58 |
| 0.00   | 0.00  | 0.00  |