

Data Path : Z:\HPCHEM1\BNA M\DATA\BM072717\  
 Data File : BM011021.D  
 Acq On : 27 Jul 2017 21:24  
 Operator : SJ/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

Manual Integrations  
 APPROVED

Sohil  
 7/28/2017 4:45:03 PM

Quant Time: Jul 28 03:51:32 2017  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jul 26 17:29:06 2017  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.39  | 152  | 187551   | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8        | 10.15 | 136  | 848611   | 20.00 | ng    | 0.00     |
| 38) Acenaphthene-d10      | 14.05 | 164  | 698788   | 20.00 | ng    | 0.00     |
| 63) Phenanthrene-d10      | 16.80 | 188  | 1994945  | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12          | 21.03 | 240  | 2192947  | 20.00 | ng    | 0.00     |
| 86) Perylene-d12          | 23.13 | 264  | 1658998  | 20.00 | ng    | 0.00     |

## System Monitoring Compounds

|                          |       |     |         |       |    |      |
|--------------------------|-------|-----|---------|-------|----|------|
| 5) 2-Fluorophenol        | 5.02  | 112 | 838815  | 75.60 | ng | 0.00 |
| 7) Phenol-d6             | 6.59  | 99  | 1259719 | 79.91 | ng | 0.00 |
| 23) Nitrobenzene-d5      | 8.54  | 82  | 1851040 | 81.87 | ng | 0.00 |
| 41) 2,4,6-Tribromophenol | 15.55 | 330 | 985175  | 87.96 | ng | 0.00 |
| 44) 2-Fluorobiphenyl     | 12.66 | 172 | 4169198 | 74.83 | ng | 0.00 |
| 78) Terphenyl-d14        | 19.47 | 244 | 9601209 | 86.73 | ng | 0.00 |

## Target Compounds

|                                |       |     |         |        |    | Qvalue |
|--------------------------------|-------|-----|---------|--------|----|--------|
| 2) 1,4-Dioxane                 | 3.03  | 88  | 191156  | 38.373 | ng | # 100  |
| 3) Pyridine                    | 3.40  | 79  | 529887  | 38.943 | ng | # 88   |
| 4) n-Nitrosodimethylamine      | 3.33  | 42  | 298452  | 43.041 | ng | # 67   |
| 6) Aniline                     | 6.74  | 93  | 798999  | 40.216 | ng | # 91   |
| 8) 2-Chlorophenol              | 6.96  | 128 | 454308  | 40.332 | ng | # 84   |
| 9) Benzaldehyde                | 6.55  | 77  | 408286  | 40.030 | ng | # 78   |
| 10) Phenol                     | 6.61  | 94  | 680485  | 39.744 | ng | # 97   |
| 11) bis(2-Chloroethyl)ether    | 6.83  | 93  | 522798  | 39.188 | ng | # 91   |
| 12) 1,3-Dichlorobenzene        | 7.27  | 146 | 545066  | 39.705 | ng | # 77   |
| 13) 1,4-Dichlorobenzene        | 7.42  | 146 | 558916  | 39.720 | ng | # 90   |
| 14) 1,2-Dichlorobenzene        | 7.73  | 146 | 525753  | 39.078 | ng | # 92   |
| 15) Benzyl Alcohol             | 7.63  | 79  | 623999  | 41.264 | ng | # 84   |
| 16) 2,2'-oxybis(1-Chloropropan | 7.93  | 45  | 502516  | 41.859 | ng | # 79   |
| 17) 2-Methylphenol             | 7.84  | 107 | 453769  | 41.467 | ng | # 75   |
| 18) Hexachloroethane           | 8.45  | 117 | 247047  | 40.259 | ng | # 79   |
| 19) n-Nitroso-di-n-propylamine | 8.20  | 70  | 600922  | 44.859 | ng | # 93   |
| 20) 3+4-Methylphenols          | 8.17  | 107 | 657610  | 43.232 | ng | # 84   |
| 22) Acetophenone               | 8.20  | 105 | 958394  | 37.676 | ng | # 90   |
| 24) Nitrobenzene               | 8.58  | 77  | 958943  | 40.928 | ng | # 85   |
| 25) Isophorone                 | 9.10  | 82  | 1566504 | 41.553 | ng | # 87   |
| 26) 2-Nitrophenol              | 9.27  | 139 | 294499  | 39.503 | ng | # 12   |
| 27) 2,4-Dimethylphenol         | 9.35  | 122 | 532246  | 40.556 | ng | # 72   |
| 28) bis(2-Chloroethoxy)methane | 9.59  | 93  | 831004  | 39.522 | ng | # 97   |
| 29) 2,4-Dichlorophenol         | 9.81  | 162 | 600138  | 40.695 | ng | # 96   |
| 30) 1,2,4-Trichlorobenzene     | 10.02 | 180 | 723732  | 39.691 | ng | # 96   |
| 31) Naphthalene                | 10.20 | 128 | 1694350 | 39.689 | ng | # 99   |
| 32) Benzoic acid               | 9.52  | 122 | 453300  | 42.971 | ng | # 62   |
| 33) 4-Chloroaniline            | 10.32 | 127 | 714605  | 41.961 | ng | # 71   |
| 34) Hexachlorobutadiene        | 10.49 | 225 | 580534  | 40.422 | ng | # 95   |
| 35) Caprolactam                | 11.12 | 113 | 199342m | 47.665 | ng | #      |
| 36) 4-Chloro-3-methylphenol    | 11.46 | 107 | 851212  | 44.701 | ng | # 74   |
| 37) 2-Methylnaphthalene        | 11.82 | 142 | 1340700 | 42.750 | ng | # 87   |
| 39) 1,2,4,5-Tetrachlorobenzene | 12.20 | 216 | 1112467 | 37.055 | ng | # 100  |
| 40) Hexachlorocyclopentadiene  | 12.19 | 237 | 645781  | 36.915 | ng | # 97   |

Data Path : Z:\HPCHEM1\BNA M\DATA\BM072717\  
 Data File : BM011021.D  
 Acq On : 27 Jul 2017 21:24  
 Operator : SJ/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

Manual Integrations  
 APPROVED


Sohil  
 7/28/2017 4:45:03 PM

Quant Time: Jul 28 03:51:32 2017  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jul 26 17:29:06 2017  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|--------------------------------|-------|------|----------|--------|-------|----------|
| 42) 2,4,6-Trichlorophenol      | 12.46 | 196  | 703097   | 38.692 | nq    | 96       |
| 43) 2,4,5-Trichlorophenol      | 12.53 | 196  | 729815   | 38.995 | nq    | # 87     |
| 45) 1,1'-Biphenyl              | 12.87 | 154  | 2247214  | 37.191 | nq    | 98       |
| 46) 2-Chloronaphthalene        | 12.90 | 162  | 1664717  | 37.393 | nq    | # 93     |
| 47) 2-Nitroaniline             | 13.13 | 65   | 684241   | 43.445 | nq    | # 70     |
| 48) Acenaphthylene             | 13.76 | 152  | 2628082  | 39.556 | nq    | 99       |
| 49) Dimethylphthalate          | 13.53 | 163  | 2398809  | 40.128 | nq    | # 99     |
| 50) 2,6-Dinitrotoluene         | 13.64 | 165  | 500184   | 41.377 | nq    | # 58     |
| 51) Acenaphthene               | 14.12 | 154  | 1616814  | 39.302 | nq    | 100      |
| 52) 3-Nitroaniline             | 13.97 | 138  | 441389   | 42.077 | nq    | # 30     |
| 53) 2,4-Dinitrophenol          | 14.17 | 184  | 360955   | 42.015 | nq    | # 81     |
| 54) Dibenzofuran               | 14.45 | 168  | 2678593  | 40.185 | nq    | # 91     |
| 55) 4-Nitrophenol              | 14.29 | 139  | 390906   | 42.415 | nq    | 88       |
| 56) 2,4-Dinitrotoluene         | 14.43 | 165  | 746544   | 43.991 | nq    | # 95     |
| 57) Fluorene                   | 15.10 | 166  | 2421729  | 41.728 | nq    | 98       |
| 58) 2,3,4,6-Tetrachlorophenol  | 14.69 | 232  | 734099   | 41.847 | nq    | # 100    |
| 59) Diethylphthalate           | 14.91 | 149  | 2571585  | 42.123 | nq    | 99       |
| 60) 4-Chlorophenyl-phenylether | 15.11 | 204  | 1455494  | 41.539 | nq    | 99       |
| 61) 4-Nitroaniline             | 15.14 | 138  | 493118   | 44.257 | nq    | # 1      |
| 62) Azobenzene                 | 15.40 | 77   | 2853854  | 43.678 | nq    | 87       |
| 64) 4,6-Dinitro-2-methylphenol | 15.20 | 198  | 543287   | 40.227 | nq    | 94       |
| 65) n-Nitrosodiphenylamine     | 15.33 | 169  | 2170359  | 38.015 | nq    | 99       |
| 66) 4-Bromophenyl-phenylether  | 16.01 | 248  | 1002485  | 39.089 | nq    | # 91     |
| 67) Hexachlorobenzene          | 16.11 | 284  | 1102912  | 38.058 | nq    | # 87     |
| 68) Atrazine                   | 16.30 | 200  | 937179   | 40.962 | nq    | 96       |
| 69) Pentachlorophenol          | 16.46 | 266  | 755313   | 41.053 | nq    | 99       |
| 70) Phenanthrene               | 16.84 | 178  | 4141664  | 39.649 | nq    | 99       |
| 71) Anthracene                 | 16.94 | 178  | 4177822  | 40.102 | nq    | 99       |
| 72) Carbazole                  | 17.22 | 167  | 3694617  | 40.552 | nq    | 99       |
| 73) Di-n-butylphthalate        | 17.80 | 149  | 4614142  | 41.588 | nq    | # 96     |
| 74) Fluoranthene               | 18.88 | 202  | 5323184  | 41.121 | nq    | 98       |
| 76) Benzidine                  | 19.08 | 184  | 2079695  | 39.642 | nq    | 98       |
| 77) Pyrene                     | 19.24 | 202  | 5513679  | 42.315 | nq    | 99       |
| 79) Butylbenzylphthalate       | 20.18 | 149  | 1971935  | 42.558 | nq    | # 69     |
| 80) Benzo(a)anthracene         | 21.01 | 228  | 4958684  | 39.601 | nq    | 99       |
| 81) 3,3'-Dichlorobenzidine     | 20.96 | 252  | 1917637  | 39.038 | nq    | # 97     |
| 82) Chrysene                   | 21.06 | 228  | 4704605  | 39.141 | nq    | 98       |
| 83) Bis(2-ethylhexyl)phthalate | 20.97 | 149  | 2784883  | 40.855 | nq    | 99       |
| 84) Di-n-octyl phthalate       | 21.82 | 149  | 4351982  | 39.337 | nq    | 99       |
| 85) Indeno(1,2,3-cd)pyrene     | 25.21 | 276  | 4013778  | 32.244 | nq    | # 96     |
| 87) Benzo(b)fluoranthene       | 22.52 | 252  | 4329845  | 40.665 | nq    | # 92     |
| 88) Benzo(k)fluoranthene       | 22.56 | 252  | 4289797  | 42.035 | nq    | # 95     |
| 89) Benzo(a)pyrene             | 23.04 | 252  | 3853087  | 39.992 | nq    | # 96     |
| 90) Dibenzo(a,h)anthracene     | 25.23 | 278  | 3306482  | 37.019 | nq    | # 91     |
| 91) Benzo(g,h,i)perylene       | 25.84 | 276  | 3192980  | 36.406 | nq    | # 84     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1





1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 7.39 min Scan# 782  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 152     | 100   |       |       |
| 150     | 147.8 | 119.5 | 179.3 |
| 115     | 67.4  | 43.0  | 64.4# |

Manual Integrations  
APPROVED




Sohil  
7/28/2017 4:45:03 PM




#2

1,4-Dioxane  
Concen: 38.373 ng  
RT: 3.03 min Scan# 42  
Delta R.T. -0.00 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 88      | 100   |       |       |
| 58      | 58.6  | 0.0   | 0.0#  |
| 43      | 29.1  | 0.0   | 0.0#  |





#3

Pvridine

Concen: 38.943 ng

RT: 3.40 min Scan# 105

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M





ClientSampleId :

SSTDCCC040

|           |       |       |        |
|-----------|-------|-------|--------|
| Tot Ion:  | 79    | Resp: | 529887 |
| Ion Ratio | Lower | Upper |        |
| 79        | 100   |       |        |
| 52        | 52.8  | 51.1  | 76.7   |
| 51        | 36.2  | 26.1  | 39.1   |

Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM



#4

n-Nitrosodimethylamine

Concen: 43.041 ng




RT: 3.33 min Scan# 92


Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 42    | Resp: | 298452 |
| Ion Ratio | Lower | Upper |        |
| 42        | 100   |       |        |
| 74        | 105.3 | 119.3 | 178.9# |
| 44        | 4.8   | 5.4   | 8.2#   |





#5

2-Fluorophenol

Concen: 75.604 ng

RT: 5.02 min Scan# 380

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M





ClientSampleId :

SSTDCCC040

|           |       |       |        |
|-----------|-------|-------|--------|
| Tot Ion:  | 112   | Resp: | 838815 |
| Ion Ratio | Lower | Upper |        |
| 112       | 100   |       |        |
| 64        | 71.4  | 38.6  | 57.8#  |
| 63        | 47.2  | 19.3  | 28.9#  |

Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM



#6

Aniline

Concen: 40.216 ng




RT: 6.74 min Scan# 672


Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 93    | Resp: | 798999 |
| Ion Ratio | Lower | Upper |        |
| 93        | 100   |       |        |
| 66        | 44.4  | 30.8  | 46.2   |
| 65        | 23.7  | 16.0  | 24.0   |





#7

Phenol-d6

Concen: 79.913 ng

RT: 6.59 min Scan# 646

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampled :

SSTDCCC040

Tot Ion: 99 Resp: 1259719

Ion Ratio Lower Upper

99 100

42 26.4 11.0 16.4#





71 62.1 23.4 35.2#

Manual Integrations

APPROVED

Sohil

7/28/2017 4:45:03 PM



#8

2-Chlorophenol

Concen: 40.332 ng

RT: 6.96 min Scan# 710

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24




Tgt Ion: 128 Resp: 454308

Ion Ratio Lower Upper


128 100

130 29.7 12.0 52.0

64 61.1 24.9 64.9







#9

Benzaldehyde

Concen: 40.030 ng

RT: 6.55 min Scan# 639

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040

Tot Ion: 77 Resp: 408286

Ion Ratio Lower Upper

77 100





105 76.8 78.8 118.8#

106 71.7 73.7 113.7#

Manual Integrations  
APPROVED

Sohil

7/28/2017 4:45:03 PM



#10

Phenol

Concen: 39.744 ng

RT: 6.61 min Scan# 650

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24




Tgt Ion: 94 Resp: 680485

Ion Ratio Lower Upper


94 100

65 39.0 18.8 58.8

66 63.9 39.9 79.9







#11

bis(2-Chloroethyl)ether

Concen: 39.188 ng

RT: 6.83 min Scan# 688

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M





ClientSampleId :

SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 93      | 100   |       |       |
| 63      | 78.5  | 49.5  | 89.5  |
| 95      | 31.6  | 13.5  | 53.5  |

Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM



#12

1,3-Dichlorobenzene

Concen: 39.705 ng




RT: 7.27 min Scan# 763


Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 60.1  | 50.9  | 76.3  |
| 75      | 61.2  | 21.4  | 32.2# |









#13  
 1,4-Dichlorobenzene  
 Concen: 39.720 ng  
 RT: 7.42 min Scan# 788  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Instrument :  
 BNA\_M  
 ClientSampled :  
 SSTDCCC040

| Tot Ion | Ratio | Resp Lower | Upper |
|---------|-------|------------|-------|
| 146     | 100   |            |       |
| 148     | 62.9  | 51.9       | 77.9  |
| 111     | 50.0  | 29.2       | 43.8# |




Manual Integrations  
 APPROVED


Sohil  
 7/28/2017 4:45:03 PM



#14  
 1,2-Dichlorobenzene  
 Concen: 39.078 ng  
 RT: 7.73 min Scan# 841  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Resp Lower | Upper |
|---------|-------|------------|-------|
| 146     | 100   |            |       |
| 148     | 64.4  | 52.3       | 78.5  |
| 111     | 50.5  | 30.6       | 45.8# |





#15

Benzyl Alcohol

Concen: 41.264 ng

RT: 7.63 min Scan# 824

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampled :

SSTDCCC040

Tot Ion: 79 Resp: 623999

Ion Ratio Lower Upper

79 100

108 58.1 65.0 97.4#





77 69.6 53.0 79.6

Manual Integrations

APPROVED

Sohil

7/28/2017 4:45:03 PM



#16

2,2'-oxybis(1-chloropropane)

Concen: 41.859 ng

RT: 7.93 min Scan# 874

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24




Tgt Ion: 45 Resp: 502516


Ion Ratio Lower Upper

45 100

77 24.7 0.0 35.2

79 19.3 0.0 32.0









#17  
2-Methylphenol  
Concen: 41.467 ng  
RT: 7.84 min Scan# 859  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampled :  
SSTDCCC040

| Tot Ion   | 107   | Resp  | 453769 |
|-----------|-------|-------|--------|
| Ion Ratio | Lower | Upper |        |
| 107       | 100   |       |        |
| 108       | 107.6 | 89.8  | 134.8  |
| 77        | 75.7  | 32.6  | 48.8#  |
| 79        | 73.4  | 32.8  | 49.2#  |




Manual Integrations  
APPROVED




Sohil  
7/28/2017 4:45:03 PM



#18  
Hexachloroethane  
Concen: 40.259 ng  
RT: 8.45 min Scan# 962  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion   | 117   | Resp  | 247047 |
|-----------|-------|-------|--------|
| Ion Ratio | Lower | Upper |        |
| 117       | 100   |       |        |
| 119       | 93.4  | 79.2  | 118.8  |
| 201       | 122.9 | 69.8  | 104.6# |









#19  
n-Nitroso-di-n-propylamine  
Concen: 44.859 ng  
RT: 8.20 min Scan# 920  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 70      | 100   |       |       |
| 42      | 50.8  | 44.5  | 66.7  |
| 101     | 8.9   | 7.4   | 11.2  |
| 130     | 14.0  | 15.3  | 22.9# |


Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#20  
3+4-Methylphenols  
Concen: 43.232 ng  
RT: 8.17 min Scan# 915  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 107     | 100   |       |       |
| 108     | 90.5  | 73.2  | 113.2 |
| 77      | 54.3  | 10.7  | 50.7# |
| 79      | 46.6  | 9.6   | 49.6  |





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.15 min Scan# 1252

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M





ClientSampleId :

SSTDCCC040

| Tot Ion | Ratio | Resp | Lower | Upper |
|---------|-------|------|-------|-------|
| 136     | 100   |      |       |       |
| 137     | 10.2  | 8.7  | 13.1  |       |
| 54      | 10.0  | 4.6  | 6.8#  |       |
| 68      | 7.2   | 3.7  | 5.5#  |       |

Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM



#22

Acetophenone

Concen: 37.676 ng




RT: 8.20 min Scan# 921


Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Resp | Lower | Upper |
|---------|-------|------|-------|-------|
| 105     | 100   |      |       |       |
| 71      | 9.1   | 0.6  | 1.0#  |       |
| 51      | 33.3  | 21.6 | 32.4# |       |
| 120     | 17.3  | 16.1 | 24.1  |       |









#23  
 Nitrobenzene-d5  
 Concen: 81.866 ng  
 RT: 8.54 min Scan# 978  
 Delta R.T. -0.00 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Instrument :  
 BNA\_M  
 ClientSampled :  
 SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 82      | 100   |       |       |
| 128     | 27.3  | 32.8  | 49.2# |
| 54      | 48.9  | 40.6  | 60.8  |




Manual Integrations  
 APPROVED

Sohil  
 7/28/2017 4:45:03 PM




#24  
 Nitrobenzene  
 Concen: 40.928 ng  
 RT: 8.58 min Scan# 985  
 Delta R.T. -0.00 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 77      | 100   |       |       |
| 123     | 29.0  | 32.6  | 49.0# |
| 65      | 14.9  | 10.9  | 16.3  |







#25

Isophorone

Concen: 41.553 ng

RT: 9.10 min Scan# 1074

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M





ClientSampled :

SSTDCCC040

| Tot Ion   | 82    | Resp  | 1566504 |
|-----------|-------|-------|---------|
| Ion Ratio | Lower | Upper |         |
| 82        | 100   |       |         |
| 95        | 9.3   | 5.8   | 8.6#    |
| 138       | 13.0  | 16.3  | 24.5#   |

Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM



#26

2-Nitrophenol

Concen: 39.503 ng




RT: 9.27 min Scan# 1103


Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

| Tgt Ion   | 139   | Resp  | 294499 |
|-----------|-------|-------|--------|
| Ion Ratio | Lower | Upper |        |
| 139       | 100   |       |        |
| 109       | 70.9  | 20.1  | 30.1#  |
| 65        | 91.7  | 31.5  | 47.3#  |









#27  
2,4-Dimethylphenol  
Concen: 40.556 ng  
RT: 9.35 min Scan# 1116  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper  |
|---------|-------|-------|--------|
| 122     | 100   |       |        |
| 107     | 150.2 | 84.7  | 127.1# |
| 121     | 58.5  | 46.6  | 69.8   |




Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#28  
bis(2-Chloroethoxy)methane  
Concen: 39.522 ng  
RT: 9.59 min Scan# 1156  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 93      | 100   |       |       |
| 95      | 31.0  | 25.5  | 38.3  |
| 123     | 8.3   | 8.6   | 13.0# |









#29  
 2,4-Dichlorophenol  
 Concen: 40.695 ng  
 RT: 9.81 min Scan# 1194  
 Delta R.T. -0.00 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTCCCC040

| Tot Ion | Ratio | Resp   | Lower | Upper |
|---------|-------|--------|-------|-------|
| 162     | 100   | 600138 |       |       |
| 164     | 63.2  |        | 42.8  | 82.8  |
| 98      | 40.6  |        | 14.5  | 54.5  |




Manual Integrations  
 APPROVED


Sohil  
 7/28/2017 4:45:03 PM



#30  
 1,2,4-Trichlorobenzene  
 Concen: 39.691 ng  
 RT: 10.02 min Scan# 1229  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Resp   | Lower | Upper |
|---------|-------|--------|-------|-------|
| 180     | 100   | 723732 |       |       |
| 182     | 101.7 |        | 77.6  | 116.4 |
| 145     | 30.5  |        | 23.4  | 35.2  |





#31

Naphthalene

Concen: 39.689 ng

RT: 10.20 min Scan# 1260

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampled :

SSTDCCC040

Tot Ion:128 Resp: 1694350

Ion Ratio Lower Upper

128 100

129 10.9 8.9 13.3


127 13.3 10.4 15.6

Manual Integrations

APPROVED

Sohil




7/28/2017 4:45:03 PM



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 42.971 ng

RT: 9.52 min Scan# 1145

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24


Tgt Ion:122 Resp: 453300

Ion Ratio Lower Upper

122 100

105 144.3 99.9 139.9#



77 149.8 73.7 113.7#




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): E









#33  
4-Chloroaniline  
Concen: 41.961 ng  
RT: 10.32 min Scan# 1280  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 127     | 100   |       |       |
| 129     | 31.4  | 25.3  | 37.9  |
| 65      | 54.3  | 17.6  | 26.4  |
| 92      | 29.6  | 13.1  | 19.7  |




Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#34  
Hexachlorobutadiene  
Concen: 40.422 ng  
RT: 10.49 min Scan# 1309  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 225     | 100   |       |       |
| 223     | 63.6  | 47.9  | 71.9  |
| 227     | 66.3  | 50.1  | 75.1  |





#35

Caprolactam

Concen: 47.665 ng/ml

RT: 11.12 min Scan# 1417

Delta R.T. 0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040

Tot Ion:113 Resp: 199342

Ion Ratio Lower Upper

113 100

55 127.6 145.9 185.9#


56 122.8 101.0 141.0

Manual Integrations

APPROVED

Sohil

7/28/2017 4:45:03 PM



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BM0

Ion 56.00 (55.70 to 56.70): BM0

150000



100000

50000

0

11.12

Time--> 11.00 11.10 11.20 11.30



#36

4-Chloro-3-methylphenol

Concen: 44.701 ng/ml

RT: 11.46 min Scan# 1474

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24


Tgt Ion:107 Resp: 851212

Ion Ratio Lower Upper

107 100

144 22.0 23.3 34.9#

142 62.5 72.6 109.0#



Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E

600000


400000


200000

0

11.46

Time--> 11.40 11.45 11.50









#37  
2-Methylnaphthalene  
Concen: 42.750 ng  
RT: 11.82 min Scan# 1536  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampled :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 142     | 100   |       |       |
| 141     | 88.0  | 71.9  | 107.9 |
| 115     | 55.6  | 25.4  | 38.0  |




Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#38  
Acenaphthene-d10  
Concen: 20.000 ng  
RT: 14.05 min Scan# 1914  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 164     | 100   |       |       |
| 162     | 101.2 | 82.4  | 123.6 |
| 160     | 46.8  | 35.8  | 53.6  |









#39  
 1,2,4,5-Tetrachlorobenzene  
 Concen: 37.055 ng  
 RT: 12.20 min Scan# 1601  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTCCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 216     | 100   |       |       |
| 214     | 77.4  | 0.0   | 0.0#  |
| 179     | 19.9  | 0.0   | 0.0#  |
| 108     | 20.5  | 0.0   | 0.0#  |




Manual Integrations  
 APPROVED

Sohil  
 7/28/2017 4:45:03 PM




#40  
 Hexachlorocyclopentadiene  
 Concen: 36.915 ng  
 RT: 12.19 min Scan# 1598  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 237     | 100   |       |       |
| 235     | 60.0  | 42.0  | 82.0  |
| 272     | 12.4  | 0.0   | 33.4  |











#41  
 2,4,6-Tribromophenol  
 Concen: 87.961 ng  
 RT: 15.55 min Scan# 2170  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

| Tot Ion | Ratio | Resp | Lower | Upper |
|---------|-------|------|-------|-------|
| 330     | 100   |      |       |       |
| 332     | 97.1  | 0.0  | 0.0   | 0.0   |
| 141     | 40.5  | 0.0  | 0.0   | 0.0   |




Manual Integrations  
 APPROVED


Sohil  
 7/28/2017 4:45:03 PM



#42  
 2,4,6-Trichlorophenol  
 Concen: 38.692 ng  
 RT: 12.46 min Scan# 1644  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Resp | Lower | Upper |
|---------|-------|------|-------|-------|
| 196     | 100   |      |       |       |
| 198     | 98.8  | 76.3 | 114.5 |       |
| 200     | 29.9  | 25.6 | 38.4  |       |









#43  
2,4,5-Trichlorophenol  
Concen: 38.995 ng  
RT: 12.53 min Scan# 1656  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Resp   | Lower | Upper |
|---------|-------|--------|-------|-------|
| 196     | 100   | 729815 |       |       |
| 198     | 97.5  | 79.4   | 119.0 |       |
| 97      | 65.6  | 26.8   | 40.2  |       |
| 132     | 23.1  | 18.6   | 28.0  |       |




Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#44  
2-Fluorobiphenyl  
Concen: 74.826 ng  
RT: 12.66 min Scan# 1678  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Resp    | Lower | Upper |
|---------|-------|---------|-------|-------|
| 172     | 100   | 4169198 |       |       |
| 171     | 34.9  | 28.5    | 42.7  |       |
| 170     | 23.1  | 18.8    | 28.2  |       |





#45

1,1'-Biphenyl

Concen: 37.191 ng

RT: 12.87 min Scan# 1714

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040

Tot Ion:154 Resp: 2247214

Ion Ratio Lower Upper

154 100

153 40.9 20.7 60.7


76 14.8 0.0 30.9

Manual Integrations

APPROVED

Sohil




7/28/2017 4:45:03 PM



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BM



#46

2-Chloronaphthalene

Concen: 37.393 ng

RT: 12.90 min Scan# 1720

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24


Tgt Ion:162 Resp: 1664717

Ion Ratio Lower Upper

162 100

127 41.1 26.9 40.3#



164 33.4 26.8 40.2




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E





#47

2-Nitroaniline

Concen: 43.445 ng

RT: 13.13 min Scan# 1758

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M





ClientSampleId :

SSTDCCC040

|           |       |       |        |
|-----------|-------|-------|--------|
| Tot Ion:  | 65    | Resp: | 684241 |
| Ion Ratio | Lower | Upper |        |
| 65        | 100   |       |        |
| 92        | 65.4  | 59.1  | 88.7   |
| 138       | 71.2  | 93.9  | 140.9# |

Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM



#48

Acenaphthylene

Concen: 39.556 ng




RT: 13.76 min Scan# 1866


Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

|           |       |       |         |
|-----------|-------|-------|---------|
| Tgt Ion:  | 152   | Resp: | 2628082 |
| Ion Ratio | Lower | Upper |         |
| 152       | 100   |       |         |
| 151       | 20.0  | 15.9  | 23.9    |
| 153       | 12.5  | 10.6  | 16.0    |









#49  
Dimethylphthalate  
Concen: 40.128 ng  
RT: 13.53 min Scan# 1826  
Delta R.T. -0.00 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 163     | 100   |       |       |
| 194     | 4.2   | 2.7   | 4.1#  |
| 164     | 10.3  | 8.2   | 12.2  |




Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#50  
2,6-Dinitrotoluene  
Concen: 41.377 ng  
RT: 13.64 min Scan# 1845  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 165     | 100   |       |       |
| 63      | 84.5  | 44.1  | 66.1# |
| 89      | 86.6  | 44.3  | 66.5# |





#51

Acenaphthene

Concen: 39.302 ng

RT: 14.12 min Scan# 1926

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040

Tot Ion:154 Resp: 1616814

Ion Ratio Lower Upper

154 100

153 113.0 90.0 135.0





152 53.3 42.6 63.8

Manual Integrations

APPROVED

Sohil

7/28/2017 4:45:03 PM



#52

3-Nitroaniline

Concen: 42.077 ng

RT: 13.97 min Scan# 1901

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24




Tgt Ion:138 Resp: 441389


Ion Ratio Lower Upper

138 100

108 14.6 7.3 10.9#

92 169.1 76.6 114.8#









#53  
2,4-Dinitrophenol  
Concen: 42.015 ng  
RT: 14.17 min Scan# 1936  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 184     | 100   |       |       |
| 63      | 93.9  | 69.2  | 103.8 |
| 154     | 92.9  | 53.3  | 79.9# |




Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#54  
Dibenzofuran  
Concen: 40.185 ng  
RT: 14.45 min Scan# 1983  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 168     | 100   |       |       |
| 139     | 41.6  | 27.3  | 40.9# |
| 169     | 13.1  | 10.6  | 16.0  |









#55  
4-Nitrophenol  
Concen: 42.415 ng  
RT: 14.29 min Scan# 1956  
Delta R.T. 0.00 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Resp  | Lower | Upper |
|---------|-------|-------|-------|-------|
| 139     | 100   |       |       |       |
| 109     | 167.5 | 130.0 | 170.0 |       |
| 65      | 163.6 | 130.0 | 170.0 |       |




Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM




#56  
2,4-Dinitrotoluene  
Concen: 43.991 ng  
RT: 14.43 min Scan# 1980  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Resp | Lower | Upper |
|---------|-------|------|-------|-------|
| 165     | 100   |      |       |       |
| 63      | 64.1  | 52.4 | 78.6  |       |
| 89      | 97.6  | 72.4 | 108.6 |       |
| 182     | 3.0   | 2.0  | 3.0#  |       |











#57  
**Fluorene**  
 Concen: 41.728 ng  
 RT: 15.10 min Scan# 2094  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTCCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 166     | 100   |       |       |
| 165     | 96.1  | 79.0  | 118.4 |
| 167     | 13.0  | 10.3  | 15.5  |




Manual Integrations  
**APPROVED**


Sohil  
 7/28/2017 4:45:03 PM



#58  
**2,3,4,6-Tetrachlorophenol**  
 Concen: 41.847 ng  
 RT: 14.69 min Scan# 2023  
 Delta R.T. -0.00 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 232     | 100   |       |       |
| 131     | 41.3  | 0.0   | 0.0#  |
| 130     | 3.0   | 0.0   | 0.0#  |
| 166     | 28.9  | 0.0   | 0.0#  |





#59

Diethylphthalate

Concen: 42.123 ng

RT: 14.91 min Scan# 2061

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040

Tot Ion:149 Resp: 2571585

Ion Ratio Lower Upper

149 100

177 23.7 18.3 27.5


150 12.4 9.8 14.8

Manual Integrations

APPROVED

Sohil

7/28/2017 4:45:03 PM






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 41.539 ng

RT: 15.11 min Scan# 2095

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24


Tgt Ion:204 Resp: 1455494

Ion Ratio Lower Upper

204 100

206 31.2 26.6 39.8

141 56.3 45.2 67.8






Abundance

Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E








#61  
4-Nitroaniline  
Concen: 44.257 ng  
RT: 15.14 min Scan# 2101  
Delta R.T. -0.00 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 138     | 100   |       |       |
| 92      | 74.9  | 16.7  | 56.7# |
| 108     | 176.6 | 37.6  | 77.6# |





Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#62  
Azobenzene  
Concen: 43.678 ng  
RT: 15.40 min Scan# 2145  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 77      | 100   |       |       |
| 182     | 20.9  | 6.5   | 46.5  |
| 105     | 16.3  | 3.8   | 43.8  |
| 51      | 32.0  | 5.5   | 45.5  |









#63  
 Phenanthrene-d10  
 Concen: 20.000 ng  
 RT: 16.80 min Scan# 2383  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 188     | 100   |       |       |
| 94      | 7.3   | 8.7   | 13.1# |
| 80      | 10.2  | 9.3   | 13.9  |




Manual Integrations  
 APPROVED


Sohil  
 7/28/2017 4:45:03 PM



#64  
 4,6-Dinitro-2-methylphenol  
 Concen: 40.227 ng  
 RT: 15.20 min Scan# 2110  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 198     | 100   |       |       |
| 51      | 52.3  | 28.8  | 68.8  |
| 105     | 46.1  | 30.5  | 70.5  |









#65  
 n-Nitrosodiphenylamine  
 Concen: 38.015 ng  
 RT: 15.33 min Scan# 2133  
 Delta R.T. -0.00 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 169     | 100   |       |       |
| 168     | 66.7  | 53.9  | 80.9  |
| 167     | 36.4  | 28.9  | 43.3  |




Manual Integrations  
 APPROVED


Sohil  
 7/28/2017 4:45:03 PM



#66  
 4-Bromophenyl-phenylether  
 Concen: 39.089 ng  
 RT: 16.01 min Scan# 2248  
 Delta R.T. -0.00 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 248     | 100   |       |       |
| 250     | 98.1  | 77.4  | 116.0 |
| 141     | 72.7  | 45.2  | 67.8  |





#67

Hexachlorobenzene

Concen: 38.058 ng

RT: 16.11 min Scan# 2265

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M





ClientSampled :

SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 284     | 100   |       |       |
| 142     | 34.6  | 20.4  | 30.6# |
| 249     | 34.7  | 23.8  | 35.6  |

Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM



#68

Atrazine

Concen: 40.962 ng




RT: 16.30 min Scan# 2297


Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 200     | 100   |       |       |
| 173     | 20.8  | 4.8   | 44.8  |
| 215     | 48.1  | 26.9  | 66.9  |






#69  
 Pentachlorophenol  
 Concen: 41.053 ng  
 RT: 16.46 min Scan# 2325  
 Delta R.T. -0.00 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040




| Tot Ion | Ratio | Resp | Lower | Upper |
|---------|-------|------|-------|-------|
| 266     | 100   |      |       |       |
| 268     | 64.2  | 52.0 | 78.0  |       |
| 264     | 60.4  | 49.0 | 73.4  |       |

Manual Integrations  
 APPROVED

Sohil  
 7/28/2017 4:45:03 PM




Abundance Ion 265.70 (265.40 to 266.40): E  
 Ion 268.00 (267.70 to 268.70): E  
 Ion 264.00 (263.70 to 264.70): E






#70  
 Phenanthrene  
 Concen: 39.649 ng  
 RT: 16.84 min Scan# 2390  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Resp | Lower | Upper |
|---------|-------|------|-------|-------|
| 178     | 100   |      |       |       |
| 176     | 18.8  | 15.2 | 22.8  |       |
| 179     | 14.8  | 12.1 | 18.1  |       |



Abundance Ion 178.00 (177.70 to 178.70): E  
 Ion 176.00 (175.70 to 176.70): E  
 Ion 179.00 (178.70 to 179.70): E









#71  
 Anthracene  
 Concen: 40.102 ng  
 RT: 16.94 min Scan# 2406  
 Delta R.T. -0.00 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDCCC040

Tot Ion:178 Resp: 4177822  
 Ion Ratio Lower Upper  
 178 100  
 176 19.0 14.6 22.0  
 179 15.4 12.6 19.0




Manual Integrations  
 APPROVED

Sohil  
 7/28/2017 4:45:03 PM




#72  
 Carbazole  
 Concen: 40.552 ng  
 RT: 17.22 min Scan# 2453  
 Delta R.T. -0.01 min  
 Lab File: BM011021.D  
 Acq: 27 Jul 2017 21:24

Tgt Ion:167 Resp: 3694617  
 Ion Ratio Lower Upper  
 167 100  
 166 21.8 17.2 25.8  
 139 13.1 10.8 16.2







#73

Di-n-butylphthalate

Concen: 41.588 ng

RT: 17.80 min Scan# 2553

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040

Tot Ion:149 Resp: 4614142

Ion Ratio Lower Upper

149 100





150 8.9 7.5 11.3

104 7.5 3.7 5.5#

Manual Integrations  
APPROVED

Sohil

7/28/2017 4:45:03 PM



#74

Fluoranthene

Concen: 41.121 ng

RT: 18.88 min Scan# 2736

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24




Tgt Ion:202 Resp: 5323184


Ion Ratio Lower Upper

202 100

101 7.2 0.0 28.7

203 17.0 0.0 37.2





#75

Chrysene-d12

Concen: 20.000 ng

RT: 21.03 min Scan# 3101

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampled :





SSTDCCC040

Tot Ion:240 Resp: 2192947

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 240 | 100   |       |       |
| 120 | 5.7   | 8.2   | 12.2# |
| 236 | 25.0  | 20.0  | 30.0  |

Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM



#76

Benzidine

Concen: 39.642 ng

RT: 19.08 min Scan# 2770




Delta R.T. -0.01 min


Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Tgt Ion:184 Resp: 2079695

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 184 | 100   |       |       |
| 185 | 14.4  | 11.3  | 16.9  |
| 183 | 12.1  | 8.9   | 13.3  |





#77

Pvrene

Concen: 42.315 ng

RT: 19.24 min Scan# 2798

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M





ClientSampleId :

SSTDCCC040

|          |       |       |         |
|----------|-------|-------|---------|
| Tot Ion: | 202   | Resp: | 5513679 |
| Ion      | Ratio | Lower | Upper   |
| 202      | 100   |       |         |
| 200      | 20.5  | 16.9  | 25.3    |
| 203      | 17.6  | 14.1  | 21.1    |

Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM



#78

Terphenyl-d14

Concen: 86.727 ng




RT: 19.47 min Scan# 2836


Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 244   | Resp: | 9601209 |
| Ion      | Ratio | Lower | Upper   |
| 244      | 100   |       |         |
| 212      | 8.0   | 4.9   | 7.3#    |
| 122      | 6.9   | 6.2   | 9.4     |









#79  
Butylbenzylphthalate  
Concen: 42.558 ng  
RT: 20.18 min Scan# 2957  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 149     | 100   |       |       |
| 91      | 91.9  | 50.1  | 75.1# |
| 206     | 24.4  | 16.2  | 24.2# |




Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#80  
Benzo(a)anthracene  
Concen: 39.601 ng  
RT: 21.01 min Scan# 3098  
Delta R.T. -0.00 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 228     | 100   |       |       |
| 226     | 26.8  | 21.7  | 32.5  |
| 229     | 19.3  | 16.0  | 24.0  |









#81  
3,3'-Dichlorobenzidine  
Concen: 39.038 ng  
RT: 20.96 min Scan# 3089  
Delta R.T. -0.00 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

Tot Ion:252 Resp: 1917637  
Ion Ratio Lower Upper  
252 100  
254 64.1 51.9 77.9  
126 8.0 9.8 14.8#




Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#82  
Chrysene  
Concen: 39.141 ng  
RT: 21.06 min Scan# 3107  
Delta R.T. -0.00 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Tgt Ion:228 Resp: 4704605  
Ion Ratio Lower Upper  
228 100  
226 28.8 24.1 36.1  
229 19.7 15.5 23.3





#83

Bis(2-ethylhexyl)phthalate

Concen: 40.855 ng

RT: 20.97 min Scan# 3091

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040

Tot Ion:149 Resp: 2784883

Ion Ratio Lower Upper

149 100

167 27.6 22.4 33.6





279 4.5 3.4 5.0

Manual Integrations

APPROVED

Sohil

7/28/2017 4:45:03 PM



#84

Di-n-octyl phthalate

Concen: 39.337 ng

RT: 21.82 min Scan# 3236

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24




Tgt Ion:149 Resp: 4351982


Ion Ratio Lower Upper

149 100

167 1.8 1.2 1.8

43 9.4 7.3 10.9









#85  
Indeno(1,2,3-cd)pyrene  
Concen: 32.244 ng  
RT: 25.21 min Scan# 3813  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

Tot Ion:276 Resp: 4013778  
Ion Ratio Lower Upper  
276 100  
138 11.7 12.0 18.0#  
277 25.6 20.0 30.0




Manual Integrations  
APPROVED


Sohil  
7/28/2017 4:45:03 PM



#86  
Perylene-d12  
Concen: 20.000 ng  
RT: 23.13 min Scan# 3459  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Tgt Ion:264 Resp: 1658998  
Ion Ratio Lower Upper  
264 100  
260 22.8 18.6 27.8  
265 21.7 17.3 25.9





#87

Benzo(b)fluoranthene

Concen: 40.665 ng

RT: 22.52 min Scan# 3354

Delta R.T. -0.00 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M





ClientSampleId :

SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 252     | 100   |       |       |
| 253     | 21.5  | 18.0  | 27.0  |
| 125     | 4.7   | 9.9   | 14.9  |

Manual Integrations  
APPROVED

Sohil  
7/28/2017 4:45:03 PM



#88

Benzo(k)fluoranthene

Concen: 42.035 ng




RT: 22.56 min Scan# 3361

Delta R.T. -0.01 min


Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 252     | 100   |       |       |
| 253     | 21.9  | 17.2  | 25.8  |
| 125     | 4.6   | 8.1   | 12.1  |











#89  
Benzo(a)pyrene  
Concen: 39.992 ng  
RT: 23.04 min Scan# 3444  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 252     | 100   |       |       |
| 253     | 22.5  | 17.6  | 26.4  |
| 125     | 5.8   | 7.4   | 11.2  |




Manual Integrations  
APPROVED




Sohil  
7/28/2017 4:45:03 PM



#90  
Dibenzo(a,h)anthracene  
Concen: 37.019 ng  
RT: 25.23 min Scan# 3815  
Delta R.T. -0.01 min  
Lab File: BM011021.D  
Acq: 27 Jul 2017 21:24

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 278     | 100   |       |       |
| 139     | 7.4   | 13.4  | 20.0  |
| 279     | 23.2  | 19.0  | 28.4  |





#91

Benzo(a,h,i)perylene

Concen: 36.406 ng

RT: 25.84 min Scan# 3920

Delta R.T. -0.01 min

Lab File: BM011021.D

Acq: 27 Jul 2017 21:24

Instrument :

BNA\_M

ClientSampleId :

SSTDCCC040

Tot Ion:276 Resp: 3192980

Ion Ratio Lower Upper

276 100

277 24.0 18.5 27.7

138 9.7 19.0 28.6#

Manual Integrations

APPROVED

Sohil

7/28/2017 4:45:03 PM

