


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM072720\
 Data File : BM026865.D
 Acq On : 27 Jul 2020 10:03
 Operator : JU/CG
 Sample : DFTPP14
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP14

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072720MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Jul 28 10:40:31 2020



AutoFind: Scans 2514, 2515, 2516; Background Corrected with Scan 2509

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.4	49712	PASS
68	69	0.00	2	1.5	1186	PASS
69	198	0.00	100	60.3	77993	PASS
70	69	0.00	2	0.7	546	PASS
127	198	10	80	51.5	66597	PASS
197	198	0.00	2	0.6	769	PASS
198	198	100	100	100.0	129338	PASS
199	198	5	9	7.0	9094	PASS
275	198	10	60	27.5	35581	PASS
365	198	1	100	4.1	5343	PASS
441	198	0.01	100	11.0	14240	PASS
442	198	50	100	73.1	94570	PASS
443	442	15	24	19.9	18821	PASS

m/z	Abundance
37.90	176.0
39.10	831.0
40.00	582.0
41.10	1463.0
42.00	441.0
43.20	1383.0
44.00	4960.0
45.00	279.0
50.80	296.0
53.00	251.0
54.00	251.0
55.00	1762.0
56.10	703.0
57.10	1260.0
59.00	186.0
60.00	194.0
63.90	167.0
65.00	234.0
67.00	793.0
68.10	446.0
69.10	844.0
70.00	374.0
71.10	815.0
73.00	1157.0
73.80	210.0
75.00	181.0
77.00	544.0
78.10	227.0
79.10	467.0
80.00	192.0
81.10	626.0
82.10	526.0
83.00	850.0
84.10	690.0
85.10	529.0
86.10	308.0
91.10	589.0
92.00	168.0
92.90	390.0
93.90	201.0
95.00	412.0
96.00	538.0
97.10	328.0
98.10	569.0
99.10	511.0
103.10	175.0
104.00	154.0
104.90	343.0
106.90	224.0
109.20	241.0
110.10	172.0
111.00	375.0
112.00	278.0
113.10	413.0
115.10	248.0
116.90	223.0
119.00	251.0
121.00	168.0
129.10	292.0
130.90	162.0
133.00	286.0
140.90	176.0
145.10	151.0
147.10	210.0
148.80	173.0
153.00	150.0
164.90	189.0
173.10	388.0
176.90	172.0
190.90	480.0
193.00	334.0
194.00	183.0
207.00	3241.0
207.90	645.0
209.00	391.0
266.90	197.0
281.00	409.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

m/z	Abundance
36.90	328.0
37.90	636.0
39.10	3697.0
40.00	814.0
41.00	1505.0
42.00	628.0
43.10	1572.0
44.00	5437.0
45.00	327.0
49.00	383.0
50.00	10556.0
51.10	38496.0
52.00	2264.0
53.00	364.0
54.00	381.0
55.10	2091.0
56.10	1819.0
57.00	4367.0
58.00	453.0
59.10	224.0
60.00	185.0
61.00	744.0
61.90	756.0
63.00	2142.0
64.00	350.0
65.00	1354.0
66.10	208.0
67.10	843.0
68.00	1308.0
69.00	59512.0
70.10	710.0
71.10	840.0
73.00	1687.0
74.00	6410.0
75.00	9059.0
76.00	3141.0
77.10	57128.0
78.10	4333.0
79.00	4745.0
80.00	3195.0
81.00	4826.0
82.10	1500.0
83.00	1687.0
84.00	992.0
85.00	1277.0
85.90	1442.0
86.90	517.0
88.00	322.0
88.80	170.0
91.00	1333.0
92.10	1343.0
93.00	7356.0
94.00	685.0
95.00	484.0
96.00	542.0
97.00	720.0
98.00	5676.0
99.00	4462.0
100.00	531.0
101.00	2423.0
101.80	184.0
102.90	794.0
104.00	1652.0
105.00	1350.0
106.00	365.0
107.00	15071.0
108.00	2515.0
110.00	25376.0
111.00	4313.0
111.90	652.0
112.90	400.0
114.90	310.0
116.10	900.0
117.00	14556.0
118.00	1065.0
119.20	323.0
120.00	340.0
121.10	312.0
122.00	1411.0
123.00	1753.0
123.90	558.0
125.00	851.0
126.10	384.0
127.00	45088.0
128.00	4117.0
129.00	18944.0
130.00	2060.0
130.90	576.0
132.00	178.0
133.10	487.0
133.90	563.0
135.00	2002.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

136.00	674.0
137.10	748.0
138.00	329.0
138.90	253.0
139.90	312.0
141.00	2572.0
142.00	1015.0
143.00	433.0
144.00	243.0
145.10	349.0
146.00	459.0
147.00	1655.0
148.00	3193.0
149.00	804.0
150.00	285.0
150.90	411.0
152.00	380.0
153.00	859.0
153.90	1029.0
154.90	1531.0
156.10	1959.0
157.10	533.0
157.70	303.0
158.90	455.0
159.90	1043.0
160.90	1111.0
161.90	397.0
163.00	177.0
164.90	1220.0
165.90	663.0
167.00	5719.0
168.00	2184.0
169.00	376.0
170.10	170.0
170.90	280.0
172.10	513.0
172.90	1102.0
173.90	1126.0
175.00	2276.0
176.00	538.0
177.00	956.0
178.00	395.0
179.00	3855.0
180.00	2629.0
180.90	1107.0
181.90	279.0
183.90	503.0
185.00	2101.0
186.00	13584.0
187.00	3955.0
188.10	523.0
189.00	1384.0
191.00	813.0
192.00	1489.0
192.90	1547.0
193.90	281.0
196.00	2766.0
196.90	562.0
198.00	81776.0
198.90	5843.0
199.90	379.0
201.30	405.0
202.10	228.0
203.00	737.0
204.10	3995.0
205.00	5718.0
206.00	22832.0
207.00	6622.0
208.00	1757.0
209.00	701.0
209.90	409.0
210.90	835.0
211.70	293.0
214.90	293.0
215.90	473.0
217.00	6267.0
218.00	899.0
220.20	347.0
221.00	3345.0
222.90	1408.0
224.00	12556.0
225.00	3441.0
225.90	252.0
226.20	228.0
227.00	6165.0
227.90	709.0
228.90	829.0
230.80	357.0
234.00	403.0
235.00	385.0
235.90	203.0
236.90	385.0
238.90	323.0
239.80	150.0
241.00	366.0
242.00	701.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

243.00	675.0
244.00	9966.0
245.10	1210.0
246.00	1982.0
246.90	711.0
249.00	420.0
249.90	180.0
250.90	167.0
251.90	181.0
252.90	224.0
253.80	390.0
255.00	50584.0
256.00	7415.0
256.90	452.0
258.00	3345.0
259.00	485.0
265.00	1608.0
265.90	192.0
266.90	207.0
271.90	215.0
273.00	1593.0
274.00	3423.0
275.00	20976.0
276.00	2466.0
276.90	1931.0
278.00	541.0
281.00	429.0
282.00	238.0
283.00	339.0
284.00	172.0
285.00	445.0
293.00	414.0
296.00	6338.0
297.00	1206.0
302.00	186.0
302.90	757.0
304.00	300.0
314.00	334.0
314.90	651.0
316.00	435.0
323.00	1951.0
323.90	367.0
327.00	410.0
328.10	162.0
332.10	179.0
332.90	318.0
334.00	1146.0
335.00	307.0
341.00	242.0
345.90	352.0
351.90	646.0
353.00	374.0
353.90	374.0
355.10	253.0
365.00	2540.0
366.10	366.0
370.80	184.0
372.00	920.0
373.00	253.0
382.80	232.0
389.90	163.0
402.00	439.0
403.10	456.0
420.90	336.0
422.00	426.0
423.00	2253.0
424.00	751.0
441.00	6811.0
442.10	44264.0
443.10	8855.0
444.00	896.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

m/z	Abundance
37.10	292.0
38.10	898.0
39.10	5647.0
40.00	803.0
41.10	1687.0
42.10	449.0
43.00	1326.0
44.00	4628.0
45.00	249.0
49.00	841.0
50.10	14960.0
51.00	57352.0
52.00	3130.0
53.00	441.0
54.00	156.0
54.20	153.0
55.10	2727.0
56.00	2772.0
57.00	4950.0
58.00	267.0
59.00	260.0
60.00	215.0
60.90	1008.0
62.00	1585.0
63.00	2810.0
64.10	659.0
65.00	1543.0
65.90	281.0
67.00	926.0
68.00	2009.0
69.00	92040.0
69.90	1074.0
71.00	807.0
73.10	1472.0
74.00	9107.0
75.00	13981.0
76.10	4443.0
77.10	92448.0
78.10	6842.0
79.00	6912.0
80.00	5191.0
81.00	7316.0
82.00	2597.0
83.00	2266.0
84.00	778.0
85.00	1593.0
85.90	2567.0
87.00	1035.0
87.90	156.0
89.10	178.0
91.00	2062.0
92.00	1748.0
93.00	11529.0
94.10	1112.0
95.10	558.0
96.10	995.0
97.00	840.0
98.00	9576.0
99.00	6279.0
99.90	591.0
101.00	3857.0
102.00	184.0
103.10	1404.0
104.00	2654.0
105.10	2399.0
106.00	673.0
107.00	25296.0
108.00	4221.0
109.10	919.0
110.00	41264.0
111.00	6333.0
112.10	1219.0
113.00	673.0
115.10	305.0
116.00	1432.0
117.00	23960.0
118.00	2151.0
119.00	375.0
120.00	412.0
121.00	322.0
121.90	1697.0
123.00	2907.0
124.00	928.0
124.90	1074.0
126.20	588.0
127.00	78928.0
128.00	6022.0
129.00	32784.0
130.00	2942.0
131.00	589.0
132.10	384.0
133.00	397.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

134.00	1147.0
135.00	2789.0
135.90	1195.0
137.00	1329.0
137.80	351.0
139.10	258.0
139.90	643.0
141.00	5051.0
142.00	1772.0
143.00	1368.0
144.00	366.0
144.90	225.0
146.00	1024.0
147.00	2614.0
148.00	5514.0
149.10	1391.0
149.90	335.0
150.90	648.0
152.10	499.0
153.00	1834.0
154.00	1119.0
155.00	2957.0
156.00	3722.0
156.90	991.0
158.00	1028.0
159.00	683.0
160.00	1483.0
161.00	2148.0
162.00	748.0
162.90	538.0
163.80	278.0
165.00	1846.0
166.00	1326.0
167.00	9945.0
168.00	4637.0
169.00	579.0
170.20	458.0
171.00	398.0
172.00	904.0
173.00	1244.0
174.00	1928.0
175.10	3879.0
176.10	1100.0
177.00	1980.0
178.00	825.0
179.00	7511.0
180.00	4953.0
181.00	2341.0
182.00	386.0
183.00	221.0
183.90	675.0
185.00	3040.0
186.00	23080.0
187.00	7108.0
188.00	698.0
189.00	2165.0
190.00	399.0
190.90	1167.0
192.00	2304.0
193.00	2713.0
194.00	717.0
194.90	243.0
196.00	4088.0
197.10	1009.0
197.90	153728.0
199.00	10628.0
199.90	919.0
201.40	501.0
201.80	531.0
203.00	1444.0
204.00	6573.0
205.00	10940.0
206.00	43088.0
207.00	8832.0
208.00	2191.0
209.10	1077.0
209.90	762.0
211.00	1831.0
211.90	369.0
215.00	610.0
216.00	774.0
217.00	12670.0
218.00	1356.0
219.00	259.0
221.00	6311.0
221.90	1357.0
223.00	2325.0
224.00	24640.0
225.00	6160.0
226.00	662.0
227.00	11006.0
228.00	1689.0
228.90	2034.0
229.90	352.0
231.00	840.0
232.10	173.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

233.00	163.0
234.00	705.0
234.90	516.0
236.00	631.0
237.00	799.0
239.00	460.0
239.90	458.0
241.00	483.0
242.10	1258.0
243.00	1332.0
244.00	16936.0
245.00	2496.0
246.00	3993.0
247.00	800.0
248.00	231.0
248.90	699.0
250.10	187.0
250.90	255.0
252.00	154.0
253.00	580.0
254.10	659.0
255.00	93592.0
256.00	13747.0
257.00	1397.0
258.00	5992.0
258.90	1059.0
259.80	243.0
264.20	280.0
265.00	2515.0
266.00	624.0
266.90	246.0
271.10	163.0
272.00	359.0
273.00	2715.0
274.00	7362.0
275.00	40792.0
276.00	5397.0
276.90	4027.0
278.00	697.0
280.90	358.0
282.90	439.0
284.00	332.0
285.00	611.0
288.80	154.0
292.10	251.0
293.00	979.0
294.20	174.0
295.10	347.0
296.00	13568.0
297.00	1948.0
301.00	240.0
301.90	265.0
303.00	1395.0
304.00	421.0
307.90	280.0
309.10	161.0
310.00	317.0
313.90	624.0
315.00	1478.0
316.00	497.0
320.90	319.0
322.00	176.0
323.10	3547.0
324.00	881.0
327.00	841.0
328.10	288.0
332.00	426.0
332.80	379.0
334.00	2727.0
334.90	607.0
341.00	442.0
346.10	901.0
352.00	1122.0
353.10	601.0
354.00	1374.0
355.10	450.0
365.00	5757.0
365.90	630.0
371.00	277.0
372.10	1625.0
373.00	495.0
383.00	507.0
384.00	185.0
390.10	247.0
391.10	207.0
392.10	229.0
401.00	189.0
402.00	873.0
403.00	847.0
403.90	432.0
421.10	668.0
422.00	1182.0
423.10	6015.0
424.00	1189.0
425.00	182.0
441.00	14661.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

442.10	102744.0
443.10	21352.0
444.10	1901.0
444.90	196.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

m/z	Abundance
37.10	400.0
38.00	787.0
39.00	4688.0
40.00	753.0
41.00	1690.0
42.00	393.0
43.10	1358.0
44.00	5182.0
44.80	283.0
48.90	401.0
50.10	14141.0
51.00	53288.0
52.00	3005.0
53.10	512.0
54.00	294.0
55.10	2566.0
56.00	2481.0
57.00	5005.0
58.00	379.0
59.10	206.0
59.90	371.0
61.00	860.0
62.00	1151.0
63.10	2778.0
64.00	714.0
65.00	1648.0
66.10	203.0
67.00	696.0
68.10	1580.0
69.00	84960.0
70.00	978.0
71.10	797.0
72.00	300.0
72.90	2033.0
74.10	8163.0
75.00	12744.0
76.00	3791.0
77.10	81728.0
78.10	6105.0
79.00	6868.0
80.00	4796.0
81.00	7011.0
82.00	2089.0
83.00	1937.0
84.00	984.0
85.00	1625.0
86.00	2221.0
87.10	1034.0
87.90	364.0
89.10	329.0
91.00	1791.0
92.00	1422.0
93.00	10334.0
94.00	756.0
95.10	648.0
96.00	979.0
97.10	832.0
98.00	8644.0
99.00	6407.0
100.00	783.0
100.90	3555.0
101.80	259.0
103.00	1066.0
104.00	2212.0
105.00	2229.0
106.00	906.0
107.00	22848.0
108.00	3813.0
110.00	38872.0
111.00	5805.0
111.90	950.0
113.10	532.0
115.00	412.0
116.10	1092.0
117.00	23792.0
118.00	1498.0
118.90	377.0
119.90	354.0
121.00	375.0
122.00	1486.0
123.00	2700.0
124.00	934.0
125.00	848.0
127.00	75776.0
128.00	5670.0
129.00	30768.0
130.00	2189.0
131.00	719.0
132.00	200.0
133.00	493.0
134.00	1186.0
135.00	2702.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

136.00	1274.0
137.00	1405.0
138.00	404.0
138.80	274.0
140.00	451.0
141.00	4116.0
142.00	1405.0
143.00	826.0
144.00	292.0
145.10	290.0
146.00	897.0
147.00	2888.0
148.00	5581.0
149.00	1153.0
150.00	316.0
151.00	613.0
152.10	476.0
153.00	1801.0
154.00	1114.0
155.00	2768.0
156.10	3281.0
157.00	953.0
157.90	804.0
159.00	710.0
160.00	1603.0
160.90	1951.0
162.10	558.0
163.00	320.0
164.10	151.0
165.00	2178.0
165.90	1490.0
167.00	10105.0
168.00	4558.0
169.00	821.0
170.00	401.0
170.90	419.0
172.00	902.0
173.00	1407.0
174.00	2232.0
175.10	3705.0
176.00	1021.0
177.00	2004.0
178.00	569.0
179.00	7409.0
180.00	4035.0
181.00	2091.0
182.00	433.0
182.80	150.0
184.00	645.0
185.00	3249.0
186.00	23992.0
187.00	7325.0
188.00	854.0
189.00	1753.0
190.10	362.0
191.00	1166.0
192.00	2383.0
193.00	3640.0
193.90	616.0
195.00	230.0
196.00	5143.0
197.10	1298.0
198.00	152512.0
199.00	10812.0
200.00	892.0
200.80	160.0
201.40	944.0
203.00	1435.0
204.00	7172.0
205.00	10967.0
206.00	44312.0
207.00	9875.0
208.00	2487.0
209.00	1236.0
209.90	677.0
210.30	861.0
211.00	1841.0
212.00	242.0
213.00	271.0
215.00	750.0
216.10	1095.0
217.00	12788.0
218.00	1756.0
218.90	173.0
219.80	275.0
221.00	5860.0
222.00	1559.0
222.90	2579.0
224.00	26424.0
225.00	6405.0
226.00	591.0
227.00	11957.0
228.00	1482.0
229.00	2665.0
229.90	381.0
231.10	820.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

232.90	296.0
234.00	784.0
235.00	861.0
236.00	670.0
237.00	845.0
239.00	492.0
239.80	300.0
240.90	527.0
242.00	1526.0
243.00	1501.0
244.00	19064.0
245.00	2699.0
246.00	4191.0
246.90	684.0
249.00	681.0
249.90	269.0
250.90	200.0
251.90	229.0
252.90	358.0
254.10	951.0
255.00	108112.0
256.00	16704.0
257.00	1442.0
258.00	7148.0
259.00	1197.0
259.90	194.0
264.00	246.0
265.00	2921.0
265.90	459.0
267.10	236.0
270.00	223.0
271.10	349.0
272.00	311.0
273.00	3027.0
274.00	8628.0
275.00	44976.0
276.00	6134.0
277.00	4718.0
277.90	588.0
278.80	201.0
280.80	377.0
282.00	152.0
283.00	466.0
284.00	312.0
284.90	759.0
288.80	275.0
292.90	1050.0
294.00	270.0
295.00	420.0
296.00	15446.0
296.90	2011.0
297.90	206.0
300.90	188.0
301.90	224.0
302.90	1708.0
304.10	512.0
308.10	200.0
309.00	159.0
313.90	800.0
315.00	1566.0
316.00	866.0
321.00	469.0
321.80	169.0
323.00	4364.0
324.10	869.0
326.00	192.0
327.00	970.0
328.10	482.0
328.90	179.0
332.10	203.0
333.10	346.0
334.00	3114.0
335.00	784.0
341.20	420.0
346.00	1181.0
347.00	192.0
352.10	1261.0
353.00	972.0
354.00	1568.0
355.00	298.0
358.90	150.0
359.20	153.0
365.00	7733.0
366.00	1224.0
371.00	432.0
372.00	2202.0
373.00	622.0
383.00	470.0
389.90	327.0
391.00	295.0
392.00	182.0
402.00	883.0
403.00	1231.0
404.00	508.0
421.00	1065.0
422.00	1160.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

423.10	7169.0
424.00	1717.0
425.00	177.0
441.10	21248.0
442.10	136704.0
443.10	26256.0
444.10	2626.0
444.90	173.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14