

Data Path : Z:\HPCHEM1\BNA M\DATA\BM072815\
 Data File : BM002107.D
 Acq On : 28 Jul 2015 23:13
 Operator : TP/UM
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02084

Quant Time: Jul 29 00:47:56 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM072715.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 29 00:46:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.413	0.388	6.1	92	0.00
3 S	1,4-Dioxane-d8	0.387	0.359	7.2	89	0.00
4	Benzaldehyde	0.759	0.767	-1.1	93	0.00
5 S	Phenol-d5	1.430	1.373	4.0	92	0.00
6	Phenol	1.462	1.441	1.4	94	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.791	0.747	5.6	90	0.00
8	Bis(2-Chloroethyl)ether	1.099	1.057	3.8	93	0.00
9 S	2-Chlorophenol-d4	1.212	1.204	0.7	95	0.00
10	2-Chlorophenol	1.215	1.203	1.0	95	0.00
11	2-Methylphenol	1.107	1.072	3.2	94	0.00
12	2,2'-oxybis(1-Chloropropane	1.126	1.039	7.7	88	0.00
13 S	4-Methylphenol-d8	1.137	1.120	1.5	95	0.00
14	Acetophenone	1.800	1.752	2.7	94	0.00
15 P	N-Nitroso-di-n-propylamine	0.872	0.873	-0.1	94	0.00
16	4-Methylphenol	1.189	1.174	1.3	96	0.00
17	Hexachloroethane	0.497	0.483	2.8	95	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
19 S	Nitrobenzene-d5	0.142	0.142	0.0	97	0.00
20	Nitrobenzene	0.334	0.321	3.9	93	0.00
21	Isophorone	0.595	0.585	1.7	94	0.00
22 S	2-Nitrophenol-d4	0.158	0.163	-3.2	98	0.00
23 C	2-Nitrophenol	0.173	0.172	0.6	95	0.00
24	2,4-Dimethylphenol	0.339	0.338	0.3	96	0.00
25	Bis(2-Chloroethoxy)methane	0.362	0.345	4.7	92	0.00
26 S	2,4-Dichlorophenol-d3	0.304	0.311	-2.3	98	0.00
27 C	2,4-Dichlorophenol	0.295	0.298	-1.0	98	0.00
28	Naphthalene	0.950	0.932	1.9	95	0.00
29 S	4-Chloroaniline-d4	0.334	0.353	-5.7	98	0.00
30	4-Chloroaniline	0.348	0.372	-6.9	99	0.00
31 C	Hexachlorobutadiene	0.220	0.221	-0.5	97	0.00
32	Caprolactam	0.083	0.081	2.4	98	0.00
33 C	4-Chloro-3-methylphenol	0.293	0.293	0.0	96	0.00
34	2-Methylnaphthalene	0.691	0.684	1.0	96	0.00
35	1-Methylnaphthalene	0.665	0.653	1.8	94	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
37	1,2,4,5-Tetrachlorobenzene	0.696	0.708	-1.7	99	0.00
38	Hexachlorocyclopentadiene	0.396	0.353	10.9	94	0.00
39 C	2,4,6-Trichlorophenol	0.408	0.425	-4.2	98	0.00
40	2,4,5-Trichlorophenol	0.441	0.453	-2.7	98	0.00
41	1,1'-Biphenyl	1.541	1.516	1.6	94	0.00
42	2-Chloronaphthalene	1.198	1.203	-0.4	96	0.00
43	2-Nitroaniline	0.299	0.298	0.3	93	0.00
44 S	Dimethylphthalate-d6	1.465	1.434	2.1	93	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM072815\
 Data File : BM002107.D
 Acq On : 28 Jul 2015 23:13
 Operator : TP/UM
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02084

Quant Time: Jul 29 00:47:56 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM072715.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 29 00:46:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.463	1.437	1.8	93	0.00
46	2,6-Dinitrotoluene	0.298	0.296	0.7	93	0.00
47 S	Acenaphthylene-d8	1.832	1.825	0.4	95	0.00
48	Acenaphthylene	1.890	1.874	0.8	94	0.00
49	3-Nitroaniline	0.265	0.283	-6.8	104	0.00
50 C	Acenaphthene	1.239	1.218	1.7	94	0.00
51	2,4-Dinitrophenol	0.175	0.108	38.3#	68	0.00
52 S	4-Nitrophenol-d4	0.245	0.234	4.5	94	0.00
53	4-Nitrophenol	0.204	0.203	0.5	98	0.00
54	Dibenzofuran	1.777	1.756	1.2	94	0.00
55	2,4-Dinitrotoluene	0.428	0.428	0.0	93	0.00
56	2,3,4,6-Tetrachlorophenol	0.371	0.387	-4.3	98	0.00
57	Diethylphthalate	1.443	1.381	4.3	90	0.00
58 S	Fluorene-d10	1.289	1.269	1.6	93	0.00
59	Fluorene	1.469	1.453	1.1	94	0.00
60	4-Chlorophenyl-phenylether	0.786	0.771	1.9	94	0.00
61	4-Nitroaniline	0.282	0.316	-12.1	105	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.122	0.093	23.8#	77	0.00
64	4,6-Dinitro-2-methylphenol	0.128	0.099	22.7#	75	0.00
65	N-Nitrosodiphenylamine	0.572	0.571	0.2	94	0.00
66	4-Bromophenyl-phenylether	0.213	0.213	0.0	93	0.00
67	Hexachlorobenzene	0.233	0.235	-0.9	95	0.00
68	Atrazine	0.221	0.220	0.5	93	0.00
69 C	Pentachlorophenol	0.119	0.116	2.5	98	0.00
70	Phenanthrene	1.036	1.028	0.8	92	0.00
71 S	Anthracene-d10	0.893	0.887	0.7	92	0.00
72	Anthracene	1.059	1.042	1.6	92	0.00
73	1,2,3,4-Tetrachlorobenzene	0.303	0.311	-2.6	96	0.00
74	Pentachlorobenzene	0.283	0.288	-1.8	96	0.00
75	Carbazole	0.905	0.888	1.9	91	0.00
76	Di-n-butylphthalate	1.079	1.047	3.0	90	0.00
77 C	Fluoranthene	1.196	1.182	1.2	92	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
79 S	Pyrene-d10	0.842	0.854	-1.4	91	0.00
80	Pyrene	1.088	1.084	0.4	90	0.00
81	Butylbenzylphthalate	0.416	0.420	-1.0	90	0.00
82	3,3'-Dichlorobenzidine	0.354	0.318	10.2	84	0.00
83	Benzo(a)anthracene	1.109	1.100	0.8	89	0.00
84	Bis(2-ethylhexyl)phthalate	0.636	0.607	4.6	86	0.00
85	Chrysene	1.037	1.030	0.7	89	0.00
86 I	Perylene-d12	1.000	1.000	0.0	91	0.00
87	Di-n-octyl phthalate	1.148	1.057	7.9	85	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM072815\
Data File : BM002107.D
Acq On : 28 Jul 2015 23:13
Operator : TP/UM
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02084

Quant Time: Jul 29 00:47:56 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM072715.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 29 00:46:08 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.124	1.073	4.5	88	0.00
89	Benzo(k)fluoranthene	1.084	1.053	2.9	89	0.00
90 S	Benzo(a)pyrene-d12	0.970	0.949	2.2	90	0.00
91 C	Benzo(a)pyrene	1.085	1.056	2.7	89	0.00
92	Indeno(1,2,3-cd)pyrene	1.250	1.248	0.2	92	0.00
93	Dibenzo(a,h)anthracene	1.070	1.056	1.3	92	0.00
94	Benzo(a,h,i)perylene	1.048	1.044	0.4	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0