

Data Path : Z:\HPCHEM1\BNA M\DATA\BM072817\  
 Data File : BM011061.D  
 Acq On : 29 Jul 2017 02:44  
 Operator : SJ/JU  
 Sample : I4466-03  
 Misc :  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 HD-02-072617

Quant Time: Jul 29 04:01:01 2017  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jul 26 17:29:06 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)  |
|-----------------------------|-------|------|----------|--------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.39  | 152  | 160967   | 20.00  | ng    | 0.00      |
| 21) Naphthalene-d8          | 10.14 | 136  | 585173   | 20.00  | ng    | -0.01     |
| 38) Acenaphthene-d10        | 14.04 | 164  | 363134   | 20.00  | ng    | -0.01     |
| 63) Phenanthrene-d10        | 16.80 | 188  | 937975   | 20.00  | ng    | -0.01     |
| 75) Chrysene-d12            | 21.02 | 240  | 871826   | 20.00  | ng    | 0.00      |
| 86) Perylene-d12            | 23.13 | 264  | 311661   | 20.00  | ng    | 0.00      |
| System Monitoring Compounds |       |      |          |        |       |           |
| 5) 2-Fluorophenol           | 5.03  | 112  | 1169918  | 122.86 | ng    | 0.00      |
| 7) Phenol-d6                | 6.59  | 99   | 1429456  | 105.66 | ng    | 0.00      |
| 23) Nitrobenzene-d5         | 8.53  | 82   | 1493705  | 95.80  | ng    | 0.00      |
| 41) 2,4,6-Tribromophenol    | 15.54 | 330  | 798343   | 137.17 | ng    | -0.01     |
| 44) 2-Fluorobiphenyl        | 12.65 | 172  | 2681625  | 92.61  | ng    | -0.01     |
| 78) Terphenyl-d14           | 19.46 | 244  | 3878864  | 88.13  | ng    | 0.00      |
| Target Compounds            |       |      |          |        |       |           |
| 31) Naphthalene             | 10.19 | 128  | 63001    | 2.140  | ng    | Qvalue 97 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM072817\  
Data File : BM011061.D  
Acq On : 29 Jul 2017 02:44  
Operator : SJ/JU  
Sample : I4466-03  
Misc :  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
HD-02-072617

Quant Time: Jul 29 04:01:01 2017  
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jul 26 17:29:06 2017  
Response via : Initial Calibration

