


Data Path : Z:\HPCHEM1\BNA M\DATA\BM072817\
 Data File : BM011031.D
 Acq On : 28 Jul 2017 08:24
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM072617.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 26 17:29:06 2017



AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	118069	PASS
68	69	0.00	2	1.8	2386	PASS
69	198	0.00	100	40.9	135501	PASS
70	69	0.00	2	0.3	348	PASS
127	198	10	80	38.1	126226	PASS
197	198	0.00	2	0.6	2065	PASS
198	198	100	100	100.0	331050	PASS
199	198	5	9	6.7	22135	PASS
275	198	10	60	29.6	97842	PASS
365	198	1	100	5.6	18690	PASS
441	443	0.01	100	87.0	37410	PASS
442	198	50	100	72.3	239232	PASS
443	442	15	24	18.0	42976	PASS