


Data Path : Z:\HPCHEM1\BNA M\DATA\BM072817\  
 Data File : BM011042.D  
 Acq On : 28 Jul 2017 15:18  
 Operator : SJ/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_M\METHODS\8270-BM072617.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Jul 26 17:29:06 2017



AutoFind: Scans 2462, 2463, 2464; Background Corrected with Scan 2455

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 39.5      | 117149  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.8       | 2525    | PASS             |
| 69          | 198          | 0.00         | 100          | 46.9      | 139141  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.7       | 941     | PASS             |
| 127         | 198          | 10           | 80           | 41.8      | 124066  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.7       | 1946    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 296725  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 20138   | PASS             |
| 275         | 198          | 10           | 60           | 28.8      | 85429   | PASS             |
| 365         | 198          | 1            | 100          | 5.0       | 14846   | PASS             |
| 441         | 443          | 0.01         | 100          | 74.9      | 29330   | PASS             |
| 442         | 198          | 50           | 100          | 65.6      | 194757  | PASS             |
| 443         | 442          | 15           | 24           | 20.1      | 39144   | PASS             |